



**Analytical Resources, LLC**  
Analytical Chemists and Consultants

18 May 2023

Ali Judkins  
Anchor QEA, LLC  
1201 3rd Ave, Suite 2600  
Seattle, WA 98101

RE: AOC5 MR Phase 1

Please find enclosed sample receipt documentation and analytical results for samples from the project referenced above.

Sample analyses were performed according to ARI's Quality Assurance Plan and any provided project specific Quality Assurance Plan. Each analytical section of this report has been approved and reviewed by an analytical peer, the appropriate Laboratory Supervisor or qualified substitute, and a technical reviewer.

Should you have any questions or problems, please feel free to contact us at your convenience.

Associated Work Order(s)  
23B0228

Associated SDG ID(s)  
N/A

-----

I certify that this data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed in the enclosed Narrative. ARI, an accredited laboratory, certifies that the report results for which ARI is accredited meets all the requirements of the accrediting body. A list of certified analyses, accreditations, and expiration dates is included in this report.

Release of the data contained in this hardcopy data package has been authorized by the Laboratory Manager or his/her designee, as verified by the following signature.

Analytical Resources, LLC

Susan Dunninghoo, Director, Client Services

*The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.*



of

23B0228

# CHAIN-OF-CUSTODY/TEST REQUEST FORM

№ 4208

Project/Client Name: AUCS MR Phase 1  
 Project Number: 210075.01.02  
 Contact Name: Amara Vanderweert  
 Sampled By: Woodward

Ship to: ARL  
 Attn: Sue Donihoo  
 Shipping Date: 2/19/23  
 Shipper: \_\_\_\_\_  
 Airbill Number: \_\_\_\_\_  
 Form filled out by: AVICC  
 Turnaround requested: Sid

Sample Collection Date (m/d/y)	Time	Sample Identification	Volume of Sample / # of Containers	Matrix	Test(s) Requested (check test(s) required)							Comments / Instructions (Jar tag number(s))
					PCBS	SMS SVCS	PAH	DH	Arcs	SMS Metals	Asmet PAHS	
2/18/23	1630	LOW23-SC1009	3	Sediment	X	X	X	X	X	X		
<del>2/19/23</del>												
<b>Total Number of Containers</b>			3	<b>Purchase Order / Statement of Work #</b> APJ-110200-AUCS-ARL								

1) Released by: Print name: <u>Amara Vanderweert</u> Signature: <u>[Signature]</u> Company: <u>Woodward</u> Date/Time: <u>2/19/23 1058</u>	1) Rec'd by: <u>[Signature]</u> Company: <u>[Company]</u> Date/Time: <u>Nov 11 1058</u>	2) Released by: Print name: Signature: Company: Date/Time:	2) Rec'd by:  Company: Date/Time:
--	--	--	--

\* Distribution: White copies accompany shipment; yellow retained by consignor.



200 1st Ave W, Suite 500  
 Seattle, WA 98119  
 206.378.1364

### To be completed by Laboratory upon sample receipt:

Date of receipt:	Laboratory W.O. #:
Condition upon receipt:	Time of receipt:
Cooler temperature:	Received by:



# Cooler Receipt Form

ARI Client: Winover/Arcon 7.5 Project Name: AACS MRI  
 COC No(s): 4208 NA 02110123 Delivered by: Fed-Ex UPS Courier Hand Delivered Other: \_\_\_\_\_  
 Assigned ARI Job No: 23B0228 Tracking No: \_\_\_\_\_ NA

**Preliminary Examination Phase:**

Were intact, properly signed and dated custody seals attached to the outside of the cooler? YES NO  
 Were custody papers included with the cooler? ..... YES NO  
 Were custody papers properly filled out (ink, signed, etc.) ..... YES NO  
 Temperature of Cooler(s) (°C) (recommended 2.0-6.0 °C for chemistry)  
 Time 13:26 0.4  
 If cooler temperature is out of compliance fill out form 00070F Temp Gun ID#: 7099208

Cooler Accepted by: [Signature] Date: 02/09/23 Time: 1055

**Complete custody forms and attach all shipping documents**

**Log-In Phase:**

Was a temperature blank included in the cooler? ..... YES NO  
 What kind of packing material was used? ... Bubble Wrap Wet Ice Gel Packs Baggies Foam Block Paper Other: \_\_\_\_\_  
 Was sufficient ice used (if appropriate)? ..... NA YES NO  
 How were bottles sealed in plastic bags? ..... Individually Grouped Not  
 Did all bottles arrive in good condition (unbroken)? ..... YES NO  
 Were all bottle labels complete and legible? ..... YES NO  
 Did the number of containers listed on COC match with the number of containers received? ..... YES NO  
 Did all bottle labels and tags agree with custody papers? ..... YES NO  
 Were all bottles used correct for the requested analyses? ..... YES NO  
 Do any of the analyses (bottles) require preservation? (attach preservation sheet, excluding VOCs) ... NA YES NO  
 Were all VOC vials free of air bubbles? ..... NA YES NO  
 Was sufficient amount of sample sent in each bottle? ..... YES NO  
 Date VOC Trip Blank was made at ARI ..... NA  
 Were the sample(s) split by ARI? NA YES Date/Time: \_\_\_\_\_ Equipment: \_\_\_\_\_ Split by: \_\_\_\_\_

Samples Logged by: [Signature] Date: 02/10/23 Time: 9:27 Labels checked by: TCS

**\*\* Notify Project Manager of discrepancies or concerns \*\***

Sample ID on Bottle	Sample ID on COC	Sample ID on Bottle	Sample ID on COC

**Additional Notes, Discrepancies, & Resolutions:**

By: \_\_\_\_\_ Date: \_\_\_\_\_



Anchor QEA, LLC

1201 3rd Ave, Suite 2600

Seattle, WA 98101

Project: AOC5 MR Phase 1

Project Number: 210075-01.02

Project Manager: Ali Judkins

**Reported:**

05/18/2023 15:13

**ANALYTICAL REPORT FOR SAMPLES**

Laboratory ID	Sample ID	Matrix	Date Sampled	Date Received
23B0228-01	LDW23-SC1009	Solid	02/08/23 16:30	02/09/23 10:55



Anchor QEA, LLC  
1201 3rd Ave, Suite 2600  
Seattle WA, 98101

Project: AOC5 MR Phase 1  
Project Number: 210075-01.02  
Project Manager: Ali Judkins

Reported:  
18-May-2023 15:13

## Case Narrative

**Client:** Anchor QEA, LLC  
**Project:** AOC5 MR Phase 1  
**Work Order:** 23B0228

### Sample receipt

One sample as listed on the preceding page was received 09-Feb-2023 10:55 under ARI work order 23B0228. For details regarding sample receipt, please refer to the Cooler Receipt Form.

### PCB Aroclors - EPA Method SW8082A

The sample(s) were extracted and analyzed within the recommended holding times.

Initial and continuing calibrations were within method requirements.

Internal standard areas were within limits.

The surrogate percent recoveries were within control limits.

The method blank(s) were clean at the reporting limits.

The blank spike and blank spike duplicate (BS/LCS and BSD/LCSD) spike recoveries and relative percent difference (RPD) were within control limits.

The reference material (SRM) percent recoveries were within control limits.

The batch BLB0427 matrix spike/matrix spike duplicate (MS/MSD) percent recoveries and relative percent difference (RPD) were within advisory control limits, reported under work order 23B0229.

Samples from this project showed extreme matrix effect with multiple failing runs. The analyst noted ICV2,CCV2,4,6 failed high for 1260 on zb5 so all associated data is reported from ZB35 as primary and a number of extracts run at dilution in an attempt to mitigate the matrix issues.

### Wet Chemistry

The sample(s) were prepared and analyzed within the recommended holding times.

Initial and continuing calibrations were within method requirements.

The method blank(s) were clean at the reporting limits.

The blank spike (BS/LCS) percent recoveries were within control limits.

The reference material (SRM) percent recoveries were within control limits.

The batch BLB0342 matrix spike (MS) percent recoveries and the duplicate (DUP) relative percent difference (RPD) were within advisory control limits, reported under work order 23B0229.

*Revised 05/18/2023 to correct calibration reference for CCAL in sequence SLB0179 and add FD00070 ICAL summary form and raw data.*



## QUALIFIERS AND NOTES

<u>Qualifier</u>	<u>Definition</u>
U	This analyte is not detected above the reporting limit (RL) or if noted, not detected above the limit of detection (LOD).
J	Estimated concentration value detected below the reporting limit.
D	The reported value is from a dilution
*	Flagged value is not within established control limits.
DET	Analyte DETECTED
ND	Analyte NOT DETECTED at or above the reporting limit
NR	Not Reported
dry	Sample results reported on a dry weight basis
RPD	Relative Percent Difference



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282337ECD7.D  
Data file 2: /230228.b/230228.b/02282337ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 23B0228-01RE1  
Client ID:  
Injection Date: 01-MAR-2023 04:32  
Report Date: 03/01/2023 12:21  
Matrix: NONE  
Dilution Factor: 3.0

SURROGATES

ZB5 Col		ZB35 Col		ZB5	ZB35	RPD	Compound/Flag	
RT	Shift Response	RT	Shift Response	on col	on col			
5.806	-0.002	114492	5.684 -0.003	50462	9.1	10.2	10.9	Tetrachloro-m-xylene
13.885	-0.008	99940	14.113 -0.007	63642	11.8	11.2	5.7	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	840131	24.7
Hexabromobiphenyl	1429847	857320	-40.0

Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	337914	7.2
Hexabromobiphenyl	513946	373697	-27.3

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)



ZB5 Col					ZB35 Col						
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	---			0.0	1	---			0.0	
Aroclor-1016	2	---			0.0	2	---			0.0	
Aroclor-1016	3	---			0.0	3	---			0.0	
Aroclor-1016	4	---			0.0	4	---			0.0	
CollAve: <3 Quant Peaks					Col2Ave: <3 Quant Peaks						
Aroclor-1221	1	---			0.0	1	---			0.0	
Aroclor-1221	2	---			0.0	2	---			0.0	
Aroclor-1221	3	---			0.0	3	---			0.0	
CollAve: <3 Quant Peaks					Col2Ave: <3 Quant Peaks						
Aroclor-1232	1	---			0.0	1	---			0.0	
Aroclor-1232	2	---			0.0	2	---			0.0	
Aroclor-1232	3	---			0.0	3	---			0.0	
Aroclor-1232	4	---			0.0	4	---			0.0	
CollAve: <3 Quant Peaks					Col2Ave: <3 Quant Peaks						
Aroclor-1242	1	---			0.0	1	---			0.0	
Aroclor-1242	2	---			0.0	2	---			0.0	
Aroclor-1242	3	---			0.0	3	---			0.0	
Aroclor-1242	4	---			0.0	4	---			0.0	
CollAve: <3 Quant Peaks					Col2Ave: <3 Quant Peaks						
Aroclor-1248	1	8.396	-0.010	27949	68.2	1	8.299	-0.008	11777	73.0	
Aroclor-1248	2	8.566	-0.015	23687	45.5	2	8.706	-0.009	11294	67.7	
Aroclor-1248	3	8.985	-0.012	73145	74.4	3	9.141	-0.027	15897	82.8	
Aroclor-1248	4	9.287	-0.008	75070	150.0	4	9.532	-0.062	16053	69.6	
Total CollAve (4 peaks):				84.5	Total Col2Ave (4 peaks):				73.3	RPD = 14	
Corrected Ave (3 peaks):				62.7	Corrected Ave (3 peaks):				70.1	RPD = 11	
<b>74.5</b>											
Aroclor-1254	1	9.287	-0.012	75070	89.0	1	9.439	-0.013	29370	114.4	
Aroclor-1254	2	9.362	-0.016	35739	94.2	2	9.957	-0.015	15191	73.5	
Aroclor-1254	3	9.657	-0.012	49919	92.0	3	10.107	-0.019	55733	124.7	
Aroclor-1254	4	9.787	-0.021	105204	99.8	4	10.354	-0.020	65613	150.6	
Aroclor-1254	5	10.115	-0.064	135513	205.0	5	10.554	-0.016	49323	185.9	
Total CollAve (5 peaks):				146.0	Total Col2Ave (5 peaks):				129.8	RPD = 11	
Corrected Ave (4 peaks):				93.7	Corrected Ave (4 peaks):				115.8	RPD = 21	
Aroclor-1260	1	11.033	-0.011	44931	145.7	1	11.643	-0.009	31037	141.2	
Aroclor-1260	2	11.348	-0.012	38466	119.4	2	11.904	-0.013	53513	95.4	
Aroclor-1260	3	11.719	-0.017	96328	112.7	3	12.424	-0.012	18763	126.1	
Aroclor-1260	4	12.121	-0.019	56103	130.4	4	12.488	-0.014	39903	105.6	
Aroclor-1260	5	12.234	-0.010	23989	129.5	NS	---			---	
Total CollAve (5 peaks):				127.5	Total Col2Ave (4 peaks):				117.1	RPD = 9	
Corrected Ave (4 peaks):				123.0	Corrected Ave (3 peaks):				109.0	RPD = 12	
Aroclor-1262	1	---			0.0	1	---			0.0	
Aroclor-1262	2	---			0.0	2	---			0.0	
Aroclor-1262	3	---			0.0	3	---			0.0	
Aroclor-1262	4	---			0.0	4	---			0.0	
CollAve: <3 Quant Peaks					Col2Ave: <3 Quant Peaks						
Aroclor-1268	1	---			0.0	1	---			0.0	
Aroclor-1268	2	---			0.0	2	---			0.0	
Aroclor-1268	3	---			0.0	3	---			0.0	
Aroclor-1268	4	---			0.0	4	---			0.0	
CollAve: <3 Quant Peaks					Col2Ave: <3 Quant Peaks						

Total PCB Area Col1 (5.907 - 13.793) = 1965990 Col1 Total PCB = 0.2 ppm\*  
Total PCB Area Col2 (5.787 - 14.020) = 944780 Col2 Total PCB = 0.2 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 23B0228-01RE1


01-MAR-2023 04:32, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 23B0228-01RE1

01-MAR-2023 04:32, 2ul



ZB-35 Manual Integration: YES







Batch: BLB0427

Prepared using: EPA 3546 (Microwave)

8082A PCB Solid 4 in Solid (Version:7 Aroclors)

Matrix: Solid

Date Prepared: 02/17/23

Balance ID: B146462114

Set Up By: CR 2/16/23

WO Comments

23B0228: <C>BPR SRM, MS, DUP <C> <M>BPR PS, MS/MSD <M> <E>BPR 8270E RM K000591, SIM PAH RM I009127 PCB RM J006840-43, 7935-36,K011477-79, MS/MSD <E>  
<H>BPR J006840-43, 7935-36, K011477-79, Dup </H> Store in freezer (except GS)  
23B0229: <C>BPR SRM, MS, DUP <C> <M>BPR PS, MS/MSD <M> <E>BPR 8270E RM H002055, SIM RM H010158, PCB RM J006840-43, 7935-36,K011477-79, MS/MSD <E>  
<H>BPR J006840-43, 7935-36, K011477-79, Dup </H> Store in freezer (except GS)

The following standards may be missing from this batch!

Designator	Description
QLS 5	QLS Spike

Analysis: 8082A PCB Solid 4

Lab Number & Container	% Solids	Initial (g)		(REQ) Acid C/U (5mL)	(REQ) Sulfur C/U (5mL)	(REQ) Silica Gel C/U (2:5)	Final Effective Vol (mL)	Vol (mL) to Lab	Extraction Comments
		Target Dry: 12.5 (Wet)	Actual						
23B0228-01 A	62.0	(20.16)	20.19	5mL	5mL	2mL	2.5	1.0	
23B0229-01 A	51.9	(24.09)	24.09	5mL	5mL	2mL	2.5	1.0	
23B0229-02 A	56.0	(22.33)	22.32	5mL	5mL	2mL	2.5	1.0	
23B0229-03 A	54.8	(22.81)	22.85	5mL	5mL	2mL	2.5	1.0	
23B0229-04 A	52.2	(23.97)	23.98	5mL	5mL	2mL	2.5	1.0	
23B0229-05 A	44.9	(27.82)	27.84	5mL	5mL	2mL	2.5	1.0	
23B0229-06 A	48.6	(25.71)	25.77	5mL	5mL	2mL	2.5	1.0	
23B0229-07 A	68.7	(18.21)	18.23	5mL	5mL	2mL	2.5	1.0	
23B0229-08 A	49.7	(25.16)	25.21	5mL	5mL	2mL	2.5	1.0	

Batch QC

Lab Number	% Solids	Initial (g)		(REQ) Acid C/U (5mL)	(REQ) Sulfur C/U (5mL)	(REQ) Silica Gel C/U (2:5)	Final Effective Vol (mL)	Vol (mL) to Lab	Extraction Comments
		Target Dry: 12.5 (Wet)	Actual						
BLB0427-BLK1	100.0	(12.50)	12.54	5mL	5mL	2mL	2.5	1.0	(10g Actual Wt.)
BLB0427-BS1	100.0	(12.50)	12.54	5mL	5mL	2mL	2.5	1.0	(10g Actual Wt.)
BLB0427-BSD1	100.0	(12.50)	12.50	5mL	5mL	2mL	2.5	1.0	(10g Actual Wt.)
BLB0427-MS1	68.7	(18.21)	18.23	5mL	5mL	2mL	2.5	1.0	Use 23B0229-07
BLB0427-MSD1	68.7	(18.21)	18.23	5mL	5mL	2mL	2.5	1.0	Use 23B0229-07
BLB0427-SRM1	100.0	(12.50) <sup>(2.50)</sup>	25.0	5mL	5mL	2mL	2.5	1.0	Use K011478

+1g DI WATER

Client ID Verified By

Date

Preparation Reviewed By

Date

Extraction Date and Time



Batch: BLB0427

Prepared using: EPA 3546 (Microwave)

8082A PCB Solid 4 in Solid (Version:7 Aroclors)

**WO Comments**  
 23B0228: <C>BPR SRM, MS, DUP <C> <M>BPR PS, MS/MSD </M> <E>BPR 8270E RM K000591, SIM PAH RM 1009127 PCB RM J006840-43, 7935-36, K011477-79, MS/MSD </E>  
 <H>BPR J006840-43, 7935-36, K011477-79, Dup </H> Store in freezer (except GS)  
 23B0229: <C>BPR SRM, MS, DUP <C> <M>BPR PS, MS/MSD </M> <E>BPR 8270E RM H002055, SIM RM H010158, PCB RM J006840-43, 7935-36, K011477-79, MS/MSD </E>  
 <H>BPR J006840-43, 7935-36, K011477-79, Dup </H> Store in freezer (except GS)

Prep Steps

Reagents Used

Surrogates & Spike Standards Used

Microwave 1 2 3 CT 2/17/23 Analyst/Date
KD 100°C Hexane Exchange (2 X 20 mL) 1 2 3 4 5 6 LO 2-20 Analyst/Date
TurboVap Pre Cleanups 1 2 3 4 5 MRS 2/20/23 Analyst/Date
TurboVap Post Cleanups 1 2 3 4 5 MRS 2/20/23 Analyst/Date
Vialing MRS 2/20/23 Analyst/Date

Station/Reagent	Standard ID
Microwave	
Analyst: CT/LO Date: 2/17/23	
Neutral Glass Wool	L000350
1:1 Hexane/Acetone	L001220
Hexane	L000889
Anhydrous Sodium Sulfate	L001285
KD	
Analyst: LO Date: 2-20-23	
Anhydrous Sodium Sulfate	
Hexane	L000889
Vialing	
Analyst: MRS Date: 2/20/23	
Hexane	L000889
Concentrated Sulfuric Acid	L001033
Silica Gel (SPE) Darts	L001054
Sodium Sulfite	K010363
Tetrabutylammonium hydrogensulfate (TBAS)	L001601

Type	Vial ID / Standard ID	Vol uL	Analyst	Witness
Surrogate	N L000773	50µL	CT	M
2µg/mL	Exp Date: 7/21/2023			
Spike	I L001587	63µL	CT	M
20µg/mL	Exp Date: 8/13/2023			

MANUALLY ENTER EXPIRATION DATES!

(V) indicates a virtual standard combining two or more physical standards. In these cases the Standard ID refers to the virtual standard, not the parent standards.

If a Standard ID is missing, but should be present, check the standard definition in Element LIMS to be sure Standard Info 6 has the correct letter or number designator matching the vial designator in the Standard ID column. If it is correct, check the batch and bench sheet in Element LIMS to be sure the correct standards are selected for surrogate(s) and spike(s).



Batch: BLB0427

Prepared using: EPA 3546 (Microwave)  
8082A PCB Solid 4 in Solid (Version:7 Aroclors)

**WO Comments**  
**23B0228:** <C>BPR SRM, MS, DUP <C> <M>BPR PS, MS/MSD <M> <E>BPR 8270E RM K000591, SIM PAH RM I009127 PCB RM J006840-43, 7935-36,K011477-79, MS/MSD <E>  
 <H>BPR J006840-43, 7935-36, K011477-79, Dup <H> Store in freezer (except GS)  
**23B0229:** <C>BPR SRM, MS, DUP <C> <M>BPR PS, MS/MSD <M> <E>BPR 8270E RM H002055, SIM RM H010158, PCB RM J006840-43, 7935-36,K011477-79, MS/MSD <E>  
 <H>BPR J006840-43, 7935-36, K011477-79, Dup <H> Store in freezer (except GS)

Prep Instructions	
<p><b>SPECIAL INSTRUCTIONS:</b></p> <ol style="list-style-type: none"> <li>1. Weigh soil/sed into beakers-lightly dry with sodium sulfate.</li> <li>2. Transfer to microwave vessel(s). Note: (do not fill vessels more than 2/3rd full. Some samples may require two vessels).</li> <li>3. Add 1:1 Hexane/Acetone until the solvent layer is 3 inches above the soil layer after homogenization.</li> <li>4. Add surr/spike.</li> <li>5. Microwave on appropriate power setting determined by # of samples.</li> <li>6. After microwave-Re-homogenize while hot then cool vessels in R-05 15 minutes. Re-homogenize while cool.</li> <li>7. Decant 1:1 Hex/Ace into Erlenmeyer flask with sodium sulfate in bottom and funnel with neutral glasswool plug.</li> <li>8. Re-homogenize and rinse with 1:1 Hexane/Acetone.</li> <li>9. Let cool and decant solvent then empty the soil into the funnel and rinse with Hexane.</li> <li>10. KD on 100° bath.</li> <li>11. Exchange (2 X with 20mL) Hexane.</li> <li>12. TurboVap.</li> <li>13. Clean-ups.</li> <li>14. TurboVap.</li> <li>15. Vial with Hexane.</li> </ol> <p>A. Need Total Solids Y <input type="checkbox"/> N <input checked="" type="checkbox"/></p> <p>B. Archive/Freeze <input checked="" type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/></p>	







Extraction Parameter: PCB Extraction Batch BLB0427

Total Solids Batch: BLB0340 Work Order(s): 23B0228 01



Screens: Soil/Sediment/Solid/Other:	Analyst/Date
<input checked="" type="checkbox"/> No Anomalies (standard soil/wet sediment/sand/gravel)= 01	UR 2/15/23
<input type="checkbox"/> Standing Water Decanted (Not shared)=	
<input type="checkbox"/> Standing Water Homogenized (Shared samples)=	
<input type="checkbox"/> Clay/Clumps (Difficult to homogenize)=	
<input type="checkbox"/> Rocks (%+size)?	
<input type="checkbox"/> Organics (Leaves/sticks/grass)=	
<input type="checkbox"/> Oily, obvious fuel/sulfur odors=	
<input type="checkbox"/> Received in 32oz jar(s)=Homogenized in Pyrex dish=	
<input checked="" type="checkbox"/> Previously Frozen = 01	UR 2/15/23
<input type="checkbox"/> Other (Details)=	
<b>Aqueous:</b>	
<input checked="" type="checkbox"/> No Anomalies	
<input type="checkbox"/> Turbid/Color=	
<input type="checkbox"/> Particulates(%)=(Note: >5%=Notify Supervisor/Lead)	
<input type="checkbox"/> Emulsions (%)=	
<input type="checkbox"/> Oily, obvious fuel/sulfur odors=	
<input type="checkbox"/> Other (Details)=	
<input type="checkbox"/> Received in 1.0L Bottle(s)=No Bottle Rinse=	
<input checked="" type="checkbox"/> Other Notes/Comments= (Note problems, concerns, corrective actions). -lost some of BSP1 and maybe BSI during sulfur chr, the stickers got stuck on the vial	NRB 2/20/23
<input checked="" type="checkbox"/> Share Samples Y/N <u>(N)</u>	UR 2/15/23
<input checked="" type="checkbox"/> Multiple Jars Y/N <u>(N)</u>	UR 2/15/23
<input type="checkbox"/> Sample Pre-Screens indicate analyte activity=	
<input type="checkbox"/> Sample weights/volumes reduced based on Pre-Screen=	

=====  
Injection Date : 2/15/2023 5:38:15 PM                   Seq. Line : 10  
Sample Name    : 23B0228 01                                Location : Vial 10  
Acq. Operator  : CR   Inj : 1  
  Inj Volume : 1 µl  
Sequence File  : C:\HPCHEM\1\SEQUENCE\021523B.S  
Method         : C:\HPCHEM\1\METHODS\SCREEN.M  
Last changed   : 7/9/2021 3:37:33 AM by TW  
SCREEN METHOD  
=====





\*\*\* End of Report \*\*\*

=====  
Injection Date : 2/15/2023 5:52:48 PM                   Seq. Line : 11  
Sample Name    : 23B0229 01                                Location : Vial 11  
Acq. Operator  : CR    Inj : 1  
  Inj Volume : 1 µl  
Sequence File  : C:\HPCHEM\1\SEQUENCE\021523B.S  
Method        : C:\HPCHEM\1\METHODS\SCREEN.M  
Last changed  : 7/9/2021 3:37:33 AM by TW  
SCREEN METHOD  
=====





\*\*\* End of Report \*\*\*

=====  
Injection Date : 2/15/2023 6:06:45 PM      Seq. Line : 12  
Sample Name : 23B0229 02                      Location : Vial 12  
Acq. Operator : CR                              Inj : 1  
   Inj Volume : 1 µl  
Sequence File : C:\HPCHEM\1\SEQUENCE\021523B.S  
Method : C:\HPCHEM\1\METHODS\SCREEN.M  
Last changed : 7/9/2021 3:37:33 AM by TW  
SCREEN METHOD  
=====





\*\*\* End of Report \*\*\*

=====  
Injection Date : 2/15/2023 6:21:23 PM                   Seq. Line : 13  
Sample Name : 23B0229 03                                    Location : Vial 13  
Acq. Operator : CR    Inj : 1  
  Inj Volume : 1 µl  
Sequence File : C:\HPCHEM\1\SEQUENCE\021523B.S  
Method : C:\HPCHEM\1\METHODS\SCREEN.M  
Last changed : 7/9/2021 3:37:33 AM by TW  
SCREEN METHOD  
=====





\*\*\* End of Report \*\*\*

=====  
Injection Date : 2/15/2023 6:35:23 PM                   Seq. Line : 14  
Sample Name : 23B0229 04                                Location : Vial 14  
Acq. Operator : CR   Inj : 1  
  Inj Volume : 1 µl  
Sequence File : C:\HPCHEM\1\SEQUENCE\021523B.S  
Method : C:\HPCHEM\1\METHODS\SCREEN.M  
Last changed : 7/9/2021 3:37:33 AM by TW  
SCREEN METHOD  
=====



\*\*\* End of Report \*\*\*

=====  
Injection Date : 2/15/2023 6:49:53 PM                   Seq. Line : 15  
Sample Name    : 23B0229 05                                Location : Vial 15  
Acq. Operator  : CR   Inj : 1  
  Inj Volume : 1 µl  
Sequence File  : C:\HPCHEM\1\SEQUENCE\021523B.S  
Method         : C:\HPCHEM\1\METHODS\SCREEN.M  
Last changed  : 7/9/2021 3:37:33 AM by TW  
SCREEN METHOD  
=====





\*\*\* End of Report \*\*\*







=====  
Injection Date : 2/15/2023 7:18:26 PM      Seq. Line : 17  
Sample Name : 23B0229 07                      Location : Vial 17  
Acq. Operator : CR                              Inj : 1  
   Inj Volume : 1 µl  
Sequence File : C:\HPCHEM\1\SEQUENCE\021523B.S  
Method : C:\HPCHEM\1\METHODS\SCREEN.M  
Last changed : 7/9/2021 3:37:33 AM by TW  
SCREEN METHOD  
=====



\*\*\* End of Report \*\*\*

=====  
Injection Date : 2/15/2023 7:32:27 PM      Seq. Line : 18  
Sample Name : 23B0229 08                      Location : Vial 18  
Acq. Operator : CR                              Inj : 1  
   Inj Volume : 1 µl  
Sequence File : C:\HPCHEM\1\SEQUENCE\021523B.S  
Method : C:\HPCHEM\1\METHODS\SCREEN.M  
Last changed : 7/9/2021 3:37:33 AM by TW  
SCREEN METHOD  
=====



\*\*\* End of Report \*\*\*



### CLEANUP BATCH SUMMARY

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Cleanup Batch: CLB0181

Cleanup Type: Sulfuric Acid

Cleanup Method: EPA 3665 Sulfuric Acid Cleanup - uL

Analysis: EPA 8082A

SAMPLE NAME	LAB SAMPLE ID	LAB FILE ID	DATE PREPARED	OBSERVATIONS
LDW23-SC1009	23B0228-01	02282337ECD7.D	02/21/2023	
Blank	BLB0427-BLK1	02282333ECD7.D	02/21/2023	
LCS	BLB0427-BS1	02282334ECD7.D	02/21/2023	
LCS Dup	BLB0427-BSD1	02282335ECD7.D	02/21/2023	
Reference	BLB0427-SRM1	02282336ECD7.D	02/21/2023	



**CLEANUP BENCH SHEET**

CLB0181

Matrix: Solid

Cleanup using: Organics - EPA 3665 Sulfuric Acid Cleanup - uL

Printed: 2/21/2023 11:13:29AM

Lab Number	Sample Container	Sample Name	Extract Container	Initial (uL)	Final (uL)	Analysis	Clean Up Date	Cleaned By	Cleanup Comments
23B0228-01	A	LDW23-SC1009	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-01	A	LDW23-SC1236	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-02	A	LDW23-SS1236	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-03	A	LDW23-SS1237	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-04	A	LDW23-SS1150	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-05	A	LDW23-SS1008	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-06	A	LDW23-SC1008	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-07	A	LDW23-SC1014	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-08	A	LDW23-SC1013	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
BLB0427-BLK1	-	Blank	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-BS1	-	LCS	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-BSD1	-	LCS Dup	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-MS1	-	Matrix Spike	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-MSD1	-	Matrix Spike Dup	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-SRM1	-	Reference	-	2.5	2.5	-	2/21/2023	NRB	



## CLEANUP BATCH SUMMARY

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Cleanup Batch: CLB0182

Cleanup Type: Sulfur

Cleanup Method: EPA 3660B Sulfur Cleanup - uL

Analysis: EPA 8082A

SAMPLE NAME	LAB SAMPLE ID	LAB FILE ID	DATE PREPARED	OBSERVATIONS
LDW23-SC1009	23B0228-01	02282337ECD7.D	02/21/2023	
Blank	BLB0427-BLK1	02282333ECD7.D	02/21/2023	
LCS	BLB0427-BS1	02282334ECD7.D	02/21/2023	
LCS Dup	BLB0427-BSD1	02282335ECD7.D	02/21/2023	
Reference	BLB0427-SRM1	02282336ECD7.D	02/21/2023	



**CLEANUP BENCH SHEET**

CLB0182

Matrix: Solid

Cleanup using: Organics - EPA 3660B Sulfur Cleanup - uL

Printed: 2/21/2023 11:14:01AM

Lab Number	Sample Container	Sample Name	Extract Container	Initial (uL)	Final (uL)	Analysis	Clean Up Date	Cleaned By	Cleanup Comments
23B0228-01	A	LDW23-SC1009	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-01	A	LDW23-SC1236	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-02	A	LDW23-SS1236	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-03	A	LDW23-SS1237	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-04	A	LDW23-SS1150	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-05	A	LDW23-SS1008	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-06	A	LDW23-SC1008	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-07	A	LDW23-SC1014	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-08	A	LDW23-SC1013	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
BLB0427-BLK1	-	Blank	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-BS1	-	LCS	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-BSD1	-	LCS Dup	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-MS1	-	Matrix Spike	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-MSD1	-	Matrix Spike Dup	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-SRM1	-	Reference	-	2.5	2.5	-	2/21/2023	NRB	



### CLEANUP BATCH SUMMARY

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Cleanup Batch: CLB0183

Cleanup Type: Silica Gel

Cleanup Method: EPA 3630C Silica Gel Cleanup - uL

Analysis: EPA 8082A

SAMPLE NAME	LAB SAMPLE ID	LAB FILE ID	DATE PREPARED	OBSERVATIONS
LDW23-SC1009	23B0228-01	02282337ECD7.D	02/21/2023	
Blank	BLB0427-BLK1	02282333ECD7.D	02/21/2023	
LCS	BLB0427-BS1	02282334ECD7.D	02/21/2023	
LCS Dup	BLB0427-BSD1	02282335ECD7.D	02/21/2023	
Reference	BLB0427-SRM1	02282336ECD7.D	02/21/2023	



**CLEANUP BENCH SHEET**

CLB0183

Matrix: Solid

Cleanup using: Organics - EPA 3630C Silica Gel Cleanup - uL

Printed: 2/21/2023 11:14:33AM

Lab Number	Sample Container	Sample Name	Extract Container	Initial (uL)	Final (uL)	Analysis	Clean Up Date	Cleaned By	Cleanup Comments
23B0228-01	A	LDW23-SC1009	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-01	A	LDW23-SC1236	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-02	A	LDW23-SS1236	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-03	A	LDW23-SS1237	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-04	A	LDW23-SS1150	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-05	A	LDW23-SS1008	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-06	A	LDW23-SC1008	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-07	A	LDW23-SC1014	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
23B0229-08	A	LDW23-SC1013	A 01	2.5	2.5	8082A PCB Solid 4	2/21/2023	NRB	
BLB0427-BLK1	-	Blank	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-BS1	-	LCS	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-BSD1	-	LCS Dup	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-MS1	-	Matrix Spike	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-MSD1	-	Matrix Spike Dup	-	2.5	2.5	-	2/21/2023	NRB	
BLB0427-SRM1	-	Reference	-	2.5	2.5	-	2/21/2023	NRB	





**Form I**  
**METHOD BLANK DATA SHEET**  
**EPA 8082A**

Blank
-------

Laboratory:	<u>Analytical Resources, LLC</u>	SDG:	<u>23B0228</u>
Client:	<u>Anchor QEA, LLC</u>	Project:	<u>AOC5 MR Phase 1</u>
Matrix:	<u>Solid</u>	Laboratory ID:	<u>BLB0427-BLK1</u>
Sampled:	<u>N/A</u>	Prepared:	<u>02/17/23 13:53</u>
Solids:		Preparation:	<u>EPA 3546 (Microwave)</u>
Batch:	<u>BLB0427</u>	Sequence:	<u>SLC0014</u>
Instrument:	<u>ECD7</u>	Column:	<u>ZB5</u>
		File ID:	<u>02282333ECD7.D</u>
		Analyzed:	<u>03/01/23 03:08</u>
		Initial/Final:	<u>12.5 g / 2.5 mL</u>
		Calibration:	<u>GB00069</u>
		Cleanups:	<u>Silica Gel, Sulfur, Sulfuric Acid</u>

CAS NO.	COMPOUND	DILUTION	CONC: (ug/kg wet)	Q	DL	RL
12674-11-2	Aroclor 1016	1	4.0	U	1.6	4.0
11104-28-2	Aroclor 1221	1	4.0	U	1.6	4.0
11141-16-5	Aroclor 1232	1	4.0	U	1.6	4.0
53469-21-9	Aroclor 1242	1	4.0	U	1.6	4.0
12672-29-6	Aroclor 1248	1	4.0	U	1.6	4.0
11097-69-1	Aroclor 1254	1	4.0	U	1.6	4.0
11096-82-5	Aroclor 1260	1	4.0	U	0.6	4.0

SURROGATES	ADDED: (ug/kg wet)	FOUND: (ug/kg wet)	% REC	QC LIMITS	Q
Decachlorobiphenyl	8.0000	7.47	93.4	40 - 126	
Tetrachlorometaxylene	8.0000	5.94	74.2	44 - 120	
Decachlorobiphenyl [2C]	8.0000	7.00	87.5	40 - 126	
Tetrachlorometaxylene [2C]	8.0000	6.43	80.3	44 - 120	

[2C] indicates second-column analyte, present if quantification on any batch samples used second column data.

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282333ECD7.D  
Data file 2: /230228.b/230228.b/02282333ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: BLB0427-BLK1  
Client ID:  
Injection Date: 01-MAR-2023 03:08  
Report Date: 03/01/2023 12:20  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

ZB5 Col		ZB35 Col			ZB5	ZB35	RPD	Compound/Flag	
RT	Shift	Response	RT	Shift	Response	on col			on col
5.807	-0.000	348911	5.686	-0.001	146228	29.7	32.1	7.9	Tetrachloro-m-xylene
13.892	-0.001	436465	14.118	-0.002	211607	37.4	35.0	6.6	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	787062	16.8
Hexabromobiphenyl	1429847	1186214	-17.0

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	310177	-1.6
Hexabromobiphenyl	513946	397133	-22.7

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	---			0.0	1	---			0.0
Aroclor-1016	2	---			0.0	2	---			0.0
Aroclor-1016	3	---			0.0	3	---			0.0
Aroclor-1016	4	---			0.0	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1221	1	---			0.0	1	---			0.0
Aroclor-1221	2	---			0.0	2	---			0.0
Aroclor-1221	3	---			0.0	3	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1232	1	---			0.0	1	---			0.0
Aroclor-1232	2	---			0.0	2	---			0.0
Aroclor-1232	3	---			0.0	3	---			0.0
Aroclor-1232	4	---			0.0	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1242	1	---			0.0	1	---			0.0
Aroclor-1242	2	---			0.0	2	---			0.0
Aroclor-1242	3	---			0.0	3	---			0.0
Aroclor-1242	4	---			0.0	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1248	1	---			0.0	1	---			0.0
Aroclor-1248	2	---			0.0	2	---			0.0
Aroclor-1248	3	---			0.0	3	---			0.0
Aroclor-1248	4	---			0.0	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1254	1	---			0.0	1	---			0.0
Aroclor-1254	2	---			0.0	2	---			0.0
Aroclor-1254	3	---			0.0	3	---			0.0
Aroclor-1254	4	---			0.0	4	---			0.0
Aroclor-1254	5	---			0.0	5	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1260	1	---			0.0	1	---			0.0
Aroclor-1260	2	---			0.0	2	---			0.0
Aroclor-1260	3	---			0.0	3	---			0.0
Aroclor-1260	4	---			0.0	4	---			0.0
Aroclor-1260	5	---			0.0	NS	---			----
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1262	1	---			0.0	1	---			0.0
Aroclor-1262	2	---			0.0	2	---			0.0
Aroclor-1262	3	---			0.0	3	---			0.0
Aroclor-1262	4	---			0.0	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1268	1	---			0.0	1	---			0.0
Aroclor-1268	2	---			0.0	2	---			0.0
Aroclor-1268	3	---			0.0	3	---			0.0
Aroclor-1268	4	---			0.0	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				

Total PCB Area Coll (5.907 - 13.793) = 469796

Coll Total PCB = 0.0 ppm\*

Total PCB Area Col2 (5.787 - 14.020) = 15554 Col2 Total PCB = 0.0 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 BLB0427-BLK1

01-MAR-2023 03:08, 2u1





**LCS / LCS DUPLICATE RECOVERY**  
**EPA 8082A**

Laboratory:	<u>Analytical Resources, LLC</u>	SDG:	<u>23B0228</u>
Client:	<u>Anchor QEA, LLC</u>	Project:	<u>AOC5 MR Phase 1</u>
Matrix:	<u>Solid</u>	Analyzed:	<u>03/01/23 03:29</u>
Batch:	<u>BLB0427</u>	Laboratory ID:	<u>BLB0427-BS1</u>
Preparation:	<u>EPA 3546 (Microwave)</u>	Sequence Name:	<u>LCS</u>
Initial/Final:	<u>12.5 g / 2.5 mL</u>		

COMPOUND	SPIKE ADDED (ug/kg wet)	LCS CONCENTRATION (ug/kg wet)	Q	LCS % REC. #	QC LIMITS REC.
Aroclor 1016	101	75.7		75.1	56 - 120
Aroclor 1260	101	107		106	58 - 120

\* Indicates values outside of QC limits

COMPOUND	SPIKE ADDED (ug/kg wet)	LCSD CONCENTRATION (ug/kg wet)	Q	LCSD % REC. #	% RPD #	QC LIMITS	
						RPD	REC.
Aroclor 1016	101	81.1		80.4	6.87	30	56 - 120
Aroclor 1260	101	111		110	3.89	30	58 - 120

\* Indicates values outside of QC limits

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282334ECD7.D  
Data file 2: /230228.b/230228.b/02282334ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: BLB0427-BS1  
Client ID:  
Injection Date: 01-MAR-2023 03:29  
Report Date: 03/01/2023 12:20  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.808	0.001	404782	5.687	0.000	160749	32.3	32.0	1.0	Tetrachloro-m-xylene
13.893	-0.000	477901	14.119	-0.001	245652	38.9	38.0	2.2	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	839905	24.7
Hexabromobiphenyl	1429847	1248668	-12.7

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	342893	8.8
Hexabromobiphenyl	513946	424424	-17.4

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	7.268	-0.001	117834	369.4	1	7.254	-0.001	69682	347.1
Aroclor-1016	2	7.652	-0.003	385165	396.1	2	7.852	-0.008	158651	389.8
Aroclor-1016	3	7.788	-0.004	166450	350.6	3	8.052	-0.006	65280	355.1
Aroclor-1016	4	8.402	-0.004	122029	397.6	4	8.304	-0.003	50842	352.5
Total CollAve (4 peaks):				378.4		Total Col2Ave (4 peaks):				361.1 RPD = 5
Corrected Ave (3 peaks):				372.0		Corrected Ave (3 peaks):				351.6 RPD = 6
Aroclor-1221	1	4.732	0.001	529	7.0	1	4.958	0.002	244	7.5
Aroclor-1221	2	6.130	-0.002	15172	112.8	2	6.299	0.002	7015	114.2
Aroclor-1221	3	6.382	-0.000	74309	237.9	3	6.622	0.000	31997	319.9
Total CollAve (3 peaks):				119.2		Total Col2Ave (3 peaks):				147.2 RPD = 21
Corrected Ave: < 3 Peaks						Corrected Ave: < 3 Peaks				
Aroclor-1232	1	4.732	0.002	529	11.8	1	4.958	0.002	244	13.9
Aroclor-1232	2	6.130	-0.001	15172	170.0	2	7.254	0.000	69682	799.3
Aroclor-1232	3	7.652	-0.004	385165	953.5	3	7.852	-0.009	158651	910.1
Aroclor-1232	4	8.575	-0.006	156341	910.6	4	8.711	-0.004	50097	998.6
Total CollAve (4 peaks):				511.5		Total Col2Ave (4 peaks):				680.5 RPD = 28
Corrected Ave (3 peaks):				364.1		Corrected Ave (3 peaks):				574.4 RPD = 45*
Aroclor-1242	1	7.268	-0.000	117834	452.7	1	7.254	0.001	69682	437.4
Aroclor-1242	2	7.652	-0.004	385165	487.3	2	7.852	-0.008	158651	473.7
Aroclor-1242	3	8.402	-0.003	122029	496.2	3	9.156	-0.012	8789	84.3
Aroclor-1242	4	8.575	-0.005	156341	430.0	4	9.581	-0.017	3378	26.6
Total CollAve (4 peaks):				466.5		Total Col2Ave (4 peaks):				255.5 RPD = 58*
Corrected Ave (3 peaks):				456.7		Corrected Ave (3 peaks):				182.8 RPD = 86*
Aroclor-1248	1	8.402	-0.004	122029	297.8	1	8.304	-0.003	50842	310.5
Aroclor-1248	2	8.575	-0.005	156341	300.1	2	8.711	-0.004	50097	295.9
Aroclor-1248	3	8.991	-0.006	161664	164.5	3	9.156	-0.013	8789	45.1
Aroclor-1248	4	9.297	0.002	136914	273.7	4	9.581	-0.014	3378	14.4
Total CollAve (4 peaks):				259.0		Total Col2Ave (4 peaks):				166.5 RPD = 43*
Corrected Ave (3 peaks):				245.3		Corrected Ave (3 peaks):				118.5 RPD = 70*
Aroclor-1254	1	9.297	-0.002	136914	162.3	1	9.447	-0.004	44947	172.5
Aroclor-1254	2	---	---	---	0.0	2	9.968	-0.004	9788	46.7
Aroclor-1254	3	9.662	-0.007	22976	42.4	3	10.143	0.018	96970	213.8
Aroclor-1254	4	9.802	-0.006	76361	72.4	4	10.368	-0.006	120137	271.7
Aroclor-1254	5	10.117	-0.061	346429	524.2	5	10.564	-0.006	169976	631.3
Total CollAve (4 peaks):				200.3		Total Col2Ave (5 peaks):				267.2 RPD = 29
Corrected Ave (3 peaks):				92.4		Corrected Ave (4 peaks):				176.1 RPD = 62*
Aroclor-1260	1	11.043	-0.001	244288	543.8	1	11.651	-0.002	118072	473.1
Aroclor-1260	2	11.359	-0.002	257717	549.1	2	11.915	-0.003	308961	485.1
Aroclor-1260	3	11.733	-0.003	636560	511.4	3	12.434	-0.002	80187	474.4
Aroclor-1260	4	12.136	-0.003	335678	535.5	4	12.500	-0.002	202823	472.4
Aroclor-1260	5	12.242	-0.001	141048	522.8	NS	---	---	---	---
Total CollAve (5 peaks):				532.5		Total Col2Ave (4 peaks):				476.3 RPD = 11
Corrected Ave (4 peaks):				528.4		Corrected Ave (3 peaks):				473.3 RPD = 11
Aroclor-1262	1	10.823	-0.006	521402	1361.2	1	11.197	-0.003	115715	319.3
Aroclor-1262	2	12.242	-0.002	141048	226.3	2	11.651	-0.001	118072	382.5
Aroclor-1262	3	12.317	-0.001	173260	258.6	3	12.434	-0.000	80187	229.0
Aroclor-1262	4	12.986	-0.001	171201	279.6	4	12.500	-0.002	202823	369.7
Total CollAve (4 peaks):				531.4		Total Col2Ave (4 peaks):				325.1 RPD = 48*
Corrected Ave (3 peaks):				254.8		Corrected Ave (3 peaks):				306.0 RPD = 18
Aroclor-1268	1	12.242	-0.004	141048	88.2	1	12.434	0.002	80187	93.8
Aroclor-1268	2	12.317	0.000	173260	109.4	2	12.500	-0.000	202823	220.8
Aroclor-1268	3	12.723	0.023	86503	63.8	3	12.891	-0.001	6252	8.0
Aroclor-1268	4	13.488	-0.002	57960	13.0	4	13.708	-0.001	21822	8.7
Total CollAve (4 peaks):				68.6		Total Col2Ave (4 peaks):				82.8 RPD = 19



Corrected Ave (3 peaks): 55.0      Corrected Ave (3 peaks): 36.8      RPD = 40

Total PCB Area Col1 (5.907 - 13.793) = 7400705      Col1 Total PCB = 0.7 ppm\*

Total PCB Area Col2 (5.787 - 14.020) = 2935143      Col2 Total PCB = 0.7 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 BLB0427-BS1


01-MAR-2023 03:29, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 BLB0427-BS1

01-MAR-2023 03:29, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282335ECD7.D  
Data file 2: /230228.b/230228.b/02282335ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: BLB0427-BSD1  
Client ID:  
Injection Date: 01-MAR-2023 03:50  
Report Date: 03/01/2023 12:20  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.808	0.000	437425	5.686	-0.001	170014	32.2	31.2	3.0	Tetrachloro-m-xylene
13.892	-0.001	513138	14.118	-0.002	264364	36.8	36.2	1.6	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	910442	35.1
Hexabromobiphenyl	1429847	1416430	-0.9

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	371249	17.8
Hexabromobiphenyl	513946	479766	-6.7

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	7.269	-0.001	138934	401.8	1	7.254	-0.001	82187	378.2
Aroclor-1016	2	7.651	-0.004	445586	422.7	2	7.853	-0.007	185014	419.8
Aroclor-1016	3	7.788	-0.003	191856	372.8	3	8.052	-0.007	75979	381.7
Aroclor-1016	4	8.403	-0.003	141085	424.1	4	8.305	-0.002	58633	375.5
Total CollAve (4 peaks):				405.3		Total Col2Ave (4 peaks):				388.8 RPD = 4
Corrected Ave (3 peaks):				399.1		Corrected Ave (3 peaks):				378.4 RPD = 5
Aroclor-1221	1	4.731	0.001	754	9.2	1	---			0.0
Aroclor-1221	2	6.130	-0.002	17380	119.2	2	6.298	0.001	8161	122.7
Aroclor-1221	3	6.382	-0.001	85973	253.9	3	6.621	-0.000	35919	331.7
Total CollAve (3 peaks):				127.5		Col2Ave: <3 Quant Peaks				
Aroclor-1232	1	4.731	0.001	754	15.5	1	---			0.0
Aroclor-1232	2	6.130	-0.001	17380	179.7	2	7.254	-0.000	82187	870.7
Aroclor-1232	3	7.651	-0.005	445586	1017.6	3	7.853	-0.008	185014	980.3
Aroclor-1232	4	8.575	-0.005	179920	966.7	4	8.712	-0.003	56683	1043.5
Total CollAve (4 peaks):				544.9		Total Col2Ave (3 peaks):				964.8 RPD = 56*
Corrected Ave (3 peaks):				387.3		Corrected Ave: < 3 Peaks				
Aroclor-1242	1	7.269	0.000	138934	492.4	1	7.254	0.001	82187	476.5
Aroclor-1242	2	7.651	-0.004	445586	520.0	2	7.853	-0.007	185014	510.3
Aroclor-1242	3	8.403	-0.002	141085	529.2	3	9.158	-0.009	9013	79.9
Aroclor-1242	4	8.575	-0.005	179920	456.5	4	9.582	-0.016	3817	27.8
Total CollAve (4 peaks):				499.6		Total Col2Ave (4 peaks):				273.6 RPD = 58*
Corrected Ave (3 peaks):				489.7		Corrected Ave (3 peaks):				194.7 RPD = 86*
Aroclor-1248	1	8.403	-0.003	141085	317.6	1	8.305	-0.002	58633	330.8
Aroclor-1248	2	8.575	-0.005	179920	318.6	2	8.712	-0.003	56683	309.3
Aroclor-1248	3	8.992	-0.005	185680	174.3	3	9.158	-0.010	9013	42.7
Aroclor-1248	4	9.297	0.002	160144	295.3	4	9.582	-0.012	3817	15.1
Total CollAve (4 peaks):				276.5		Total Col2Ave (4 peaks):				174.5 RPD = 45*
Corrected Ave (3 peaks):				262.4		Corrected Ave (3 peaks):				122.4 RPD = 73*
Aroclor-1254	1	9.297	-0.002	160144	175.2	1	9.447	-0.004	51612	182.9
Aroclor-1254	2	---			0.0	2	9.967	-0.004	11332	49.9
Aroclor-1254	3	9.662	-0.007	24190	41.2	3	10.143	0.017	113266	230.6
Aroclor-1254	4	9.801	-0.007	87207	76.3	4	10.368	-0.006	139864	292.1
Aroclor-1254	5	10.117	-0.061	402080	561.3	5	10.564	-0.005	196760	674.9
Total CollAve (4 peaks):				213.5		Total Col2Ave (5 peaks):				286.1 RPD = 29
Corrected Ave (3 peaks):				97.5		Corrected Ave (4 peaks):				188.9 RPD = 64*
Aroclor-1260	1	11.042	-0.002	289833	568.8	1	11.650	-0.003	140639	498.5
Aroclor-1260	2	11.359	-0.002	309392	581.1	2	11.916	-0.002	346725	481.6
Aroclor-1260	3	11.734	-0.002	766204	542.6	3	12.434	-0.001	96676	506.0
Aroclor-1260	4	12.136	-0.004	389548	547.9	4	12.500	-0.002	240427	495.4
Aroclor-1260	5	12.242	-0.001	161521	527.7	NS	---			----
Total CollAve (5 peaks):				553.6		Total Col2Ave (4 peaks):				495.4 RPD = 11
Corrected Ave (4 peaks):				546.8		Corrected Ave (3 peaks):				491.8 RPD = 11
Aroclor-1262	1	10.823	-0.006	622450	1432.5	1	11.197	-0.003	136288	332.7
Aroclor-1262	2	12.242	-0.002	161521	228.4	2	11.650	-0.001	140639	403.1
Aroclor-1262	3	12.318	-0.001	197825	260.3	3	12.434	0.000	96676	244.2
Aroclor-1262	4	12.986	-0.001	204047	293.7	4	12.500	-0.002	240427	387.7
Total CollAve (4 peaks):				553.7		Total Col2Ave (4 peaks):				341.9 RPD = 47*
Corrected Ave (3 peaks):				260.8		Corrected Ave (3 peaks):				321.5 RPD = 21
Aroclor-1268	1	12.242	-0.005	161521	89.0	1	12.434	0.002	96676	100.1
Aroclor-1268	2	12.318	0.001	197825	110.1	2	12.500	-0.000	240427	231.5
Aroclor-1268	3	12.721	0.021	99538	64.7	3	12.890	-0.002	7925	8.9
Aroclor-1268	4	13.486	-0.004	70605	13.9	4	13.707	-0.002	26402	9.3
Total CollAve (4 peaks):				69.4		Total Col2Ave (4 peaks):				87.5 RPD = 23
Corrected Ave (3 peaks):				55.9		Corrected Ave (3 peaks):				39.4 RPD = 35

Total PCB Area Col1 (5.907 - 13.793) = 8641117 Col1 Total PCB = 0.8 ppm\*

Total PCB Area Col2 (5.787 - 14.020) = 3435750 Col2 Total PCB = 0.8 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 BLB0427-BSD1


01-MAR-2023 03:50, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 BLB0427-BSD1

01-MAR-2023 03:50, 2ul





## STANDARD REFERENCE MATERIAL RECOVERY

### EPA 8082A

**Laboratory:** Analytical Resources, LLC

**SDG:** 23B0228

**Client:** Anchor QEA, LLC

**Project:** AOC5 MR Phase 1

**Matrix:** Solid

**Laboratory ID:** BLB0427-SRM1

**Batch:** BLB0427

**Initial/Final:** 2.5 g / 2.5 mL

**Preparation:** EPA 3546 (Microwave)

**Analyzed:** 03/01/2023 4:11

**Standard ID:** K011478

**Expires:** 06/11/2023

**Standard Lot#:** PSRM0169

**Description:** Puget Sound reference-SRM

ANALYTE	TRUE (ug/kg wet)	FOUND (ug/kg wet)	MDL	MRL	Q	SRM % REC.	QC LIMITS REC.
Aroclor 1260	108.00	133	2.9	20.0		123	38 - 167
Aroclor 1260 [2C]	108.00	131	2.9	20.0		122	38 - 167

\* Values outside of QC limits

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282336ECD7.D  
Data file 2: /230228.b/230228.b/02282336ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: BLB0427-SRM1  
Client ID:  
Injection Date: 01-MAR-2023 04:11  
Report Date: 03/01/2023 12:20  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

ZB5 Col		ZB35 Col			ZB5	ZB35	RPD	Compound/Flag	
RT	Shift	Response	RT	Shift	Response	on col			on col
5.806	-0.001	353977	5.685	-0.003	151921	28.0	30.3	7.9	Tetrachloro-m-xylene
13.887	-0.006	375307	14.115	-0.005	212554	33.5	32.8	1.8	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	845866	25.5
Hexabromobiphenyl	1429847	1138921	-20.3

Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	341218	8.2
Hexabromobiphenyl	513946	424905	-17.3

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)



ZB5 Col					ZB35 Col						
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.237	-0.032	25337	78.9	1	7.263	0.008	8514	42.6	
Aroclor-1016	2	7.649	-0.007	14426	14.7	2	7.849	-0.010	6346	15.7	
Aroclor-1016	3	7.797	0.005	10049	21.0	3	8.048	-0.010	2618	14.3	
Aroclor-1016	4	8.400	-0.005	13034	42.2	4	8.302	-0.006	7249	50.5	
Total CollAve (4 peaks):				39.2	Total Col2Ave (4 peaks):				30.8	RPD = 24	
Corrected Ave (3 peaks):				26.0	Corrected Ave (3 peaks):				24.2	RPD = 7	
Aroclor-1221	1	4.781	0.051	1950	25.7	1	4.940	-0.016	603	18.7	
Aroclor-1221	2	6.157	0.025	1720	12.7	2	6.346	0.049	13920	227.7	
Aroclor-1221	3	6.391	0.009	4014	12.8	3	6.638	0.016	5361	53.9	
Total CollAve (3 peaks):				17.1	Total Col2Ave (3 peaks):				100.1	RPD = 142*	
Corrected Ave: < 3 Peaks					Corrected Ave: < 3 Peaks						
Aroclor-1232	1	4.781	0.051	1950	43.0	1	4.940	-0.016	603	34.5	
Aroclor-1232	2	6.157	0.026	1720	19.1	2	7.263	0.009	8514	98.1	
Aroclor-1232	3	7.649	-0.007	14426	35.5	3	7.849	-0.011	6346	36.6	
Aroclor-1232	4	8.570	-0.010	11647	67.4	4	8.709	-0.006	5502	110.2	
Total CollAve (4 peaks):				41.2	Total Col2Ave (4 peaks):				69.9	RPD = 52*	
Corrected Ave (3 peaks):				32.5	Corrected Ave (3 peaks):				56.4	RPD = 54*	
Aroclor-1242	1	7.237	-0.031	25337	96.7	1	7.263	0.009	8514	53.7	
Aroclor-1242	2	7.649	-0.007	14426	18.1	2	7.849	-0.010	6346	19.0	
Aroclor-1242	3	8.400	-0.005	13034	52.6	3	9.149	-0.019	7148	68.9	
Aroclor-1242	4	8.570	-0.010	11647	31.8	4	9.537	-0.061	12325	97.5	
Total CollAve (4 peaks):				49.8	Total Col2Ave (4 peaks):				59.8	RPD = 18	
Corrected Ave (3 peaks):				34.2	Corrected Ave (3 peaks):				47.2	RPD = 32	
Aroclor-1248	1	8.400	-0.006	13034	31.6	1	8.302	-0.005	7249	44.5	
Aroclor-1248	2	8.570	-0.010	11647	22.2	2	8.709	-0.006	5502	32.7	
Aroclor-1248	3	8.989	-0.007	35763	36.1	3	9.149	-0.020	7148	36.9	
Aroclor-1248	4	9.291	-0.004	47809	94.9	4	9.537	-0.057	12325	53.0	
Total CollAve (4 peaks):				46.2	Total Col2Ave (4 peaks):				41.7	RPD = 10	
Corrected Ave (3 peaks):				30.0	Corrected Ave (3 peaks):				38.0	RPD = 24	
Aroclor-1254	1	9.291	-0.008	47809	56.3	1	9.442	-0.009	19190	74.0	
Aroclor-1254	2	9.366	-0.012	18377	48.1	2	9.961	-0.011	9483	45.5	
Aroclor-1254	3	9.663	-0.006	23658	43.3	3	10.114	-0.011	37173	82.3	
Aroclor-1254	4	9.793	-0.015	65690	61.9	4	10.362	-0.013	47579	108.1	
Aroclor-1254	5	10.113	-0.065	107857	162.1	5	10.558	-0.012	48551	181.2	
Total CollAve (5 peaks):				74.3	Total Col2Ave (5 peaks):				98.2	RPD = 28	
Corrected Ave (4 peaks):				52.4	Corrected Ave (4 peaks):				77.5	RPD = 39	
Aroclor-1260	1	11.036	-0.008	61703	150.6	1	11.646	-0.007	34888	139.6	
Aroclor-1260	2	11.350	-0.011	51150	119.5	2	11.908	-0.010	76125	119.4	
Aroclor-1260	3	11.723	-0.013	158186	139.3	3	12.426	-0.009	25334	149.7	
Aroclor-1260	4	12.124	-0.015	77396	135.4	4	12.492	-0.010	50204	116.8	
Aroclor-1260	5	12.236	-0.007	29601	120.3	NS	---			----	
Total CollAve (5 peaks):				133.0	Total Col2Ave (4 peaks):				131.4	RPD = 1	
Corrected Ave (4 peaks):				128.6	Corrected Ave (3 peaks):				125.3	RPD = 3	
Aroclor-1262	1	10.813	-0.016	141926	406.2	1	11.192	-0.008	30384	83.7	
Aroclor-1262	2	12.236	-0.008	29601	52.1	2	11.646	-0.006	34888	112.9	
Aroclor-1262	3	12.311	-0.008	36981	60.5	3	12.426	-0.008	25334	72.3	
Aroclor-1262	4	12.978	-0.009	37424	67.0	4	12.492	-0.010	50204	91.4	
Total CollAve (4 peaks):				146.4	Total Col2Ave (4 peaks):				90.1	RPD = 48*	
Corrected Ave (3 peaks):				59.9	Corrected Ave (3 peaks):				82.5	RPD = 32	
Aroclor-1268	1	12.236	-0.011	29601	20.3	1	12.426	-0.006	25334	29.6	
Aroclor-1268	2	12.311	-0.006	36981	25.6	2	12.492	-0.008	50204	54.6	
Aroclor-1268	3	12.714	0.015	17661	14.3	3	12.889	-0.003	1845	2.3	
Aroclor-1268	4	13.481	-0.009	10629	2.6	4	13.701	-0.007	7710	3.1	
Total CollAve (4 peaks):				15.7	Total Col2Ave (4 peaks):				22.4	RPD = 35	

Corrected Ave (3 peaks): 12.4      Corrected Ave (3 peaks): 11.7      RPD = 6

Total PCB Area Col1 (5.907 - 13.793) = 1886614      Col1 Total PCB = 0.2 ppm\*

Total PCB Area Col2 (5.787 - 14.020) = 898081      Col2 Total PCB = 0.2 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 BLB0427-SRM1

01-MAR-2023 04:11, 2u1











**INITIAL CALIBRATION DATA**  
**EPA 8082A**

Laboratory:	Analytical Resources, LLC	SDG:	23B0228
Client:	Anchor QEA, LLC	Project:	AOC5 MR Phase 1
Calibration:	GB00069	Instrument:	ECD7
Calibration Date:	02/24/2023	Column (1):	ZB5

COMPOUND	Mean RRF	RRF RSD	Linear COD	Quad COD	Limit Type & Limit	Q
Aroclor 1016	4.936617E-02	5.3			RSD (20)	
Aroclor-1016 (1)	3.038517E-02	7.4			RSD (20)	
Aroclor-1016 (2)	9.263078E-02	3.5			RSD (20)	
Aroclor-1016 (3)	0.045218	12.9			RSD (20)	
Aroclor-1016 (4)	2.923074E-02	3.5			RSD (20)	
Aroclor 1221		0.0			RSD (20)	
Aroclor-1221 (1)		0.0			RSD (20)	
Aroclor-1221 (2)		0.0			RSD (20)	
Aroclor-1221 (3)		0.0			RSD (20)	
Aroclor 1232		0.0			RSD (20)	
Aroclor-1232 (1)		0.0			RSD (20)	
Aroclor-1232 (2)		0.0			RSD (20)	
Aroclor-1232 (3)		0.0			RSD (20)	
Aroclor-1232 (4)		0.0			RSD (20)	
Aroclor 1242		0.0			RSD (20)	
Aroclor-1242 (1)		0.0			RSD (20)	
Aroclor-1242 (2)		0.0			RSD (20)	
Aroclor-1242 (3)		0.0			RSD (20)	
Aroclor-1242 (4)		0.0			RSD (20)	
Aroclor 1248		0.0			RSD (20)	
Aroclor-1248 (1)		0.0			RSD (20)	
Aroclor-1248 (2)		0.0			RSD (20)	
Aroclor-1248 (3)		0.0			RSD (20)	
Aroclor-1248 (4)		0.0			RSD (20)	
Aroclor 1254		0.0			RSD (20)	
Aroclor-1254 (1)		0.0			RSD (20)	
Aroclor-1254 (2)		0.0			RSD (20)	
Aroclor-1254 (3)		0.0			RSD (20)	
Aroclor-1254 (4)		0.0			RSD (20)	
Aroclor-1254 (5)		0.0			RSD (20)	
Aroclor 1260	3.920913E-02	4.7			RSD (20)	



**INITIAL CALIBRATION DATA**  
**EPA 8082A**

Laboratory:	Analytical Resources, LLC	SDG:	23B0228
Client:	Anchor QEA, LLC	Project:	AOC5 MR Phase 1
Calibration:	GB00069	Instrument:	ECD7
Calibration Date:	02/24/2023	Column (1):	ZB5

COMPOUND	Mean RRF	RRF RSD	Linear COD	Quad COD	Limit Type & Limit	Q
Aroclor-1260 (1)	2.877854E-02	4.7			RSD (20)	
Aroclor-1260 (2)	3.006895E-02	5.0			RSD (20)	
Aroclor-1260 (3)	7.975167E-02	4.6			RSD (20)	
Aroclor-1260 (4)	4.015993E-02	5.8			RSD (20)	
Aroclor-1260 (5)	1.728656E-02	5.0			RSD (20)	
Aroclor 1262		0.0			RSD (20)	
Aroclor-1262 (1)		0.0			RSD (20)	
Aroclor-1262 (2)		0.0			RSD (20)	
Aroclor-1262 (3)		0.0			RSD (20)	
Aroclor-1262 (4)		0.0			RSD (20)	
Aroclor 1268		0.0			RSD (20)	
Aroclor-1268 (1)		0.0			RSD (20)	
Aroclor-1268 (2)		0.0			RSD (20)	
Aroclor-1268 (3)		0.0			RSD (20)	
Aroclor-1268 (4)		0.0			RSD (20)	
Decachlorobiphenyl	0.7878687	4.2			RSD (20)	
Tetrachlorometaxylene	1.194488	3.9			RSD (20)	











**INITIAL CALIBRATION DATA**  
**EPA 8082A**

Laboratory:	Analytical Resources, LLC	SDG:	23B0228
Client:	Anchor QEA, LLC	Project:	AOC5 MR Phase 1
Calibration:	GB00069	Instrument:	ECD7
Calibration Date:	02/24/2023	Column (2):	ZB35

COMPOUND	Mean RRF	RRF RSD	Linear COD	Quad COD	Limit Type & Limit	Q
Aroclor 1016 [2C]	5.458565E-02	6.3			RSD (20)	
Aroclor-1016 (1) [2C]	4.683127E-02	8.5			RSD (20)	
Aroclor-1016 (2) [2C]	9.496755E-02	8.0			RSD (20)	
Aroclor-1016 (3) [2C]	4.289222E-02	7.9			RSD (20)	
Aroclor-1016 (4) [2C]	3.365154E-02	10.3			RSD (20)	
Aroclor 1221 [2C]		0.0			RSD (20)	
Aroclor-1221 (1) [2C]		0.0			RSD (20)	
Aroclor-1221 (2) [2C]		0.0			RSD (20)	
Aroclor-1221 (3) [2C]		0.0			RSD (20)	
Aroclor 1232 [2C]		0.0			RSD (20)	
Aroclor-1232 (1) [2C]		0.0			RSD (20)	
Aroclor-1232 (2) [2C]		0.0			RSD (20)	
Aroclor-1232 (3) [2C]		0.0			RSD (20)	
Aroclor-1232 (4) [2C]		0.0			RSD (20)	
Aroclor 1242 [2C]		0.0			RSD (20)	
Aroclor-1242 (1) [2C]		0.0			RSD (20)	
Aroclor-1242 (2) [2C]		0.0			RSD (20)	
Aroclor-1242 (3) [2C]		0.0			RSD (20)	
Aroclor-1242 (4) [2C]		0.0			RSD (20)	
Aroclor 1248 [2C]		0.0			RSD (20)	
Aroclor-1248 (1) [2C]		0.0			RSD (20)	
Aroclor-1248 (2) [2C]		0.0			RSD (20)	
Aroclor-1248 (3) [2C]		0.0			RSD (20)	
Aroclor-1248 (4) [2C]		0.0			RSD (20)	
Aroclor 1254 [2C]		0.0			RSD (20)	
Aroclor-1254 (1) [2C]		0.0			RSD (20)	
Aroclor-1254 (2) [2C]		0.0			RSD (20)	
Aroclor-1254 (3) [2C]		0.0			RSD (20)	
Aroclor-1254 (4) [2C]		0.0			RSD (20)	
Aroclor-1254 (5) [2C]		0.0			RSD (20)	
Aroclor 1260 [2C]	6.996878E-02	6.4			RSD (20)	



**INITIAL CALIBRATION DATA**  
**EPA 8082A**

Laboratory:	Analytical Resources, LLC	SDG:	23B0228
Client:	Anchor QEA, LLC	Project:	AOC5 MR Phase 1
Calibration:	GB00069	Instrument:	ECD7
Calibration Date:	02/24/2023	Column (2):	ZB35

COMPOUND	Mean RRF	RRF RSD	Linear COD	Quad COD	Limit Type & Limit	Q
Aroclor-1260 (1) [2C]	4.704064E-02	8.4			RSD (20)	
Aroclor-1260 (2) [2C]	0.1200523	7.6			RSD (20)	
Aroclor-1260 (3) [2C]	3.185902E-02	6.0			RSD (20)	
Aroclor-1260 (4) [2C]	8.092314E-02	5.1			RSD (20)	
Aroclor 1262 [2C]		0.0			RSD (20)	
Aroclor-1262 (1) [2C]		0.0			RSD (20)	
Aroclor-1262 (2) [2C]		0.0			RSD (20)	
Aroclor-1262 (3) [2C]		0.0			RSD (20)	
Aroclor-1262 (4) [2C]		0.0			RSD (20)	
Aroclor 1268 [2C]		0.0			RSD (20)	
Aroclor-1268 (1) [2C]		0.0			RSD (20)	
Aroclor-1268 (2) [2C]		0.0			RSD (20)	
Aroclor-1268 (3) [2C]		0.0			RSD (20)	
Aroclor-1268 (4) [2C]		0.0			RSD (20)	
Decachlorobiphenyl [2C]	1.218271	3.9			RSD (20)	
Tetrachlorometaxylene [2C]	1.173721	3.9			RSD (20)	



## GC LOG SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b

	Inject	Date/Time	Filename	DF	LabID	ClientID
1	24-FEB-2023	10:51	02242301ECD7.D	1	IB	
2	24-FEB-2023	11:12	02242302ECD7.D	1	0.25PPMAR1660	
3	24-FEB-2023	11:33	02242303ECD7.D	1	0.02PPMAR1660	
4	24-FEB-2023	11:54	02242304ECD7.D	1	0.05PPMAR1660	
5	24-FEB-2023	12:15	02242305ECD7.D	1	1.0PPMAR1660	
6	24-FEB-2023	12:36	02242306ECD7.D	1	0.1PPMAR1660	
7	24-FEB-2023	12:57	02242307ECD7.D	1	0.5PPMAR1660	
8	24-FEB-2023	13:18	02242308ECD7.D	1	0.25PPMAR1242	
9	24-FEB-2023	13:39	02242309ECD7.D	1	0.25PPMAR1248	
10	24-FEB-2023	14:00	02242310ECD7.D	1	0.25PPMAR1254	
11	24-FEB-2023	14:21	02242311ECD7.D	1	0.25PPMAR2162	
12	24-FEB-2023	14:42	02242312ECD7.D	1	0.25PPMAR3268	
13	24-FEB-2023	15:03	02242313ECD7.D	1	AR1660SCV	
14	24-FEB-2023	15:24	02242314ECD7.D	1	AR1242SCV	
15	24-FEB-2023	15:45	02242315ECD7.D	1	AR1248SCV	
16	24-FEB-2023	16:06	02242316ECD7.D	1	AR1254SCV	
17	24-FEB-2023	16:27	02242317ECD7.D	1	AR2162SCV	
18	24-FEB-2023	16:48	02242318ECD7.D	1	AR3268SCV	
19	24-FEB-2023	17:09	02242319ECD7.D	1	DDTS	
20	24-FEB-2023	17:30	02242320ECD7.D	1	DDT BD	

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b

ARI Job No.: IB Method: PCB.m Instrument: ecd7.i Date: 24-FEB-2023

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1051	02242301ECD7.D	IB		1	NO MANUAL INTEGRATION
1112	02242302ECD7.D	0.25PPMAR1660		1	NO MANUAL INTEGRATION
1133	02242303ECD7.D	0.02PPMAR1660		1	NO MANUAL INTEGRATION
1154	02242304ECD7.D	0.05PPMAR1660		1	NO MANUAL INTEGRATION
1215	02242305ECD7.D	1.0PPMAR1660		1	NO MANUAL INTEGRATION
1236	02242306ECD7.D	0.1PPMAR1660		1	NO MANUAL INTEGRATION
1257	02242307ECD7.D	0.5PPMAR1660		1	NO MANUAL INTEGRATION
1318	02242308ECD7.D	0.25PPMAR1242		1	NO MANUAL INTEGRATION
1339	02242309ECD7.D	0.25PPMAR1248		1	NO MANUAL INTEGRATION
1400	02242310ECD7.D	0.25PPMAR1254		1	NO MANUAL INTEGRATION
1421	02242311ECD7.D	0.25PPMAR2162		1	NO MANUAL INTEGRATION
1442	02242312ECD7.D	0.25PPMAR3268		1	NO MANUAL INTEGRATION
1503	02242313ECD7.D	AR1660SCV		1	NO MANUAL INTEGRATION
1524	02242314ECD7.D	AR1242SCV		1	NO MANUAL INTEGRATION
1545	02242315ECD7.D	AR1248SCV		1	NO MANUAL INTEGRATION
1606	02242316ECD7.D	AR1254SCV		1	NO MANUAL INTEGRATION
1627	02242317ECD7.D	AR2162SCV		1	NO MANUAL INTEGRATION



MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1648	02242318ECD7.D	AR3268SCV		1	NO MANUAL INTEGRATION
1709	02242319ECD7.D	DDTS		1	NO MANUAL INTEGRATION
1730	02242320ECD7.D	DDT BD		1	NO MANUAL INTEGRATION
1751	02242321ECD7.D			1	NO MANUAL INTEGRATION
1812	02242322ECD7.D			1	NO MANUAL INTEGRATION
1833	02242323ECD7.D			1	NO MANUAL INTEGRATION
1854	02242324ECD7.D			1	NO MANUAL INTEGRATION
1915	02242325ECD7.D			1	NO MANUAL INTEGRATION
1936	02242326ECD7.D			1	NO MANUAL INTEGRATION
1957	02242327ECD7.D			1	NO MANUAL INTEGRATION
2018	02242328ECD7.D			1	NO MANUAL INTEGRATION
2039	02242329ECD7.D			1	NO MANUAL INTEGRATION
2059	02242330ECD7.D			1	NO MANUAL INTEGRATION
2120	02242331ECD7.D			1	NO MANUAL INTEGRATION
2141	02242332ECD7.D			1	NO MANUAL INTEGRATION
2202	02242333ECD7.D			1	NO MANUAL INTEGRATION
2223	02242334ECD7.D			1	NO MANUAL INTEGRATION
2244	02242335ECD7.D			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
2305	02242336ECD7.D			1	NO MANUAL INTEGRATION
2326	02242337ECD7.D			1	NO MANUAL INTEGRATION
2347	02242338ECD7.D			1	NO MANUAL INTEGRATION
0008	02242339ECD7.D			1	NO MANUAL INTEGRATION
0029	02242340ECD7.D			1	NO MANUAL INTEGRATION
0050	02242341ECD7.D			1	NO MANUAL INTEGRATION
0111	02242342ECD7.D			1	NO MANUAL INTEGRATION
0132	02242343ECD7.D			1	NO MANUAL INTEGRATION
0153	02242344ECD7.D			1	NO MANUAL INTEGRATION
0214	02242345ECD7.D			1	NO MANUAL INTEGRATION
0235	02242346ECD7.D			1	NO MANUAL INTEGRATION
0256	02242347ECD7.D			1	NO MANUAL INTEGRATION
0317	02242348ECD7.D			1	NO MANUAL INTEGRATION
0338	02242349ECD7.D			1	NO MANUAL INTEGRATION
0359	02242350ECD7.D			1	NO MANUAL INTEGRATION
0420	02242351ECD7.D			1	NO MANUAL INTEGRATION
0441	02242352ECD7.D			1	NO MANUAL INTEGRATION
0502	02242353ECD7.D			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
0523	02242354ECD7.D			1	NO MANUAL INTEGRATION
0544	02242355ECD7.D			1	NO MANUAL INTEGRATION
0605	02242356ECD7.D			1	NO MANUAL INTEGRATION
0626	02242357ECD7.D			1	NO MANUAL INTEGRATION
0647	02242358ECD7.D			1	NO MANUAL INTEGRATION
0708	02242359ECD7.D			1	NO MANUAL INTEGRATION
0729	02242360ECD7.D			1	NO MANUAL INTEGRATION
0750	02242361ECD7.D			1	NO MANUAL INTEGRATION
0811	02242362ECD7.D			1	NO MANUAL INTEGRATION
0832	02242363ECD7.D			1	NO MANUAL INTEGRATION
0853	02242364ECD7.D			1	NO MANUAL INTEGRATION
0914	02242365ECD7.D			1	NO MANUAL INTEGRATION
0935	02242366ECD7.D			1	NO MANUAL INTEGRATION
0956	02242367ECD7.D			1	NO MANUAL INTEGRATION
1017	02242368ECD7.D			1	NO MANUAL INTEGRATION
1038	02242369ECD7.D			1	NO MANUAL INTEGRATION
1059	02242370ECD7.D			1	NO MANUAL INTEGRATION
1120	02242371ECD7.D			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1141	02242372ECD7.D			1	NO MANUAL INTEGRATION
1202	02242373ECD7.D			1	NO MANUAL INTEGRATION
1223	02242374ECD7.D			1	NO MANUAL INTEGRATION
1244	02242375ECD7.D			1	NO MANUAL INTEGRATION
1305	02242376ECD7.D			1	NO MANUAL INTEGRATION
1326	02242377ECD7.D			1	NO MANUAL INTEGRATION
1347	02242378ECD7.D			1	NO MANUAL INTEGRATION
1408	02242379ECD7.D			1	NO MANUAL INTEGRATION
1429	02242380ECD7.D			1	NO MANUAL INTEGRATION
1450	02242381ECD7.D			1	NO MANUAL INTEGRATION
1511	02242382ECD7.D			1	NO MANUAL INTEGRATION
1532	02242383ECD7.D			1	NO MANUAL INTEGRATION
1553	02242384ECD7.D			1	NO MANUAL INTEGRATION
1051	02242301ECD7.D IB			1	NO MANUAL INTEGRATION
1112	02242302ECD7.D 0.25PPMAR1660			1	NO MANUAL INTEGRATION
1133	02242303ECD7.D 0.02PPMAR1660			1	Aroclor-1016 [2C],
1154	02242304ECD7.D 0.05PPMAR1660			1	NO MANUAL INTEGRATION
1215	02242305ECD7.D 1.0PPMAR1660			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1236	02242306ECD7.D	0.1PPMAR1660		1	NO MANUAL INTEGRATION
1257	02242307ECD7.D	0.5PPMAR1660		1	NO MANUAL INTEGRATION
1318	02242308ECD7.D	0.25PPMAR1242		1	NO MANUAL INTEGRATION
1339	02242309ECD7.D	0.25PPMAR1248		1	NO MANUAL INTEGRATION
1400	02242310ECD7.D	0.25PPMAR1254		1	NO MANUAL INTEGRATION
1421	02242311ECD7.D	0.25PPMAR2162		1	NO MANUAL INTEGRATION
1442	02242312ECD7.D	0.25PPMAR3268		1	NO MANUAL INTEGRATION
1503	02242313ECD7.D	AR1660SCV		1	NO MANUAL INTEGRATION
1524	02242314ECD7.D	AR1242SCV		1	NO MANUAL INTEGRATION
1545	02242315ECD7.D	AR1248SCV		1	NO MANUAL INTEGRATION
1606	02242316ECD7.D	AR1254SCV		1	NO MANUAL INTEGRATION
1627	02242317ECD7.D	AR2162SCV		1	NO MANUAL INTEGRATION
1648	02242318ECD7.D	AR3268SCV		1	NO MANUAL INTEGRATION
1709	02242319ECD7.D	DDTS		1	NO MANUAL INTEGRATION
1730	02242320ECD7.D	DDT BD		1	NO MANUAL INTEGRATION
1751	02242321ECD7.D			1	NO MANUAL INTEGRATION
1812	02242322ECD7.D			1	NO MANUAL INTEGRATION
1833	02242323ECD7.D			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1854	02242324ECD7.D			1	NO MANUAL INTEGRATION
1915	02242325ECD7.D			1	NO MANUAL INTEGRATION
1936	02242326ECD7.D			1	NO MANUAL INTEGRATION
1957	02242327ECD7.D			1	NO MANUAL INTEGRATION
2018	02242328ECD7.D			1	NO MANUAL INTEGRATION
2038	02242329ECD7.D			1	NO MANUAL INTEGRATION
2059	02242330ECD7.D			1	NO MANUAL INTEGRATION
2120	02242331ECD7.D			1	NO MANUAL INTEGRATION
2141	02242332ECD7.D			1	NO MANUAL INTEGRATION
2202	02242333ECD7.D			1	NO MANUAL INTEGRATION
2223	02242334ECD7.D			1	NO MANUAL INTEGRATION
2244	02242335ECD7.D			1	NO MANUAL INTEGRATION
2305	02242336ECD7.D			1	NO MANUAL INTEGRATION
2326	02242337ECD7.D			1	NO MANUAL INTEGRATION
2347	02242338ECD7.D			1	NO MANUAL INTEGRATION
0008	02242339ECD7.D			1	NO MANUAL INTEGRATION
0029	02242340ECD7.D			1	NO MANUAL INTEGRATION
0050	02242341ECD7.D			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
0111	02242342ECD7.D			1	NO MANUAL INTEGRATION
0132	02242343ECD7.D			1	NO MANUAL INTEGRATION
0153	02242344ECD7.D			1	NO MANUAL INTEGRATION
0214	02242345ECD7.D			1	NO MANUAL INTEGRATION
0235	02242346ECD7.D			1	NO MANUAL INTEGRATION
0256	02242347ECD7.D			1	NO MANUAL INTEGRATION
0317	02242348ECD7.D			1	NO MANUAL INTEGRATION
0338	02242349ECD7.D			1	NO MANUAL INTEGRATION
0359	02242350ECD7.D			1	NO MANUAL INTEGRATION
0420	02242351ECD7.D			1	NO MANUAL INTEGRATION
0441	02242352ECD7.D			1	NO MANUAL INTEGRATION
0502	02242353ECD7.D			1	NO MANUAL INTEGRATION
0523	02242354ECD7.D			1	NO MANUAL INTEGRATION
0544	02242355ECD7.D			1	NO MANUAL INTEGRATION
0605	02242356ECD7.D			1	NO MANUAL INTEGRATION
0626	02242357ECD7.D			1	NO MANUAL INTEGRATION
0647	02242358ECD7.D			1	NO MANUAL INTEGRATION
0708	02242359ECD7.D			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
0729	02242360ECD7.D			1	NO MANUAL INTEGRATION
0750	02242361ECD7.D			1	NO MANUAL INTEGRATION
0811	02242362ECD7.D			1	NO MANUAL INTEGRATION
0832	02242363ECD7.D			1	NO MANUAL INTEGRATION
0853	02242364ECD7.D			1	NO MANUAL INTEGRATION
0914	02242365ECD7.D			1	NO MANUAL INTEGRATION
0935	02242366ECD7.D			1	NO MANUAL INTEGRATION
0956	02242367ECD7.D			1	NO MANUAL INTEGRATION
1017	02242368ECD7.D			1	NO MANUAL INTEGRATION
1038	02242369ECD7.D			1	NO MANUAL INTEGRATION
1059	02242370ECD7.D			1	NO MANUAL INTEGRATION
1120	02242371ECD7.D			1	NO MANUAL INTEGRATION
1141	02242372ECD7.D			1	NO MANUAL INTEGRATION
1202	02242373ECD7.D			1	NO MANUAL INTEGRATION
1223	02242374ECD7.D			1	NO MANUAL INTEGRATION
1244	02242375ECD7.D			1	NO MANUAL INTEGRATION
1305	02242376ECD7.D			1	NO MANUAL INTEGRATION
1326	02242377ECD7.D			1	NO MANUAL INTEGRATION



MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1347	02242378ECD7.D			1	NO MANUAL INTEGRATION
1408	02242379ECD7.D			1	NO MANUAL INTEGRATION
1429	02242380ECD7.D			1	NO MANUAL INTEGRATION
1450	02242381ECD7.D			1	NO MANUAL INTEGRATION
1511	02242382ECD7.D			1	NO MANUAL INTEGRATION
1532	02242383ECD7.D			1	NO MANUAL INTEGRATION
1553	02242384ECD7.D			1	NO MANUAL INTEGRATION

Security Status Report

Date: 28-Feb-2023 09:27

02242301ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242302ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242303ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242304ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242305ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242306ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242307ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242308ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242309ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242310ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242311ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242312ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242313ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242314ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242315ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242316ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242317ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242318ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242319ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27
02242320ECD7.D	Data Locked	richardl, 28-Feb-2023 09:27

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m  
 Last Edit : 24-Feb-2023 15:31 richardl  
 Curve Type : Average

Calibration File Names:

Level 1: \\target\share\chem4\ecd7.i\230224.b\02242303ECD7.D  
 Level 2: \\target\share\chem4\ecd7.i\230224.b\02242304ECD7.D  
 Level 3: \\target\share\chem4\ecd7.i\230224.b\02242306ECD7.D  
 Level 4: \\target\share\chem4\ecd7.i\230224.b\02242302ECD7.D  
 Level 5: \\target\share\chem4\ecd7.i\230224.b\02242307ECD7.D  
 Level 6: \\target\share\chem4\ecd7.i\230224.b\02242305ECD7.D  
 Level 7: \\target\share\chem4\ecd7.i\230224.b\02242312ECD7.D

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
2 Aroclor-1221 (1)	+++++ 0.00716	+++++	+++++	+++++	+++++	+++++	0.00716	0.000
(2)	+++++ 0.01281	+++++	+++++	+++++	+++++	+++++	0.01281	0.000
(3)	+++++ 0.02975	+++++	+++++	+++++	+++++	+++++	0.02975	0.000
3 Aroclor-1242 (1)	+++++ 0.02479	+++++	+++++	+++++	+++++	+++++	0.02479	0.000
(2)	+++++ 0.07529	+++++	+++++	+++++	+++++	+++++	0.07529	0.000
(3)	+++++ 0.02343	+++++	+++++	+++++	+++++	+++++	0.02343	0.000
(4)	+++++ 0.03463	+++++	+++++	+++++	+++++	+++++	0.03463	0.000
4 Aroclor-1232 (1)	+++++ 0.00429	+++++	+++++	+++++	+++++	+++++	0.00429	0.000

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m  
 Last Edit : 24-Feb-2023 15:31 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
(2)	++++ 0.00850	++++	++++	++++	++++	++++	0.00850	0.000
(3)	++++ 0.03848	++++	++++	++++	++++	++++	0.03848	0.000
(4)	++++ 0.01635	++++	++++	++++	++++	++++	0.01635	0.000
7 Aroclor-1016(1)	0.03172 ++++	0.03253	0.03142	0.03141	0.02856	0.02667	0.03039	7.449
(2)	0.09239 ++++	0.09246	0.09222	0.09849	0.09174	0.08849	0.09263	3.499
(3)	0.05165 ++++	0.05037	0.04823	0.04393	0.03991	0.03721	0.04522	12.936
(4)	0.03002 ++++	0.02894	0.02959	0.03058	0.02852	0.02774	0.02923	3.542
6 Aroclor-1248(1)	++++ 0.03903	++++	++++	++++	++++	++++	0.03903	0.000
(2)	++++ 0.04961	++++	++++	++++	++++	++++	0.04961	0.000
(3)	++++ 0.09360	++++	++++	++++	++++	++++	0.09360	0.000
(4)	++++ 0.04765	++++	++++	++++	++++	++++	0.04765	0.000

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m  
 Last Edit : 24-Feb-2023 15:31 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
8 Aroclor-1254 (1)	++++ 0.08033	++++	++++	++++	++++	++++	0.08033	0.000
(2)	++++ 0.03613	++++	++++	++++	++++	++++	0.03613	0.000
(3)	++++ 0.05165	++++	++++	++++	++++	++++	0.05165	0.000
(4)	++++ 0.10042	++++	++++	++++	++++	++++	0.10042	0.000
(5)	++++ 0.06294	++++	++++	++++	++++	++++	0.06294	0.000
9 Aroclor-1260 (1)	0.02926 ++++	0.02920	0.02841	0.03096	0.02737	0.02746	0.02878	4.677
(2)	0.02967 ++++	0.03006	0.03011	0.03291	0.02910	0.02857	0.03007	5.029
(3)	0.08088 ++++	0.08045	0.07954	0.08575	0.07515	0.07674	0.07975	4.627
(4)	0.03905 ++++	0.03887	0.03955	0.04485	0.03942	0.03922	0.04016	5.753
(5)	0.01783 ++++	0.01715	0.01679	0.01875	0.01664	0.01655	0.01729	4.953
10 Aroclor-1262 (1)	++++ 0.02454	++++	++++	++++	++++	++++	0.02454	0.000

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m  
 Last Edit : 24-Feb-2023 15:31 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
(2)	++++ 0.03993	++++	++++	++++	++++	++++	0.03993	0.000
(3)	++++ 0.04293	++++	++++	++++	++++	++++	0.04293	0.000
(4)	++++ 0.03923	++++	++++	++++	++++	++++	0.03923	0.000
11 Aroclor-1268(1)	++++ 0.10250	++++	++++	++++	++++	++++	0.10250	0.000
(2)	++++ 0.10151	++++	++++	++++	++++	++++	0.10151	0.000
(3)	++++ 0.08686	++++	++++	++++	++++	++++	0.08686	0.000
(4)	++++ 0.28598	++++	++++	++++	++++	++++	0.28598	0.000
42 2,4-DDE	++++ ++++	++++	++++	++++	++++	++++	++++	++++
43 2,4-DDD	++++ ++++	++++	++++	++++	++++	++++	++++	++++
44 2,4-DDT	++++ ++++	++++	++++	++++	++++	++++	++++	++++
46 4,4-DDE	++++ ++++	++++	++++	++++	++++	++++	++++	++++

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m  
 Last Edit : 24-Feb-2023 15:31 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
47 4,4-DDD	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
48 4,4-DDT	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
49 Hexachlorobutadiene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
50 Hexachlorobenzene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
1 Tetrachloro-m-xylene	1.16827	1.24402	1.18546	1.20509	1.12295	1.24114	1.19449	3.860
13 Decachlorobiphenyl	0.82901	0.80558	0.77587	0.78808	0.73125	0.79742	0.78787	4.189

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m  
 Last Edit : 24-Feb-2023 15:29 richardl  
 Curve Type : Average

Calibration File Names:

Level 1: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242303ECD7.D  
 Level 2: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242304ECD7.D  
 Level 3: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242306ECD7.D  
 Level 4: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242302ECD7.D  
 Level 5: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242307ECD7.D  
 Level 6: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242305ECD7.D  
 Level 7: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242312ECD7.D

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
1 Aroclor-1221 [2C] (1)	+++++	+++++	+++++	+++++	+++++	+++++	0.00757	0.000
(2)	0.01433						0.01433	0.000
(3)	0.02333						0.02333	0.000
4 Aroclor-1232 [2C] (1)	+++++	+++++	+++++	+++++	+++++	+++++	0.00409	0.000
(2)	0.02034						0.02034	0.000
(3)	0.04067						0.04067	0.000
(4)	0.01170						0.01170	0.000
3 Aroclor-1242 [2C] (1)	+++++	+++++	+++++	+++++	+++++	+++++	0.03717	0.000



ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m  
 Last Edit : 24-Feb-2023 15:29 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
(2)	++++ 0.07813	++++	++++	++++	++++	++++	0.07813	0.000
(3)	++++ 0.02431	++++	++++	++++	++++	++++	0.02431	0.000
(4)	++++ 0.02962	++++	++++	++++	++++	++++	0.02962	0.000
6 Aroclor-1248 [2C] (1)	++++ 0.03820	++++	++++	++++	++++	++++	0.03820	0.000
(2)	++++ 0.03949	++++	++++	++++	++++	++++	0.03949	0.000
(3)	++++ 0.04545	++++	++++	++++	++++	++++	0.04545	0.000
(4)	++++ 0.05457	++++	++++	++++	++++	++++	0.05457	0.000
7 Aroclor-1016 [2C] (1)	0.05071 ++++	0.05022	0.04868	0.04733	0.04326	0.04080	0.04683	8.503
(2)	0.08143 ++++	0.09407	0.10159	0.10259	0.09651	0.09362	0.09497	8.025
(3)	0.04006 ++++	0.04718	0.04613	0.04410	0.04062	0.03926	0.04289	7.857
(4)	0.03181 ++++	0.03802	0.03707	0.03450	0.03115	0.02936	0.03365	10.251

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m  
 Last Edit : 24-Feb-2023 15:29 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
8 Aroclor-1254 [2C] (1)	++++ 0.06081	++++	++++	++++	++++	++++	0.06081	0.000
(2)	++++ 0.04892	++++	++++	++++	++++	++++	0.04892	0.000
(3)	++++ 0.10584	++++	++++	++++	++++	++++	0.10584	0.000
(4)	++++ 0.10317	++++	++++	++++	++++	++++	0.10317	0.000
(5)	++++ 0.06282	++++	++++	++++	++++	++++	0.06282	0.000
10 Aroclor-1262 [2C] (1)	++++ 0.06831	++++	++++	++++	++++	++++	0.06831	0.000
(2)	++++ 0.05818	++++	++++	++++	++++	++++	0.05818	0.000
(3)	++++ 0.06601	++++	++++	++++	++++	++++	0.06601	0.000
(4)	++++ 0.10341	++++	++++	++++	++++	++++	0.10341	0.000
9 Aroclor-1260 [2C] (1)	0.05286 ++++	0.04911	0.04696	0.04801	0.04329	0.04201	0.04704	8.422
(2)	0.12976 ++++	0.12431	0.12095	0.12664	0.11320	0.10545	0.12005	7.605

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m  
 Last Edit : 24-Feb-2023 15:29 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
(3)	0.03524 +++++	0.03147	0.02937	0.03208	0.03102	0.03198	0.03186	6.045
(4)	0.08632 +++++	0.08237	0.08044	0.08393	0.07718	0.07531	0.08092	5.126
11 Aroclor-1268 [2C] (1)	+++++ 0.16109	+++++	+++++	+++++	+++++	+++++	0.16109	0.000
(2)	+++++ 0.17318	+++++	+++++	+++++	+++++	+++++	0.17318	0.000
(3)	+++++ 0.14787	+++++	+++++	+++++	+++++	+++++	0.14787	0.000
(4)	+++++ 0.47260	+++++	+++++	+++++	+++++	+++++	0.47260	0.000
41 2,4-DDE [2C]	+++++ +++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
42 2,4-DDD [2C]	+++++ +++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
44 4,4-DDE [2C]	+++++ +++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
45 4,4-DDD/2,4-DDT [2C]	+++++ +++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
46 4,4-DDT [2C]	+++++ +++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m  
 Last Edit : 24-Feb-2023 15:29 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
48 Hexachlorobutadiene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
49 Hexachlorobenzene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
\$ 2 Tetrachloro-m-xylene [2C]	1.21526 +++++	1.19545	1.17555	1.21907	1.12560	1.11139	1.17372	3.897
\$ 13 Decachlorobiphenyl [2C]	1.17066 +++++	1.20406	1.20549	1.31040	1.21104	1.20797	1.21827	3.898

ARI Labs, Inc.  
RETENTION TIME SUMMARY REPORT

Method File: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Batch File: \\target\share\chem4\ecd7.i\230224.b  
Inst ID: ecd7.i

ID:	RT01	RT02	RT03	RT04	RT05	RT06
FILENAME:	02242302ECD7	02242303ECD7	02242304ECD7	02242305ECD7	02242306ECD7	02242307ECD7
INJ. DATE:	24-FEB-2023	24-FEB-2023	24-FEB-2023	24-FEB-2023	24-FEB-2023	24-FEB-2023
INJ. TIME:	11:12	11:33	11:54	12:15	12:36	12:57

Compound	RT01	RT02	RT03	RT04	RT05	RT06	EXPEC RT	RT WINDOW	AVG RT	STD DEV
* 41 IS-BNB	3.493	3.492	3.492	3.492	3.491	3.491	3.493	3.393-3.593	3.492	0.001
§ 1 Tetrachloro-m-xylene	5.811	5.809	5.809	5.813	5.809	5.810	5.811	5.711-5.911	5.810	0.002
2 Aroclor-1221	+++++	+++++	+++++	+++++	+++++	+++++	4.732	4.632-4.832	+++++	+++++
3 Aroclor-1242	+++++	+++++	+++++	+++++	+++++	+++++	7.269	7.169-7.369	+++++	+++++
4 Aroclor-1232	+++++	+++++	+++++	+++++	+++++	+++++	4.732	4.632-4.832	+++++	+++++
7 Aroclor-1016	7.272	7.272	7.272	7.270	7.271	7.270	7.272	7.172-7.372	7.271	0.001
6 Aroclor-1248	+++++	+++++	+++++	+++++	+++++	+++++	8.403	8.303-8.503	+++++	+++++
8 Aroclor-1254	+++++	+++++	+++++	+++++	+++++	+++++	9.295	9.195-9.395	+++++	+++++
9 Aroclor-1260	11.046	11.047	11.046	11.044	11.045	11.044	11.046	10.946-11.146	11.045	0.001
10 Aroclor-1262	+++++	+++++	+++++	+++++	+++++	+++++	10.824	10.724-10.924	+++++	+++++
11 Aroclor-1268	+++++	+++++	+++++	+++++	+++++	+++++	12.243	12.143-12.343	+++++	+++++
§ 13 Decachlorobiphenyl	13.897	13.893	13.893	13.899	13.892	13.898	13.897	13.797-13.997	13.895	0.003
* 12 IS-HBBP	14.269	14.268	14.268	14.267	14.268	14.268	14.269	14.169-14.369	14.268	0.001
42 2,4-DDE	+++++	+++++	+++++	+++++	+++++	+++++	9.260	9.210-9.310	+++++	+++++
43 2,4-DDD	+++++	+++++	+++++	+++++	+++++	+++++	9.801	9.751-9.851	+++++	+++++
44 2,4-DDT	+++++	+++++	+++++	+++++	+++++	+++++	10.293	10.243-10.343	+++++	+++++
46 4,4-DDE	+++++	+++++	+++++	+++++	+++++	+++++	9.683	9.583-9.783	+++++	+++++

Reviewer 1 \_\_\_\_\_ Date: \_\_\_\_\_  
Reviewer 2 \_\_\_\_\_ Date: \_\_\_\_\_

ARI Labs, Inc.  
RETENTION TIME SUMMARY REPORT

Method File: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
 Batch File: \\target\share\chem4\ecd7.i\230224.b  
 Inst ID: ecd7.i

Compound	RT01	RT02	RT03	RT04	RT05	RT06	EXPEC RT	RT WINDOW	AVG RT	STD DEV
47 4,4-DDD	+++++	+++++	+++++	+++++	+++++	+++++	10.254	10.154-10.354	+++++	+++++
48 4,4-DDT	+++++	+++++	+++++	+++++	+++++	+++++	10.754	10.654-10.854	+++++	+++++
49 Hexachlorobutadiene	+++++	+++++	+++++	+++++	+++++	+++++	1.842	1.742-1.942	+++++	+++++
50 Hexachlorobenzene	+++++	+++++	+++++	+++++	+++++	+++++	6.708	6.608-6.808	+++++	+++++

ARI Labs, Inc.
RETENTION TIME SUMMARY REPORT

Method File: \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m
Batch File: \\target\share\chem4\ecd7.i\230224.b\230224.b
Inst ID: ecd7.i

ID: RT01 RT02 RT03 RT04 RT05 RT06
FILENAME: 02242302ECD7 02242303ECD7 02242304ECD7 02242305ECD7 02242306ECD7 02242307ECD7
INJ. DATE: 24-FEB-2023 24-FEB-2023 24-FEB-2023 24-FEB-2023 24-FEB-2023 24-FEB-2023
INJ. TIME: 11:12 11:33 11:54 12:15 12:36 12:57

Table with 11 columns: Compound, RT01, RT02, RT03, RT04, RT05, RT06, EXPEC RT, RT WINDOW, AVG RT, STD DEV. Rows include various chemical compounds like Aroclor-1221, Aroclor-1232, etc., with their respective retention times and standard deviations.

Reviewer 1 \_\_\_\_\_ Date: \_\_\_\_\_
Reviewer 2 \_\_\_\_\_ Date: \_\_\_\_\_

ARI Labs, Inc.  
RETENTION TIME SUMMARY REPORT

Method File: \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m  
 Batch File: \\target\share\chem4\ecd7.i\230224.b\230224.b  
 Inst ID: ecd7.i

Compound	RT01	RT02	RT03	RT04	RT05	RT06	EXPEC RT	RT WINDOW	AVG RT	STD DEV
46 4,4-DDT [2C]	+++++	+++++	+++++	+++++	+++++	+++++	11.092	10.992-11.192	+++++	+++++
48 Hexachlorobutadiene	+++++	+++++	+++++	+++++	+++++	+++++	1.703	1.603-1.803	+++++	+++++
49 Hexachlorobenzene	+++++	+++++	+++++	+++++	+++++	+++++	7.178	7.078-7.278	+++++	+++++



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242301ECD7.D  
Data file 2: /230224.b/230224.b/02242301ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: IB  
Client ID:  
Injection Date: 24-FEB-2023 10:51  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

ZB5 Col		ZB35 Col			ZB5	ZB35	RPD	Compound/Flag	
RT	Shift	Response	RT	Shift	Response	on col			on col
5.826	0.019	382217	5.683	-0.002	180378	33.8	36.5	7.7	Tetrachloro-m-xylene
13.904	0.011	534110	14.120	0.001	295605	35.3	37.2	5.2	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	756896	12.3
Hexabromobiphenyl	1429847	1534275	7.3

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	336543	6.8
Hexabromobiphenyl	513946	521508	1.5

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	---			0.0	1	---			0.0
Aroclor-1016	2	---			0.0	2	---			0.0
Aroclor-1016	3	---			0.0	3	---			0.0
Aroclor-1016	4	---			0.0	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1221	1	---			0.0	1	---			0.0
Aroclor-1221	2	---			0.0	2	6.321	0.025	1873	31.1
Aroclor-1221	3	---			0.0	3	6.633	0.012	314	3.2
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1232	1	---			0.0	1	---			0.0
Aroclor-1232	2	---			0.0	2	---			0.0
Aroclor-1232	3	7.698	0.043	2193	6.0	3	---			0.0
Aroclor-1232	4	8.505	-0.076	11525	74.5	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1242	1	---			0.0	1	---			0.0
Aroclor-1242	2	7.698	0.042	2193	3.1	2	---			0.0
Aroclor-1242	3	---			0.0	3	---			0.0
Aroclor-1242	4	8.505	-0.074	11525	35.2	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1248	1	---			0.0	1	---			0.0
Aroclor-1248	2	---			0.0	2	---			0.0
Aroclor-1248	3	---			0.0	3	---			0.0
Aroclor-1248	4	---			0.0	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1254	1	---			0.0	1	---			0.0
Aroclor-1254	2	---			0.0	2	---			0.0
Aroclor-1254	3	9.596	-0.072	31424	64.3	3	---			0.0
Aroclor-1254	4	---			0.0	4	---			0.0
Aroclor-1254	5	10.167	-0.010	18361	30.8	5	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1260	1	11.098	0.054	6994	12.7	1	---			0.0
Aroclor-1260	2	---			0.0	2	---			0.0
Aroclor-1260	3	11.706	-0.027	7806	5.1	3	---			0.0
Aroclor-1260	4	---			0.0	4	---			0.0
Aroclor-1260	5	---			0.0	NS	---			----
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1262	1	10.824	-0.005	16873	35.8	1	---			0.0
Aroclor-1262	2	---			0.0	2	---			0.0
Aroclor-1262	3	---			0.0	3	---			0.0
Aroclor-1262	4	13.040	0.053	14031	18.6	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1268	1	---			0.0	1	---			0.0
Aroclor-1268	2	---			0.0	2	---			0.0
Aroclor-1268	3	12.709	0.009	6037	3.6	3	12.891	-0.001	659	0.7
Aroclor-1268	4	13.499	0.010	12396	2.3	4	13.710	0.001	1848	0.6
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				

Total PCB Area Coll (5.906 - 13.793) = 260205

Coll Total PCB = 0.0 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 18252 Col2 Total PCB = 0.0 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 IB


24-FEB-2023 10:51, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 IB

24-FEB-2023 10:51, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242302ECD7.D  
Data file 2: /230224.b/230224.b/02242302ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.25PPMAR1660  
Client ID:  
Injection Date: 24-FEB-2023 11:12  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.811	0.005	405980	5.687	0.002	192160	40.4	41.5	2.9	Tetrachloro-m-xylene
13.897	0.004	563414	14.120	0.001	336737	40.0	43.0	7.3	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	673778	0.0
Hexabromobiphenyl	1429847	1429847	0.0
Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	315256	0.0
Hexabromobiphenyl	513946	513946	0.0

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col ZB35 Col

Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.272	0.001	66125	258.4	1	7.255	-0.001	46626	252.6	
Aroclor-1016	2	7.654	-0.000	207370	265.8	2	7.855	-0.001	101071	270.1	
Aroclor-1016	3	7.792	0.002	92507	242.9	3	8.055	0.001	43448	257.1	
Aroclor-1016	4	8.406	0.001	64388	261.5	4	8.306	-0.000	33986	256.3	
Total CollAve (4 peaks):				257.2	Total Col2Ave (4 peaks):				259.0	RPD = 1	
Corrected Ave (3 peaks):				254.3	Corrected Ave (3 peaks):				255.3	RPD = 0	

CalAmt %D: 2.9

CalAmt %D: 3.6

Aroclor-1260	1	11.046	0.001	138355	269.0	1	11.653	0.001	77114	255.2	
Aroclor-1260	2	11.363	0.002	147051	273.6	2	11.918	0.001	203401	263.7	
Aroclor-1260	3	11.736	0.003	383171	268.8	3	12.435	-0.000	51517	251.7	
Aroclor-1260	4	12.141	0.002	200399	279.2	4	12.502	0.001	134797	259.3	
Aroclor-1260	5	12.247	0.003	83796	271.2	NS	---			----	
Total CollAve (5 peaks):				272.4	Total Col2Ave (4 peaks):				257.5	RPD = 6	
Corrected Ave (4 peaks):				270.7	Corrected Ave (3 peaks):				255.4	RPD = 6	

CalAmt %D: 8.9

CalAmt %D: 3.0

Total PCB Area Coll (5.906 - 13.793) = 4024419 Coll Total PCB = 0.5 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 1889311 Col2 Total PCB = 0.5 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.25PPMAR1660


24-FEB-2023 11:12, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.25PPMAR1660

24-FEB-2023 11:12, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242303ECD7.D  
Data file 2: /230224.b/230224.b/02242303ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.02PPMAR1660  
Client ID:  
Injection Date: 24-FEB-2023 11:33  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.809	0.002	29768	5.688	0.003	14932	3.1	3.3	5.7	Tetrachloro-m-xylene
13.893	0.000	45992	14.120	0.000	23950	3.4	3.1	9.1	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	637010	-5.5
Hexabromobiphenyl	1429847	1386953	-3.0

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	307177	-2.6
Hexabromobiphenyl	513946	511463	-0.5

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)



ZB5 Col						ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.272	0.001	5052	20.9	1	7.256	0.000	3894	21.7	
Aroclor-1016	2	7.659	0.005	14714	19.9	2	7.864	0.008	6253	17.1	
Aroclor-1016	3	7.795	0.005	8226	22.8	3	8.060	0.006	3076	18.7	
Aroclor-1016	4	8.407	0.002	4780	20.5	4	8.309	0.002	2443	18.9	
Total CollAve (4 peaks):				21.1	Total Col2Ave (4 peaks):				19.1	RPD = 10	
Corrected Ave (3 peaks):				20.5	Corrected Ave (3 peaks):				18.2	RPD = 11	
CalAmt %D:				5.3	CalAmt %D:				-4.5		
Aroclor-1260	1	11.047	0.003	10147	20.3	1	11.656	0.003	6759	22.5	
Aroclor-1260	2	11.364	0.003	10287	19.7	2	11.922	0.005	16592	21.6	
Aroclor-1260	3	11.740	0.006	28043	20.3	3	12.438	0.002	4506	22.1	
Aroclor-1260	4	12.145	0.006	13540	19.4	4	12.505	0.004	11037	21.3	
Aroclor-1260	5	12.246	0.002	6182	20.6	NS	---			----	
Total CollAve (5 peaks):				20.1	Total Col2Ave (4 peaks):				21.9	RPD = 9	
Corrected Ave (4 peaks):				19.9	Corrected Ave (3 peaks):				21.7	RPD = 8	
CalAmt %D:				0.4	CalAmt %D:				9.4		

Total PCB Area Coll (5.906 - 13.793) = 324832 Coll Total PCB = 0.0 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 157149 Col2 Total PCB = 0.0 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.02PPMAR1660


24-FEB-2023 11:33, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.02PPMAR1660

24-FEB-2023 11:33, 2ul




ZB-35 Manual Integration: YES


Manual Peak Adjustment, ZB-35

Datafile: ecd7.i/230224.b/230224.b/02242303ECD7.D Injection Date: 24-FEB-2023

Manual Integration (After)



Processed Integration (Before)



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242304ECD7.D  
Data file 2: /230224.b/230224.b/02242304ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.05PPMAR1660  
Client ID:  
Injection Date: 24-FEB-2023 11:54  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

ZB5 Col		ZB35 Col		ZB5	ZB35	RPD	Compound/Flag		
RT	Shift Response	RT	Shift Response	on col	on col				
5.809	0.003	78493	5.688	0.003	36772	8.3	8.1	2.2	Tetrachloro-m-xylene
13.893	-0.000	113544	14.119	-0.000	62745	8.2	7.9	3.4	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	630965	-6.4
Hexabromobiphenyl	1429847	1409464	-1.4

Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	307599	-2.4
Hexabromobiphenyl	513946	521112	1.4

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col ZB35 Col

Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.272	0.002	12829	53.5	1	7.256	0.000	9654	53.6	
Aroclor-1016	2	7.660	0.006	36461	49.9	2	7.864	0.008	18085	49.5	
Aroclor-1016	3	7.795	0.005	19865	55.7	3	8.063	0.008	9071	55.0	
Aroclor-1016	4	8.408	0.003	11411	49.5	4	8.310	0.003	7309	56.5	
Total CollAve (4 peaks):				52.2	Total Col2Ave (4 peaks):				53.7	RPD = 3	
Corrected Ave (3 peaks):				51.0	Corrected Ave (3 peaks):				52.7	RPD = 3	

CalAmt %D: 4.3

CalAmt %D: 7.3

Aroclor-1260	1	11.046	0.002	25727	50.7	1	11.655	0.002	15996	52.2	
Aroclor-1260	2	11.363	0.002	26482	50.0	2	11.922	0.004	40487	51.8	
Aroclor-1260	3	11.739	0.005	70871	50.4	3	12.437	0.002	10248	49.4	
Aroclor-1260	4	12.143	0.004	34239	48.4	4	12.506	0.004	26828	50.9	
Aroclor-1260	5	12.246	0.002	15109	49.6	NS	---			----	
Total CollAve (5 peaks):				49.8	Total Col2Ave (4 peaks):				51.1	RPD = 2	
Corrected Ave (4 peaks):				49.6	Corrected Ave (3 peaks):				50.7	RPD = 2	

CalAmt %D: -0.3

CalAmt %D: 2.1

Total PCB Area Coll (5.906 - 13.793) = 758292 Coll Total PCB = 0.1 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 386383 Col2 Total PCB = 0.1 ppm\*

\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242305ECD7.D  
Data file 2: /230224.b/230224.b/02242305ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 1.0PPMAR1660  
Client ID:  
Injection Date: 24-FEB-2023 12:15  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.813	0.006	1641874	5.688	0.003	709674	166.2	151.5	9.3	Tetrachloro-m-xylene
13.899	0.006	2344583	14.122	0.002	1300114	161.9	158.6	2.1	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	661440	-1.8
Hexabromobiphenyl	1429847	1470100	2.8
Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	319272	1.3
Hexabromobiphenyl	513946	538138	4.7

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	7.270	-0.000	220519	877.8	1	7.254	-0.001	162833	871.2
Aroclor-1016	2	7.652	-0.002	731607	955.3	2	7.852	-0.004	373610	985.8
Aroclor-1016	3	7.789	-0.001	307629	822.8	3	8.051	-0.003	156666	915.2
Aroclor-1016	4	8.404	-0.001	229387	949.1	4	8.305	-0.002	117186	872.6
Total CollAve (4 peaks):				901.3		Total Col2Ave (4 peaks):				911.2 RPD = 1
Corrected Ave (3 peaks):				883.3		Corrected Ave (3 peaks):				886.3 RPD = 0
CalAmt %D:				-9.9		CalAmt %D:				-8.9
Aroclor-1260	1	11.044	-0.000	504641	954.2	1	11.652	-0.000	282606	893.1
Aroclor-1260	2	11.360	-0.001	524931	950.0	2	11.917	-0.000	709329	878.4
Aroclor-1260	3	11.734	-0.000	1410270	962.3	3	12.434	-0.001	215124	1003.8
Aroclor-1260	4	12.137	-0.002	720770	976.7	4	12.501	-0.001	506566	930.6
Aroclor-1260	5	12.243	-0.001	304211	957.7	NS	---			----
Total CollAve (5 peaks):				960.2		Total Col2Ave (4 peaks):				926.5 RPD = 4
Corrected Ave (4 peaks):				956.0		Corrected Ave (3 peaks):				900.7 RPD = 6
CalAmt %D:				-4.0		CalAmt %D:				-7.4

Total PCB Area Coll (5.906 - 13.793) = 14454279 Coll Total PCB = 1.8 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 7029563 Col2 Total PCB = 1.8 ppm\*

\* Quantitated against AR1660 0.25ppm in Ical


PCB-Form 10 Mod.



# PCB Dual Column Chromatograms

ECD7-ZB5 1.0PPMAR1660


24-FEB-2023 12:15, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 1.0PPMAR1660

24-FEB-2023 12:15, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242306ECD7.D  
Data file 2: /230224.b/230224.b/02242306ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.1PPMAR1660  
Client ID:  
Injection Date: 24-FEB-2023 12:36  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

ZB5 Col		ZB35 Col		ZB5	ZB35	RPD	Compound/Flag		
RT	Shift Response	RT	Shift Response	on col	on col				
5.809	0.002	155528	5.688	0.003	74628	15.9	16.0	0.9	Tetrachloro-m-xylene
13.892	-0.001	227253	14.119	-0.000	128496	15.8	15.8	0.5	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	655979	-2.6
Hexabromobiphenyl	1429847	1464509	2.4
Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	317418	0.7
Hexabromobiphenyl	513946	532962	3.7

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col				
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	7.271	0.000	25761	103.4	1	7.255	-0.000	19315	103.9
Aroclor-1016	2	7.657	0.003	75616	99.6	2	7.863	0.007	40308	107.0
Aroclor-1016	3	7.794	0.004	39547	106.7	3	8.059	0.005	18304	107.6
Aroclor-1016	4	8.406	0.001	24260	101.2	4	8.309	0.002	14708	110.2
Total CollAve (4 peaks):				102.7		Total Col2Ave (4 peaks):				107.2 RPD = 4
Corrected Ave (3 peaks):				101.4		Corrected Ave (3 peaks):				106.2 RPD = 5
CalAmt %D:				2.7		CalAmt %D:				7.2
Aroclor-1260	1	11.045	0.000	52009	98.7	1	11.655	0.002	31282	99.8
Aroclor-1260	2	11.362	0.001	55116	100.1	2	11.920	0.003	80574	100.7
Aroclor-1260	3	11.738	0.004	145604	99.7	3	12.437	0.002	19566	92.2
Aroclor-1260	4	12.141	0.002	72408	98.5	4	12.503	0.001	53588	99.4
Aroclor-1260	5	12.245	0.001	30745	97.2	NS	---			----
Total CollAve (5 peaks):				98.8		Total Col2Ave (4 peaks):				98.0 RPD = 1
Corrected Ave (4 peaks):				98.5		Corrected Ave (3 peaks):				97.1 RPD = 1
CalAmt %D:				-1.2		CalAmt %D:				-2.0

Total PCB Area Coll (5.906 - 13.793) = 1555762 Coll Total PCB = 0.2 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 764924 Col2 Total PCB = 0.2 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.1PPMAR1660

24-FEB-2023 12:36, 2u1



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242307ECD7.D  
Data file 2: /230224.b/230224.b/02242307ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.5PPMAR1660  
Client ID:  
Injection Date: 24-FEB-2023 12:57  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.810	0.004	724614	5.688	0.003	359257	75.2	76.7	2.0	Tetrachloro-m-xylene
13.898	0.005	1056911	14.120	0.000	650153	74.3	79.5	6.9	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	645275	-4.2
Hexabromobiphenyl	1429847	1445345	1.1

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	319170	1.2
Hexabromobiphenyl	513946	536853	4.5

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col						
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.270	0.000	115193	470.0	1	7.256	0.000	86287	461.8	
Aroclor-1016	2	7.654	0.000	369991	495.2	2	7.856	0.000	192524	508.1	
Aroclor-1016	3	7.790	0.000	160952	441.3	3	8.055	0.000	81039	473.6	
Aroclor-1016	4	8.405	0.000	115032	487.9	4	8.307	0.000	62136	462.8	
Total CollAve (4 peaks):				473.6		Total Col2Ave (4 peaks):				476.6	RPD = 1
Corrected Ave (3 peaks):				466.4		Corrected Ave (3 peaks):				466.1	RPD = 0

CalAmt %D: -5.3

CalAmt %D: -4.7

Aroclor-1260	1	11.044	0.000	247212	475.5	1	11.653	0.000	145247	460.1	
Aroclor-1260	2	11.361	0.000	262877	483.9	2	11.918	0.000	379838	471.5	
Aroclor-1260	3	11.734	0.000	678830	471.1	3	12.436	0.000	104092	486.9	
Aroclor-1260	4	12.139	0.000	356067	490.7	4	12.502	0.000	258953	476.9	
Aroclor-1260	5	12.244	0.000	150280	481.2	NS	---			----	
Total CollAve (5 peaks):				480.5		Total Col2Ave (4 peaks):				473.8	RPD = 1
Corrected Ave (4 peaks):				477.9		Corrected Ave (3 peaks):				469.5	RPD = 2

CalAmt %D: -3.9

CalAmt %D: -5.2

Total PCB Area Coll (5.906 - 13.793) = 7134169 Coll Total PCB = 0.9 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 3589735 Col2 Total PCB = 0.9 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.5PPMAR1660


24-FEB-2023 12:57, 2u1



ZB-5 Manual Integration: NO

ECD7-ZB35 0.5PPMAR1660

24-FEB-2023 12:57, 2u1



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242308ECD7.D  
Data file 2: /230224.b/230224.b/02242308ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1242.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.25PPMAR1242  
Client ID:  
Injection Date: 24-FEB-2023 13:18  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.809	0.003	434187	5.688	0.003	214306	46.0	46.5	1.1	Tetrachloro-m-xylene
13.894	0.000	515867	14.119	-0.001	312943	35.6	38.5	7.6	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	632576	-6.1
Hexabromobiphenyl	1429847	1469715	2.8

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	314129	-0.4
Hexabromobiphenyl	513946	534294	4.0

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)



ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1242	1	7.271	0.000	49009	250.0	1	7.255	0.000	36487	250.0
Aroclor-1242	2	7.656	0.000	148833	250.0	2	7.858	0.000	76699	250.0
Aroclor-1242	3	8.405	0.000	46308	250.0	3	9.167	0.000	23866	250.0
Aroclor-1242	4	8.579	0.000	68453	250.0	4	9.597	0.000	29080	250.0
Total Col1Ave (4 peaks):				250.0	Total Col2Ave (4 peaks):				250.0	RPD = 0
Corrected Ave (3 peaks):				250.0	Corrected Ave (3 peaks):				250.0	RPD = 0

Total PCB Area Col1 (5.906 - 13.793) = 1221467 Col1 Total PCB = 0.2 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 572067 Col2 Total PCB = 0.2 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.25PPMAR1242


24-FEB-2023 13:18, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.25PPMAR1242

24-FEB-2023 13:18, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242309ECD7.D  
Data file 2: /230224.b/230224.b/02242309ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1248.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.25PPMAR1248  
Client ID:  
Injection Date: 24-FEB-2023 13:39  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.809	0.003	349513	5.688	0.003	176615	36.6	37.9	3.4	Tetrachloro-m-xylene
13.894	0.001	523008	14.121	0.001	322054	36.4	39.3	7.5	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	639911	-5.0
Hexabromobiphenyl	1429847	1458696	2.0
Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	317938	0.9
Hexabromobiphenyl	513946	538760	4.8

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col ZB35 Col

Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1248	1	8.405	0.000	78055	250.0	1	8.308	0.000	37951	250.0
Aroclor-1248	2	8.580	0.000	99216	250.0	2	8.714	0.000	39239	250.0
Aroclor-1248	3	8.999	0.000	187178	250.0	3	9.166	0.000	45157	250.0
Aroclor-1248	4	9.295	0.000	95291	250.0	4	9.590	0.000	54216	250.0
Total Col1Ave (4 peaks):				250.0	Total Col2Ave (4 peaks):				250.0	RPD = 0
Corrected Ave (3 peaks):				250.0	Corrected Ave (3 peaks):				250.0	RPD = 0

Total PCB Area Col1 (5.906 - 13.793) = 1565180 Col1 Total PCB = 0.2 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 754991 Col2 Total PCB = 0.2 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.25PPMAR1248


24-FEB-2023 13:39, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.25PPMAR1248

24-FEB-2023 13:39, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242310ECD7.D  
Data file 2: /230224.b/230224.b/02242310ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1254.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.25PPMAR1254  
Client ID:  
Injection Date: 24-FEB-2023 14:00  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.808	0.002	352587	5.687	0.002	177502	37.3	38.6	3.4	Tetrachloro-m-xylene
13.895	0.002	532500	14.119	0.000	325903	37.0	40.2	8.2	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	633407	-6.0
Hexabromobiphenyl	1429847	1460265	2.1

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	313673	-0.5
Hexabromobiphenyl	513946	532442	3.6

\* Standard Areas taken from Initial Cal Level 3

Initial Calibration Date: 24-FEB-2023

<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col ZB35 Col

Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1254	1	9.299	0.000	159011	250.0	1	9.449	0.000	59603	250.0	
Aroclor-1254	2	9.377	0.000	71516	250.0	2	9.970	0.000	47949	250.0	
Aroclor-1254	3	9.668	0.000	102230	250.0	3	10.124	0.000	103745	250.0	
Aroclor-1254	4	9.807	0.000	198777	250.0	4	10.373	0.000	101135	250.0	
Aroclor-1254	5	10.176	0.000	124586	250.0	5	10.569	0.000	61577	250.0	
Total CollAve (5 peaks):				250.0	Total Col2Ave (5 peaks):				250.0	RPD = 0	
Corrected Ave (4 peaks):				250.0	Corrected Ave (4 peaks):				250.0	RPD = 0	

Total PCB Area Coll (5.906 - 13.793) = 2179224 Coll Total PCB = 0.3 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 1022156 Col2 Total PCB = 0.3 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.25PPMAR1254


24-FEB-2023 14:00, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.25PPMAR1254

24-FEB-2023 14:00, 2ul



ZB-35 Manual Integration: NO



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242311ECD7.D  
Data file 2: /230224.b/230224.b/02242311ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR2162.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.25PPMAR2162  
Client ID:  
Injection Date: 24-FEB-2023 14:21  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.807	0.001	362236	5.686	0.000	177349	38.4	39.2	2.1	Tetrachloro-m-xylene
13.894	0.001	523254	14.119	-0.000	321034	36.0	39.2	8.3	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	632433	-6.1
Hexabromobiphenyl	1429847	1474039	3.1

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	308453	-2.2
Hexabromobiphenyl	513946	538177	4.7

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col ZB35 Col

Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1221	1	4.731	0.000	14160	250.0	1	4.956	0.000	7300	250.0	
Aroclor-1221	2	6.132	0.000	25324	250.0	2	6.296	0.000	13816	250.0	
Aroclor-1221	3	6.382	0.000	58795	250.0	3	6.622	0.000	22491	250.0	
Total CollAve (3 peaks):				250.0	Total Col2Ave (3 peaks):				250.0	RPD = 0	
Corrected Ave: < 3 Peaks					Corrected Ave: < 3 Peaks						

Aroclor-1262	1	10.829	0.000	113046	250.0	1	11.200	0.000	114880	250.0	
Aroclor-1262	2	12.244	0.000	183948	250.0	2	11.652	0.000	97844	250.0	
Aroclor-1262	3	12.319	0.000	197749	250.0	3	12.434	0.000	111015	250.0	
Aroclor-1262	4	12.987	0.000	180727	250.0	4	12.502	0.000	173913	250.0	
Total CollAve (4 peaks):				250.0	Total Col2Ave (4 peaks):				250.0	RPD = 0	
Corrected Ave (3 peaks):				250.0	Corrected Ave (3 peaks):				250.0	RPD = 0	

Total PCB Area Coll (5.906 - 13.793) = 3105316 Coll Total PCB = 0.4 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 1573107 Col2 Total PCB = 0.4 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.25PPMAR2162


24-FEB-2023 14:21, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.25PPMAR2162

24-FEB-2023 14:21, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242312ECD7.D  
Data file 2: /230224.b/230224.b/02242312ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR3268.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.25PPMAR3268  
Client ID:  
Injection Date: 24-FEB-2023 14:42  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.806	0.000	366416	5.685	0.000	179450	38.0	38.9	2.4	Tetrachloro-m-xylene
13.893	0.000	778191	14.119	0.000	477889	53.0	57.5	8.3	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	645602	-4.2
Hexabromobiphenyl	1429847	1492154	4.4

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	314042	-0.4
Hexabromobiphenyl	513946	545458	6.1

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col				
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1232	1	4.730	0.000	8647	250.0	1	4.956	0.000	4017	250.0
Aroclor-1232	2	6.131	0.000	17148	250.0	2	7.254	0.000	19962	250.0
Aroclor-1232	3	7.656	0.000	77627	250.0	3	7.861	0.000	39913	250.0
Aroclor-1232	4	8.581	0.000	32993	250.0	4	8.715	0.000	11487	250.0
Total CollAve (4 peaks):				250.0		Total Col2Ave (4 peaks):				250.0 RPD = 0
Corrected Ave (3 peaks):				250.0		Corrected Ave (3 peaks):				250.0 RPD = 0
Aroclor-1268	1	12.247	0.000	477974	250.0	1	12.432	0.000	274595	250.0
Aroclor-1268	2	12.317	0.000	473326	250.0	2	12.500	0.000	295194	250.0
Aroclor-1268	3	12.699	0.000	405011	250.0	3	12.892	0.000	252048	250.0
Aroclor-1268	4	13.490	0.000	1333528	250.0	4	13.709	0.000	805579	250.0
Total CollAve (4 peaks):				250.0		Total Col2Ave (4 peaks):				250.0 RPD = 0
Corrected Ave (3 peaks):				250.0		Corrected Ave (3 peaks):				250.0 RPD = 0

Total PCB Area Coll (5.906 - 13.793) = 3998414 Coll Total PCB = 0.5 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 2300029 Col2 Total PCB = 0.6 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.25PPMAR3268


24-FEB-2023 14:42, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.25PPMAR3268

24-FEB-2023 14:42, 2ul



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242313ECD7.D  
Data file 2: /230224.b/230224.b/02242313ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1660SCV  
Client ID:  
Injection Date: 24-FEB-2023 15:03  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.807	0.001	337070	5.686	0.001	165848	34.9	35.8	2.3	Tetrachloro-m-xylene
13.895	0.002	515407	14.119	-0.000	316730	34.3	37.3	8.4	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	645975	-4.1
Hexabromobiphenyl	1429847	1524245	6.6

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	316115	0.3
Hexabromobiphenyl	513946	556950	8.4

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col						
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.269	-0.002	59491	242.5	1	7.254	-0.002	44576	240.9	
Aroclor-1016	2	7.655	0.001	181090	242.1	2	7.857	0.002	95386	254.2	
Aroclor-1016	3	7.790	0.000	88470	242.3	3	8.056	0.002	42160	248.8	
Aroclor-1016	4	8.404	-0.001	57980	245.6	4	8.307	0.000	32197	242.1	
Total CollAve (4 peaks):				243.1	Total Col2Ave (4 peaks):				246.5	RPD = 1	
Corrected Ave (3 peaks):				242.3	Corrected Ave (3 peaks):				243.9	RPD = 1	
Aroclor-1221	1	4.731	0.000	464	8.0	1	---			0.0	
Aroclor-1221	2	6.130	-0.002	9233	89.2	2	6.300	0.004	5379	95.0	
Aroclor-1221	3	6.382	-0.001	42570	177.2	3	6.623	0.001	20952	227.2	
Total CollAve (3 peaks):				91.5	Col2Ave: <3 Quant Peaks						
Aroclor-1232	1	4.731	0.001	464	13.4	1	---			0.0	
Aroclor-1232	2	6.130	-0.001	9233	134.5	2	7.254	-0.000	44576	554.6	
Aroclor-1232	3	7.655	-0.001	181090	582.9	3	7.857	-0.003	95386	593.5	
Aroclor-1232	4	8.580	-0.001	79916	605.2	4	8.713	-0.002	29795	644.2	
Total CollAve (4 peaks):				334.0	Total Col2Ave (3 peaks):				597.4	RPD = 57*	
Corrected Ave (3 peaks):				243.6	Corrected Ave: < 3 Peaks						
Aroclor-1242	1	7.269	-0.002	59491	297.2	1	7.254	-0.002	44576	303.5	
Aroclor-1242	2	7.655	-0.001	181090	297.9	2	7.857	-0.000	95386	309.0	
Aroclor-1242	3	8.404	-0.001	57980	306.5	3	9.115	-0.052	18754	195.2	
Aroclor-1242	4	8.580	0.000	79916	285.8	4	9.697	0.100	1355	11.6	
Total CollAve (4 peaks):				296.8	Total Col2Ave (4 peaks):				204.8	RPD = 37	
Corrected Ave (3 peaks):				293.6	Corrected Ave (3 peaks):				170.1	RPD = 53*	
Aroclor-1248	1	8.404	-0.001	57980	184.0	1	8.307	-0.001	32197	213.3	
Aroclor-1248	2	8.580	-0.001	79916	199.5	2	8.713	-0.001	29795	190.9	
Aroclor-1248	3	8.993	-0.006	71805	95.0	3	9.115	-0.050	18754	104.4	
Aroclor-1248	4	9.300	0.006	47348	123.1	4	---			0.0	
Total CollAve (4 peaks):				150.4	Total Col2Ave (3 peaks):				169.6	RPD = 12	
Corrected Ave (3 peaks):				134.0	Corrected Ave: < 3 Peaks						
Aroclor-1254	1	9.300	0.002	47348	73.0	1	9.451	0.001	22438	93.4	
Aroclor-1254	2	---			0.0	2	9.972	0.001	2694	13.9	
Aroclor-1254	3	9.670	0.002	5461	13.1	3	10.147	0.024	52914	126.5	
Aroclor-1254	4	9.807	-0.000	18944	23.4	4	10.370	-0.003	70430	172.8	
Aroclor-1254	5	10.121	-0.056	154170	303.3	5	10.568	-0.000	98525	396.9	
Total CollAve (4 peaks):				103.2	Total Col2Ave (5 peaks):				160.7	RPD = 44*	
Corrected Ave (3 peaks):				36.5	Corrected Ave (4 peaks):				101.7	RPD = 94*	
Aroclor-1260	1	11.044	0.000	149195	272.1	1	11.653	0.000	82210	251.0	
Aroclor-1260	2	11.361	-0.000	153832	268.5	2	11.919	0.001	222226	265.9	
Aroclor-1260	3	11.736	0.002	396660	261.0	3	12.435	-0.000	59148	266.7	
Aroclor-1260	4	12.140	0.001	190448	248.9	4	12.504	0.002	147180	261.2	
Aroclor-1260	5	12.244	-0.000	91385	277.5	NS	---			----	
Total CollAve (5 peaks):				265.6	Total Col2Ave (4 peaks):				261.2	RPD = 2	
Corrected Ave (4 peaks):				262.6	Corrected Ave (3 peaks):				259.4	RPD = 1	
Aroclor-1262	1	10.827	-0.002	220238	471.0	1	11.199	-0.001	84479	177.6	
Aroclor-1262	2	12.244	0.000	91385	120.1	2	11.653	0.002	82210	203.0	
Aroclor-1262	3	12.320	0.001	113066	138.2	3	12.435	0.002	59148	128.7	
Aroclor-1262	4	12.988	0.001	102156	136.7	4	12.504	0.002	147180	204.4	
Total CollAve (4 peaks):				216.5	Total Col2Ave (4 peaks):				178.4	RPD = 19	
Corrected Ave (3 peaks):				131.7	Corrected Ave (3 peaks):				169.8	RPD = 25	
Aroclor-1268	1	12.244	-0.003	91385	46.8	1	12.435	0.003	59148	52.7	
Aroclor-1268	2	12.320	0.003	113066	58.5	2	12.504	0.004	147180	122.1	
Aroclor-1268	3	12.726	0.027	46633	28.2	3	12.893	0.001	2874	2.8	
Aroclor-1268	4	13.489	-0.000	25567	4.7	4	13.709	-0.000	13041	4.0	
Total CollAve (4 peaks):				34.5	Total Col2Ave (4 peaks):				45.4	RPD = 27	
Corrected Ave (3 peaks):				26.6	Corrected Ave (3 peaks):				19.8	RPD = 29	



Total PCB Area Col1 (5.906 - 13.793) = 3743076 Col1 Total PCB = 0.5 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 1897008 Col2 Total PCB = 0.5 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR1660SCV


24-FEB-2023 15:03, 2u1



ZB-5 Manual Integration: NO

ECD7-ZB35 AR1660SCV

24-FEB-2023 15:03, 2u1



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242314ECD7.D  
Data file 2: /230224.b/230224.b/02242314ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1242SCV  
Client ID:  
Injection Date: 24-FEB-2023 15:24  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.808	0.002	354283	5.686	0.001	172455	33.6	34.5	2.6	Tetrachloro-m-xylene
13.895	0.002	567088	14.120	0.001	347430	37.0	40.3	8.6	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	705650	4.7
Hexabromobiphenyl	1429847	1555683	8.8

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	340433	8.0
Hexabromobiphenyl	513946	565609	10.1

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	7.270	-0.000	39927	149.0	1	7.256	0.000	32417	162.7
Aroclor-1016	2	7.653	-0.001	132339	162.0	2	7.856	0.001	69235	171.3
Aroclor-1016	3	7.791	0.001	59310	148.7	3	8.055	0.000	29473	161.5
Aroclor-1016	4	8.405	0.000	42537	165.0	4	8.307	-0.000	22792	159.2
Total CollAve (4 peaks):				156.2		Total Col2Ave (4 peaks):				163.7 RPD = 5
Corrected Ave (3 peaks):				153.2		Corrected Ave (3 peaks):				161.1 RPD = 5
Aroclor-1221	1	4.733	0.002	319	5.0	1	---			0.0
Aroclor-1221	2	6.131	-0.001	6534	57.8	2	6.319	0.022	4365	71.6
Aroclor-1221	3	6.384	0.001	29664	113.0	3	6.624	0.002	14916	150.2
Total CollAve (3 peaks):				58.6		Col2Ave: <3 Quant Peaks				
Aroclor-1232	1	4.733	0.003	319	8.4	1	---			0.0
Aroclor-1232	2	6.131	0.000	6534	87.2	2	7.256	0.002	32417	374.5
Aroclor-1232	3	7.653	-0.003	132339	389.9	3	7.856	-0.004	69235	400.0
Aroclor-1232	4	8.579	-0.002	69445	481.4	4	8.714	-0.001	22167	445.0
Total CollAve (4 peaks):				241.7		Total Col2Ave (3 peaks):				406.5 RPD = 51*
Corrected Ave (3 peaks):				161.8		Corrected Ave: < 3 Peaks				
Aroclor-1242	1	7.270	-0.001	39927	182.6	1	7.256	0.000	32417	205.0
Aroclor-1242	2	7.653	-0.003	132339	199.3	2	7.856	-0.002	69235	208.2
Aroclor-1242	3	8.405	-0.000	42537	205.9	3	9.164	-0.004	23068	223.0
Aroclor-1242	4	8.579	-0.000	69445	227.4	4	9.587	-0.010	31021	246.1
Total CollAve (4 peaks):				203.8		Total Col2Ave (4 peaks):				220.6 RPD = 8
Corrected Ave (3 peaks):				195.9		Corrected Ave (3 peaks):				212.1 RPD = 8
Aroclor-1248	1	8.405	0.000	42537	123.5	1	8.307	-0.001	22792	140.2
Aroclor-1248	2	8.579	-0.001	69445	158.7	2	8.714	-0.000	22167	131.9
Aroclor-1248	3	9.001	0.003	91942	111.4	3	9.164	-0.002	23068	119.3
Aroclor-1248	4	9.294	-0.000	38711	92.1	4	9.587	-0.003	31021	133.6
Total CollAve (4 peaks):				121.4		Total Col2Ave (4 peaks):				131.2 RPD = 8
Corrected Ave (3 peaks):				109.0		Corrected Ave (3 peaks):				128.3 RPD = 16
Aroclor-1254	1	9.294	-0.005	38711	54.6	1	9.450	0.001	13131	50.7
Aroclor-1254	2	9.377	-0.000	17371	54.5	2	9.970	0.000	8340	40.1
Aroclor-1254	3	9.668	-0.000	16373	35.9	3	10.123	-0.000	16364	36.3
Aroclor-1254	4	9.807	-0.001	27490	31.0	4	10.382	0.009	16062	36.6
Aroclor-1254	5	10.175	-0.001	20494	36.9	5	10.572	0.004	4818	18.0
Total CollAve (5 peaks):				42.6		Total Col2Ave (5 peaks):				36.4 RPD = 16
Corrected Ave (4 peaks):				39.6		Corrected Ave (4 peaks):				32.8 RPD = 19
Aroclor-1260	1	11.048	0.003	794	1.4	1	11.665	0.012	1652	5.0
Aroclor-1260	2	11.366	0.005	814	1.4	2	11.926	0.008	842	1.0
Aroclor-1260	3	11.739	0.006	1848	1.2	3	12.438	0.002	483	2.1
Aroclor-1260	4	12.145	0.006	1372	1.8	4	12.506	0.004	790	1.4
Aroclor-1260	5	---			0.0	NS	---			---
Total CollAve (4 peaks):				1.4		Total Col2Ave (4 peaks):				2.4 RPD = 49*
Corrected Ave (3 peaks):				1.3		Corrected Ave (3 peaks):				1.5 RPD = 12
Aroclor-1262	1	10.832	0.003	13157	27.6	1	11.121	-0.079	6113	12.7
Aroclor-1262	2	12.145	-0.098	1372	1.8	2	11.665	0.013	1652	4.0
Aroclor-1262	3	---			0.0	3	12.438	0.004	483	1.0
Aroclor-1262	4	13.038	0.051	842	1.1	4	12.506	0.004	790	1.1
Total CollAve (3 peaks):				10.1		Total Col2Ave (4 peaks):				4.7 RPD = 73*
Corrected Ave: < 3 Peaks						Corrected Ave (3 peaks):				2.0
Aroclor-1268	1	---			0.0	1	12.438	0.006	483	0.4
Aroclor-1268	2	---			0.0	2	12.506	0.006	790	0.6
Aroclor-1268	3	12.617	-0.082	5851	3.5	3	12.899	0.007	491	0.5
Aroclor-1268	4	13.500	0.010	1745	0.3	4	13.714	0.005	379	0.1
CollAve: <3 Quant Peaks						Col2Ave:				0.4

Total PCB Area Col1 (5.906 - 13.793) = 1149784 Col1 Total PCB = 0.1 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 572210 Col2 Total PCB = 0.1 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR1242SCV


24-FEB-2023 15:24, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 AR1242SCV

24-FEB-2023 15:24, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242315ECD7.D  
Data file 2: /230224.b/230224.b/02242315ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1248SCV  
Client ID:  
Injection Date: 24-FEB-2023 15:45  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.809	0.002	336655	5.687	0.002	168719	34.9	36.4	4.2	Tetrachloro-m-xylene
13.894	0.001	499162	14.118	-0.001	308317	33.1	36.3	9.2	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	646554	-4.0
Hexabromobiphenyl	1429847	1529451	7.0

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	316066	0.3
Hexabromobiphenyl	513946	557213	8.4

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	7.271	0.000	19773	80.5	1	7.254	-0.002	16926	91.5
Aroclor-1016	2	7.653	-0.001	88099	117.7	2	7.857	0.001	45733	121.9
Aroclor-1016	3	7.794	0.003	35915	98.3	3	8.060	0.005	8078	47.7
Aroclor-1016	4	8.406	0.001	77842	329.5	4	8.307	0.000	37348	280.9
Total CollAve (4 peaks):				156.5		Total Col2Ave (4 peaks):				135.5 RPD = 14
Corrected Ave (3 peaks):				98.8		Corrected Ave (3 peaks):				87.0 RPD = 13
Aroclor-1221	1	---			0.0	1	---			0.0
Aroclor-1221	2	6.133	0.001	680	6.6	2	6.326	0.030	1966	34.7
Aroclor-1221	3	6.384	0.002	3390	14.1	3	6.631	0.009	1571	17.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1232	1	---			0.0	1	---			0.0
Aroclor-1232	2	6.133	0.002	680	9.9	2	7.254	-0.000	16926	210.6
Aroclor-1232	3	7.653	-0.002	88099	283.3	3	7.857	-0.004	45733	284.6
Aroclor-1232	4	8.581	-0.000	99572	753.4	4	8.714	-0.001	38224	826.6
Total CollAve (3 peaks):				348.9		Total Col2Ave (3 peaks):				440.6 RPD = 23
Corrected Ave: < 3 Peaks						Corrected Ave: < 3 Peaks				
Aroclor-1242	1	7.271	-0.000	19773	98.7	1	7.254	-0.002	16926	115.3
Aroclor-1242	2	7.653	-0.003	88099	144.8	2	7.857	-0.001	45733	148.2
Aroclor-1242	3	8.406	0.000	77842	411.2	3	9.165	-0.002	45021	468.7
Aroclor-1242	4	8.581	0.001	99572	355.8	4	9.590	-0.008	53613	458.1
Total CollAve (4 peaks):				252.6		Total Col2Ave (4 peaks):				297.6 RPD = 16
Corrected Ave (3 peaks):				199.8		Corrected Ave (3 peaks):				240.5 RPD = 19
Aroclor-1248	1	8.406	0.000	77842	246.8	1	8.307	-0.001	37348	247.5
Aroclor-1248	2	8.581	0.000	99572	248.3	2	8.714	-0.000	38224	245.0
Aroclor-1248	3	8.998	-0.000	186857	247.0	3	9.165	-0.000	45021	250.7
Aroclor-1248	4	9.294	-0.000	98398	255.5	4	9.590	-0.001	53613	248.7
Total CollAve (4 peaks):				249.4		Total Col2Ave (4 peaks):				248.0 RPD = 1
Corrected Ave (3 peaks):				247.4		Corrected Ave (3 peaks):				247.0 RPD = 0
Aroclor-1254	1	9.294	-0.004	98398	151.6	1	9.450	0.001	21823	90.8
Aroclor-1254	2	9.377	-0.001	49616	169.9	2	9.971	0.001	19450	100.6
Aroclor-1254	3	9.669	0.001	40230	96.4	3	10.124	0.000	36574	87.5
Aroclor-1254	4	9.808	0.001	68500	84.4	4	10.389	0.016	35100	86.1
Aroclor-1254	5	10.183	0.007	47365	93.1	5	10.573	0.004	5676	22.9
Total CollAve (5 peaks):				119.1		Total Col2Ave (5 peaks):				77.6 RPD = 42*
Corrected Ave (4 peaks):				106.4		Corrected Ave (4 peaks):				71.8 RPD = 39
Aroclor-1260	1	11.047	0.003	1670	3.0	1	11.662	0.009	2055	6.3
Aroclor-1260	2	11.362	0.001	1111	1.9	2	11.924	0.007	1466	1.8
Aroclor-1260	3	11.739	0.005	2107	1.4	3	12.434	-0.002	573	2.6
Aroclor-1260	4	12.144	0.005	1379	1.8	4	12.505	0.003	1003	1.8
Aroclor-1260	5	12.251	0.006	698	2.1	NS	---			----
Total CollAve (5 peaks):				2.1		Total Col2Ave (4 peaks):				3.1 RPD = 41*
Corrected Ave (4 peaks):				1.8		Corrected Ave (3 peaks):				2.0 RPD = 12
Aroclor-1262	1	10.833	0.005	15355	32.7	1	11.122	-0.079	7225	15.2
Aroclor-1262	2	12.251	0.007	698	0.9	2	11.662	0.011	2055	5.1
Aroclor-1262	3	12.321	0.002	836	1.0	3	12.434	0.000	573	1.2
Aroclor-1262	4	12.991	0.004	1043	1.4	4	12.505	0.003	1003	1.4
Total CollAve (4 peaks):				9.0		Total Col2Ave (4 peaks):				5.7 RPD = 45*
Corrected Ave (3 peaks):				1.1		Corrected Ave (3 peaks):				2.6 RPD = 80*
Aroclor-1268	1	12.251	0.004	698	0.4	1	12.434	0.002	573	0.5
Aroclor-1268	2	12.321	0.004	836	0.4	2	12.505	0.005	1003	0.8
Aroclor-1268	3	12.700	0.001	2449	1.5	3	12.892	0.001	721	0.7
Aroclor-1268	4	13.493	0.003	7547	1.4	4	13.708	-0.001	2265	0.7
Total CollAve (4 peaks):				0.9		Total Col2Ave (4 peaks):				0.7 RPD = 29
Corrected Ave (3 peaks):				0.7		Corrected Ave (3 peaks):				0.6 RPD = 13



Total PCB Area Col1 (5.906 - 13.793) = 1574335 Col1 Total PCB = 0.2 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 746330 Col2 Total PCB = 0.2 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR1248SCV


24-FEB-2023 15:45, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 AR1248SCV

24-FEB-2023 15:45, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242316ECD7.D  
Data file 2: /230224.b/230224.b/02242316ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1254SCV  
Client ID:  
Injection Date: 24-FEB-2023 16:06  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.806	-0.000	354312	5.686	0.001	174604	36.1	37.1	2.6	Tetrachloro-m-xylene
13.895	0.002	540961	14.119	-0.000	329134	34.6	37.9	9.0	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	656887	-2.5
Hexabromobiphenyl	1429847	1585505	10.9

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	320936	1.8
Hexabromobiphenyl	513946	570006	10.9

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.270	0.000	565	2.3	1	7.255	-0.001	387	2.1	
Aroclor-1016	2	7.656	0.002	1875	2.5	2	7.854	-0.002	860	2.3	
Aroclor-1016	3	7.792	0.002	1106	3.0	3	8.098	0.043	578	3.4	
Aroclor-1016	4	8.405	0.000	29924	124.7	4	8.307	0.000	21985	162.9	
Total CollAve (4 peaks):				33.1	Total Col2Ave (4 peaks):				42.6	RPD = 25	
Corrected Ave (3 peaks):				2.6	Corrected Ave (3 peaks):				2.6	RPD = 0	
Aroclor-1221	1	---			0.0	1	---			0.0	
Aroclor-1221	2	---			0.0	2	6.325	0.028	1947	33.9	
Aroclor-1221	3	---			0.0	3	6.637	0.015	368	3.9	
CollAve: <3 Quant Peaks					Col2Ave: <3 Quant Peaks						
Aroclor-1232	1	---			0.0	1	---			0.0	
Aroclor-1232	2	---			0.0	2	7.255	0.001	387	4.7	
Aroclor-1232	3	7.656	0.000	1875	5.9	3	7.854	-0.007	860	5.3	
Aroclor-1232	4	8.583	0.002	12327	91.8	4	8.715	0.000	15013	319.7	
CollAve: <3 Quant Peaks					Col2Ave: 109.9						
Aroclor-1242	1	7.270	-0.000	565	2.8	1	7.255	-0.001	387	2.6	
Aroclor-1242	2	7.656	0.000	1875	3.0	2	7.854	-0.004	860	2.7	
Aroclor-1242	3	8.405	-0.000	29924	155.6	3	9.169	0.002	21933	224.9	
Aroclor-1242	4	8.583	0.003	12327	43.4	4	9.545	-0.053	34065	286.6	
Total CollAve (4 peaks):				51.2	Total Col2Ave (4 peaks):				129.2	RPD = 87*	
Corrected Ave (3 peaks):				16.4	Corrected Ave (3 peaks):				76.7	RPD = 130*	
Aroclor-1248	1	8.405	0.000	29924	93.4	1	8.307	-0.001	21985	143.5	
Aroclor-1248	2	8.583	0.002	12327	30.3	2	8.715	0.001	15013	94.8	
Aroclor-1248	3	8.992	-0.007	145580	189.4	3	9.169	0.004	21933	120.3	
Aroclor-1248	4	9.298	0.003	155450	397.3	4	9.545	-0.046	34065	155.6	
Total CollAve (4 peaks):				177.6	Total Col2Ave (4 peaks):				128.5	RPD = 32	
Corrected Ave (3 peaks):				104.3	Corrected Ave (3 peaks):				119.5	RPD = 14	
Aroclor-1254	1	9.298	-0.001	155450	235.7	1	9.450	0.001	58639	240.4	
Aroclor-1254	2	9.377	-0.001	69801	235.3	2	9.971	0.000	47008	239.5	
Aroclor-1254	3	9.668	-0.000	100839	237.8	3	10.124	0.000	100062	235.7	
Aroclor-1254	4	9.807	0.000	190544	231.1	4	10.373	0.000	99535	240.5	
Aroclor-1254	5	10.176	-0.000	122321	236.7	5	10.570	0.001	61549	244.2	
Total CollAve (5 peaks):				235.3	Total Col2Ave (5 peaks):				240.1	RPD = 2	
Corrected Ave (4 peaks):				234.7	Corrected Ave (4 peaks):				239.0	RPD = 2	
Aroclor-1260	1	11.043	-0.002	12288	21.5	1	11.661	0.008	29062	86.7	
Aroclor-1260	2	11.361	-0.001	13660	22.9	2	11.921	0.003	22238	26.0	
Aroclor-1260	3	11.736	0.002	37632	23.8	3	12.441	0.005	3555	15.7	
Aroclor-1260	4	12.141	0.002	27105	34.1	4	12.503	0.001	13126	22.8	
Aroclor-1260	5	12.320	0.076	2381	6.9	NS	---			---	
Total CollAve (5 peaks):				21.9	Total Col2Ave (4 peaks):				37.8	RPD = 53*	
Corrected Ave (4 peaks):				18.8	Corrected Ave (3 peaks):				21.5	RPD = 13	
Aroclor-1262	1	10.827	-0.002	220626	453.6	1	11.281	0.081	13562	27.9	
Aroclor-1262	2	12.320	0.076	2381	3.0	2	11.661	0.009	29062	70.1	
Aroclor-1262	3	---			0.0	3	12.441	0.007	3555	7.6	
Aroclor-1262	4	12.989	0.002	3225	4.1	4	12.503	0.001	13126	17.8	
Total CollAve (3 peaks):				153.6	Total Col2Ave (4 peaks):				30.8	RPD = 133*	
Corrected Ave: < 3 Peaks					Corrected Ave (3 peaks):				17.7		
Aroclor-1268	1	12.320	0.074	2381	1.2	1	12.441	0.009	3555	3.1	
Aroclor-1268	2	---			0.0	2	12.503	0.003	13126	10.6	
Aroclor-1268	3	12.701	0.002	2939	1.7	3	12.892	0.000	772	0.7	
Aroclor-1268	4	13.493	0.003	9164	1.6	4	13.707	-0.002	2801	0.8	
Total CollAve (3 peaks):				1.5	Total Col2Ave (4 peaks):				3.8	RPD = 87*	
Corrected Ave: < 3 Peaks					Corrected Ave (3 peaks):				1.6		

Total PCB Area Col1 (5.906 - 13.793) = 2118645 Col1 Total PCB = 0.3 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 1007601 Col2 Total PCB = 0.3 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR1254SCV


24-FEB-2023 16:06, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 AR1254SCV

24-FEB-2023 16:06, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242317ECD7.D  
Data file 2: /230224.b/230224.b/02242317ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR2162SCV  
Client ID:  
Injection Date: 24-FEB-2023 16:27  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.807	0.000	356001	5.685	0.000	170882	36.0	36.6	1.7	Tetrachloro-m-xylene
13.895	0.002	533971	14.119	0.000	326235	34.4	37.9	9.5	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	661953	-1.8
Hexabromobiphenyl	1429847	1574993	10.2

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	317807	0.8
Hexabromobiphenyl	513946	565951	10.1

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.269	-0.001	7175	28.5	1	7.256	0.000	3727	20.0	
Aroclor-1016	2	7.659	0.005	12893	16.8	2	7.863	0.007	5834	15.5	
Aroclor-1016	3	7.794	0.004	6936	18.5	3	8.063	0.009	2963	17.4	
Aroclor-1016	4	8.408	0.003	3610	14.9	4	8.308	0.002	2045	15.3	
Total CollAve (4 peaks):				19.7	Total Col2Ave (4 peaks):				17.0	RPD = 14	
Corrected Ave (3 peaks):				16.8	Corrected Ave (3 peaks):				16.1	RPD = 4	
Aroclor-1221	1	4.730	-0.000	15803	266.6	1	4.955	-0.001	7909	262.9	
Aroclor-1221	2	6.131	-0.001	26946	254.1	2	6.296	-0.000	14303	251.2	
Aroclor-1221	3	6.382	-0.000	62477	253.8	3	6.622	0.000	23612	254.7	
Total CollAve (3 peaks):				258.2	Total Col2Ave (3 peaks):				256.3	RPD = 1	
Corrected Ave: < 3 Peaks					Corrected Ave: < 3 Peaks						
Aroclor-1232	1	4.730	0.000	15803	445.6	1	4.955	-0.001	7909	486.4	
Aroclor-1232	2	6.131	0.000	26946	383.1	2	7.256	0.002	3727	46.1	
Aroclor-1232	3	7.659	0.003	12893	40.5	3	7.863	0.002	5834	36.1	
Aroclor-1232	4	8.583	0.003	2684	19.8	4	8.716	0.002	1189	25.6	
Total CollAve (4 peaks):				222.3	Total Col2Ave (4 peaks):				148.5	RPD = 40	
Corrected Ave (3 peaks):				147.8	Corrected Ave (3 peaks):				35.9	RPD = 122*	
Aroclor-1242	1	7.269	-0.001	7175	35.0	1	7.256	0.000	3727	25.2	
Aroclor-1242	2	7.659	0.003	12893	20.7	2	7.863	0.005	5834	18.8	
Aroclor-1242	3	8.408	0.002	3610	18.6	3	9.175	0.008	1082	11.2	
Aroclor-1242	4	8.583	0.004	2684	9.4	4	9.543	-0.054	1390	11.8	
Total CollAve (4 peaks):				20.9	Total Col2Ave (4 peaks):				16.8	RPD = 22	
Corrected Ave (3 peaks):				16.2	Corrected Ave (3 peaks):				13.9	RPD = 15	
Aroclor-1248	1	8.408	0.002	3610	11.2	1	8.308	0.001	2045	13.5	
Aroclor-1248	2	8.583	0.003	2684	6.5	2	8.716	0.002	1189	7.6	
Aroclor-1248	3	8.994	-0.005	24440	31.6	3	9.175	0.009	1082	6.0	
Aroclor-1248	4	9.302	0.008	26328	66.8	4	9.543	-0.048	1390	6.4	
Total CollAve (4 peaks):				29.0	Total Col2Ave (4 peaks):				8.4	RPD = 110*	
Corrected Ave (3 peaks):				16.4	Corrected Ave (3 peaks):				6.7	RPD = 85*	
Aroclor-1254	1	9.302	0.004	26328	39.6	1	9.452	0.003	9571	39.6	
Aroclor-1254	2	---			0.0	2	9.972	0.002	1733	8.9	
Aroclor-1254	3	9.670	0.002	3721	8.7	3	10.147	0.023	49218	117.1	
Aroclor-1254	4	9.808	0.000	9653	11.6	4	10.370	-0.002	59603	145.4	
Aroclor-1254	5	10.120	-0.056	131179	251.9	5	10.569	0.001	79533	318.7	
Total CollAve (4 peaks):				78.0	Total Col2Ave (5 peaks):				125.9	RPD = 47*	
Corrected Ave (3 peaks):				20.0	Corrected Ave (4 peaks):				77.8	RPD = 118*	
Aroclor-1260	1	11.044	-0.000	223208	394.0	1	11.652	-0.001	104071	312.7	
Aroclor-1260	2	11.361	-0.001	190166	321.2	2	11.919	0.002	251579	296.2	
Aroclor-1260	3	11.737	0.003	458281	291.9	3	12.435	-0.001	113645	504.2	
Aroclor-1260	4	12.141	0.002	149720	189.4	4	12.501	-0.001	182951	319.6	
Aroclor-1260	5	12.244	0.000	196033	576.0	NS	---			----	
Total CollAve (5 peaks):				354.5	Total Col2Ave (4 peaks):				358.2	RPD = 1	
Corrected Ave (4 peaks):				299.1	Corrected Ave (3 peaks):				309.5	RPD = 3	
Aroclor-1262	1	10.828	-0.001	121431	251.3	1	11.201	0.000	121335	251.1	
Aroclor-1262	2	12.244	0.000	196033	249.3	2	11.652	0.000	104071	252.9	
Aroclor-1262	3	12.319	0.001	211092	249.8	3	12.435	0.001	113645	243.4	
Aroclor-1262	4	12.988	0.001	183455	237.5	4	12.501	-0.001	182951	250.1	
Total CollAve (4 peaks):				247.0	Total Col2Ave (4 peaks):				249.3	RPD = 1	
Corrected Ave (3 peaks):				245.5	Corrected Ave (3 peaks):				248.2	RPD = 1	
Aroclor-1268	1	12.244	-0.002	196033	97.1	1	12.435	0.003	113645	99.7	
Aroclor-1268	2	12.319	0.002	211092	105.6	2	12.501	0.001	182951	149.3	
Aroclor-1268	3	12.723	0.024	77240	45.2	3	12.891	-0.000	7755	7.4	
Aroclor-1268	4	13.488	-0.002	65479	11.6	4	13.709	0.000	35146	10.5	
Total CollAve (4 peaks):				64.9	Total Col2Ave (4 peaks):				66.7	RPD = 3	



Corrected Ave (3 peaks): 51.3      Corrected Ave (3 peaks): 39.2      RPD = 27

Total PCB Area Col1 (5.906 - 13.793) = 3239932      Col1 Total PCB = 0.4 ppm\*  
Total PCB Area Col2 (5.785 - 14.019) = 1655522      Col2 Total PCB = 0.4 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR2162SCV


24-FEB-2023 16:27, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 AR2162SCV

24-FEB-2023 16:27, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242318ECD7.D  
Data file 2: /230224.b/230224.b/02242318ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR3268SCV  
Client ID:  
Injection Date: 24-FEB-2023 16:48  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

ZB5 Col		ZB35 Col		ZB5	ZB35	RPD	Compound/Flag		
RT	Shift Response	RT	Shift Response	on col	on col				
5.806	0.000	363331	5.685	0.000	176204	37.1	38.2	2.9	Tetrachloro-m-xylene
13.894	0.001	800845	14.118	-0.001	488290	51.3	56.4	9.5	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	656592	-2.6
Hexabromobiphenyl	1429847	1584453	10.8

Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	314741	-0.2
Hexabromobiphenyl	513946	568346	10.6

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	7.270	-0.001	28327	113.6	1	7.254	-0.001	20651	112.1
Aroclor-1016	2	7.657	0.003	80668	106.1	2	7.861	0.005	41326	110.6
Aroclor-1016	3	7.793	0.003	40661	109.6	3	8.060	0.005	20446	121.2
Aroclor-1016	4	8.407	0.002	24680	102.9	4	8.308	0.001	13576	102.5
Total CollAve (4 peaks):				108.0	Total Col2Ave (4 peaks):				111.6	RPD = 3
Corrected Ave (3 peaks):				106.2	Corrected Ave (3 peaks):				108.4	RPD = 2
Aroclor-1221	1	4.729	-0.001	8535	145.1	1	4.956	-0.000	3965	133.1
Aroclor-1221	2	6.132	-0.000	15523	147.6	2	6.297	0.001	8689	154.1
Aroclor-1221	3	6.382	-0.000	45872	187.9	3	6.622	0.001	22272	242.6
Total CollAve (3 peaks):				160.2	Total Col2Ave (3 peaks):				176.6	RPD = 10
Corrected Ave: < 3 Peaks					Corrected Ave: < 3 Peaks					
Aroclor-1232	1	4.729	-0.001	8535	242.6	1	4.956	0.000	3965	246.2
Aroclor-1232	2	6.132	0.001	15523	222.5	2	7.254	0.000	20651	258.1
Aroclor-1232	3	7.657	0.001	80668	255.4	3	7.861	0.001	41326	258.3
Aroclor-1232	4	8.582	0.001	34784	259.2	4	8.714	-0.001	12504	271.5
Total CollAve (4 peaks):				244.9	Total Col2Ave (4 peaks):				258.5	RPD = 5
Corrected Ave (3 peaks):				240.2	Corrected Ave (3 peaks):				254.2	RPD = 6
Aroclor-1242	1	7.270	-0.001	28327	139.2	1	7.254	-0.001	20651	141.2
Aroclor-1242	2	7.657	0.001	80668	130.5	2	7.861	0.003	41326	134.4
Aroclor-1242	3	8.407	0.001	24680	128.4	3	9.170	0.003	12830	134.1
Aroclor-1242	4	8.582	0.003	34784	122.4	4	9.600	0.003	14836	127.3
Total CollAve (4 peaks):				130.1	Total Col2Ave (4 peaks):				134.3	RPD = 3
Corrected Ave (3 peaks):				127.1	Corrected Ave (3 peaks):				132.0	RPD = 4
Aroclor-1248	1	8.407	0.001	24680	77.0	1	8.308	0.000	13576	90.3
Aroclor-1248	2	8.582	0.001	34784	85.4	2	8.714	-0.000	12504	80.5
Aroclor-1248	3	8.996	-0.003	83592	108.8	3	9.170	0.004	12830	71.8
Aroclor-1248	4	9.292	-0.003	39603	101.3	4	9.600	0.010	14836	69.1
Total CollAve (4 peaks):				93.1	Total Col2Ave (4 peaks):				77.9	RPD = 18
Corrected Ave (3 peaks):				87.9	Corrected Ave (3 peaks):				73.8	RPD = 17
Aroclor-1254	1	9.292	-0.007	39603	60.1	1	9.452	0.003	4590	19.2
Aroclor-1254	2	9.377	-0.000	11450	38.6	2	9.973	0.003	2892	15.0
Aroclor-1254	3	9.674	0.005	6387	15.1	3	10.131	0.007	6052	14.5
Aroclor-1254	4	9.813	0.006	10162	12.3	4	10.390	0.017	5324	13.1
Aroclor-1254	5	10.189	0.012	6862	13.3	5	10.572	0.004	1891	7.7
Total CollAve (5 peaks):				27.9	Total Col2Ave (5 peaks):				13.9	RPD = 67*
Corrected Ave (4 peaks):				19.8	Corrected Ave (4 peaks):				12.6	RPD = 45*
Aroclor-1260	1	11.046	0.002	87033	152.7	1	11.645	-0.008	62543	187.1
Aroclor-1260	2	11.362	0.001	6300	10.6	2	11.920	0.003	28552	33.5
Aroclor-1260	3	11.738	0.004	54524	34.5	3	12.432	-0.004	285450	1261.2
Aroclor-1260	4	12.144	0.005	1727	2.2	4	12.499	-0.002	306992	534.0
Aroclor-1260	5	12.246	0.002	502931	1469.0	NS	---			----
Total CollAve (5 peaks):				333.8	Total Col2Ave (4 peaks):				503.9	RPD = 41*
Corrected Ave (4 peaks):				50.0	Corrected Ave (3 peaks):				251.5	RPD = 134*
Aroclor-1262	1	10.832	0.004	3395	7.0	1	11.201	0.001	44255	91.2
Aroclor-1262	2	12.246	0.002	502931	635.9	2	11.645	-0.007	62543	151.3
Aroclor-1262	3	12.318	-0.000	497006	584.5	3	12.432	-0.002	285450	608.7
Aroclor-1262	4	12.987	-0.000	202197	260.2	4	12.499	-0.003	306992	417.9
Total CollAve (4 peaks):				371.9	Total Col2Ave (4 peaks):				317.3	RPD = 16
Corrected Ave (3 peaks):				283.9	Corrected Ave (3 peaks):				220.1	RPD = 25
Aroclor-1268	1	12.246	-0.001	502931	247.7	1	12.432	-0.000	285450	249.4
Aroclor-1268	2	12.318	0.002	497006	247.2	2	12.499	-0.001	306992	249.5
Aroclor-1268	3	12.699	-0.000	422793	245.8	3	12.892	0.000	260893	248.4
Aroclor-1268	4	13.490	0.000	1386953	244.9	4	13.709	-0.000	829733	247.1
Total CollAve (4 peaks):				246.4	Total Col2Ave (4 peaks):				248.6	RPD = 1

Corrected Ave (3 peaks): 246.0      Corrected Ave (3 peaks): 248.3      RPD = 1

Total PCB Area Col1 (5.906 - 13.793) = 4180607      Col1 Total PCB = 0.5 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 2376912      Col2 Total PCB = 0.6 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR3268SCV


24-FEB-2023 16:48, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 AR3268SCV

24-FEB-2023 16:48, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
8082 DDT SCREEN REPORT

Data file 1: /230224.b/02242319ECD7.D

ARI ID: DDTS

RT	ZB5 Col Shift Response	ZB35 Col Shift Response	RT	ZB5 on col	ZB35 on col	RPD	Compound/Flag
9.261	0.000 694353	9.912 0.000 580269	0.100	0.100	0.0	2,4-DDE	
0.000	-10.293 0	10.672 0.000 673479	0.000	0.200#	----	2,4-DDT	
9.686	0.000 1191406	10.212 0.000 433373	0.100	0.100	0.0	4,4-DDE	
10.259	0.000 1721760	10.672 0.000 673479	0.100	0.200#	66.7*	4,4-DDD	

# Indicates value is from co-eluting peaks

\* Indicates RPD > 40%

Analytical Resources Inc.  
8082 DDT SCREEN REPORT

Data file 1: /230224.b/02242320ECD7.D

ARI ID: DDT BD

RT	ZB5 Col Shift Response		RT	ZB35 Col Shift Response		ZB5 on col	ZB35 on col	RPD	Compound/Flag
9.285	0.023	4923	9.921	0.009	9972	0.001	0.002	84.3*	2,4-DDE
0.000	-10.293	0	10.677	0.004	249094	0.000	0.074#	----	2,4-DDT
9.692	0.006	12128	10.221	0.009	528	0.001	0.000	156.7*	4,4-DDE
10.265	0.006	410017	10.677	0.004	249094	0.023	0.074#	103.6*	4,4-DDD

# Indicates value is from co-eluting peaks

\* Indicates RPD > 40%





## GC LOG SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b

	Inject	Date/Time	Filename	DF	LabID	ClientID
1	24-FEB-2023	10:51	02242301ECD7.D	1	IB	
2	24-FEB-2023	11:12	02242302ECD7.D	1	0.25PPMAR1660	
3	24-FEB-2023	11:33	02242303ECD7.D	1	0.02PPMAR1660	
4	24-FEB-2023	11:54	02242304ECD7.D	1	0.05PPMAR1660	
5	24-FEB-2023	12:15	02242305ECD7.D	1	1.0PPMAR1660	
6	24-FEB-2023	12:36	02242306ECD7.D	1	0.1PPMAR1660	
7	24-FEB-2023	12:57	02242307ECD7.D	1	0.5PPMAR1660	
8	24-FEB-2023	13:18	02242308ECD7.D	1	0.25PPMAR1242	
9	24-FEB-2023	13:39	02242309ECD7.D	1	0.25PPMAR1248	
10	24-FEB-2023	14:00	02242310ECD7.D	1	0.25PPMAR1254	
11	24-FEB-2023	14:21	02242311ECD7.D	1	0.25PPMAR2162	
12	24-FEB-2023	14:42	02242312ECD7.D	1	0.25PPMAR3268	
13	24-FEB-2023	15:03	02242313ECD7.D	1	AR1660SCV	
14	24-FEB-2023	15:24	02242314ECD7.D	1	AR1242SCV	
15	24-FEB-2023	15:45	02242315ECD7.D	1	AR1248SCV	
16	24-FEB-2023	16:06	02242316ECD7.D	1	AR1254SCV	
17	24-FEB-2023	16:27	02242317ECD7.D	1	AR2162SCV	
18	24-FEB-2023	16:48	02242318ECD7.D	1	AR3268SCV	
19	24-FEB-2023	17:09	02242319ECD7.D	1	DDTS	
20	24-FEB-2023	17:30	02242320ECD7.D	1	DDT BD	

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b

ARI Job No.: IB Method: PCB.m Instrument: ecd7.i Date: 24-FEB-2023

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1051	02242301ECD7.D	IB		1	NO MANUAL INTEGRATION
1112	02242302ECD7.D	0.25PPMAR1660		1	NO MANUAL INTEGRATION
1133	02242303ECD7.D	0.02PPMAR1660		1	NO MANUAL INTEGRATION
1154	02242304ECD7.D	0.05PPMAR1660		1	NO MANUAL INTEGRATION
1215	02242305ECD7.D	1.0PPMAR1660		1	NO MANUAL INTEGRATION
1236	02242306ECD7.D	0.1PPMAR1660		1	NO MANUAL INTEGRATION
1257	02242307ECD7.D	0.5PPMAR1660		1	NO MANUAL INTEGRATION
1318	02242308ECD7.D	0.25PPMAR1242		1	NO MANUAL INTEGRATION
1339	02242309ECD7.D	0.25PPMAR1248		1	NO MANUAL INTEGRATION
1400	02242310ECD7.D	0.25PPMAR1254		1	NO MANUAL INTEGRATION
1421	02242311ECD7.D	0.25PPMAR2162		1	NO MANUAL INTEGRATION
1442	02242312ECD7.D	0.25PPMAR3268		1	NO MANUAL INTEGRATION
1503	02242313ECD7.D	AR1660SCV		1	NO MANUAL INTEGRATION
1524	02242314ECD7.D	AR1242SCV		1	NO MANUAL INTEGRATION
1545	02242315ECD7.D	AR1248SCV		1	NO MANUAL INTEGRATION
1606	02242316ECD7.D	AR1254SCV		1	NO MANUAL INTEGRATION
1627	02242317ECD7.D	AR2162SCV		1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1648	02242318ECD7.D	AR3268SCV		1	NO MANUAL INTEGRATION
1709	02242319ECD7.D	DDTS		1	NO MANUAL INTEGRATION
1730	02242320ECD7.D	DDT BD		1	NO MANUAL INTEGRATION
1751	02242321ECD7.D			1	NO MANUAL INTEGRATION
1812	02242322ECD7.D			1	NO MANUAL INTEGRATION
1833	02242323ECD7.D			1	NO MANUAL INTEGRATION
1854	02242324ECD7.D			1	NO MANUAL INTEGRATION
1915	02242325ECD7.D			1	NO MANUAL INTEGRATION
1936	02242326ECD7.D			1	NO MANUAL INTEGRATION
1957	02242327ECD7.D			1	NO MANUAL INTEGRATION
2018	02242328ECD7.D			1	NO MANUAL INTEGRATION
2039	02242329ECD7.D			1	NO MANUAL INTEGRATION
2059	02242330ECD7.D			1	NO MANUAL INTEGRATION
2120	02242331ECD7.D			1	NO MANUAL INTEGRATION
2141	02242332ECD7.D			1	NO MANUAL INTEGRATION
2202	02242333ECD7.D			1	NO MANUAL INTEGRATION
2223	02242334ECD7.D			1	NO MANUAL INTEGRATION
2244	02242335ECD7.D			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
2305	02242336ECD7.D			1	NO MANUAL INTEGRATION
2326	02242337ECD7.D			1	NO MANUAL INTEGRATION
2347	02242338ECD7.D			1	NO MANUAL INTEGRATION
0008	02242339ECD7.D			1	NO MANUAL INTEGRATION
0029	02242340ECD7.D			1	NO MANUAL INTEGRATION
0050	02242341ECD7.D			1	NO MANUAL INTEGRATION
0111	02242342ECD7.D			1	NO MANUAL INTEGRATION
0132	02242343ECD7.D			1	NO MANUAL INTEGRATION
0153	02242344ECD7.D			1	NO MANUAL INTEGRATION
0214	02242345ECD7.D			1	NO MANUAL INTEGRATION
0235	02242346ECD7.D			1	NO MANUAL INTEGRATION
0256	02242347ECD7.D			1	NO MANUAL INTEGRATION
0317	02242348ECD7.D			1	NO MANUAL INTEGRATION
0338	02242349ECD7.D			1	NO MANUAL INTEGRATION
0359	02242350ECD7.D			1	NO MANUAL INTEGRATION
0420	02242351ECD7.D			1	NO MANUAL INTEGRATION
0441	02242352ECD7.D			1	NO MANUAL INTEGRATION
0502	02242353ECD7.D			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
0523	02242354ECD7.D			1	NO MANUAL INTEGRATION
0544	02242355ECD7.D			1	NO MANUAL INTEGRATION
0605	02242356ECD7.D			1	NO MANUAL INTEGRATION
0626	02242357ECD7.D			1	NO MANUAL INTEGRATION
0647	02242358ECD7.D			1	NO MANUAL INTEGRATION
0708	02242359ECD7.D			1	NO MANUAL INTEGRATION
0729	02242360ECD7.D			1	NO MANUAL INTEGRATION
0750	02242361ECD7.D			1	NO MANUAL INTEGRATION
0811	02242362ECD7.D			1	NO MANUAL INTEGRATION
0832	02242363ECD7.D			1	NO MANUAL INTEGRATION
0853	02242364ECD7.D			1	NO MANUAL INTEGRATION
0914	02242365ECD7.D			1	NO MANUAL INTEGRATION
0935	02242366ECD7.D			1	NO MANUAL INTEGRATION
0956	02242367ECD7.D			1	NO MANUAL INTEGRATION
1017	02242368ECD7.D			1	NO MANUAL INTEGRATION
1038	02242369ECD7.D			1	NO MANUAL INTEGRATION
1059	02242370ECD7.D			1	NO MANUAL INTEGRATION
1120	02242371ECD7.D			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1141	02242372ECD7.D			1	NO MANUAL INTEGRATION
1202	02242373ECD7.D			1	NO MANUAL INTEGRATION
1223	02242374ECD7.D			1	NO MANUAL INTEGRATION
1244	02242375ECD7.D			1	NO MANUAL INTEGRATION
1305	02242376ECD7.D			1	NO MANUAL INTEGRATION
1326	02242377ECD7.D			1	NO MANUAL INTEGRATION
1347	02242378ECD7.D			1	NO MANUAL INTEGRATION
1408	02242379ECD7.D			1	NO MANUAL INTEGRATION
1429	02242380ECD7.D			1	NO MANUAL INTEGRATION
1450	02242381ECD7.D			1	NO MANUAL INTEGRATION
1511	02242382ECD7.D			1	NO MANUAL INTEGRATION
1532	02242383ECD7.D			1	NO MANUAL INTEGRATION
1553	02242384ECD7.D			1	NO MANUAL INTEGRATION
1051	02242301ECD7.D IB			1	NO MANUAL INTEGRATION
1112	02242302ECD7.D 0.25PPMAR1660			1	NO MANUAL INTEGRATION
1133	02242303ECD7.D 0.02PPMAR1660			1	Aroclor-1016 [2C],
1154	02242304ECD7.D 0.05PPMAR1660			1	NO MANUAL INTEGRATION
1215	02242305ECD7.D 1.0PPMAR1660			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1236	02242306ECD7.D	0.1PPMAR1660		1	NO MANUAL INTEGRATION
1257	02242307ECD7.D	0.5PPMAR1660		1	NO MANUAL INTEGRATION
1318	02242308ECD7.D	0.25PPMAR1242		1	NO MANUAL INTEGRATION
1339	02242309ECD7.D	0.25PPMAR1248		1	NO MANUAL INTEGRATION
1400	02242310ECD7.D	0.25PPMAR1254		1	NO MANUAL INTEGRATION
1421	02242311ECD7.D	0.25PPMAR2162		1	NO MANUAL INTEGRATION
1442	02242312ECD7.D	0.25PPMAR3268		1	NO MANUAL INTEGRATION
1503	02242313ECD7.D	AR1660SCV		1	NO MANUAL INTEGRATION
1524	02242314ECD7.D	AR1242SCV		1	NO MANUAL INTEGRATION
1545	02242315ECD7.D	AR1248SCV		1	NO MANUAL INTEGRATION
1606	02242316ECD7.D	AR1254SCV		1	NO MANUAL INTEGRATION
1627	02242317ECD7.D	AR2162SCV		1	NO MANUAL INTEGRATION
1648	02242318ECD7.D	AR3268SCV		1	NO MANUAL INTEGRATION
1709	02242319ECD7.D	DDTS		1	NO MANUAL INTEGRATION
1730	02242320ECD7.D	DDT BD		1	NO MANUAL INTEGRATION
1751	02242321ECD7.D			1	NO MANUAL INTEGRATION
1812	02242322ECD7.D			1	NO MANUAL INTEGRATION
1833	02242323ECD7.D			1	NO MANUAL INTEGRATION



MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1854	02242324ECD7.D			1	NO MANUAL INTEGRATION
1915	02242325ECD7.D			1	NO MANUAL INTEGRATION
1936	02242326ECD7.D			1	NO MANUAL INTEGRATION
1957	02242327ECD7.D			1	NO MANUAL INTEGRATION
2018	02242328ECD7.D			1	NO MANUAL INTEGRATION
2038	02242329ECD7.D			1	NO MANUAL INTEGRATION
2059	02242330ECD7.D			1	NO MANUAL INTEGRATION
2120	02242331ECD7.D			1	NO MANUAL INTEGRATION
2141	02242332ECD7.D			1	NO MANUAL INTEGRATION
2202	02242333ECD7.D			1	NO MANUAL INTEGRATION
2223	02242334ECD7.D			1	NO MANUAL INTEGRATION
2244	02242335ECD7.D			1	NO MANUAL INTEGRATION
2305	02242336ECD7.D			1	NO MANUAL INTEGRATION
2326	02242337ECD7.D			1	NO MANUAL INTEGRATION
2347	02242338ECD7.D			1	NO MANUAL INTEGRATION
0008	02242339ECD7.D			1	NO MANUAL INTEGRATION
0029	02242340ECD7.D			1	NO MANUAL INTEGRATION
0050	02242341ECD7.D			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
0111	02242342ECD7.D			1	NO MANUAL INTEGRATION
0132	02242343ECD7.D			1	NO MANUAL INTEGRATION
0153	02242344ECD7.D			1	NO MANUAL INTEGRATION
0214	02242345ECD7.D			1	NO MANUAL INTEGRATION
0235	02242346ECD7.D			1	NO MANUAL INTEGRATION
0256	02242347ECD7.D			1	NO MANUAL INTEGRATION
0317	02242348ECD7.D			1	NO MANUAL INTEGRATION
0338	02242349ECD7.D			1	NO MANUAL INTEGRATION
0359	02242350ECD7.D			1	NO MANUAL INTEGRATION
0420	02242351ECD7.D			1	NO MANUAL INTEGRATION
0441	02242352ECD7.D			1	NO MANUAL INTEGRATION
0502	02242353ECD7.D			1	NO MANUAL INTEGRATION
0523	02242354ECD7.D			1	NO MANUAL INTEGRATION
0544	02242355ECD7.D			1	NO MANUAL INTEGRATION
0605	02242356ECD7.D			1	NO MANUAL INTEGRATION
0626	02242357ECD7.D			1	NO MANUAL INTEGRATION
0647	02242358ECD7.D			1	NO MANUAL INTEGRATION
0708	02242359ECD7.D			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
0729	02242360ECD7.D			1	NO MANUAL INTEGRATION
0750	02242361ECD7.D			1	NO MANUAL INTEGRATION
0811	02242362ECD7.D			1	NO MANUAL INTEGRATION
0832	02242363ECD7.D			1	NO MANUAL INTEGRATION
0853	02242364ECD7.D			1	NO MANUAL INTEGRATION
0914	02242365ECD7.D			1	NO MANUAL INTEGRATION
0935	02242366ECD7.D			1	NO MANUAL INTEGRATION
0956	02242367ECD7.D			1	NO MANUAL INTEGRATION
1017	02242368ECD7.D			1	NO MANUAL INTEGRATION
1038	02242369ECD7.D			1	NO MANUAL INTEGRATION
1059	02242370ECD7.D			1	NO MANUAL INTEGRATION
1120	02242371ECD7.D			1	NO MANUAL INTEGRATION
1141	02242372ECD7.D			1	NO MANUAL INTEGRATION
1202	02242373ECD7.D			1	NO MANUAL INTEGRATION
1223	02242374ECD7.D			1	NO MANUAL INTEGRATION
1244	02242375ECD7.D			1	NO MANUAL INTEGRATION
1305	02242376ECD7.D			1	NO MANUAL INTEGRATION
1326	02242377ECD7.D			1	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230224.b\230224.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1347	02242378ECD7.D			1	NO MANUAL INTEGRATION
1408	02242379ECD7.D			1	NO MANUAL INTEGRATION
1429	02242380ECD7.D			1	NO MANUAL INTEGRATION
1450	02242381ECD7.D			1	NO MANUAL INTEGRATION
1511	02242382ECD7.D			1	NO MANUAL INTEGRATION
1532	02242383ECD7.D			1	NO MANUAL INTEGRATION
1553	02242384ECD7.D			1	NO MANUAL INTEGRATION

Security Status Report

Date: 28-Feb-2023 10:53

02242301ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242302ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242303ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242304ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242305ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242306ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242307ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242308ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242309ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242310ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242311ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242312ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242313ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242314ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242315ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242316ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242317ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242318ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242319ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53
02242320ECD7.D	Data Locked	richardl, 28-Feb-2023 10:53

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m  
 Last Edit : 24-Feb-2023 15:31 richardl  
 Curve Type : Average

Calibration File Names:

Level 1: \\target\share\chem4\ecd7.i\230224.b\02242303ECD7.D  
 Level 2: \\target\share\chem4\ecd7.i\230224.b\02242304ECD7.D  
 Level 3: \\target\share\chem4\ecd7.i\230224.b\02242306ECD7.D  
 Level 4: \\target\share\chem4\ecd7.i\230224.b\02242302ECD7.D  
 Level 5: \\target\share\chem4\ecd7.i\230224.b\02242307ECD7.D  
 Level 6: \\target\share\chem4\ecd7.i\230224.b\02242305ECD7.D  
 Level 7: \\target\share\chem4\ecd7.i\230224.b\02242312ECD7.D

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
2 Aroclor-1221 (1)	+++++ 0.00716	+++++	+++++	+++++	+++++	+++++	0.00716	0.000
(2)	+++++ 0.01281	+++++	+++++	+++++	+++++	+++++	0.01281	0.000
(3)	+++++ 0.02975	+++++	+++++	+++++	+++++	+++++	0.02975	0.000
3 Aroclor-1242 (1)	+++++ 0.02479	+++++	+++++	+++++	+++++	+++++	0.02479	0.000
(2)	+++++ 0.07529	+++++	+++++	+++++	+++++	+++++	0.07529	0.000
(3)	+++++ 0.02343	+++++	+++++	+++++	+++++	+++++	0.02343	0.000
(4)	+++++ 0.03463	+++++	+++++	+++++	+++++	+++++	0.03463	0.000
4 Aroclor-1232 (1)	+++++ 0.00429	+++++	+++++	+++++	+++++	+++++	0.00429	0.000

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m  
 Last Edit : 24-Feb-2023 15:31 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
(2)	++++ 0.00850	++++	++++	++++	++++	++++	0.00850	0.000
(3)	++++ 0.03848	++++	++++	++++	++++	++++	0.03848	0.000
(4)	++++ 0.01635	++++	++++	++++	++++	++++	0.01635	0.000
7 Aroclor-1016(1)	0.03172 ++++	0.03253	0.03142	0.03141	0.02856	0.02667	0.03039	7.449
(2)	0.09239 ++++	0.09246	0.09222	0.09849	0.09174	0.08849	0.09263	3.499
(3)	0.05165 ++++	0.05037	0.04823	0.04393	0.03991	0.03721	0.04522	12.936
(4)	0.03002 ++++	0.02894	0.02959	0.03058	0.02852	0.02774	0.02923	3.542
6 Aroclor-1248(1)	++++ 0.03903	++++	++++	++++	++++	++++	0.03903	0.000
(2)	++++ 0.04961	++++	++++	++++	++++	++++	0.04961	0.000
(3)	++++ 0.09360	++++	++++	++++	++++	++++	0.09360	0.000
(4)	++++ 0.04765	++++	++++	++++	++++	++++	0.04765	0.000

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m  
 Last Edit : 24-Feb-2023 15:31 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
8 Aroclor-1254 (1)	++++ 0.08033	++++	++++	++++	++++	++++	0.08033	0.000
(2)	++++ 0.03613	++++	++++	++++	++++	++++	0.03613	0.000
(3)	++++ 0.05165	++++	++++	++++	++++	++++	0.05165	0.000
(4)	++++ 0.10042	++++	++++	++++	++++	++++	0.10042	0.000
(5)	++++ 0.06294	++++	++++	++++	++++	++++	0.06294	0.000
9 Aroclor-1260 (1)	0.02926 ++++	0.02920	0.02841	0.03096	0.02737	0.02746	0.02878	4.677
(2)	0.02967 ++++	0.03006	0.03011	0.03291	0.02910	0.02857	0.03007	5.029
(3)	0.08088 ++++	0.08045	0.07954	0.08575	0.07515	0.07674	0.07975	4.627
(4)	0.03905 ++++	0.03887	0.03955	0.04485	0.03942	0.03922	0.04016	5.753
(5)	0.01783 ++++	0.01715	0.01679	0.01875	0.01664	0.01655	0.01729	4.953
10 Aroclor-1262 (1)	++++ 0.02454	++++	++++	++++	++++	++++	0.02454	0.000



ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m  
 Last Edit : 24-Feb-2023 15:31 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
(2)	++++ 0.03993	++++	++++	++++	++++	++++	0.03993	0.000
(3)	++++ 0.04293	++++	++++	++++	++++	++++	0.04293	0.000
(4)	++++ 0.03923	++++	++++	++++	++++	++++	0.03923	0.000
11 Aroclor-1268(1)	++++ 0.10250	++++	++++	++++	++++	++++	0.10250	0.000
(2)	++++ 0.10151	++++	++++	++++	++++	++++	0.10151	0.000
(3)	++++ 0.08686	++++	++++	++++	++++	++++	0.08686	0.000
(4)	++++ 0.28598	++++	++++	++++	++++	++++	0.28598	0.000
42 2,4-DDE	++++ ++++	++++	++++	++++	++++	++++	++++	++++
43 2,4-DDD	++++ ++++	++++	++++	++++	++++	++++	++++	++++
44 2,4-DDT	++++ ++++	++++	++++	++++	++++	++++	++++	++++
46 4,4-DDE	++++ ++++	++++	++++	++++	++++	++++	++++	++++

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m  
 Last Edit : 24-Feb-2023 15:31 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	250.000 Level 7	RRF	% RSD
47 4,4-DDD	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
48 4,4-DDT	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
49 Hexachlorobutadiene	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
50 Hexachlorobenzene	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
1 Tetrachloro-m-xylene	1.16827	1.24402	1.18546	1.20509	1.12295	1.24114		1.19449	3.860
13 Decachlorobiphenyl	0.82901	0.80558	0.77587	0.78808	0.73125	0.79742		0.78787	4.189

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m  
 Last Edit : 24-Feb-2023 15:29 richardl  
 Curve Type : Average

Calibration File Names:

Level 1: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242303ECD7.D  
 Level 2: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242304ECD7.D  
 Level 3: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242306ECD7.D  
 Level 4: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242302ECD7.D  
 Level 5: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242307ECD7.D  
 Level 6: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242305ECD7.D  
 Level 7: \\target\share\chem4\ecd7.i\230224.b\230224.b\02242312ECD7.D

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
1 Aroclor-1221 [2C] (1)	+++++ 0.00757	+++++	+++++	+++++	+++++	+++++	0.00757	0.000
(2)	+++++ 0.01433	+++++	+++++	+++++	+++++	+++++	0.01433	0.000
(3)	+++++ 0.02333	+++++	+++++	+++++	+++++	+++++	0.02333	0.000
4 Aroclor-1232 [2C] (1)	+++++ 0.00409	+++++	+++++	+++++	+++++	+++++	0.00409	0.000
(2)	+++++ 0.02034	+++++	+++++	+++++	+++++	+++++	0.02034	0.000
(3)	+++++ 0.04067	+++++	+++++	+++++	+++++	+++++	0.04067	0.000
(4)	+++++ 0.01170	+++++	+++++	+++++	+++++	+++++	0.01170	0.000
3 Aroclor-1242 [2C] (1)	+++++ 0.03717	+++++	+++++	+++++	+++++	+++++	0.03717	0.000

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m  
 Last Edit : 24-Feb-2023 15:29 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
(2)	++++ 0.07813	++++	++++	++++	++++	++++	0.07813	0.000
(3)	++++ 0.02431	++++	++++	++++	++++	++++	0.02431	0.000
(4)	++++ 0.02962	++++	++++	++++	++++	++++	0.02962	0.000
6 Aroclor-1248 [2C] (1)	++++ 0.03820	++++	++++	++++	++++	++++	0.03820	0.000
(2)	++++ 0.03949	++++	++++	++++	++++	++++	0.03949	0.000
(3)	++++ 0.04545	++++	++++	++++	++++	++++	0.04545	0.000
(4)	++++ 0.05457	++++	++++	++++	++++	++++	0.05457	0.000
7 Aroclor-1016 [2C] (1)	0.05071 ++++	0.05022	0.04868	0.04733	0.04326	0.04080	0.04683	8.503
(2)	0.08143 ++++	0.09407	0.10159	0.10259	0.09651	0.09362	0.09497	8.025
(3)	0.04006 ++++	0.04718	0.04613	0.04410	0.04062	0.03926	0.04289	7.857
(4)	0.03181 ++++	0.03802	0.03707	0.03450	0.03115	0.02936	0.03365	10.251

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m  
 Last Edit : 24-Feb-2023 15:29 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
8 Aroclor-1254 [2C] (1)	++++ 0.06081	++++	++++	++++	++++	++++	0.06081	0.000
(2)	++++ 0.04892	++++	++++	++++	++++	++++	0.04892	0.000
(3)	++++ 0.10584	++++	++++	++++	++++	++++	0.10584	0.000
(4)	++++ 0.10317	++++	++++	++++	++++	++++	0.10317	0.000
(5)	++++ 0.06282	++++	++++	++++	++++	++++	0.06282	0.000
10 Aroclor-1262 [2C] (1)	++++ 0.06831	++++	++++	++++	++++	++++	0.06831	0.000
(2)	++++ 0.05818	++++	++++	++++	++++	++++	0.05818	0.000
(3)	++++ 0.06601	++++	++++	++++	++++	++++	0.06601	0.000
(4)	++++ 0.10341	++++	++++	++++	++++	++++	0.10341	0.000
9 Aroclor-1260 [2C] (1)	0.05286 ++++	0.04911	0.04696	0.04801	0.04329	0.04201	0.04704	8.422
(2)	0.12976 ++++	0.12431	0.12095	0.12664	0.11320	0.10545	0.12005	7.605

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m  
 Last Edit : 24-Feb-2023 15:29 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
(3)	0.03524 ++++	0.03147	0.02937	0.03208	0.03102	0.03198	0.03186	6.045
(4)	0.08632 ++++	0.08237	0.08044	0.08393	0.07718	0.07531	0.08092	5.126
11 Aroclor-1268 [2C] (1)	++++ 0.16109	++++	++++	++++	++++	++++	0.16109	0.000
(2)	++++ 0.17318	++++	++++	++++	++++	++++	0.17318	0.000
(3)	++++ 0.14787	++++	++++	++++	++++	++++	0.14787	0.000
(4)	++++ 0.47260	++++	++++	++++	++++	++++	0.47260	0.000
41 2,4-DDE [2C]	++++ ++++	++++	++++	++++	++++	++++	++++	++++
42 2,4-DDD [2C]	++++ ++++	++++	++++	++++	++++	++++	++++	++++
44 4,4-DDE [2C]	++++ ++++	++++	++++	++++	++++	++++	++++	++++
45 4,4-DDD/2,4-DDT [2C]	++++ ++++	++++	++++	++++	++++	++++	++++	++++
46 4,4-DDT [2C]	++++ ++++	++++	++++	++++	++++	++++	++++	++++

ARI Labs, Inc.

INITIAL CALIBRATION DATA

Start Cal Date : 24-FEB-2023 11:12  
 End Cal Date : 24-FEB-2023 14:42  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 4.14  
 Integrator : HP Genie  
 Method file : \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m  
 Last Edit : 24-Feb-2023 15:29 richardl  
 Curve Type : Average

Compound	20.000 Level 1	50.000 Level 2	100.000 Level 3	250.000 Level 4	500.000 Level 5	1000.000 Level 6	RRF	% RSD
48 Hexachlorobutadiene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
49 Hexachlorobenzene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
\$ 2 Tetrachloro-m-xylene [2C]	1.21526 +++++	1.19545	1.17555	1.21907	1.12560	1.11139	1.17372	3.897
\$ 13 Decachlorobiphenyl [2C]	1.17066 +++++	1.20406	1.20549	1.31040	1.21104	1.20797	1.21827	3.898

ARI Labs, Inc.  
RETENTION TIME SUMMARY REPORT

Method File: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Batch File: \\target\share\chem4\ecd7.i\230224.b  
Inst ID: ecd7.i

ID:	RT01	RT02	RT03	RT04	RT05	RT06
FILENAME:	02242302ECD7	02242303ECD7	02242304ECD7	02242305ECD7	02242306ECD7	02242307ECD7
INJ. DATE:	24-FEB-2023	24-FEB-2023	24-FEB-2023	24-FEB-2023	24-FEB-2023	24-FEB-2023
INJ. TIME:	11:12	11:33	11:54	12:15	12:36	12:57

Compound	RT01	RT02	RT03	RT04	RT05	RT06	EXPEC RT	RT WINDOW	AVG RT	STD DEV
* 41 IS-BNB	3.493	3.492	3.492	3.492	3.491	3.491	3.493	3.393-3.593	3.492	0.001
§ 1 Tetrachloro-m-xylene	5.811	5.809	5.809	5.813	5.809	5.810	5.811	5.711-5.911	5.810	0.002
2 Aroclor-1221	+++++	+++++	+++++	+++++	+++++	+++++	4.732	4.632-4.832	+++++	+++++
3 Aroclor-1242	+++++	+++++	+++++	+++++	+++++	+++++	7.269	7.169-7.369	+++++	+++++
4 Aroclor-1232	+++++	+++++	+++++	+++++	+++++	+++++	4.732	4.632-4.832	+++++	+++++
7 Aroclor-1016	7.272	7.272	7.272	7.270	7.271	7.270	7.272	7.172-7.372	7.271	0.001
6 Aroclor-1248	+++++	+++++	+++++	+++++	+++++	+++++	8.403	8.303-8.503	+++++	+++++
8 Aroclor-1254	+++++	+++++	+++++	+++++	+++++	+++++	9.295	9.195-9.395	+++++	+++++
9 Aroclor-1260	11.046	11.047	11.046	11.044	11.045	11.044	11.046	10.946-11.146	11.045	0.001
10 Aroclor-1262	+++++	+++++	+++++	+++++	+++++	+++++	10.824	10.724-10.924	+++++	+++++
11 Aroclor-1268	+++++	+++++	+++++	+++++	+++++	+++++	12.243	12.143-12.343	+++++	+++++
§ 13 Decachlorobiphenyl	13.897	13.893	13.893	13.899	13.892	13.898	13.897	13.797-13.997	13.895	0.003
* 12 IS-HBBP	14.269	14.268	14.268	14.267	14.268	14.268	14.269	14.169-14.369	14.268	0.001
42 2,4-DDE	+++++	+++++	+++++	+++++	+++++	+++++	9.260	9.210-9.310	+++++	+++++
43 2,4-DDD	+++++	+++++	+++++	+++++	+++++	+++++	9.801	9.751-9.851	+++++	+++++
44 2,4-DDT	+++++	+++++	+++++	+++++	+++++	+++++	10.293	10.243-10.343	+++++	+++++
46 4,4-DDE	+++++	+++++	+++++	+++++	+++++	+++++	9.683	9.583-9.783	+++++	+++++

Reviewer 1 \_\_\_\_\_ Date: \_\_\_\_\_  
Reviewer 2 \_\_\_\_\_ Date: \_\_\_\_\_



ARI Labs, Inc.  
RETENTION TIME SUMMARY REPORT

Method File: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
 Batch File: \\target\share\chem4\ecd7.i\230224.b  
 Inst ID: ecd7.i

Compound	RT01	RT02	RT03	RT04	RT05	RT06	EXPEC RT	RT WINDOW	AVG RT	STD DEV
47 4,4-DDD	+++++	+++++	+++++	+++++	+++++	+++++	10.254	10.154-10.354	+++++	+++++
48 4,4-DDT	+++++	+++++	+++++	+++++	+++++	+++++	10.754	10.654-10.854	+++++	+++++
49 Hexachlorobutadiene	+++++	+++++	+++++	+++++	+++++	+++++	1.842	1.742-1.942	+++++	+++++
50 Hexachlorobenzene	+++++	+++++	+++++	+++++	+++++	+++++	6.708	6.608-6.808	+++++	+++++

ARI Labs, Inc.
RETENTION TIME SUMMARY REPORT

Method File: \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m
Batch File: \\target\share\chem4\ecd7.i\230224.b\230224.b
Inst ID: ecd7.i

ID: RT01 RT02 RT03 RT04 RT05 RT06
FILENAME: 02242302ECD7 02242303ECD7 02242304ECD7 02242305ECD7 02242306ECD7 02242307ECD7
INJ. DATE: 24-FEB-2023 24-FEB-2023 24-FEB-2023 24-FEB-2023 24-FEB-2023 24-FEB-2023
INJ. TIME: 11:12 11:33 11:54 12:15 12:36 12:57

Table with 11 columns: Compound, RT01, RT02, RT03, RT04, RT05, RT06, EXPEC RT, RT WINDOW, AVG RT, STD DEV. Rows include various chemical compounds like Aroclor-1221, Aroclor-1232, etc., with their respective retention times and standard deviations.

Reviewer 1 \_\_\_\_\_ Date: \_\_\_\_\_
Reviewer 2 \_\_\_\_\_ Date: \_\_\_\_\_

ARI Labs, Inc.  
RETENTION TIME SUMMARY REPORT

Method File: \\target\share\chem4\ecd7.i\230224.b\PCB.m\PCB2.m  
 Batch File: \\target\share\chem4\ecd7.i\230224.b\230224.b  
 Inst ID: ecd7.i

Compound	RT01	RT02	RT03	RT04	RT05	RT06	EXPEC RT	RT WINDOW	AVG RT	STD DEV
46 4,4-DDT [2C]	+++++	+++++	+++++	+++++	+++++	+++++	11.092	10.992-11.192	+++++	+++++
48 Hexachlorobutadiene	+++++	+++++	+++++	+++++	+++++	+++++	1.703	1.603-1.803	+++++	+++++
49 Hexachlorobenzene	+++++	+++++	+++++	+++++	+++++	+++++	7.178	7.078-7.278	+++++	+++++

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242301ECD7.D  
Data file 2: /230224.b/230224.b/02242301ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: IB  
Client ID:  
Injection Date: 24-FEB-2023 10:51  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

ZB5 Col		ZB35 Col			ZB5	ZB35	RPD	Compound/Flag	
RT	Shift	Response	RT	Shift	Response	on col			on col
5.826	0.019	382217	5.683	-0.002	180378	33.8	36.5	7.7	Tetrachloro-m-xylene
13.904	0.011	534110	14.120	0.001	295605	35.3	37.2	5.2	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	756896	12.3
Hexabromobiphenyl	1429847	1534275	7.3

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	336543	6.8
Hexabromobiphenyl	513946	521508	1.5

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	---			0.0	1	---			0.0
Aroclor-1016	2	---			0.0	2	---			0.0
Aroclor-1016	3	---			0.0	3	---			0.0
Aroclor-1016	4	---			0.0	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1221	1	---			0.0	1	---			0.0
Aroclor-1221	2	---			0.0	2	6.321	0.025	1873	31.1
Aroclor-1221	3	---			0.0	3	6.633	0.012	314	3.2
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1232	1	---			0.0	1	---			0.0
Aroclor-1232	2	---			0.0	2	---			0.0
Aroclor-1232	3	7.698	0.043	2193	6.0	3	---			0.0
Aroclor-1232	4	8.505	-0.076	11525	74.5	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1242	1	---			0.0	1	---			0.0
Aroclor-1242	2	7.698	0.042	2193	3.1	2	---			0.0
Aroclor-1242	3	---			0.0	3	---			0.0
Aroclor-1242	4	8.505	-0.074	11525	35.2	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1248	1	---			0.0	1	---			0.0
Aroclor-1248	2	---			0.0	2	---			0.0
Aroclor-1248	3	---			0.0	3	---			0.0
Aroclor-1248	4	---			0.0	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1254	1	---			0.0	1	---			0.0
Aroclor-1254	2	---			0.0	2	---			0.0
Aroclor-1254	3	9.596	-0.072	31424	64.3	3	---			0.0
Aroclor-1254	4	---			0.0	4	---			0.0
Aroclor-1254	5	10.167	-0.010	18361	30.8	5	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1260	1	11.098	0.054	6994	12.7	1	---			0.0
Aroclor-1260	2	---			0.0	2	---			0.0
Aroclor-1260	3	11.706	-0.027	7806	5.1	3	---			0.0
Aroclor-1260	4	---			0.0	4	---			0.0
Aroclor-1260	5	---			0.0	NS	---			----
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1262	1	10.824	-0.005	16873	35.8	1	---			0.0
Aroclor-1262	2	---			0.0	2	---			0.0
Aroclor-1262	3	---			0.0	3	---			0.0
Aroclor-1262	4	13.040	0.053	14031	18.6	4	---			0.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1268	1	---			0.0	1	---			0.0
Aroclor-1268	2	---			0.0	2	---			0.0
Aroclor-1268	3	12.709	0.009	6037	3.6	3	12.891	-0.001	659	0.7
Aroclor-1268	4	13.499	0.010	12396	2.3	4	13.710	0.001	1848	0.6
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				

Total PCB Area Coll (5.906 - 13.793) = 260205

Coll Total PCB = 0.0 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 18252 Col2 Total PCB = 0.0 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 IB


24-FEB-2023 10:51, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 IB

24-FEB-2023 10:51, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242302ECD7.D  
Data file 2: /230224.b/230224.b/02242302ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.25PPMAR1660  
Client ID:  
Injection Date: 24-FEB-2023 11:12  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.811	0.005	405980	5.687	0.002	192160	40.4	41.5	2.9	Tetrachloro-m-xylene
13.897	0.004	563414	14.120	0.001	336737	40.0	43.0	7.3	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	673778	0.0
Hexabromobiphenyl	1429847	1429847	0.0

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	315256	0.0
Hexabromobiphenyl	513946	513946	0.0

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)



ZB5 Col ZB35 Col

Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.272	0.001	66125	258.4	1	7.255	-0.001	46626	252.6	
Aroclor-1016	2	7.654	-0.000	207370	265.8	2	7.855	-0.001	101071	270.1	
Aroclor-1016	3	7.792	0.002	92507	242.9	3	8.055	0.001	43448	257.1	
Aroclor-1016	4	8.406	0.001	64388	261.5	4	8.306	-0.000	33986	256.3	
Total CollAve (4 peaks):				257.2	Total Col2Ave (4 peaks):				259.0	RPD = 1	
Corrected Ave (3 peaks):				254.3	Corrected Ave (3 peaks):				255.3	RPD = 0	

CalAmt %D: 2.9

CalAmt %D: 3.6

Aroclor-1260	1	11.046	0.001	138355	269.0	1	11.653	0.001	77114	255.2	
Aroclor-1260	2	11.363	0.002	147051	273.6	2	11.918	0.001	203401	263.7	
Aroclor-1260	3	11.736	0.003	383171	268.8	3	12.435	-0.000	51517	251.7	
Aroclor-1260	4	12.141	0.002	200399	279.2	4	12.502	0.001	134797	259.3	
Aroclor-1260	5	12.247	0.003	83796	271.2	NS	---			----	
Total CollAve (5 peaks):				272.4	Total Col2Ave (4 peaks):				257.5	RPD = 6	
Corrected Ave (4 peaks):				270.7	Corrected Ave (3 peaks):				255.4	RPD = 6	

CalAmt %D: 8.9

CalAmt %D: 3.0

Total PCB Area Coll (5.906 - 13.793) = 4024419 Coll Total PCB = 0.5 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 1889311 Col2 Total PCB = 0.5 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.25PPMAR1660

24-FEB-2023 11:12, 2ul



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242303ECD7.D  
Data file 2: /230224.b/230224.b/02242303ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.02PPMAR1660  
Client ID:  
Injection Date: 24-FEB-2023 11:33  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

ZB5 Col		ZB35 Col		ZB5	ZB35	RPD	Compound/Flag		
RT	Shift Response	RT	Shift Response	on col	on col				
5.809	0.002	29768	5.688	0.003	14932	3.1	3.3	5.7	Tetrachloro-m-xylene
13.893	0.000	45992	14.120	0.000	23950	3.4	3.1	9.1	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	637010	-5.5
Hexabromobiphenyl	1429847	1386953	-3.0

Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	307177	-2.6
Hexabromobiphenyl	513946	511463	-0.5

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.272	0.001	5052	20.9	1	7.256	0.000	3894	21.7	
Aroclor-1016	2	7.659	0.005	14714	19.9	2	7.864	0.008	6253	17.1	
Aroclor-1016	3	7.795	0.005	8226	22.8	3	8.060	0.006	3076	18.7	
Aroclor-1016	4	8.407	0.002	4780	20.5	4	8.309	0.002	2443	18.9	
Total CollAve (4 peaks):				21.1	Total Col2Ave (4 peaks):				19.1	RPD = 10	
Corrected Ave (3 peaks):				20.5	Corrected Ave (3 peaks):				18.2	RPD = 11	
CalAmt %D:				5.3	CalAmt %D:				-4.5		
Aroclor-1260	1	11.047	0.003	10147	20.3	1	11.656	0.003	6759	22.5	
Aroclor-1260	2	11.364	0.003	10287	19.7	2	11.922	0.005	16592	21.6	
Aroclor-1260	3	11.740	0.006	28043	20.3	3	12.438	0.002	4506	22.1	
Aroclor-1260	4	12.145	0.006	13540	19.4	4	12.505	0.004	11037	21.3	
Aroclor-1260	5	12.246	0.002	6182	20.6	NS	---			----	
Total CollAve (5 peaks):				20.1	Total Col2Ave (4 peaks):				21.9	RPD = 9	
Corrected Ave (4 peaks):				19.9	Corrected Ave (3 peaks):				21.7	RPD = 8	
CalAmt %D:				0.4	CalAmt %D:				9.4		

Total PCB Area Coll (5.906 - 13.793) = 324832 Coll Total PCB = 0.0 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 157149 Col2 Total PCB = 0.0 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.02PPMAR1660


24-FEB-2023 11:33, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.02PPMAR1660

24-FEB-2023 11:33, 2ul




ZB-35 Manual Integration: YES


Manual Peak Adjustment, ZB-35

Datafile: ecd7.i/230224.b/230224.b/02242303ECD7.D Injection Date: 24-FEB-2023

Manual Integration (After)



Processed Integration (Before)



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242304ECD7.D  
Data file 2: /230224.b/230224.b/02242304ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.05PPMAR1660  
Client ID:  
Injection Date: 24-FEB-2023 11:54  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

ZB5 Col		ZB35 Col		ZB5	ZB35	RPD	Compound/Flag		
RT	Shift Response	RT	Shift Response	on col	on col				
5.809	0.003	78493	5.688	0.003	36772	8.3	8.1	2.2	Tetrachloro-m-xylene
13.893	-0.000	113544	14.119	-0.000	62745	8.2	7.9	3.4	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	630965	-6.4
Hexabromobiphenyl	1429847	1409464	-1.4

Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	307599	-2.4
Hexabromobiphenyl	513946	521112	1.4

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.272	0.002	12829	53.5	1	7.256	0.000	9654	53.6	
Aroclor-1016	2	7.660	0.006	36461	49.9	2	7.864	0.008	18085	49.5	
Aroclor-1016	3	7.795	0.005	19865	55.7	3	8.063	0.008	9071	55.0	
Aroclor-1016	4	8.408	0.003	11411	49.5	4	8.310	0.003	7309	56.5	
Total CollAve (4 peaks):				52.2	Total Col2Ave (4 peaks):				53.7	RPD = 3	
Corrected Ave (3 peaks):				51.0	Corrected Ave (3 peaks):				52.7	RPD = 3	
CalAmt %D:				4.3	CalAmt %D:				7.3		
Aroclor-1260	1	11.046	0.002	25727	50.7	1	11.655	0.002	15996	52.2	
Aroclor-1260	2	11.363	0.002	26482	50.0	2	11.922	0.004	40487	51.8	
Aroclor-1260	3	11.739	0.005	70871	50.4	3	12.437	0.002	10248	49.4	
Aroclor-1260	4	12.143	0.004	34239	48.4	4	12.506	0.004	26828	50.9	
Aroclor-1260	5	12.246	0.002	15109	49.6	NS	---			----	
Total CollAve (5 peaks):				49.8	Total Col2Ave (4 peaks):				51.1	RPD = 2	
Corrected Ave (4 peaks):				49.6	Corrected Ave (3 peaks):				50.7	RPD = 2	
CalAmt %D:				-0.3	CalAmt %D:				2.1		

Total PCB Area Coll (5.906 - 13.793) = 758292 Coll Total PCB = 0.1 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 386383 Col2 Total PCB = 0.1 ppm\*

\* Quantitated against AR1660 0.25ppm in Ical


PCB-Form 10 Mod.



# PCB Dual Column Chromatograms

ECD7-ZB5 0.05PPMAR1660


24-FEB-2023 11:54, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.05PPMAR1660

24-FEB-2023 11:54, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242305ECD7.D  
Data file 2: /230224.b/230224.b/02242305ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 1.0PPMAR1660  
Client ID:  
Injection Date: 24-FEB-2023 12:15  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.813	0.006	1641874	5.688	0.003	709674	166.2	151.5	9.3	Tetrachloro-m-xylene
13.899	0.006	2344583	14.122	0.002	1300114	161.9	158.6	2.1	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	661440	-1.8
Hexabromobiphenyl	1429847	1470100	2.8
Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	319272	1.3
Hexabromobiphenyl	513946	538138	4.7

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col						
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.270	-0.000	220519	877.8	1	7.254	-0.001	162833	871.2	
Aroclor-1016	2	7.652	-0.002	731607	955.3	2	7.852	-0.004	373610	985.8	
Aroclor-1016	3	7.789	-0.001	307629	822.8	3	8.051	-0.003	156666	915.2	
Aroclor-1016	4	8.404	-0.001	229387	949.1	4	8.305	-0.002	117186	872.6	
Total CollAve (4 peaks):				901.3		Total Col2Ave (4 peaks):				911.2	RPD = 1
Corrected Ave (3 peaks):				883.3		Corrected Ave (3 peaks):				886.3	RPD = 0

CalAmt %D: -9.9

CalAmt %D: -8.9

Aroclor-1260	1	11.044	-0.000	504641	954.2	1	11.652	-0.000	282606	893.1	
Aroclor-1260	2	11.360	-0.001	524931	950.0	2	11.917	-0.000	709329	878.4	
Aroclor-1260	3	11.734	-0.000	1410270	962.3	3	12.434	-0.001	215124	1003.8	
Aroclor-1260	4	12.137	-0.002	720770	976.7	4	12.501	-0.001	506566	930.6	
Aroclor-1260	5	12.243	-0.001	304211	957.7	NS	---			----	
Total CollAve (5 peaks):				960.2		Total Col2Ave (4 peaks):				926.5	RPD = 4
Corrected Ave (4 peaks):				956.0		Corrected Ave (3 peaks):				900.7	RPD = 6

CalAmt %D: -4.0

CalAmt %D: -7.4

Total PCB Area Coll (5.906 - 13.793) = 14454279 Coll Total PCB = 1.8 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 7029563 Col2 Total PCB = 1.8 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

PCB Dual Column Chromatograms

ECD7-ZB5 1.0PPMAR1660


24-FEB-2023 12:15, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 1.0PPMAR1660

24-FEB-2023 12:15, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242306ECD7.D  
Data file 2: /230224.b/230224.b/02242306ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.1PPMAR1660  
Client ID:  
Injection Date: 24-FEB-2023 12:36  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

ZB5 Col		ZB35 Col		ZB5	ZB35	RPD	Compound/Flag		
RT	Shift Response	RT	Shift Response	on col	on col				
5.809	0.002	155528	5.688	0.003	74628	15.9	16.0	0.9	Tetrachloro-m-xylene
13.892	-0.001	227253	14.119	-0.000	128496	15.8	15.8	0.5	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	655979	-2.6
Hexabromobiphenyl	1429847	1464509	2.4
Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	317418	0.7
Hexabromobiphenyl	513946	532962	3.7

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	7.271	0.000	25761	103.4	1	7.255	-0.000	19315	103.9
Aroclor-1016	2	7.657	0.003	75616	99.6	2	7.863	0.007	40308	107.0
Aroclor-1016	3	7.794	0.004	39547	106.7	3	8.059	0.005	18304	107.6
Aroclor-1016	4	8.406	0.001	24260	101.2	4	8.309	0.002	14708	110.2
Total CollAve (4 peaks):				102.7		Total Col2Ave (4 peaks):				107.2 RPD = 4
Corrected Ave (3 peaks):				101.4		Corrected Ave (3 peaks):				106.2 RPD = 5
CalAmt %D:				2.7		CalAmt %D:				7.2
Aroclor-1260	1	11.045	0.000	52009	98.7	1	11.655	0.002	31282	99.8
Aroclor-1260	2	11.362	0.001	55116	100.1	2	11.920	0.003	80574	100.7
Aroclor-1260	3	11.738	0.004	145604	99.7	3	12.437	0.002	19566	92.2
Aroclor-1260	4	12.141	0.002	72408	98.5	4	12.503	0.001	53588	99.4
Aroclor-1260	5	12.245	0.001	30745	97.2	NS	---			----
Total CollAve (5 peaks):				98.8		Total Col2Ave (4 peaks):				98.0 RPD = 1
Corrected Ave (4 peaks):				98.5		Corrected Ave (3 peaks):				97.1 RPD = 1
CalAmt %D:				-1.2		CalAmt %D:				-2.0

Total PCB Area Coll (5.906 - 13.793) = 1555762 Coll Total PCB = 0.2 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 764924 Col2 Total PCB = 0.2 ppm\*

\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242307ECD7.D  
Data file 2: /230224.b/230224.b/02242307ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.5PPMAR1660  
Client ID:  
Injection Date: 24-FEB-2023 12:57  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.810	0.004	724614	5.688	0.003	359257	75.2	76.7	2.0	Tetrachloro-m-xylene
13.898	0.005	1056911	14.120	0.000	650153	74.3	79.5	6.9	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	645275	-4.2
Hexabromobiphenyl	1429847	1445345	1.1
Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	319170	1.2
Hexabromobiphenyl	513946	536853	4.5

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)



ZB5 Col					ZB35 Col						
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.270	0.000	115193	470.0	1	7.256	0.000	86287	461.8	
Aroclor-1016	2	7.654	0.000	369991	495.2	2	7.856	0.000	192524	508.1	
Aroclor-1016	3	7.790	0.000	160952	441.3	3	8.055	0.000	81039	473.6	
Aroclor-1016	4	8.405	0.000	115032	487.9	4	8.307	0.000	62136	462.8	
Total CollAve (4 peaks):				473.6		Total Col2Ave (4 peaks):				476.6	RPD = 1
Corrected Ave (3 peaks):				466.4		Corrected Ave (3 peaks):				466.1	RPD = 0

CalAmt %D: -5.3

CalAmt %D: -4.7

Aroclor-1260	1	11.044	0.000	247212	475.5	1	11.653	0.000	145247	460.1	
Aroclor-1260	2	11.361	0.000	262877	483.9	2	11.918	0.000	379838	471.5	
Aroclor-1260	3	11.734	0.000	678830	471.1	3	12.436	0.000	104092	486.9	
Aroclor-1260	4	12.139	0.000	356067	490.7	4	12.502	0.000	258953	476.9	
Aroclor-1260	5	12.244	0.000	150280	481.2	NS	---			----	
Total CollAve (5 peaks):				480.5		Total Col2Ave (4 peaks):				473.8	RPD = 1
Corrected Ave (4 peaks):				477.9		Corrected Ave (3 peaks):				469.5	RPD = 2

CalAmt %D: -3.9

CalAmt %D: -5.2

Total PCB Area Coll (5.906 - 13.793) = 7134169 Coll Total PCB = 0.9 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 3589735 Col2 Total PCB = 0.9 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.5PPMAR1660


24-FEB-2023 12:57, 2u1



ZB-5 Manual Integration: NO

ECD7-ZB35 0.5PPMAR1660

24-FEB-2023 12:57, 2u1



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242308ECD7.D  
Data file 2: /230224.b/230224.b/02242308ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1242.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.25PPMAR1242  
Client ID:  
Injection Date: 24-FEB-2023 13:18  
Report Date: 02/28/2023 09:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.809	0.003	434187	5.688	0.003	214306	46.0	46.5	1.1	Tetrachloro-m-xylene
13.894	0.000	515867	14.119	-0.001	312943	35.6	38.5	7.6	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	632576	-6.1
Hexabromobiphenyl	1429847	1469715	2.8

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	314129	-0.4
Hexabromobiphenyl	513946	534294	4.0

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1242	1	7.271	0.000	49009	250.0	1	7.255	0.000	36487	250.0
Aroclor-1242	2	7.656	0.000	148833	250.0	2	7.858	0.000	76699	250.0
Aroclor-1242	3	8.405	0.000	46308	250.0	3	9.167	0.000	23866	250.0
Aroclor-1242	4	8.579	0.000	68453	250.0	4	9.597	0.000	29080	250.0
Total Col1Ave (4 peaks):				250.0	Total Col2Ave (4 peaks):				250.0	RPD = 0
Corrected Ave (3 peaks):				250.0	Corrected Ave (3 peaks):				250.0	RPD = 0

Total PCB Area Col1 (5.906 - 13.793) = 1221467 Col1 Total PCB = 0.2 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 572067 Col2 Total PCB = 0.2 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.25PPMAR1242


24-FEB-2023 13:18, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.25PPMAR1242

24-FEB-2023 13:18, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242309ECD7.D  
Data file 2: /230224.b/230224.b/02242309ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1248.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.25PPMAR1248  
Client ID:  
Injection Date: 24-FEB-2023 13:39  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.809	0.003	349513	5.688	0.003	176615	36.6	37.9	3.4	Tetrachloro-m-xylene
13.894	0.001	523008	14.121	0.001	322054	36.4	39.3	7.5	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	639911	-5.0
Hexabromobiphenyl	1429847	1458696	2.0

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	317938	0.9
Hexabromobiphenyl	513946	538760	4.8

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023

<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col				
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1248	1	8.405	0.000	78055	250.0	1	8.308	0.000	37951	250.0
Aroclor-1248	2	8.580	0.000	99216	250.0	2	8.714	0.000	39239	250.0
Aroclor-1248	3	8.999	0.000	187178	250.0	3	9.166	0.000	45157	250.0
Aroclor-1248	4	9.295	0.000	95291	250.0	4	9.590	0.000	54216	250.0
Total CollAve (4 peaks):				250.0	Total Col2Ave (4 peaks):				250.0	RPD = 0
Corrected Ave (3 peaks):				250.0	Corrected Ave (3 peaks):				250.0	RPD = 0

Total PCB Area Coll (5.906 - 13.793) = 1565180 Coll Total PCB = 0.2 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 754991 Col2 Total PCB = 0.2 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.25PPMAR1248


24-FEB-2023 13:39, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.25PPMAR1248

24-FEB-2023 13:39, 2ul



ZB-35 Manual Integration: NO



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242310ECD7.D  
Data file 2: /230224.b/230224.b/02242310ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR1254.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.25PPMAR1254  
Client ID:  
Injection Date: 24-FEB-2023 14:00  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.808	0.002	352587	5.687	0.002	177502	37.3	38.6	3.4	Tetrachloro-m-xylene
13.895	0.002	532500	14.119	0.000	325903	37.0	40.2	8.2	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	633407	-6.0
Hexabromobiphenyl	1429847	1460265	2.1

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	313673	-0.5
Hexabromobiphenyl	513946	532442	3.6

\* Standard Areas taken from Initial Cal Level 3

Initial Calibration Date: 24-FEB-2023

<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col ZB35 Col

Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1254	1	9.299	0.000	159011	250.0	1	9.449	0.000	59603	250.0
Aroclor-1254	2	9.377	0.000	71516	250.0	2	9.970	0.000	47949	250.0
Aroclor-1254	3	9.668	0.000	102230	250.0	3	10.124	0.000	103745	250.0
Aroclor-1254	4	9.807	0.000	198777	250.0	4	10.373	0.000	101135	250.0
Aroclor-1254	5	10.176	0.000	124586	250.0	5	10.569	0.000	61577	250.0
Total CollAve (5 peaks):				250.0		Total Col2Ave (5 peaks):				250.0 RPD = 0
Corrected Ave (4 peaks):				250.0		Corrected Ave (4 peaks):				250.0 RPD = 0

Total PCB Area Coll (5.906 - 13.793) = 2179224 Coll Total PCB = 0.3 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 1022156 Col2 Total PCB = 0.3 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.25PPMAR1254


24-FEB-2023 14:00, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.25PPMAR1254

24-FEB-2023 14:00, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242311ECD7.D  
Data file 2: /230224.b/230224.b/02242311ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR2162.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.25PPMAR2162  
Client ID:  
Injection Date: 24-FEB-2023 14:21  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.807	0.001	362236	5.686	0.000	177349	38.4	39.2	2.1	Tetrachloro-m-xylene
13.894	0.001	523254	14.119	-0.000	321034	36.0	39.2	8.3	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	632433	-6.1
Hexabromobiphenyl	1429847	1474039	3.1

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	308453	-2.2
Hexabromobiphenyl	513946	538177	4.7

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col				
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1221	1	4.731	0.000	14160	250.0	1	4.956	0.000	7300	250.0
Aroclor-1221	2	6.132	0.000	25324	250.0	2	6.296	0.000	13816	250.0
Aroclor-1221	3	6.382	0.000	58795	250.0	3	6.622	0.000	22491	250.0
Total CollAve (3 peaks):				250.0		Total Col2Ave (3 peaks):				250.0 RPD = 0
Corrected Ave: < 3 Peaks						Corrected Ave: < 3 Peaks				

Aroclor-1262	1	10.829	0.000	113046	250.0	1	11.200	0.000	114880	250.0
Aroclor-1262	2	12.244	0.000	183948	250.0	2	11.652	0.000	97844	250.0
Aroclor-1262	3	12.319	0.000	197749	250.0	3	12.434	0.000	111015	250.0
Aroclor-1262	4	12.987	0.000	180727	250.0	4	12.502	0.000	173913	250.0
Total CollAve (4 peaks):				250.0		Total Col2Ave (4 peaks):				250.0 RPD = 0
Corrected Ave (3 peaks):				250.0		Corrected Ave (3 peaks):				250.0 RPD = 0

Total PCB Area Coll (5.906 - 13.793) = 3105316 Coll Total PCB = 0.4 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 1573107 Col2 Total PCB = 0.4 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 0.25PPMAR2162

24-FEB-2023 14:21, 2ul



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242312ECD7.D  
Data file 2: /230224.b/230224.b/02242312ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: AR3268.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: 0.25PPMAR3268  
Client ID:  
Injection Date: 24-FEB-2023 14:42  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.806	0.000	366416	5.685	0.000	179450	38.0	38.9	2.4	Tetrachloro-m-xylene
13.893	0.000	778191	14.119	0.000	477889	53.0	57.5	8.3	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	645602	-4.2
Hexabromobiphenyl	1429847	1492154	4.4

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	314042	-0.4
Hexabromobiphenyl	513946	545458	6.1

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col				
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1232	1	4.730	0.000	8647	250.0	1	4.956	0.000	4017	250.0
Aroclor-1232	2	6.131	0.000	17148	250.0	2	7.254	0.000	19962	250.0
Aroclor-1232	3	7.656	0.000	77627	250.0	3	7.861	0.000	39913	250.0
Aroclor-1232	4	8.581	0.000	32993	250.0	4	8.715	0.000	11487	250.0
Total CollAve (4 peaks):				250.0		Total Col2Ave (4 peaks):				250.0 RPD = 0
Corrected Ave (3 peaks):				250.0		Corrected Ave (3 peaks):				250.0 RPD = 0
Aroclor-1268	1	12.247	0.000	477974	250.0	1	12.432	0.000	274595	250.0
Aroclor-1268	2	12.317	0.000	473326	250.0	2	12.500	0.000	295194	250.0
Aroclor-1268	3	12.699	0.000	405011	250.0	3	12.892	0.000	252048	250.0
Aroclor-1268	4	13.490	0.000	1333528	250.0	4	13.709	0.000	805579	250.0
Total CollAve (4 peaks):				250.0		Total Col2Ave (4 peaks):				250.0 RPD = 0
Corrected Ave (3 peaks):				250.0		Corrected Ave (3 peaks):				250.0 RPD = 0

Total PCB Area Col1 (5.906 - 13.793) = 3998414 Col1 Total PCB = 0.5 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 2300029 Col2 Total PCB = 0.6 ppm\*

\* Quantitated against AR1660 0.25ppm in Ical


PCB-Form 10 Mod.



# PCB Dual Column Chromatograms

ECD7-ZB5 0.25PPMAR3268


24-FEB-2023 14:42, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 0.25PPMAR3268

24-FEB-2023 14:42, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242313ECD7.D  
Data file 2: /230224.b/230224.b/02242313ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1660SCV  
Client ID:  
Injection Date: 24-FEB-2023 15:03  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.807	0.001	337070	5.686	0.001	165848	34.9	35.8	2.3	Tetrachloro-m-xylene
13.895	0.002	515407	14.119	-0.000	316730	34.3	37.3	8.4	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	645975	-4.1
Hexabromobiphenyl	1429847	1524245	6.6

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	316115	0.3
Hexabromobiphenyl	513946	556950	8.4

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col						
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.269	-0.002	59491	242.5	1	7.254	-0.002	44576	240.9	
Aroclor-1016	2	7.655	0.001	181090	242.1	2	7.857	0.002	95386	254.2	
Aroclor-1016	3	7.790	0.000	88470	242.3	3	8.056	0.002	42160	248.8	
Aroclor-1016	4	8.404	-0.001	57980	245.6	4	8.307	0.000	32197	242.1	
Total CollAve (4 peaks):				243.1	Total Col2Ave (4 peaks):				246.5	RPD = 1	
Corrected Ave (3 peaks):				242.3	Corrected Ave (3 peaks):				243.9	RPD = 1	
Aroclor-1221	1	4.731	0.000	464	8.0	1	---			0.0	
Aroclor-1221	2	6.130	-0.002	9233	89.2	2	6.300	0.004	5379	95.0	
Aroclor-1221	3	6.382	-0.001	42570	177.2	3	6.623	0.001	20952	227.2	
Total CollAve (3 peaks):				91.5	Col2Ave: <3 Quant Peaks						
Aroclor-1232	1	4.731	0.001	464	13.4	1	---			0.0	
Aroclor-1232	2	6.130	-0.001	9233	134.5	2	7.254	-0.000	44576	554.6	
Aroclor-1232	3	7.655	-0.001	181090	582.9	3	7.857	-0.003	95386	593.5	
Aroclor-1232	4	8.580	-0.001	79916	605.2	4	8.713	-0.002	29795	644.2	
Total CollAve (4 peaks):				334.0	Total Col2Ave (3 peaks):				597.4	RPD = 57*	
Corrected Ave (3 peaks):				243.6	Corrected Ave: < 3 Peaks						
Aroclor-1242	1	7.269	-0.002	59491	297.2	1	7.254	-0.002	44576	303.5	
Aroclor-1242	2	7.655	-0.001	181090	297.9	2	7.857	-0.000	95386	309.0	
Aroclor-1242	3	8.404	-0.001	57980	306.5	3	9.115	-0.052	18754	195.2	
Aroclor-1242	4	8.580	0.000	79916	285.8	4	9.697	0.100	1355	11.6	
Total CollAve (4 peaks):				296.8	Total Col2Ave (4 peaks):				204.8	RPD = 37	
Corrected Ave (3 peaks):				293.6	Corrected Ave (3 peaks):				170.1	RPD = 53*	
Aroclor-1248	1	8.404	-0.001	57980	184.0	1	8.307	-0.001	32197	213.3	
Aroclor-1248	2	8.580	-0.001	79916	199.5	2	8.713	-0.001	29795	190.9	
Aroclor-1248	3	8.993	-0.006	71805	95.0	3	9.115	-0.050	18754	104.4	
Aroclor-1248	4	9.300	0.006	47348	123.1	4	---			0.0	
Total CollAve (4 peaks):				150.4	Total Col2Ave (3 peaks):				169.6	RPD = 12	
Corrected Ave (3 peaks):				134.0	Corrected Ave: < 3 Peaks						
Aroclor-1254	1	9.300	0.002	47348	73.0	1	9.451	0.001	22438	93.4	
Aroclor-1254	2	---			0.0	2	9.972	0.001	2694	13.9	
Aroclor-1254	3	9.670	0.002	5461	13.1	3	10.147	0.024	52914	126.5	
Aroclor-1254	4	9.807	-0.000	18944	23.4	4	10.370	-0.003	70430	172.8	
Aroclor-1254	5	10.121	-0.056	154170	303.3	5	10.568	-0.000	98525	396.9	
Total CollAve (4 peaks):				103.2	Total Col2Ave (5 peaks):				160.7	RPD = 44*	
Corrected Ave (3 peaks):				36.5	Corrected Ave (4 peaks):				101.7	RPD = 94*	
Aroclor-1260	1	11.044	0.000	149195	272.1	1	11.653	0.000	82210	251.0	
Aroclor-1260	2	11.361	-0.000	153832	268.5	2	11.919	0.001	222226	265.9	
Aroclor-1260	3	11.736	0.002	396660	261.0	3	12.435	-0.000	59148	266.7	
Aroclor-1260	4	12.140	0.001	190448	248.9	4	12.504	0.002	147180	261.2	
Aroclor-1260	5	12.244	-0.000	91385	277.5	NS	---			----	
Total CollAve (5 peaks):				265.6	Total Col2Ave (4 peaks):				261.2	RPD = 2	
Corrected Ave (4 peaks):				262.6	Corrected Ave (3 peaks):				259.4	RPD = 1	
Aroclor-1262	1	10.827	-0.002	220238	471.0	1	11.199	-0.001	84479	177.6	
Aroclor-1262	2	12.244	0.000	91385	120.1	2	11.653	0.002	82210	203.0	
Aroclor-1262	3	12.320	0.001	113066	138.2	3	12.435	0.002	59148	128.7	
Aroclor-1262	4	12.988	0.001	102156	136.7	4	12.504	0.002	147180	204.4	
Total CollAve (4 peaks):				216.5	Total Col2Ave (4 peaks):				178.4	RPD = 19	
Corrected Ave (3 peaks):				131.7	Corrected Ave (3 peaks):				169.8	RPD = 25	
Aroclor-1268	1	12.244	-0.003	91385	46.8	1	12.435	0.003	59148	52.7	
Aroclor-1268	2	12.320	0.003	113066	58.5	2	12.504	0.004	147180	122.1	
Aroclor-1268	3	12.726	0.027	46633	28.2	3	12.893	0.001	2874	2.8	
Aroclor-1268	4	13.489	-0.000	25567	4.7	4	13.709	-0.000	13041	4.0	
Total CollAve (4 peaks):				34.5	Total Col2Ave (4 peaks):				45.4	RPD = 27	
Corrected Ave (3 peaks):				26.6	Corrected Ave (3 peaks):				19.8	RPD = 29	

Total PCB Area Col1 (5.906 - 13.793) = 3743076 Col1 Total PCB = 0.5 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 1897008 Col2 Total PCB = 0.5 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR1660SCV


24-FEB-2023 15:03, 2u1



ZB-5 Manual Integration: NO

ECD7-ZB35 AR1660SCV

24-FEB-2023 15:03, 2u1



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242314ECD7.D  
Data file 2: /230224.b/230224.b/02242314ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1242SCV  
Client ID:  
Injection Date: 24-FEB-2023 15:24  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.808	0.002	354283	5.686	0.001	172455	33.6	34.5	2.6	Tetrachloro-m-xylene
13.895	0.002	567088	14.120	0.001	347430	37.0	40.3	8.6	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	705650	4.7
Hexabromobiphenyl	1429847	1555683	8.8
Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	340433	8.0
Hexabromobiphenyl	513946	565609	10.1

\* Standard Areas taken from Initial Cal Level 3

Initial Calibration Date: 24-FEB-2023

<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	7.270	-0.000	39927	149.0	1	7.256	0.000	32417	162.7
Aroclor-1016	2	7.653	-0.001	132339	162.0	2	7.856	0.001	69235	171.3
Aroclor-1016	3	7.791	0.001	59310	148.7	3	8.055	0.000	29473	161.5
Aroclor-1016	4	8.405	0.000	42537	165.0	4	8.307	-0.000	22792	159.2
Total CollAve (4 peaks):				156.2		Total Col2Ave (4 peaks):				163.7 RPD = 5
Corrected Ave (3 peaks):				153.2		Corrected Ave (3 peaks):				161.1 RPD = 5
Aroclor-1221	1	4.733	0.002	319	5.0	1	---			0.0
Aroclor-1221	2	6.131	-0.001	6534	57.8	2	6.319	0.022	4365	71.6
Aroclor-1221	3	6.384	0.001	29664	113.0	3	6.624	0.002	14916	150.2
Total CollAve (3 peaks):				58.6		Col2Ave: <3 Quant Peaks				
Aroclor-1232	1	4.733	0.003	319	8.4	1	---			0.0
Aroclor-1232	2	6.131	0.000	6534	87.2	2	7.256	0.002	32417	374.5
Aroclor-1232	3	7.653	-0.003	132339	389.9	3	7.856	-0.004	69235	400.0
Aroclor-1232	4	8.579	-0.002	69445	481.4	4	8.714	-0.001	22167	445.0
Total CollAve (4 peaks):				241.7		Total Col2Ave (3 peaks):				406.5 RPD = 51*
Corrected Ave (3 peaks):				161.8		Corrected Ave: < 3 Peaks				
Aroclor-1242	1	7.270	-0.001	39927	182.6	1	7.256	0.000	32417	205.0
Aroclor-1242	2	7.653	-0.003	132339	199.3	2	7.856	-0.002	69235	208.2
Aroclor-1242	3	8.405	-0.000	42537	205.9	3	9.164	-0.004	23068	223.0
Aroclor-1242	4	8.579	-0.000	69445	227.4	4	9.587	-0.010	31021	246.1
Total CollAve (4 peaks):				203.8		Total Col2Ave (4 peaks):				220.6 RPD = 8
Corrected Ave (3 peaks):				195.9		Corrected Ave (3 peaks):				212.1 RPD = 8
Aroclor-1248	1	8.405	0.000	42537	123.5	1	8.307	-0.001	22792	140.2
Aroclor-1248	2	8.579	-0.001	69445	158.7	2	8.714	-0.000	22167	131.9
Aroclor-1248	3	9.001	0.003	91942	111.4	3	9.164	-0.002	23068	119.3
Aroclor-1248	4	9.294	-0.000	38711	92.1	4	9.587	-0.003	31021	133.6
Total CollAve (4 peaks):				121.4		Total Col2Ave (4 peaks):				131.2 RPD = 8
Corrected Ave (3 peaks):				109.0		Corrected Ave (3 peaks):				128.3 RPD = 16
Aroclor-1254	1	9.294	-0.005	38711	54.6	1	9.450	0.001	13131	50.7
Aroclor-1254	2	9.377	-0.000	17371	54.5	2	9.970	0.000	8340	40.1
Aroclor-1254	3	9.668	-0.000	16373	35.9	3	10.123	-0.000	16364	36.3
Aroclor-1254	4	9.807	-0.001	27490	31.0	4	10.382	0.009	16062	36.6
Aroclor-1254	5	10.175	-0.001	20494	36.9	5	10.572	0.004	4818	18.0
Total CollAve (5 peaks):				42.6		Total Col2Ave (5 peaks):				36.4 RPD = 16
Corrected Ave (4 peaks):				39.6		Corrected Ave (4 peaks):				32.8 RPD = 19
Aroclor-1260	1	11.048	0.003	794	1.4	1	11.665	0.012	1652	5.0
Aroclor-1260	2	11.366	0.005	814	1.4	2	11.926	0.008	842	1.0
Aroclor-1260	3	11.739	0.006	1848	1.2	3	12.438	0.002	483	2.1
Aroclor-1260	4	12.145	0.006	1372	1.8	4	12.506	0.004	790	1.4
Aroclor-1260	5	---			0.0	NS	---			---
Total CollAve (4 peaks):				1.4		Total Col2Ave (4 peaks):				2.4 RPD = 49*
Corrected Ave (3 peaks):				1.3		Corrected Ave (3 peaks):				1.5 RPD = 12
Aroclor-1262	1	10.832	0.003	13157	27.6	1	11.121	-0.079	6113	12.7
Aroclor-1262	2	12.145	-0.098	1372	1.8	2	11.665	0.013	1652	4.0
Aroclor-1262	3	---			0.0	3	12.438	0.004	483	1.0
Aroclor-1262	4	13.038	0.051	842	1.1	4	12.506	0.004	790	1.1
Total CollAve (3 peaks):				10.1		Total Col2Ave (4 peaks):				4.7 RPD = 73*
Corrected Ave: < 3 Peaks						Corrected Ave (3 peaks):				2.0
Aroclor-1268	1	---			0.0	1	12.438	0.006	483	0.4
Aroclor-1268	2	---			0.0	2	12.506	0.006	790	0.6
Aroclor-1268	3	12.617	-0.082	5851	3.5	3	12.899	0.007	491	0.5
Aroclor-1268	4	13.500	0.010	1745	0.3	4	13.714	0.005	379	0.1
CollAve: <3 Quant Peaks						Col2Ave:				0.4

Total PCB Area Col1 (5.906 - 13.793) = 1149784 Col1 Total PCB = 0.1 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 572210 Col2 Total PCB = 0.1 ppm\*

\* Quantitated against AR1660 0.25ppm in Ical


PCB-Form 10 Mod.



# PCB Dual Column Chromatograms

ECD7-ZB5 AR1242SCV


24-FEB-2023 15:24, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 AR1242SCV

24-FEB-2023 15:24, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242315ECD7.D  
Data file 2: /230224.b/230224.b/02242315ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1248SCV  
Client ID:  
Injection Date: 24-FEB-2023 15:45  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.809	0.002	336655	5.687	0.002	168719	34.9	36.4	4.2	Tetrachloro-m-xylene
13.894	0.001	499162	14.118	-0.001	308317	33.1	36.3	9.2	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	646554	-4.0
Hexabromobiphenyl	1429847	1529451	7.0

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	316066	0.3
Hexabromobiphenyl	513946	557213	8.4

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	7.271	0.000	19773	80.5	1	7.254	-0.002	16926	91.5
Aroclor-1016	2	7.653	-0.001	88099	117.7	2	7.857	0.001	45733	121.9
Aroclor-1016	3	7.794	0.003	35915	98.3	3	8.060	0.005	8078	47.7
Aroclor-1016	4	8.406	0.001	77842	329.5	4	8.307	0.000	37348	280.9
Total CollAve (4 peaks):				156.5		Total Col2Ave (4 peaks):				135.5 RPD = 14
Corrected Ave (3 peaks):				98.8		Corrected Ave (3 peaks):				87.0 RPD = 13
Aroclor-1221	1	---			0.0	1	---			0.0
Aroclor-1221	2	6.133	0.001	680	6.6	2	6.326	0.030	1966	34.7
Aroclor-1221	3	6.384	0.002	3390	14.1	3	6.631	0.009	1571	17.0
CollAve: <3 Quant Peaks						Col2Ave: <3 Quant Peaks				
Aroclor-1232	1	---			0.0	1	---			0.0
Aroclor-1232	2	6.133	0.002	680	9.9	2	7.254	-0.000	16926	210.6
Aroclor-1232	3	7.653	-0.002	88099	283.3	3	7.857	-0.004	45733	284.6
Aroclor-1232	4	8.581	-0.000	99572	753.4	4	8.714	-0.001	38224	826.6
Total CollAve (3 peaks):				348.9		Total Col2Ave (3 peaks):				440.6 RPD = 23
Corrected Ave: < 3 Peaks						Corrected Ave: < 3 Peaks				
Aroclor-1242	1	7.271	-0.000	19773	98.7	1	7.254	-0.002	16926	115.3
Aroclor-1242	2	7.653	-0.003	88099	144.8	2	7.857	-0.001	45733	148.2
Aroclor-1242	3	8.406	0.000	77842	411.2	3	9.165	-0.002	45021	468.7
Aroclor-1242	4	8.581	0.001	99572	355.8	4	9.590	-0.008	53613	458.1
Total CollAve (4 peaks):				252.6		Total Col2Ave (4 peaks):				297.6 RPD = 16
Corrected Ave (3 peaks):				199.8		Corrected Ave (3 peaks):				240.5 RPD = 19
Aroclor-1248	1	8.406	0.000	77842	246.8	1	8.307	-0.001	37348	247.5
Aroclor-1248	2	8.581	0.000	99572	248.3	2	8.714	-0.000	38224	245.0
Aroclor-1248	3	8.998	-0.000	186857	247.0	3	9.165	-0.000	45021	250.7
Aroclor-1248	4	9.294	-0.000	98398	255.5	4	9.590	-0.001	53613	248.7
Total CollAve (4 peaks):				249.4		Total Col2Ave (4 peaks):				248.0 RPD = 1
Corrected Ave (3 peaks):				247.4		Corrected Ave (3 peaks):				247.0 RPD = 0
Aroclor-1254	1	9.294	-0.004	98398	151.6	1	9.450	0.001	21823	90.8
Aroclor-1254	2	9.377	-0.001	49616	169.9	2	9.971	0.001	19450	100.6
Aroclor-1254	3	9.669	0.001	40230	96.4	3	10.124	0.000	36574	87.5
Aroclor-1254	4	9.808	0.001	68500	84.4	4	10.389	0.016	35100	86.1
Aroclor-1254	5	10.183	0.007	47365	93.1	5	10.573	0.004	5676	22.9
Total CollAve (5 peaks):				119.1		Total Col2Ave (5 peaks):				77.6 RPD = 42*
Corrected Ave (4 peaks):				106.4		Corrected Ave (4 peaks):				71.8 RPD = 39
Aroclor-1260	1	11.047	0.003	1670	3.0	1	11.662	0.009	2055	6.3
Aroclor-1260	2	11.362	0.001	1111	1.9	2	11.924	0.007	1466	1.8
Aroclor-1260	3	11.739	0.005	2107	1.4	3	12.434	-0.002	573	2.6
Aroclor-1260	4	12.144	0.005	1379	1.8	4	12.505	0.003	1003	1.8
Aroclor-1260	5	12.251	0.006	698	2.1	NS	---			----
Total CollAve (5 peaks):				2.1		Total Col2Ave (4 peaks):				3.1 RPD = 41*
Corrected Ave (4 peaks):				1.8		Corrected Ave (3 peaks):				2.0 RPD = 12
Aroclor-1262	1	10.833	0.005	15355	32.7	1	11.122	-0.079	7225	15.2
Aroclor-1262	2	12.251	0.007	698	0.9	2	11.662	0.011	2055	5.1
Aroclor-1262	3	12.321	0.002	836	1.0	3	12.434	0.000	573	1.2
Aroclor-1262	4	12.991	0.004	1043	1.4	4	12.505	0.003	1003	1.4
Total CollAve (4 peaks):				9.0		Total Col2Ave (4 peaks):				5.7 RPD = 45*
Corrected Ave (3 peaks):				1.1		Corrected Ave (3 peaks):				2.6 RPD = 80*
Aroclor-1268	1	12.251	0.004	698	0.4	1	12.434	0.002	573	0.5
Aroclor-1268	2	12.321	0.004	836	0.4	2	12.505	0.005	1003	0.8
Aroclor-1268	3	12.700	0.001	2449	1.5	3	12.892	0.001	721	0.7
Aroclor-1268	4	13.493	0.003	7547	1.4	4	13.708	-0.001	2265	0.7
Total CollAve (4 peaks):				0.9		Total Col2Ave (4 peaks):				0.7 RPD = 29
Corrected Ave (3 peaks):				0.7		Corrected Ave (3 peaks):				0.6 RPD = 13

Total PCB Area Col1 (5.906 - 13.793) = 1574335 Col1 Total PCB = 0.2 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 746330 Col2 Total PCB = 0.2 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR1248SCV


24-FEB-2023 15:45, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 AR1248SCV

24-FEB-2023 15:45, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242316ECD7.D  
Data file 2: /230224.b/230224.b/02242316ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1254SCV  
Client ID:  
Injection Date: 24-FEB-2023 16:06  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.806	-0.000	354312	5.686	0.001	174604	36.1	37.1	2.6	Tetrachloro-m-xylene
13.895	0.002	540961	14.119	-0.000	329134	34.6	37.9	9.0	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	656887	-2.5
Hexabromobiphenyl	1429847	1585505	10.9

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	320936	1.8
Hexabromobiphenyl	513946	570006	10.9

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.270	0.000	565	2.3	1	7.255	-0.001	387	2.1	
Aroclor-1016	2	7.656	0.002	1875	2.5	2	7.854	-0.002	860	2.3	
Aroclor-1016	3	7.792	0.002	1106	3.0	3	8.098	0.043	578	3.4	
Aroclor-1016	4	8.405	0.000	29924	124.7	4	8.307	0.000	21985	162.9	
Total CollAve (4 peaks):				33.1	Total Col2Ave (4 peaks):				42.6	RPD = 25	
Corrected Ave (3 peaks):				2.6	Corrected Ave (3 peaks):				2.6	RPD = 0	
Aroclor-1221	1	---			0.0	1	---			0.0	
Aroclor-1221	2	---			0.0	2	6.325	0.028	1947	33.9	
Aroclor-1221	3	---			0.0	3	6.637	0.015	368	3.9	
CollAve: <3 Quant Peaks					Col2Ave: <3 Quant Peaks						
Aroclor-1232	1	---			0.0	1	---			0.0	
Aroclor-1232	2	---			0.0	2	7.255	0.001	387	4.7	
Aroclor-1232	3	7.656	0.000	1875	5.9	3	7.854	-0.007	860	5.3	
Aroclor-1232	4	8.583	0.002	12327	91.8	4	8.715	0.000	15013	319.7	
CollAve: <3 Quant Peaks					Col2Ave: 109.9						
Aroclor-1242	1	7.270	-0.000	565	2.8	1	7.255	-0.001	387	2.6	
Aroclor-1242	2	7.656	0.000	1875	3.0	2	7.854	-0.004	860	2.7	
Aroclor-1242	3	8.405	-0.000	29924	155.6	3	9.169	0.002	21933	224.9	
Aroclor-1242	4	8.583	0.003	12327	43.4	4	9.545	-0.053	34065	286.6	
Total CollAve (4 peaks):				51.2	Total Col2Ave (4 peaks):				129.2	RPD = 87*	
Corrected Ave (3 peaks):				16.4	Corrected Ave (3 peaks):				76.7	RPD = 130*	
Aroclor-1248	1	8.405	0.000	29924	93.4	1	8.307	-0.001	21985	143.5	
Aroclor-1248	2	8.583	0.002	12327	30.3	2	8.715	0.001	15013	94.8	
Aroclor-1248	3	8.992	-0.007	145580	189.4	3	9.169	0.004	21933	120.3	
Aroclor-1248	4	9.298	0.003	155450	397.3	4	9.545	-0.046	34065	155.6	
Total CollAve (4 peaks):				177.6	Total Col2Ave (4 peaks):				128.5	RPD = 32	
Corrected Ave (3 peaks):				104.3	Corrected Ave (3 peaks):				119.5	RPD = 14	
Aroclor-1254	1	9.298	-0.001	155450	235.7	1	9.450	0.001	58639	240.4	
Aroclor-1254	2	9.377	-0.001	69801	235.3	2	9.971	0.000	47008	239.5	
Aroclor-1254	3	9.668	-0.000	100839	237.8	3	10.124	0.000	100062	235.7	
Aroclor-1254	4	9.807	0.000	190544	231.1	4	10.373	0.000	99535	240.5	
Aroclor-1254	5	10.176	-0.000	122321	236.7	5	10.570	0.001	61549	244.2	
Total CollAve (5 peaks):				235.3	Total Col2Ave (5 peaks):				240.1	RPD = 2	
Corrected Ave (4 peaks):				234.7	Corrected Ave (4 peaks):				239.0	RPD = 2	
Aroclor-1260	1	11.043	-0.002	12288	21.5	1	11.661	0.008	29062	86.7	
Aroclor-1260	2	11.361	-0.001	13660	22.9	2	11.921	0.003	22238	26.0	
Aroclor-1260	3	11.736	0.002	37632	23.8	3	12.441	0.005	3555	15.7	
Aroclor-1260	4	12.141	0.002	27105	34.1	4	12.503	0.001	13126	22.8	
Aroclor-1260	5	12.320	0.076	2381	6.9	NS	---			---	
Total CollAve (5 peaks):				21.9	Total Col2Ave (4 peaks):				37.8	RPD = 53*	
Corrected Ave (4 peaks):				18.8	Corrected Ave (3 peaks):				21.5	RPD = 13	
Aroclor-1262	1	10.827	-0.002	220626	453.6	1	11.281	0.081	13562	27.9	
Aroclor-1262	2	12.320	0.076	2381	3.0	2	11.661	0.009	29062	70.1	
Aroclor-1262	3	---			0.0	3	12.441	0.007	3555	7.6	
Aroclor-1262	4	12.989	0.002	3225	4.1	4	12.503	0.001	13126	17.8	
Total CollAve (3 peaks):				153.6	Total Col2Ave (4 peaks):				30.8	RPD = 133*	
Corrected Ave: < 3 Peaks					Corrected Ave (3 peaks):				17.7		
Aroclor-1268	1	12.320	0.074	2381	1.2	1	12.441	0.009	3555	3.1	
Aroclor-1268	2	---			0.0	2	12.503	0.003	13126	10.6	
Aroclor-1268	3	12.701	0.002	2939	1.7	3	12.892	0.000	772	0.7	
Aroclor-1268	4	13.493	0.003	9164	1.6	4	13.707	-0.002	2801	0.8	
Total CollAve (3 peaks):				1.5	Total Col2Ave (4 peaks):				3.8	RPD = 87*	
Corrected Ave: < 3 Peaks					Corrected Ave (3 peaks):				1.6		

Total PCB Area Col1 (5.906 - 13.793) = 2118645 Col1 Total PCB = 0.3 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 1007601 Col2 Total PCB = 0.3 ppm\*

\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.





Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242317ECD7.D  
Data file 2: /230224.b/230224.b/02242317ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR2162SCV  
Client ID:  
Injection Date: 24-FEB-2023 16:27  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.807	0.000	356001	5.685	0.000	170882	36.0	36.6	1.7	Tetrachloro-m-xylene
13.895	0.002	533971	14.119	0.000	326235	34.4	37.9	9.5	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	661953	-1.8
Hexabromobiphenyl	1429847	1574993	10.2

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	317807	0.8
Hexabromobiphenyl	513946	565951	10.1

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.269	-0.001	7175	28.5	1	7.256	0.000	3727	20.0	
Aroclor-1016	2	7.659	0.005	12893	16.8	2	7.863	0.007	5834	15.5	
Aroclor-1016	3	7.794	0.004	6936	18.5	3	8.063	0.009	2963	17.4	
Aroclor-1016	4	8.408	0.003	3610	14.9	4	8.308	0.002	2045	15.3	
Total CollAve (4 peaks):				19.7	Total Col2Ave (4 peaks):				17.0	RPD = 14	
Corrected Ave (3 peaks):				16.8	Corrected Ave (3 peaks):				16.1	RPD = 4	
Aroclor-1221	1	4.730	-0.000	15803	266.6	1	4.955	-0.001	7909	262.9	
Aroclor-1221	2	6.131	-0.001	26946	254.1	2	6.296	-0.000	14303	251.2	
Aroclor-1221	3	6.382	-0.000	62477	253.8	3	6.622	0.000	23612	254.7	
Total CollAve (3 peaks):				258.2	Total Col2Ave (3 peaks):				256.3	RPD = 1	
Corrected Ave: < 3 Peaks					Corrected Ave: < 3 Peaks						
Aroclor-1232	1	4.730	0.000	15803	445.6	1	4.955	-0.001	7909	486.4	
Aroclor-1232	2	6.131	0.000	26946	383.1	2	7.256	0.002	3727	46.1	
Aroclor-1232	3	7.659	0.003	12893	40.5	3	7.863	0.002	5834	36.1	
Aroclor-1232	4	8.583	0.003	2684	19.8	4	8.716	0.002	1189	25.6	
Total CollAve (4 peaks):				222.3	Total Col2Ave (4 peaks):				148.5	RPD = 40	
Corrected Ave (3 peaks):				147.8	Corrected Ave (3 peaks):				35.9	RPD = 122*	
Aroclor-1242	1	7.269	-0.001	7175	35.0	1	7.256	0.000	3727	25.2	
Aroclor-1242	2	7.659	0.003	12893	20.7	2	7.863	0.005	5834	18.8	
Aroclor-1242	3	8.408	0.002	3610	18.6	3	9.175	0.008	1082	11.2	
Aroclor-1242	4	8.583	0.004	2684	9.4	4	9.543	-0.054	1390	11.8	
Total CollAve (4 peaks):				20.9	Total Col2Ave (4 peaks):				16.8	RPD = 22	
Corrected Ave (3 peaks):				16.2	Corrected Ave (3 peaks):				13.9	RPD = 15	
Aroclor-1248	1	8.408	0.002	3610	11.2	1	8.308	0.001	2045	13.5	
Aroclor-1248	2	8.583	0.003	2684	6.5	2	8.716	0.002	1189	7.6	
Aroclor-1248	3	8.994	-0.005	24440	31.6	3	9.175	0.009	1082	6.0	
Aroclor-1248	4	9.302	0.008	26328	66.8	4	9.543	-0.048	1390	6.4	
Total CollAve (4 peaks):				29.0	Total Col2Ave (4 peaks):				8.4	RPD = 110*	
Corrected Ave (3 peaks):				16.4	Corrected Ave (3 peaks):				6.7	RPD = 85*	
Aroclor-1254	1	9.302	0.004	26328	39.6	1	9.452	0.003	9571	39.6	
Aroclor-1254	2	---			0.0	2	9.972	0.002	1733	8.9	
Aroclor-1254	3	9.670	0.002	3721	8.7	3	10.147	0.023	49218	117.1	
Aroclor-1254	4	9.808	0.000	9653	11.6	4	10.370	-0.002	59603	145.4	
Aroclor-1254	5	10.120	-0.056	131179	251.9	5	10.569	0.001	79533	318.7	
Total CollAve (4 peaks):				78.0	Total Col2Ave (5 peaks):				125.9	RPD = 47*	
Corrected Ave (3 peaks):				20.0	Corrected Ave (4 peaks):				77.8	RPD = 118*	
Aroclor-1260	1	11.044	-0.000	223208	394.0	1	11.652	-0.001	104071	312.7	
Aroclor-1260	2	11.361	-0.001	190166	321.2	2	11.919	0.002	251579	296.2	
Aroclor-1260	3	11.737	0.003	458281	291.9	3	12.435	-0.001	113645	504.2	
Aroclor-1260	4	12.141	0.002	149720	189.4	4	12.501	-0.001	182951	319.6	
Aroclor-1260	5	12.244	0.000	196033	576.0	NS	---			----	
Total CollAve (5 peaks):				354.5	Total Col2Ave (4 peaks):				358.2	RPD = 1	
Corrected Ave (4 peaks):				299.1	Corrected Ave (3 peaks):				309.5	RPD = 3	
Aroclor-1262	1	10.828	-0.001	121431	251.3	1	11.201	0.000	121335	251.1	
Aroclor-1262	2	12.244	0.000	196033	249.3	2	11.652	0.000	104071	252.9	
Aroclor-1262	3	12.319	0.001	211092	249.8	3	12.435	0.001	113645	243.4	
Aroclor-1262	4	12.988	0.001	183455	237.5	4	12.501	-0.001	182951	250.1	
Total CollAve (4 peaks):				247.0	Total Col2Ave (4 peaks):				249.3	RPD = 1	
Corrected Ave (3 peaks):				245.5	Corrected Ave (3 peaks):				248.2	RPD = 1	
Aroclor-1268	1	12.244	-0.002	196033	97.1	1	12.435	0.003	113645	99.7	
Aroclor-1268	2	12.319	0.002	211092	105.6	2	12.501	0.001	182951	149.3	
Aroclor-1268	3	12.723	0.024	77240	45.2	3	12.891	-0.000	7755	7.4	
Aroclor-1268	4	13.488	-0.002	65479	11.6	4	13.709	0.000	35146	10.5	
Total CollAve (4 peaks):				64.9	Total Col2Ave (4 peaks):				66.7	RPD = 3	

Corrected Ave (3 peaks): 51.3      Corrected Ave (3 peaks): 39.2      RPD = 27

Total PCB Area Col1 (5.906 - 13.793) = 3239932      Col1 Total PCB = 0.4 ppm\*  
Total PCB Area Col2 (5.785 - 14.019) = 1655522      Col2 Total PCB = 0.4 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR2162SCV


24-FEB-2023 16:27, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 AR2162SCV

24-FEB-2023 16:27, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230224.b/02242318ECD7.D  
Data file 2: /230224.b/230224.b/02242318ECD7.D  
Method: \\target\share\chem4\ecd7.i\230224.b\PCB.m  
Compound Sublist: PCB.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR3268SCV  
Client ID:  
Injection Date: 24-FEB-2023 16:48  
Report Date: 02/28/2023 09:51  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.806	0.000	363331	5.685	0.000	176204	37.1	38.2	2.9	Tetrachloro-m-xylene
13.894	0.001	800845	14.118	-0.001	488290	51.3	56.4	9.5	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	656592	-2.6
Hexabromobiphenyl	1429847	1584453	10.8

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	314741	-0.2
Hexabromobiphenyl	513946	568346	10.6

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col						
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.270	-0.001	28327	113.6	1	7.254	-0.001	20651	112.1	
Aroclor-1016	2	7.657	0.003	80668	106.1	2	7.861	0.005	41326	110.6	
Aroclor-1016	3	7.793	0.003	40661	109.6	3	8.060	0.005	20446	121.2	
Aroclor-1016	4	8.407	0.002	24680	102.9	4	8.308	0.001	13576	102.5	
Total CollAve (4 peaks):				108.0	Total Col2Ave (4 peaks):				111.6	RPD = 3	
Corrected Ave (3 peaks):				106.2	Corrected Ave (3 peaks):				108.4	RPD = 2	
Aroclor-1221	1	4.729	-0.001	8535	145.1	1	4.956	-0.000	3965	133.1	
Aroclor-1221	2	6.132	-0.000	15523	147.6	2	6.297	0.001	8689	154.1	
Aroclor-1221	3	6.382	-0.000	45872	187.9	3	6.622	0.001	22272	242.6	
Total CollAve (3 peaks):				160.2	Total Col2Ave (3 peaks):				176.6	RPD = 10	
Corrected Ave: < 3 Peaks					Corrected Ave: < 3 Peaks						
Aroclor-1232	1	4.729	-0.001	8535	242.6	1	4.956	0.000	3965	246.2	
Aroclor-1232	2	6.132	0.001	15523	222.5	2	7.254	0.000	20651	258.1	
Aroclor-1232	3	7.657	0.001	80668	255.4	3	7.861	0.001	41326	258.3	
Aroclor-1232	4	8.582	0.001	34784	259.2	4	8.714	-0.001	12504	271.5	
Total CollAve (4 peaks):				244.9	Total Col2Ave (4 peaks):				258.5	RPD = 5	
Corrected Ave (3 peaks):				240.2	Corrected Ave (3 peaks):				254.2	RPD = 6	
Aroclor-1242	1	7.270	-0.001	28327	139.2	1	7.254	-0.001	20651	141.2	
Aroclor-1242	2	7.657	0.001	80668	130.5	2	7.861	0.003	41326	134.4	
Aroclor-1242	3	8.407	0.001	24680	128.4	3	9.170	0.003	12830	134.1	
Aroclor-1242	4	8.582	0.003	34784	122.4	4	9.600	0.003	14836	127.3	
Total CollAve (4 peaks):				130.1	Total Col2Ave (4 peaks):				134.3	RPD = 3	
Corrected Ave (3 peaks):				127.1	Corrected Ave (3 peaks):				132.0	RPD = 4	
Aroclor-1248	1	8.407	0.001	24680	77.0	1	8.308	0.000	13576	90.3	
Aroclor-1248	2	8.582	0.001	34784	85.4	2	8.714	-0.000	12504	80.5	
Aroclor-1248	3	8.996	-0.003	83592	108.8	3	9.170	0.004	12830	71.8	
Aroclor-1248	4	9.292	-0.003	39603	101.3	4	9.600	0.010	14836	69.1	
Total CollAve (4 peaks):				93.1	Total Col2Ave (4 peaks):				77.9	RPD = 18	
Corrected Ave (3 peaks):				87.9	Corrected Ave (3 peaks):				73.8	RPD = 17	
Aroclor-1254	1	9.292	-0.007	39603	60.1	1	9.452	0.003	4590	19.2	
Aroclor-1254	2	9.377	-0.000	11450	38.6	2	9.973	0.003	2892	15.0	
Aroclor-1254	3	9.674	0.005	6387	15.1	3	10.131	0.007	6052	14.5	
Aroclor-1254	4	9.813	0.006	10162	12.3	4	10.390	0.017	5324	13.1	
Aroclor-1254	5	10.189	0.012	6862	13.3	5	10.572	0.004	1891	7.7	
Total CollAve (5 peaks):				27.9	Total Col2Ave (5 peaks):				13.9	RPD = 67*	
Corrected Ave (4 peaks):				19.8	Corrected Ave (4 peaks):				12.6	RPD = 45*	
Aroclor-1260	1	11.046	0.002	87033	152.7	1	11.645	-0.008	62543	187.1	
Aroclor-1260	2	11.362	0.001	6300	10.6	2	11.920	0.003	28552	33.5	
Aroclor-1260	3	11.738	0.004	54524	34.5	3	12.432	-0.004	285450	1261.2	
Aroclor-1260	4	12.144	0.005	1727	2.2	4	12.499	-0.002	306992	534.0	
Aroclor-1260	5	12.246	0.002	502931	1469.0	NS	---			----	
Total CollAve (5 peaks):				333.8	Total Col2Ave (4 peaks):				503.9	RPD = 41*	
Corrected Ave (4 peaks):				50.0	Corrected Ave (3 peaks):				251.5	RPD = 134*	
Aroclor-1262	1	10.832	0.004	3395	7.0	1	11.201	0.001	44255	91.2	
Aroclor-1262	2	12.246	0.002	502931	635.9	2	11.645	-0.007	62543	151.3	
Aroclor-1262	3	12.318	-0.000	497006	584.5	3	12.432	-0.002	285450	608.7	
Aroclor-1262	4	12.987	-0.000	202197	260.2	4	12.499	-0.003	306992	417.9	
Total CollAve (4 peaks):				371.9	Total Col2Ave (4 peaks):				317.3	RPD = 16	
Corrected Ave (3 peaks):				283.9	Corrected Ave (3 peaks):				220.1	RPD = 25	
Aroclor-1268	1	12.246	-0.001	502931	247.7	1	12.432	-0.000	285450	249.4	
Aroclor-1268	2	12.318	0.002	497006	247.2	2	12.499	-0.001	306992	249.5	
Aroclor-1268	3	12.699	-0.000	422793	245.8	3	12.892	0.000	260893	248.4	
Aroclor-1268	4	13.490	0.000	1386953	244.9	4	13.709	-0.000	829733	247.1	
Total CollAve (4 peaks):				246.4	Total Col2Ave (4 peaks):				248.6	RPD = 1	

Corrected Ave (3 peaks): 246.0      Corrected Ave (3 peaks): 248.3      RPD = 1

Total PCB Area Col1 (5.906 - 13.793) = 4180607      Col1 Total PCB = 0.5 ppm\*

Total PCB Area Col2 (5.785 - 14.019) = 2376912      Col2 Total PCB = 0.6 ppm\*

\* Quantitated against AR1660 0.25ppm in Ical


PCB-Form 10 Mod.



# PCB Dual Column Chromatograms

ECD7-ZB5 AR3268SCV


24-FEB-2023 16:48, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 AR3268SCV

24-FEB-2023 16:48, 2ul



ZB-35 Manual Integration: NO

Analytical Resources Inc.  
8082 DDT SCREEN REPORT

Data file 1: /230224.b/02242319ECD7.D

ARI ID: DDTS

RT	ZB5 Col Shift Response	ZB35 Col Shift Response	RT	ZB5 on col	ZB35 on col	RPD	Compound/Flag
9.261	0.000 694353	9.912 0.000 580269	0.100	0.100	0.0	2,4-DDE	
0.000	-10.293 0	10.672 0.000 673479	0.000	0.200#	----	2,4-DDT	
9.686	0.000 1191406	10.212 0.000 433373	0.100	0.100	0.0	4,4-DDE	
10.259	0.000 1721760	10.672 0.000 673479	0.100	0.200#	66.7*	4,4-DDD	

# Indicates value is from co-eluting peaks

\* Indicates RPD > 40%

Analytical Resources Inc.  
8082 DDT SCREEN REPORT

Data file 1: /230224.b/02242320ECD7.D

ARI ID: DDT BD

RT	ZB5 Col Shift Response		ZB35 Col Shift Response		ZB5 on col	ZB35 on col	RPD	Compound/Flag	
9.285	0.023	4923	9.921	0.009	9972	0.001	0.002	84.3*	2,4-DDE
0.000	-10.293	0	10.677	0.004	249094	0.000	0.074#	----	2,4-DDT
9.692	0.006	12128	10.221	0.009	528	0.001	0.000	156.7*	4,4-DDE
10.265	0.006	410017	10.677	0.004	249094	0.023	0.074#	103.6*	4,4-DDD

# Indicates value is from co-eluting peaks

\* Indicates RPD > 40%



**SECOND-SOURCE CALIBRATION VERIFICATION**  
**EPA 8082A**

**Laboratory:** Analytical Resources, LLC

**SDG:** 23B0228

**Client:** Anchor QEA, LLC

**Project:** AOC5 MR Phase 1

**Calibration:** GB00069

**Laboratory ID:** SLB0342-SCV1

**Sequence:** SLB0342

**Sequence Name:** AR1660SCV1

**Standard ID:** L002065

<b>ANALYTE</b>	<b>EXPECTED (ug/L)</b>	<b>FOUND (ug/L)</b>	<b>% DRIFT</b>	<b>QC LIMIT</b>
Aroclor 1016	250.00	243	-2.7	20.00
Aroclor 1016 [2C]	250.00	246	-1.4	20.00
Aroclor 1260	250.00	266	6.2	20.00
Aroclor 1260 [2C]	250.00	261	4.5	20.00
Decachlorobiphenyl	40.000	34.3	-14.2	20.00
Tetrachlorometaxylene	40.000	34.9	-12.6	20.00
Decachlorobiphenyl [2C]	40.000	37.3	-6.6	20.00
Tetrachlorometaxylene [2C]	40.000	35.8	-10.6	20.00

\* Indicates values outside of QC limits  
[2C] indicates second-column analyte.



**SECOND-SOURCE CALIBRATION VERIFICATION**  
**EPA 8082A**

**Laboratory:** Analytical Resources, LLC

**SDG:** 23B0228

**Client:** Anchor QEA, LLC

**Project:** AOC5 MR Phase 1

**Calibration:** GB00069

**Laboratory ID:** SLB0342-SCV2

**Sequence:** SLB0342

**Sequence Name:** AR1242SCV2

**Standard ID:** K007656

<b>ANALYTE</b>	<b>EXPECTED (ug/L)</b>	<b>FOUND (ug/L)</b>	<b>% DRIFT</b>	<b>QC LIMIT</b>
Aroclor 1242	250.00	204	-18.5	20.00
Aroclor 1242 [2C]	250.00	221	-11.8	20.00
Decachlorobiphenyl	40.000	37.0	-7.5	20.00
Tetrachlorometaxylene	40.000	33.6	-15.9	20.00
Decachlorobiphenyl [2C]	40.000	40.3	0.8	20.00
Tetrachlorometaxylene [2C]	40.000	34.5	-13.7	20.00

\* Indicates values outside of QC limits

[2C] indicates second-column analyte.



**SECOND-SOURCE CALIBRATION VERIFICATION**  
**EPA 8082A**

**Laboratory:** Analytical Resources, LLC

**SDG:** 23B0228

**Client:** Anchor QEA, LLC

**Project:** AOC5 MR Phase 1

**Calibration:** GB00069

**Laboratory ID:** SLB0342-SCV3

**Sequence:** SLB0342

**Sequence Name:** AR1248SCV3

**Standard ID:** L002066

<b>ANALYTE</b>	<b>EXPECTED (ug/L)</b>	<b>FOUND (ug/L)</b>	<b>% DRIFT</b>	<b>QC LIMIT</b>
Aroclor 1248	250.00	249	-0.2	20.00
Aroclor 1248 [2C]	250.00	248	-0.8	20.00
Decachlorobiphenyl	40.000	33.1	-17.2	20.00
Tetrachlorometaxylene	40.000	34.9	-12.8	20.00
Decachlorobiphenyl [2C]	40.000	36.3	-9.2	20.00
Tetrachlorometaxylene [2C]	40.000	36.4	-9.0	20.00

\* Indicates values outside of QC limits  
[2C] indicates second-column analyte.



**SECOND-SOURCE CALIBRATION VERIFICATION**  
**EPA 8082A**

**Laboratory:** Analytical Resources, LLC

**SDG:** 23B0228

**Client:** Anchor QEA, LLC

**Project:** AOC5 MR Phase 1

**Calibration:** GB00069

**Laboratory ID:** SLB0342-SCV4

**Sequence:** SLB0342

**Sequence Name:** AR1254SCV4

**Standard ID:** L002067

<b>ANALYTE</b>	<b>EXPECTED (ug/L)</b>	<b>FOUND (ug/L)</b>	<b>% DRIFT</b>	<b>QC LIMIT</b>
Aroclor 1254	250.00	235	-5.9	20.00
Aroclor 1254 [2C]	250.00	240	-4.0	20.00
Decachlorobiphenyl	40.000	34.6	-13.4	20.00
Tetrachlorometaxylene	40.000	36.1	-9.7	20.00
Decachlorobiphenyl [2C]	40.000	37.9	-5.2	20.00
Tetrachlorometaxylene [2C]	40.000	37.1	-7.3	20.00

\* Indicates values outside of QC limits

[2C] indicates second-column analyte.



**SECOND-SOURCE CALIBRATION VERIFICATION**  
**EPA 8082A**

**Laboratory:** Analytical Resources, LLC

**SDG:** 23B0228

**Client:** Anchor QEA, LLC

**Project:** AOC5 MR Phase 1

**Calibration:** GB00069

**Laboratory ID:** SLB0342-SCV5

**Sequence:** SLB0342

**Sequence Name:** AR2162SCV5

**Standard ID:** L002068

<b>ANALYTE</b>	<b>EXPECTED (ug/L)</b>	<b>FOUND (ug/L)</b>	<b>% DRIFT</b>	<b>QC LIMIT</b>
Aroclor 1221	250.00	258	3.3	20.00
Aroclor 1221 [2C]	250.00	256	2.5	20.00
Aroclor 1262	250.00	247	-1.2	20.00
Aroclor 1262 [2C]	250.00	249	-0.3	20.00
Decachlorobiphenyl	40.000	34.4	-13.9	20.00
Tetrachlorometaxylene	40.000	36.0	-10.0	20.00
Decachlorobiphenyl [2C]	40.000	37.9	-5.4	20.00
Tetrachlorometaxylene [2C]	40.000	36.6	-8.4	20.00

\* Indicates values outside of QC limits  
[2C] indicates second-column analyte.





**SECOND-SOURCE CALIBRATION VERIFICATION**  
**EPA 8082A**

**Laboratory:** Analytical Resources, LLC

**SDG:** 23B0228

**Client:** Anchor QEA, LLC

**Project:** AOC5 MR Phase 1

**Calibration:** GB00069

**Laboratory ID:** SLB0342-SCV6

**Sequence:** SLB0342

**Sequence Name:** AR3268SCV6

**Standard ID:** L002069

<b>ANALYTE</b>	<b>EXPECTED (ug/L)</b>	<b>FOUND (ug/L)</b>	<b>% DRIFT</b>	<b>QC LIMIT</b>
Aroclor 1232	250.00	245	-2.0	20.00
Aroclor 1232 [2C]	250.00	259	3.4	20.00
Aroclor 1268	250.00	246	-1.4	20.00
Aroclor 1268 [2C]	250.00	249	-0.6	20.00
Decachlorobiphenyl	40.000	51.3	28.3	20.00
Tetrachlorometaxylene	40.000	37.1	-7.3	20.00
Decachlorobiphenyl [2C]	40.000	56.4	41.0	20.00
Tetrachlorometaxylene [2C]	40.000	38.2	-4.6	20.00

\* Indicates values outside of QC limits  
[2C] indicates second-column analyte.



**INITIAL CALIBRATION CHECK**  
**EPA 8082A**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor OEA, LLC

Project: AOC5 MR Phase 1

Instrument ID: ECD7

Calibration: GB00069

Lab File ID: 02282310ECD7.D

Calibration Date: 02/24/2023

Sequence: SLC0014

Injection Date: 02/28/23

Lab Sample ID: SLC0014-ICV1

Injection Time: 19:04

Sequence Name: AR1254ICV1

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR			% DRIFT/DIFF	
		STD	ICV	ICAL	ICV	MIN	ICV	LIMIT
Aroclor 1254	A	250.00	284	0.0662949	0.0748899		13.7	+/-20
Aroclor-1254 (1)	A	250.00	282	0.0803331	0.0904834			
Aroclor-1254 (2)	A	250.00	292	0.0361302	0.0421510			
Aroclor-1254 (3)	A	250.00	291	0.0516471	0.0600635			
Aroclor-1254 (4)	A	250.00	278	0.1004230	0.1118357			
Aroclor-1254 (5)	A	250.00	278	0.0629414	0.0699159			
Aroclor 1254 [2C]	A	250.00	263	0.0763106	0.0803775		5.4	+/-20
Aroclor-1254 (1) [2C]	A	250.00	264	0.0608052	0.0641940			
Aroclor-1254 (2) [2C]	A	250.00	266	0.0489162	0.0521536			
Aroclor-1254 (3) [2C]	A	250.00	266	0.1058376	0.1127019			
Aroclor-1254 (4) [2C]	A	250.00	258	0.1031750	0.1067184			
Aroclor-1254 (5) [2C]	A	250.00	263	0.0628191	0.0661198			
Decachlorobiphenyl	A	40.000	37.4	0.7878687	0.7377254		-6.5	+/-20
Tetrachlorometaxylene	A	40.000	40.1	1.1944880	1.1986300		0.3	+/-20
Decachlorobiphenyl [2C]	A	40.000	41.4	1.2182710	1.2609770		3.5	+/-20
Tetrachlorometaxylene [2C]	A	40.000	40.4	1.1737210	1.1848750		1.0	+/-20

\* Values outside of QC limits

\* Values outside of QC limits

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282310ECD7.D  
Data file 2: /230228.b/230228.b/02282310ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: AR1254.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1254ICV1  
Client ID:  
Injection Date: 28-FEB-2023 19:04  
Report Date: 03/01/2023 12:20  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.809	0.002	447919	5.688	0.001	178473	40.1	40.4	0.6	Tetrachloro-m-xylene
13.894	0.001	610235	14.119	-0.001	294184	37.5	41.4	10.0	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	747385	10.9
Hexabromobiphenyl	1429847	1654369	15.7

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	301252	-4.4
Hexabromobiphenyl	513946	466597	-9.2

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1254	1	9.300	0.001	211331	281.6	1	9.451	0.000	60433	263.9
Aroclor-1254	2	9.378	-0.000	98447	291.7	2	9.971	-0.001	49098	266.5
Aroclor-1254	3	9.670	0.001	140283	290.7	3	10.125	-0.000	106099	266.2
Aroclor-1254	4	9.809	0.001	261201	278.4	4	10.375	0.001	100466	258.6
Aroclor-1254	5	10.179	0.001	163294	277.7	5	10.570	0.000	62246	263.1
Total CollAve (5 peaks):				284.0		Total Col2Ave (5 peaks):				263.7 RPD = 7
Corrected Ave (4 peaks):				282.1		Corrected Ave (4 peaks):				263.0 RPD = 7
CalAmt %D:				13.6		CalAmt %D:				5.5

Total PCB Area Col1 (5.907 - 13.793) = 3085449 Col1 Total PCB = 0.3 ppm\*

Total PCB Area Col2 (5.787 - 14.020) = 1027352 Col2 Total PCB = 0.3 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR1254ICV1


28-FEB-2023 19:04, 2ul



ZB-5 Manual Integration: YES

ECD7-ZB35 AR1254ICV1

28-FEB-2023 19:04, 2ul




ZB-35 Manual Integration: NO


# Manual Peak Adjustment, ZB-5

Datafile: ecd7.i/230228.b/02282310ECD7.D Injection Date: 28-FEB-2023 19:04

## Manual Integration (After)



## Processed Integration (Before)





INITIAL CALIBRATION CHECK  
EPA 8082A

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Instrument ID: ECD7

Calibration: GB00069

Lab File ID: 02282311ECD7.D

Calibration Date: 02/24/2023

Sequence: SLC0014

Injection Date: 02/28/23

Lab Sample ID: SLC0014-ICV2

Injection Time: 19:25

Sequence Name: AR1660ICV2

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR			% DRIFT/DIFF	
		STD	ICV	ICAL	ICV	MIN	ICV	LIMIT
Aroclor 1016	A	250.00	280	0.0493662	0.0552775		12.0	+/-20
Aroclor-1016 (1)	A	250.00	277	0.0303852	0.0337077		10.8	
Aroclor-1016 (2)	A	250.00	278	0.0926308	0.1031854		11.2	
Aroclor-1016 (3)	A	250.00	284	0.0452180	0.0513640		13.6	
Aroclor-1016 (4)	A	250.00	281	0.0292307	0.0328528		12.4	
Aroclor 1016 [2C]	A	250.00	267	0.0545857	0.0586892		6.6	+/-20
Aroclor-1016 (1) [2C]	A	250.00	258	0.0468313	0.0483572		3.2	
Aroclor-1016 (2) [2C]	A	250.00	278	0.0949676	0.1054261		11.2	
Aroclor-1016 (3) [2C]	A	250.00	260	0.0428922	0.0445645		4.0	
Aroclor-1016 (4) [2C]	A	250.00	270	0.0336515	0.0364091		8.0	
Aroclor 1260	A	250.00	308	0.0392091	0.0484773		23.3	+/-20 *
Aroclor-1260 (1)	A	250.00	282	0.0287785	0.0324665		12.8	
Aroclor-1260 (2)	A	250.00	318	0.0300690	0.0382204		27.2	
Aroclor-1260 (3)	A	250.00	310	0.0797517	0.0989819		24.0	
Aroclor-1260 (4)	A	250.00	318	0.0401599	0.0510785		27.2	
Aroclor-1260 (5)	A	250.00	313	0.0172866	0.0216391		25.2	
Aroclor 1260 [2C]	A	250.00	252	0.0699688	0.0716171		0.7	+/-20
Aroclor-1260 (1) [2C]	A	250.00	237	0.0470406	0.0445304		-5.2	
Aroclor-1260 (2) [2C]	A	250.00	260	0.1200523	0.1249385		4.0	
Aroclor-1260 (3) [2C]	A	250.00	246	0.0318590	0.0313749		-1.6	
Aroclor-1260 (4) [2C]	A	250.00	264	0.0809231	0.0856245		5.6	
Decachlorobiphenyl	A	40.000	40.5	0.7878687	0.7977915		1.3	+/-20
Tetrachlorometaxylene	A	40.000	41.8	1.1944880	1.2496600		4.5	+/-20
Decachlorobiphenyl [2C]	A	40.000	40.4	1.2182710	1.2317770		1.0	+/-20
Tetrachlorometaxylene [2C]	A	40.000	42.2	1.1737210	1.2392490		5.5	+/-20

\* Values outside of QC limits

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282311ECD7.D  
Data file 2: /230228.b/230228.b/02282311ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1660ICV2  
Client ID:  
Injection Date: 28-FEB-2023 19:25  
Report Date: 03/01/2023 12:20  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.808	0.000	469361	5.687	-0.000	189615	41.8	42.2	0.9	Tetrachloro-m-xylene
13.893	-0.001	628838	14.118	-0.002	286648	40.5	40.4	0.1	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	751182	11.5
Hexabromobiphenyl	1429847	1576447	10.3

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	306016	-2.9
Hexabromobiphenyl	513946	465422	-9.4

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)



ZB5 Col					ZB35 Col						
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.269	0.000	79127	277.3	1	7.254	-0.001	46244	258.1	
Aroclor-1016	2	7.655	-0.000	242222	278.5	2	7.858	-0.001	100819	277.5	
Aroclor-1016	3	7.792	0.000	120574	284.0	3	8.057	-0.002	42617	259.7	
Aroclor-1016	4	8.405	-0.001	77120	281.0	4	8.308	-0.000	34818	270.5	
Total CollAve (4 peaks):				280.2		Total Col2Ave (4 peaks):				266.5	RPD = 5
Corrected Ave (3 peaks):				278.9		Corrected Ave (3 peaks):				262.8	RPD = 6

CalAmt %D: 12.1

CalAmt %D: 6.6

Aroclor-1260	1	11.045	0.001	159943	282.0	1	11.653	-0.000	64767	236.7	
Aroclor-1260	2	11.362	0.001	188289	317.8	2	11.919	0.001	181716	260.2	
Aroclor-1260	3	11.736	0.001	487624	310.3	3	12.436	0.001	45633	246.2	
Aroclor-1260	4	12.140	0.000	251633	318.0	4	12.502	0.000	124536	264.5	
Aroclor-1260	5	12.244	0.001	106603	312.9	NS	---			----	
Total CollAve (5 peaks):				308.2		Total Col2Ave (4 peaks):				251.9	RPD = 20
Corrected Ave (4 peaks):				305.8		Corrected Ave (3 peaks):				247.7	RPD = 21

CalAmt %D: 23.3

CalAmt %D: 0.8

Total PCB Area Coll (5.907 - 13.793) = 5090549 Coll Total PCB = 0.6 ppm\*

Total PCB Area Col2 (5.787 - 14.020) = 1776656 Col2 Total PCB = 0.5 ppm\*

\* Quantitated against AR1660 0.25ppm in Ical


PCB-Form 10 Mod.




Manual Peak Adjustment, ZB-35

Datafile: ecd7.i/230228.b/230228.b/02282311ECD7.D Injection Date: 28-FEB-2023

Manual Integration (After)



Processed Integration (Before)





**SECOND-SOURCE  
CONTINUING CALIBRATION CHECK  
EPA 8082A**

Laboratory:	<u>Analytical Resources, LLC</u>	SDG:	<u>23B0228</u>
Client:	<u>Anchor QEA, LLC</u>	Project:	<u>AOC5 MR Phase 1</u>
Instrument ID:	<u>ECD7</u>	Calibration:	<u>GB00069</u>
Lab File ID:	<u>02242313ECD7.D</u>	Calibration Date:	<u>02/24/2023</u>
Sequence:	<u>SLB0342</u>	Injection Date:	<u>02/24/23</u>
Lab Sample ID:	<u>SLB0342-SCV1</u>	Injection Time:	<u>15:03</u>
Sequence Name:	<u>AR1660SCV1</u>		

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR (RRF)			% DRIFT/DIFF	
		STD	CCV	ICAL	CCV	MIN	CCV	LIMIT
Aroclor 1016	A	250.00	243	0.0493662	0.0479314		-2.7	+/-20
Aroclor 1016 [2C]	A	250.00	246	0.0545857	0.0542382		-1.4	+/-20
Aroclor 1260	A	250.00	266	0.0392091	0.0412121		6.2	+/-20
Aroclor 1260 [2C]	A	250.00	261	0.0699688	0.0733659		4.5	+/-20
Decachlorobiphenyl	A	40.000	34.3	0.7878687	0.6762784		-14.2	+/-20
Tetrachlorometaxylene	A	40.000	34.9	1.1944880	1.0436010		-12.6	+/-20
Decachlorobiphenyl [2C]	A	40.000	37.3	1.2182710	1.1373730		-6.6	+/-20
Tetrachlorometaxylene [2C]	A	40.000	35.8	1.1737210	1.0492890		-10.6	+/-20

\* Values outside of QC limits

\* Values outside of QC limits



**SECOND-SOURCE  
CONTINUING CALIBRATION CHECK  
EPA 8082A**

Laboratory:	<u>Analytical Resources, LLC</u>	SDG:	<u>23B0228</u>
Client:	<u>Anchor QEA, LLC</u>	Project:	<u>AOC5 MR Phase 1</u>
Instrument ID:	<u>ECD7</u>	Calibration:	<u>GB00069</u>
Lab File ID:	<u>02242314ECD7.D</u>	Calibration Date:	<u>02/24/2023</u>
Sequence:	<u>SLB0342</u>	Injection Date:	<u>02/24/23</u>
Lab Sample ID:	<u>SLB0342-SCV2</u>	Injection Time:	<u>15:24</u>
Sequence Name:	<u>AR1242SCV2</u>		

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR (RRF)			% DRIFT/DIFF	
		STD	CCV	ICAL	CCV	MIN	CCV	LIMIT
Aroclor 1242	A	250.00	204	0.0395340	0.0322254		-18.5	+/-20
Aroclor 1242 [2C]	A	250.00	221	0.0423092	0.0365983		-11.8	+/-20
Decachlorobiphenyl	A	40.000	37.0	0.7878687	0.7290534		-7.5	+/-20
Tetrachlorometaxylene	A	40.000	33.6	1.1944880	1.0041320		-15.9	+/-20
Decachlorobiphenyl [2C]	A	40.000	40.3	1.2182710	1.2285170		0.8	+/-20
Tetrachlorometaxylene [2C]	A	40.000	34.5	1.1737210	1.0131510		-13.7	+/-20

\* Values outside of QC limits

\* Values outside of QC limits



**SECOND-SOURCE  
CONTINUING CALIBRATION CHECK  
EPA 8082A**

Laboratory:	<u>Analytical Resources, LLC</u>	SDG:	<u>23B0228</u>
Client:	<u>Anchor QEA, LLC</u>	Project:	<u>AOC5 MR Phase 1</u>
Instrument ID:	<u>ECD7</u>	Calibration:	<u>GB00069</u>
Lab File ID:	<u>02242315ECD7.D</u>	Calibration Date:	<u>02/24/2023</u>
Sequence:	<u>SLB0342</u>	Injection Date:	<u>02/24/23</u>
Lab Sample ID:	<u>SLB0342-SCV3</u>	Injection Time:	<u>15:45</u>
Sequence Name:	<u>AR1248SCV3</u>		

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR (RRF)			% DRIFT/DIFF	
		STD	CCV	ICAL	CCV	MIN	CCV	LIMIT
Aroclor 1248	A	250.00	249	0.0574755	0.0572474		-0.2	+/-20
Aroclor 1248 [2C]	A	250.00	248	0.0444270	0.0440936		-0.8	+/-20
Decachlorobiphenyl	A	40.000	33.1	0.7878687	0.6527336		-17.2	+/-20
Tetrachlorometaxylene	A	40.000	34.9	1.1944880	1.0413820		-12.8	+/-20
Decachlorobiphenyl [2C]	A	40.000	36.3	1.2182710	1.1066400		-9.2	+/-20
Tetrachlorometaxylene [2C]	A	40.000	36.4	1.1737210	1.0676190		-9.0	+/-20

\* Values outside of QC limits

\* Values outside of QC limits



**SECOND-SOURCE  
CONTINUING CALIBRATION CHECK  
EPA 8082A**

Laboratory:	<u>Analytical Resources, LLC</u>	SDG:	<u>23B0228</u>
Client:	<u>Anchor QEA, LLC</u>	Project:	<u>AOC5 MR Phase 1</u>
Instrument ID:	<u>ECD7</u>	Calibration:	<u>GB00069</u>
Lab File ID:	<u>02242316ECD7.D</u>	Calibration Date:	<u>02/24/2023</u>
Sequence:	<u>SLB0342</u>	Injection Date:	<u>02/24/23</u>
Lab Sample ID:	<u>SLB0342-SCV4</u>	Injection Time:	<u>16:06</u>
Sequence Name:	<u>AR1254SCV4</u>		

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR (RRF)			% DRIFT/DIFF	
		STD	CCV	ICAL	CCV	MIN	CCV	LIMIT
Aroclor 1254	A	250.00	235	0.0662949	0.0622529		-5.9	+/-20
Aroclor 1254 [2C]	A	250.00	240	0.0763106	0.0731447		-4.0	+/-20
Decachlorobiphenyl	A	40.000	34.6	0.7878687	0.6823832		-13.4	+/-20
Tetrachlorometaxylene	A	40.000	36.1	1.1944880	1.0787610		-9.7	+/-20
Decachlorobiphenyl [2C]	A	40.000	37.9	1.2182710	1.1548440		-5.2	+/-20
Tetrachlorometaxylene [2C]	A	40.000	37.1	1.1737210	1.0880920		-7.3	+/-20

\* Values outside of QC limits

\* Values outside of QC limits



**SECOND-SOURCE  
CONTINUING CALIBRATION CHECK  
EPA 8082A**

Laboratory:	<u>Analytical Resources, LLC</u>	SDG:	<u>23B0228</u>
Client:	<u>Anchor QEA, LLC</u>	Project:	<u>AOC5 MR Phase 1</u>
Instrument ID:	<u>ECD7</u>	Calibration:	<u>GB00069</u>
Lab File ID:	<u>02242317ECD7.D</u>	Calibration Date:	<u>02/24/2023</u>
Sequence:	<u>SLB0342</u>	Injection Date:	<u>02/24/23</u>
Lab Sample ID:	<u>SLB0342-SCV5</u>	Injection Time:	<u>16:27</u>
Sequence Name:	<u>AR2162SCV5</u>		

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR (RRF)			% DRIFT/DIFF	
		STD	CCV	ICAL	CCV	MIN	CCV	LIMIT
Aroclor 1221	A	250.00	258	0.0165758	0.0169561		3.3	+/-20
Aroclor 1221 [2C]	A	250.00	256	0.0150798	0.0153801		2.5	+/-20
Aroclor 1262	A	250.00	247	0.0366596	0.0361658		-1.2	+/-20
Aroclor 1262 [2C]	A	250.00	249	0.0739760	0.0737876		-0.3	+/-20
Decachlorobiphenyl	A	40.000	34.4	0.7878687	0.6780614		-13.9	+/-20
Tetrachlorometaxylene	A	40.000	36.0	1.1944880	1.0756080		-10.0	+/-20
Decachlorobiphenyl [2C]	A	40.000	37.9	1.2182710	1.1528740		-5.4	+/-20
Tetrachlorometaxylene [2C]	A	40.000	36.6	1.1737210	1.0753820		-8.4	+/-20

\* Values outside of QC limits

\* Values outside of QC limits





**SECOND-SOURCE  
CONTINUING CALIBRATION CHECK  
EPA 8082A**

Laboratory:	<u>Analytical Resources, LLC</u>	SDG:	<u>23B0228</u>
Client:	<u>Anchor QEA, LLC</u>	Project:	<u>AOC5 MR Phase 1</u>
Instrument ID:	<u>ECD7</u>	Calibration:	<u>GB00069</u>
Lab File ID:	<u>02242318ECD7.D</u>	Calibration Date:	<u>02/24/2023</u>
Sequence:	<u>SLB0342</u>	Injection Date:	<u>02/24/23</u>
Lab Sample ID:	<u>SLB0342-SCV6</u>	Injection Time:	<u>16:48</u>
Sequence Name:	<u>AR3268SCV6</u>		

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR (RRF)			% DRIFT/DIFF	
		STD	CCV	ICAL	CCV	MIN	CCV	LIMIT
Aroclor 1232	A	250.00	245	0.0169039	0.0169981		-2.0	+/-20
Aroclor 1232 [2C]	A	250.00	259	0.0192023	0.0199392		3.4	+/-20
Aroclor 1268	A	250.00	246	0.1442124	0.1418626		-1.4	+/-20
Aroclor 1268 [2C]	A	250.00	249	0.2386862	0.2369075		-0.6	+/-20
Decachlorobiphenyl	A	40.000	51.3	0.7878687	1.0108790		28.3	+/-20
Tetrachlorometaxylene	A	40.000	37.1	1.1944880	1.1067180		-7.3	+/-20
Decachlorobiphenyl [2C]	A	40.000	56.4	1.2182710	1.7182840		41.0	+/-20
Tetrachlorometaxylene [2C]	A	40.000	38.2	1.1737210	1.1196760		-4.6	+/-20

\* Values outside of QC limits



**CONTINUING CALIBRATION CHECK**  
**EPA 8082A**

Laboratory:	<u>Analytical Resources, LLC</u>	SDG:	<u>23B0228</u>
Client:	<u>Anchor QEA, LLC</u>	Project:	<u>AOC5 MR Phase 1</u>
Instrument ID:	<u>ECD7</u>	Calibration:	<u>GB00069</u>
Lab File ID:	<u>02282313ECD7.D</u>	Calibration Date:	<u>02/24/2023</u>
Sequence:	<u>SLC0014</u>	Injection Date:	<u>02/28/23</u>
Lab Sample ID:	<u>SLC0014-CCV1</u>	Injection Time:	<u>20:07</u>
Sequence Name:	<u>AR1248CCV1</u>		

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR (RRF)			% DRIFT/DIFF	
		STD	CCV	ICAL	CCV	MIN	CCV	LIMIT
Aroclor 1248	A	250.00	269	0.0574755	0.0623062		7.6	+/-20
Aroclor-1248 (1)	A	250.00	262		0.0409869			
Aroclor-1248 (2)	A	250.00	263		0.0521196			
Aroclor-1248 (3)	A	250.00	278		0.1040716			
Aroclor-1248 (4)	A	250.00	273		0.0520464			
Aroclor 1248 [2C]	A	250.00	251	0.0444270	0.0445547		0.5	+/-20
Aroclor-1248 (1) [2C]	A	250.00	256		0.0391330			
Aroclor-1248 (2) [2C]	A	250.00	254		0.0400822			
Aroclor-1248 (3) [2C]	A	250.00	248		0.0450484			
Aroclor-1248 (4) [2C]	A	250.00	247		0.0539552			
Decachlorobiphenyl	A	40.000	36.6	0.7878687	0.7215080		-8.5	+/-20
Tetrachlorometaxylene	A	40.000	37.8	1.1944880	1.1297030		-5.5	+/-20
Decachlorobiphenyl [2C]	A	40.000	40.0	1.2182710	1.2199110		0.0	+/-20
Tetrachlorometaxylene [2C]	A	40.000	38.5	1.1737210	1.1298520		-3.8	+/-20

\* Values outside of QC limits

\* Values outside of QC limits

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282313ECD7.D  
Data file 2: /230228.b/230228.b/02282313ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: AR1248.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1248CCV1  
Client ID:  
Injection Date: 28-FEB-2023 20:07  
Report Date: 03/01/2023 12:20  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.808	0.001	442341	5.687	-0.000	182079	37.8	38.5	1.8	Tetrachloro-m-xylene
13.894	0.001	593120	14.118	-0.001	303850	36.6	40.1	8.9	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	783110	16.2
Hexabromobiphenyl	1429847	1644112	15.0

Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	322306	2.2
Hexabromobiphenyl	513946	498151	-3.1

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1248	1	8.406	0.000	100304	262.5	1	8.307	0.000	39415	256.1	
Aroclor-1248	2	8.581	0.000	127548	262.6	2	8.715	0.000	40371	253.7	
Aroclor-1248	3	8.997	0.000	254686	278.0	3	9.169	0.000	45373	247.8	
Aroclor-1248	4	9.295	0.000	127369	273.1	4	9.594	0.000	54344	247.2	
Total CollAve (4 peaks):				269.0	Total Col2Ave (4 peaks):				251.2	RPD = 7	
Corrected Ave (3 peaks):				266.1	Corrected Ave (3 peaks):				249.6	RPD = 6	
CalAmt %D:				7.6	CalAmt %D:				0.5		

Total PCB Area Col1 (5.907 - 13.793) = 1982415 Col1 Total PCB = 0.2 ppm\*

Total PCB Area Col2 (5.787 - 14.020) = 774039 Col2 Total PCB = 0.2 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR1248CCV1


28-FEB-2023 20:07, 2ul



ZB-5 Manual Integration: YES

ECD7-ZB35 AR1248CCV1

28-FEB-2023 20:07, 2ul




ZB-35 Manual Integration: NO

# Manual Peak Adjustment, ZB-5


Datafile: ecd7.i/230228.b/02282313ECD7.D

Injection Date: 28-FEB-2023 20:07

## Manual Integration (After)



## Processed Integration (Before)





**CONTINUING CALIBRATION CHECK**  
**EPA 8082A**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Instrument ID: ECD7

Calibration: GB00069

Lab File ID: 02282314ECD7.D

Calibration Date: 02/24/2023

Sequence: SLC0014

Injection Date: 02/28/23

Lab Sample ID: SLC0014-CCV2

Injection Time: 20:28

Sequence Name: AR1660CCV2

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR (RRF)			% DRIFT/DIFF	
		STD	CCV	ICAL	CCV	MIN	CCV	LIMIT
Aroclor 1016	A	250.00	269	0.0493662	0.0531918		7.6	+/-20
Aroclor-1016 (1)	A	250.00	264	0.0303852	0.0320347		5.6	
Aroclor-1016 (2)	A	250.00	270	0.0926308	0.0998609		8.0	
Aroclor-1016 (3)	A	250.00	274	0.0452180	0.0495154		9.6	
Aroclor-1016 (4)	A	250.00	268	0.0292307	0.0313563		7.2	
Aroclor 1016 [2C]	A	250.00	261	0.0545857	0.0576031		4.5	+/-20
Aroclor-1016 (1) [2C]	A	250.00	252	0.0468313	0.0472870		0.8	
Aroclor-1016 (2) [2C]	A	250.00	272	0.0949676	0.1035346		8.8	
Aroclor-1016 (3) [2C]	A	250.00	254	0.0428922	0.0436436		1.6	
Aroclor-1016 (4) [2C]	A	250.00	267	0.0336515	0.0359471		6.8	
Aroclor 1260	A	250.00	302	0.0392091	0.0475954		20.8	+/-20 *
Aroclor-1260 (1)	A	250.00	277	0.0287785	0.0319038		10.8	
Aroclor-1260 (2)	A	250.00	310	0.0300690	0.0373331		24.0	
Aroclor-1260 (3)	A	250.00	306	0.0797517	0.0976402		22.4	
Aroclor-1260 (4)	A	250.00	310	0.0401599	0.0498416		24.0	
Aroclor-1260 (5)	A	250.00	307	0.0172866	0.0212581		22.8	
Aroclor 1260 [2C]	A	250.00	245	0.0699688	0.0695777		-2.2	+/-20
Aroclor-1260 (1) [2C]	A	250.00	230	0.0470406	0.0433241		-8.0	
Aroclor-1260 (2) [2C]	A	250.00	252	0.1200523	0.1210413		0.8	
Aroclor-1260 (3) [2C]	A	250.00	238	0.0318590	0.0303454		-4.8	
Aroclor-1260 (4) [2C]	A	250.00	258	0.0809231	0.0835999		3.2	
Decachlorobiphenyl	A	40.000	40.1	0.7878687	0.7892449		0.3	+/-20
Tetrachlorometaxylene	A	40.000	40.8	1.1944880	1.2188610		2.0	+/-20
Decachlorobiphenyl [2C]	A	40.000	40.6	1.2182710	1.2361950		1.5	+/-20
Tetrachlorometaxylene [2C]	A	40.000	41.3	1.1737210	1.2124060		3.3	+/-20

\* Values outside of QC limits

\* Values outside of QC limits

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282314ECD7.D  
Data file 2: /230228.b/230228.b/02282314ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1660CCV2  
Client ID:  
Injection Date: 28-FEB-2023 20:28  
Report Date: 03/01/2023 12:20  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.808	0.001	465648	5.688	0.000	189813	40.8	41.3	1.2	Tetrachloro-m-xylene
13.893	-0.000	625756	14.119	-0.001	297494	40.1	40.6	1.3	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	764071	13.4
Hexabromobiphenyl	1429847	1585708	10.9

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	313118	-0.7
Hexabromobiphenyl	513946	481306	-6.4

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)



ZB5 Col					ZB35 Col						
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.270	0.001	76490	263.6	1	7.255	-0.000	46270	252.4	
Aroclor-1016	2	7.657	0.002	238440	269.5	2	7.860	0.000	101308	272.6	
Aroclor-1016	3	7.792	0.001	118229	273.8	3	8.059	0.000	42705	254.4	
Aroclor-1016	4	8.406	-0.000	74870	268.2	4	8.307	-0.001	35174	267.1	
Total CollAve (4 peaks):				268.8	Total Col2Ave (4 peaks):				261.6	RPD = 3	
Corrected Ave (3 peaks):				267.1	Corrected Ave (3 peaks):				258.0	RPD = 3	
CalAmt %D:				7.5	CalAmt %D:				4.6		
Aroclor-1260	1	11.045	0.000	158094	277.1	1	11.654	0.001	65163	230.2	
Aroclor-1260	2	11.362	0.001	184998	310.4	2	11.918	0.001	182056	252.1	
Aroclor-1260	3	11.737	0.001	483840	306.1	3	12.436	0.001	45642	238.1	
Aroclor-1260	4	12.141	0.001	246982	310.3	4	12.503	0.001	125741	258.3	
Aroclor-1260	5	12.245	0.002	105341	307.4	NS	---			----	
Total CollAve (5 peaks):				302.3	Total Col2Ave (4 peaks):				244.7	RPD = 21	
Corrected Ave (4 peaks):				300.2	Corrected Ave (3 peaks):				240.1	RPD = 22	
CalAmt %D:				20.9	CalAmt %D:				-2.1		

Total PCB Area Coll (5.907 - 13.793) = 4921192 Coll Total PCB = 0.5 ppm\*

Total PCB Area Col2 (5.787 - 14.020) = 1783996 Col2 Total PCB = 0.5 ppm\*

\* Quantitated against AR1660 0.25ppm in Ical


PCB-Form 10 Mod.




Manual Peak Adjustment, ZB-35

Datafile: ecd7.i/230228.b/230228.b/02282314ECD7.D Injection Date: 28-FEB-2023

Manual Integration (After)



Processed Integration (Before)





**CONTINUING CALIBRATION CHECK**  
**EPA 8082A**

Laboratory:	<u>Analytical Resources, LLC</u>	SDG:	<u>23B0228</u>
Client:	<u>Anchor QEA, LLC</u>	Project:	<u>AOC5 MR Phase 1</u>
Instrument ID:	<u>ECD7</u>	Calibration:	<u>GB00069</u>
Lab File ID:	<u>02282331ECD7.D</u>	Calibration Date:	<u>02/24/2023</u>
Sequence:	<u>SLC0014</u>	Injection Date:	<u>03/01/23</u>
Lab Sample ID:	<u>SLC0014-CCV3</u>	Injection Time:	<u>02:26</u>
Sequence Name:	<u>AR1242CCV3</u>		

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR (RRF)			% DRIFT/DIFF	
		STD	CCV	ICAL	CCV	MIN	CCV	LIMIT
Aroclor 1242	A	250.00	250	0.0395340	0.0396605		0.1	+/-20
Aroclor-1242 (1)	A	250.00	249		0.0246910			
Aroclor-1242 (2)	A	250.00	251		0.0757041			
Aroclor-1242 (3)	A	250.00	250		0.0234345			
Aroclor-1242 (4)	A	250.00	251		0.0348122			
Aroclor 1242 [2C]	A	250.00	248	0.0423092	0.0421656		-0.8	+/-20
Aroclor-1242 (1) [2C]	A	250.00	251		0.0372722			
Aroclor-1242 (2) [2C]	A	250.00	251		0.0785934			
Aroclor-1242 (3) [2C]	A	250.00	246		0.0239032			
Aroclor-1242 (4) [2C]	A	250.00	244		0.0288938			
Decachlorobiphenyl	A	40.000	39.8	0.7878687	0.7842420		-0.5	+/-20
Tetrachlorometaxylene	A	40.000	44.9	1.1944880	1.3409470		12.3	+/-20
Decachlorobiphenyl [2C]	A	40.000	42.1	1.2182710	1.2828530		5.3	+/-20
Tetrachlorometaxylene [2C]	A	40.000	46.8	1.1737210	1.3739480		17.0	+/-20

\* Values outside of QC limits

\* Values outside of QC limits

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282331ECD7.D  
Data file 2: /230228.b/230228.b/02282331ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: AR1242.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1242CCV3  
Client ID:  
Injection Date: 01-MAR-2023 02:26  
Report Date: 03/01/2023 12:20  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

ZB5 Col		ZB35 Col			ZB5	ZB35	RPD	Compound/Flag	
RT	Shift	Response	RT	Shift	Response	on col			on col
5.807	-0.000	520986	5.687	-0.000	217825	44.9	46.8	4.2	Tetrachloro-m-xylene
13.892	-0.001	352496	14.118	-0.001	222500	39.8	42.1	5.6	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	777042	15.3
Hexabromobiphenyl	1429847	898947	-37.1

Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	317079	0.6
Hexabromobiphenyl	513946	346883	-32.5

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col						ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1242	1	7.269	0.000	59956	249.0	1	7.253	0.000	36932	250.7	
Aroclor-1242	2	7.656	0.000	183829	251.4	2	7.859	0.000	77876	251.5	
Aroclor-1242	3	8.405	0.000	56905	250.1	3	9.167	0.000	23685	245.8	
Aroclor-1242	4	8.580	0.000	84533	251.3	4	9.598	0.000	28630	243.8	
Total Col1Ave (4 peaks):				250.4	Total Col2Ave (4 peaks):				248.0	RPD = 1	
Corrected Ave (3 peaks):				250.1	Corrected Ave (3 peaks):				246.8	RPD = 1	
CalAmt %D:				0.2	CalAmt %D:				-0.8		

Total PCB Area Col1 (5.907 - 13.793) = 1461012      Col1 Total PCB = 0.2 ppm\*

Total PCB Area Col2 (5.787 - 14.020) = 562198      Col2 Total PCB = 0.1 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR1242CCV3


01-MAR-2023 02:26, 2ul



ZB-5 Manual Integration: NO

ECD7-ZB35 AR1242CCV3

01-MAR-2023 02:26, 2ul





**CONTINUING CALIBRATION CHECK**  
**EPA 8082A**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Instrument ID: ECD7

Calibration: GB00069

Lab File ID: 02282332ECD7.D

Calibration Date: 02/24/2023

Sequence: SLC0014

Injection Date: 03/01/23

Lab Sample ID: SLC0014-CCV4

Injection Time: 02:47

Sequence Name: AR1660CCV4

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR (RRF)			% DRIFT/DIFF	
		STD	CCV	ICAL	CCV	MIN	CCV	LIMIT
Aroclor 1016	A	250.00	265	0.0493662	0.0524970		5.9	+/-20
Aroclor-1016 (1)	A	250.00	258	0.0303852	0.0314097		3.2	
Aroclor-1016 (2)	A	250.00	267	0.0926308	0.0990067		6.8	
Aroclor-1016 (3)	A	250.00	266	0.0452180	0.0482150		6.4	
Aroclor-1016 (4)	A	250.00	268	0.0292307	0.0313565		7.2	
Aroclor 1016 [2C]	A	250.00	254	0.0545857	0.0560912		1.5	+/-20
Aroclor-1016 (1) [2C]	A	250.00	247	0.0468313	0.0462735		-1.2	
Aroclor-1016 (2) [2C]	A	250.00	267	0.0949676	0.1015811		6.8	
Aroclor-1016 (3) [2C]	A	250.00	245	0.0428922	0.0420645		-2.0	
Aroclor-1016 (4) [2C]	A	250.00	256	0.0336515	0.0344458		2.4	
Aroclor 1260	A	250.00	383	0.0392091	0.0596066		53.2	+/-20 *
Aroclor-1260 (1)	A	250.00	366	0.0287785	0.0421980		46.4	
Aroclor-1260 (2)	A	250.00	379	0.0300690	0.0455431		51.6	
Aroclor-1260 (3)	A	250.00	369	0.0797517	0.1177419		47.6	
Aroclor-1260 (4)	A	250.00	406	0.0401599	0.0652154		62.4	
Aroclor-1260 (5)	A	250.00	395	0.0172866	0.0273346		58.0	
Aroclor 1260 [2C]	A	250.00	297	0.0699688	0.0837623		18.6	+/-20
Aroclor-1260 (1) [2C]	A	250.00	288	0.0470406	0.0541160		15.2	
Aroclor-1260 (2) [2C]	A	250.00	303	0.1200523	0.1454699		21.2	
Aroclor-1260 (3) [2C]	A	250.00	291	0.0318590	0.0371272		16.4	
Aroclor-1260 (4) [2C]	A	250.00	304	0.0809231	0.0983362		21.6	
Decachlorobiphenyl	A	40.000	45.3	0.7878687	0.8917978		13.3	+/-20
Tetrachlorometaxylene	A	40.000	39.7	1.1944880	1.1844690		-0.8	+/-20
Decachlorobiphenyl [2C]	A	40.000	44.0	1.2182710	1.3404950		10.0	+/-20
Tetrachlorometaxylene [2C]	A	40.000	41.8	1.1737210	1.2272110		4.5	+/-20

\* Values outside of QC limits

\* Values outside of QC limits



Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282332ECD7.D  
Data file 2: /230228.b/230228.b/02282332ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1660CCV4  
Client ID:  
Injection Date: 01-MAR-2023 02:47  
Report Date: 03/01/2023 12:20  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

ZB5 Col		ZB35 Col			ZB5	ZB35	RPD	Compound/Flag	
RT	Shift	Response	RT	Shift	Response	on col			on col
5.807	-0.001	466455	5.686	-0.001	195329	39.7	41.8	5.3	Tetrachloro-m-xylene
13.892	-0.001	461529	14.118	-0.002	245956	45.3	44.0	2.8	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Column 1			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	787619	16.9
Hexabromobiphenyl	1429847	1035053	-27.6
Column 2			
Standard Cpnd	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	318330	1.0
Hexabromobiphenyl	513946	366963	-28.6

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col						
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1016	1	7.269	-0.001	77309	258.4	1	7.253	-0.002	46032	247.0	
Aroclor-1016	2	7.654	-0.001	243686	267.2	2	7.857	-0.003	101051	267.4	
Aroclor-1016	3	7.791	-0.001	118672	266.6	3	8.056	-0.002	41845	245.2	
Aroclor-1016	4	8.405	-0.001	77178	268.2	4	8.306	-0.002	34266	255.9	
Total CollAve (4 peaks):				265.1		Total Col2Ave (4 peaks):				253.9	RPD = 4
Corrected Ave (3 peaks):				264.1		Corrected Ave (3 peaks):				249.4	RPD = 6

CalAmt %D: 6.0

CalAmt %D: 1.6

Aroclor-1260	1	11.044	-0.000	136491	366.6	1	11.653	-0.000	62058	287.6	
Aroclor-1260	2	11.361	0.000	147311	378.7	2	11.917	-0.001	166819	302.9	
Aroclor-1260	3	11.735	-0.000	380841	369.1	3	12.435	0.000	42576	291.3	
Aroclor-1260	4	12.139	-0.001	210942	406.0	4	12.501	-0.001	112768	303.8	
Aroclor-1260	5	12.244	0.001	88415	395.3	NS	---			----	
Total CollAve (5 peaks):				383.1		Total Col2Ave (4 peaks):				296.4	RPD = 26
Corrected Ave (4 peaks):				377.4		Corrected Ave (3 peaks):				294.0	RPD = 25

CalAmt %D: 53.2

CalAmt %D: 18.6

Total PCB Area Coll (5.907 - 13.793) = 4504656 Coll Total PCB = 0.5 ppm\*

Total PCB Area Col2 (5.787 - 14.020) = 1698889 Col2 Total PCB = 0.4 ppm\*

\* Quantitated against AR1660 0.25ppm in Ical


PCB-Form 10 Mod.




Manual Peak Adjustment, ZB-35

Datafile: ecd7.i/230228.b/230228.b/02282332ECD7.D Injection Date: 01-MAR-2023

Manual Integration (After)



Processed Integration (Before)





**CONTINUING CALIBRATION CHECK**  
**EPA 8082A**

Laboratory:	<u>Analytical Resources, LLC</u>	SDG:	<u>23B0228</u>
Client:	<u>Anchor QEA, LLC</u>	Project:	<u>AOC5 MR Phase 1</u>
Instrument ID:	<u>ECD7</u>	Calibration:	<u>GB00069</u>
Lab File ID:	<u>02282348ECD7.D</u>	Calibration Date:	<u>02/24/2023</u>
Sequence:	<u>SLC0014</u>	Injection Date:	<u>03/01/23</u>
Lab Sample ID:	<u>SLC0014-CCV5</u>	Injection Time:	<u>08:24</u>
Sequence Name:	<u>AR1254CCV5</u>		

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR (RRF)			% DRIFT/DIFF	
		STD	CCV	ICAL	CCV	MIN	CCV	LIMIT
Aroclor 1254	A	250.00	245	0.0662949	0.0646060		-2.2	+/-20
Aroclor-1254 (1)	A	250.00	247		0.0793086			
Aroclor-1254 (2)	A	250.00	250		0.0361281			
Aroclor-1254 (3)	A	250.00	245		0.0506642			
Aroclor-1254 (4)	A	250.00	238		0.0956681			
Aroclor-1254 (5)	A	250.00	243		0.0612610			
Aroclor 1254 [2C]	A	250.00	243	0.0763106	0.0741904		-2.8	+/-20
Aroclor-1254 (1) [2C]	A	250.00	248		0.0603001			
Aroclor-1254 (2) [2C]	A	250.00	246		0.0482340			
Aroclor-1254 (3) [2C]	A	250.00	245		0.1038543			
Aroclor-1254 (4) [2C]	A	250.00	240		0.0991212			
Aroclor-1254 (5) [2C]	A	250.00	236		0.0594422			
Decachlorobiphenyl	A	40.000	38.4	0.7878687	0.7555757		-4.0	+/-20
Tetrachlorometaxylene	A	40.000	37.5	1.1944880	1.1209210		-6.3	+/-20
Decachlorobiphenyl [2C]	A	40.000	42.7	1.2182710	1.3011340		6.8	+/-20
Tetrachlorometaxylene [2C]	A	40.000	39.5	1.1737210	1.1593320		-1.3	+/-20

\* Values outside of QC limits

\* Values outside of QC limits

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282348ECD7.D  
Data file 2: /230228.b/230228.b/02282348ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: AR1254.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1254CCV5  
Client ID:  
Injection Date: 01-MAR-2023 08:24  
Report Date: 03/01/2023 09:00  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.808	-0.000	436115	5.687	0.000	183801	37.5	39.5	5.1	Tetrachloro-m-xylene
13.893	-0.001	396137	14.119	0.000	249929	38.4	42.7	10.8	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	778137	15.5
Hexabromobiphenyl	1429847	1048570	-26.7

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	317081	0.6
Hexabromobiphenyl	513946	384171	-25.3

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col						
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount	
Aroclor-1254	1	9.299	-0.001	192853	246.8	1	9.451	0.000	59750	247.9	
Aroclor-1254	2	9.378	0.000	87852	250.0	2	9.972	0.000	47794	246.5	
Aroclor-1254	3	9.669	-0.001	123199	245.2	3	10.125	0.000	102907	245.3	
Aroclor-1254	4	9.808	-0.001	232634	238.2	4	10.374	0.000	98217	240.2	
Aroclor-1254	5	10.178	-0.001	148967	243.3	5	10.570	0.000	58900	236.6	
Total CollAve (5 peaks):				244.7		Total Col2Ave (5 peaks):				243.3	RPD = 1
Corrected Ave (4 peaks):				243.4		Corrected Ave (4 peaks):				242.1	RPD = 1
CalAmt %D:				-2.1		CalAmt %D:				-2.7	

Total PCB Area Col1 (5.908 - 13.794) = 2555616 Col1 Total PCB = 0.3 ppm\*

Total PCB Area Col2 (5.787 - 14.019) = 986433 Col2 Total PCB = 0.3 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR1254CCV5

01-MAR-2023 08:24, 2ul







**CONTINUING CALIBRATION CHECK**  
**EPA 8082A**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Instrument ID: ECD7

Calibration: GB00069

Lab File ID: 02282349ECD7.D

Calibration Date: 02/24/2023

Sequence: SLC0014

Injection Date: 03/01/23

Lab Sample ID: SLC0014-CCV6

Injection Time: 08:45

Sequence Name: AR1660CCV6

COMPOUND	TYPE	CONC. (ug/L)		RESPONSE FACTOR (RRF)			% DRIFT/DIFF	
		STD	CCV	ICAL	CCV	MIN	CCV	LIMIT
Aroclor 1016	A	250.00	267	0.0493662	0.0529550		6.6	+/-20
Aroclor-1016 (1)	A	250.00	260	0.0303852	0.0316067		4.0	
Aroclor-1016 (2)	A	250.00	270	0.0926308	0.1001731		8.0	
Aroclor-1016 (3)	A	250.00	270	0.0452180	0.0488706		8.0	
Aroclor-1016 (4)	A	250.00	266	0.0292307	0.0311698		6.4	
Aroclor 1016 [2C]	A	250.00	256	0.0545857	0.0566479		2.4	+/-20
Aroclor-1016 (1) [2C]	A	250.00	249	0.0468313	0.0466155		-0.4	
Aroclor-1016 (2) [2C]	A	250.00	272	0.0949676	0.1032238		8.8	
Aroclor-1016 (3) [2C]	A	250.00	244	0.0428922	0.0419403		-2.4	
Aroclor-1016 (4) [2C]	A	250.00	259	0.0336515	0.0348121		3.6	
Aroclor 1260	A	250.00	335	0.0392091	0.0526811		34.1	+/-20 *
Aroclor-1260 (1)	A	250.00	330	0.0287785	0.0379438		32.0	
Aroclor-1260 (2)	A	250.00	352	0.0300690	0.0422977		40.8	
Aroclor-1260 (3)	A	250.00	336	0.0797517	0.1070851		34.4	
Aroclor-1260 (4)	A	250.00	335	0.0401599	0.0537628		34.0	
Aroclor-1260 (5)	A	250.00	323	0.0172866	0.0223159		29.2	
Aroclor 1260 [2C]	A	250.00	272	0.0699688	0.0769061		8.6	+/-20
Aroclor-1260 (1) [2C]	A	250.00	263	0.0470406	0.0494367		5.2	
Aroclor-1260 (2) [2C]	A	250.00	278	0.1200523	0.1337765		11.2	
Aroclor-1260 (3) [2C]	A	250.00	266	0.0318590	0.0339438		6.4	
Aroclor-1260 (4) [2C]	A	250.00	279	0.0809231	0.0904670		11.6	
Decachlorobiphenyl	A	40.000	38.7	0.7878687	0.7622857		-3.3	+/-20
Tetrachlorometaxylene	A	40.000	40.1	1.1944880	1.1963830		0.3	+/-20
Decachlorobiphenyl [2C]	A	40.000	42.8	1.2182710	1.3020620		7.0	+/-20
Tetrachlorometaxylene [2C]	A	40.000	42.1	1.1737210	1.2367140		5.3	+/-20

\* Values outside of QC limits

Analytical Resources Inc.  
Dual Column 608/8082 PCB Quantitation Report

Data file 1: /230228.b/02282349ECD7.D  
Data file 2: /230228.b/230228.b/02282349ECD7.D  
Method: \\target\share\chem4\ecd7.i\230228.b\PCB.m  
Compound Sublist: AR1660.sub  
Instrument, Inj. Vol.: ecd7.i, 2ul  
Quant Method: Internal Std

ARI ID: AR1660CCV6  
Client ID:  
Injection Date: 01-MAR-2023 08:45  
Report Date: 03/01/2023 12:50  
Matrix: NONE  
Dilution Factor: 1.0

SURROGATES

RT	ZB5 Col Shift	Response	RT	ZB35 Col Shift	Response	ZB5 on col	ZB35 on col	RPD	Compound/Flag
5.807	0.000	481038	5.687	0.000	200869	40.1	42.1	5.1	Tetrachloro-m-xylene
13.893	0.000	498212	14.120	0.000	275782	38.7	42.8	9.9	Decachlorobiphenyl

\* Indicates RPD > 40%

INTERNAL STANDARD SUMMARY

Standard Cpnd	Column 1		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	673778	804154	19.3
Hexabromobiphenyl	1429847	1307153	-8.6

Standard Cpnd	Column 2		
	Standard Area*	Sample Area	%D
Bromo-Nitrobenzene	315256	324843	3.0
Hexabromobiphenyl	513946	423608	-17.6

\* Standard Areas taken from Initial Cal Level 3  
Initial Calibration Date: 24-FEB-2023  
<- Indicates standard response outside Limits (-50 to +100%)

ZB5 Col					ZB35 Col					
Aroclor	Peak#	RT	Shift	Area	Amount	Peak#	RT	Shift	Area	Amount
Aroclor-1016	1	7.269	0.000	79427	260.1	1	7.255	0.000	47321	248.8
Aroclor-1016	2	7.655	0.000	251733	270.4	2	7.859	0.000	104786	271.7
Aroclor-1016	3	7.792	0.000	122811	270.2	3	8.059	0.000	42575	244.5
Aroclor-1016	4	8.406	0.000	78329	266.6	4	8.308	0.000	35339	258.6
Total CollAve (4 peaks):				266.8		Total Col2Ave (4 peaks):				255.9 RPD = 4
Corrected Ave (3 peaks):				265.6		Corrected Ave (3 peaks):				250.6 RPD = 6

CalAmt %D: 6.7

CalAmt %D: 2.4

Aroclor-1260	1	11.044	0.000	154995	329.6	1	11.653	0.000	65443	262.7
Aroclor-1260	2	11.360	0.000	172780	351.7	2	11.918	0.000	177090	278.6
Aroclor-1260	3	11.736	0.000	437427	335.7	3	12.435	0.000	44934	266.4
Aroclor-1260	4	12.140	0.000	219613	334.7	4	12.502	0.000	119758	279.5
Aroclor-1260	5	12.243	0.000	91157	322.7	NS	---			----
Total CollAve (5 peaks):				334.9		Total Col2Ave (4 peaks):				271.8 RPD = 21
Corrected Ave (4 peaks):				330.7		Corrected Ave (3 peaks):				269.2 RPD = 20

CalAmt %D: 34.0

CalAmt %D: 8.7

Total PCB Area Coll (5.907 - 13.793) = 4749889 Coll Total PCB = 0.5 ppm\*

Total PCB Area Col2 (5.787 - 14.020) = 1777486 Col2 Total PCB = 0.5 ppm\*


\* Quantitated against AR1660 0.25ppm in Ical

PCB-Form 10 Mod.

# PCB Dual Column Chromatograms

ECD7-ZB5 AR1660CCV6


01-MAR-2023 08:45, 2ul




Manual Peak Adjustment, ZB-35

Datafile: ecd7.i/230228.b/230228.b/02282349ECD7.D Injection Date: 01-MAR-2023

Manual Integration (After)



Processed Integration (Before)





Dual Column  
ANALYSIS BATCH (SEQUENCE) SUMMARY  
EPA 8082A

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor OEA, LLC

Project: AOC5 MR Phase 1

Sequence: SLB0342

Instrument: ECD7

Calibration: GB00069

Sample Name	Lab Sample ID	Column 1 File ID	Column 2 File ID	Matrix	Analysis Date/Time
Cal Standard	SLB0342-CAL1	02242302ECD7.D	02242302ECD7.D	NA	02/24/23 11:12
Cal Standard	SLB0342-CAL2	02242303ECD7.D	02242303ECD7.D	NA	02/24/23 11:33
Cal Standard	SLB0342-CAL3	02242304ECD7.D	02242304ECD7.D	NA	02/24/23 11:54
Cal Standard	SLB0342-CAL4	02242305ECD7.D	02242305ECD7.D	NA	02/24/23 12:15
Cal Standard	SLB0342-CAL5	02242306ECD7.D	02242306ECD7.D	NA	02/24/23 12:36
Cal Standard	SLB0342-CAL6	02242307ECD7.D	02242307ECD7.D	NA	02/24/23 12:57
Cal Standard	SLB0342-CAL7	02242308ECD7.D	02242308ECD7.D	NA	02/24/23 13:18
Cal Standard	SLB0342-CAL8	02242309ECD7.D	02242309ECD7.D	NA	02/24/23 13:39
Cal Standard	SLB0342-CAL9	02242310ECD7.D	02242310ECD7.D	NA	02/24/23 14:00
Cal Standard	SLB0342-CALA	02242311ECD7.D	02242311ECD7.D	NA	02/24/23 14:21
Cal Standard	SLB0342-CALB	02242312ECD7.D	02242312ECD7.D	NA	02/24/23 14:42
Secondary Cal Check	SLB0342-SCV1	02242313ECD7.D	02242313ECD7.D	NA	02/24/23 15:03
Secondary Cal Check	SLB0342-SCV2	02242314ECD7.D	02242314ECD7.D	NA	02/24/23 15:24
Secondary Cal Check	SLB0342-SCV3	02242315ECD7.D	02242315ECD7.D	NA	02/24/23 15:45
Secondary Cal Check	SLB0342-SCV4	02242316ECD7.D	02242316ECD7.D	NA	02/24/23 16:06
Secondary Cal Check	SLB0342-SCV5	02242317ECD7.D	02242317ECD7.D	NA	02/24/23 16:27
Secondary Cal Check	SLB0342-SCV6	02242318ECD7.D	02242318ECD7.D	NA	02/24/23 16:48



Dual Column  
ANALYSIS BATCH (SEQUENCE) SUMMARY  
EPA 8082A

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor OEA, LLC

Project: AOC5 MR Phase 1

Sequence: SLC0014

Instrument: ECD7

Calibration: GB00069

Sample Name	Lab Sample ID	Column 1 File ID	Column 2 File ID	Matrix	Analysis Date/Time
Initial Cal Check	SLC0014-ICV1	02282310ECD7.D	02282310ECD7.D	NA	02/28/23 19:04
Initial Cal Check	SLC0014-ICV2	02282311ECD7.D	02282311ECD7.D	NA	02/28/23 19:25
Calibration Check	SLC0014-CCV1	02282313ECD7.D	02282313ECD7.D	NA	02/28/23 20:07
Calibration Check	SLC0014-CCV2	02282314ECD7.D	02282314ECD7.D	NA	02/28/23 20:28
Calibration Check	SLC0014-CCV3	02282331ECD7.D	02282331ECD7.D	NA	03/01/23 02:26
Calibration Check	SLC0014-CCV4	02282332ECD7.D	02282332ECD7.D	NA	03/01/23 02:47
Blank	BLB0427-BLK1	02282333ECD7.D	02282333ECD7.D	Solid	03/01/23 03:08
LCS	BLB0427-BS1	02282334ECD7.D	02282334ECD7.D	Solid	03/01/23 03:29
LCS Dup	BLB0427-BSD1	02282335ECD7.D	02282335ECD7.D	Solid	03/01/23 03:50
Reference	BLB0427-SRM1	02282336ECD7.D	02282336ECD7.D	Solid	03/01/23 04:11
LDW23-SC1009	23B0228-01	02282337ECD7.D	02282337ECD7.D	Solid	03/01/23 04:32
Calibration Check	SLC0014-CCV5	02282348ECD7.D	02282348ECD7.D	NA	03/01/23 08:24
Calibration Check	SLC0014-CCV6	02282349ECD7.D	02282349ECD7.D	NA	03/01/23 08:45







ANALYSIS SEQUENCE

SLC0014

Instrument: ECD7  
Calibration ID: GB00069

Printed: 3/1/2023 2:29:55PM

Lab Number	Analysis	Container	Order	Position	STD ID	ISTD ID	Client	Comments
SLC0014-CCV3	QC		22		L000860	L000844		
SLC0014-CCV4	QC		23		L000856	L000844		
BLB0427-BLK1	QC		24			L000844		
BLB0427-BS1	QC		25			L000844		
BLB0427-BSD1	QC		26			L000844		
BLB0427-SRM1	QC		27			L000844		
23B0228-01	8082A PCB Solid 4	A 01	28			L000844	Anchor QEA, LLC	
23B0229-01	8082A PCB Solid 4	A 01	29			L000844	Anchor QEA, LLC	
23B0229-02	8082A PCB Solid 4	A 01	30			L000844	Anchor QEA, LLC	
23B0229-03	8082A PCB Solid 4	A 01	31			L000844	Anchor QEA, LLC	
23B0229-04	8082A PCB Solid 4	A 01	32			L000844	Anchor QEA, LLC	
23B0229-05	8082A PCB Solid 4	A 01	33			L000844	Anchor QEA, LLC	
23B0229-06	8082A PCB Solid 4	A 01	34			L000844	Anchor QEA, LLC	
23B0229-07	8082A PCB Solid 4	A 01	35			L000844	Anchor QEA, LLC	
BLB0427-MS1	QC		36			L000844		
BLB0427-MSD1	QC		37			L000844		
23B0229-08	8082A PCB Solid 4	A 01	38			L000844	Anchor QEA, LLC	
SLC0014-CCV5	QC		39		L000862	L000844		
SLC0014-CCV6	QC		40		L000856	L000844		

Samples Loaded By \_\_\_\_\_ Date \_\_\_\_\_

Data Processed By \_\_\_\_\_ Date \_\_\_\_\_

## GC LOG SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230228.b

	Inject	Date/Time	Filename	DF	LabID	ClientID
1	28-FEB-2023	19:04	02282310ECD7.D	1	AR1254ICV1	
2	28-FEB-2023	19:25	02282311ECD7.D	1	AR1660ICV2	
3	28-FEB-2023	19:46	02282312ECD7.D	10	23A0556-01RE1	
4	28-FEB-2023	20:07	02282313ECD7.D	1	AR1248CCV1	
5	28-FEB-2023	20:28	02282314ECD7.D	1	AR1660CCV2	
6	28-FEB-2023	20:49	02282315ECD7.D	1	BLB0391-BLK1	
7	28-FEB-2023	21:10	02282316ECD7.D	1	BLB0391-BS1	
8	28-FEB-2023	21:31	02282317ECD7.D	1	BLB0391-BSD1	
9	28-FEB-2023	21:53	02282318ECD7.D	1	BLB0391-SRM1	
10	28-FEB-2023	22:14	02282319ECD7.D	1	23A0420-01	
11	28-FEB-2023	22:35	02282320ECD7.D	3	23A0420-02RE1	
12	28-FEB-2023	22:56	02282321ECD7.D	3	23A0420-03RE1	
13	28-FEB-2023	23:17	02282322ECD7.D	1	23A0420-04	
14	28-FEB-2023	23:38	02282323ECD7.D	1	BLB0391-MS1	
15	28-FEB-2023	23:59	02282324ECD7.D	1	BLB0391-MSD1	
16	01-MAR-2023	00:20	02282325ECD7.D	15	23A0420-05RE2	
17	01-MAR-2023	00:41	02282326ECD7.D	25	23A0420-06RE2	
18	01-MAR-2023	01:02	02282327ECD7.D	3	23A0420-07RE1	
19	01-MAR-2023	01:23	02282328ECD7.D	1	23A0420-08	
20	01-MAR-2023	01:44	02282329ECD7.D	5	23A0420-09RE1	
21	01-MAR-2023	02:05	02282330ECD7.D	3	23A0276-01RE1	
22	01-MAR-2023	02:26	02282331ECD7.D	1	AR1242CCV3	
23	01-MAR-2023	02:47	02282332ECD7.D	1	AR1660CCV4	
24	01-MAR-2023	03:08	02282333ECD7.D	1	BLB0427-BLK1	
25	01-MAR-2023	03:29	02282334ECD7.D	1	BLB0427-BS1	
26	01-MAR-2023	03:50	02282335ECD7.D	1	BLB0427-BSD1	
27	01-MAR-2023	04:11	02282336ECD7.D	1	BLB0427-SRM1	
28	01-MAR-2023	04:32	02282337ECD7.D	3	23B0228-01RE1	
29	01-MAR-2023	04:53	02282338ECD7.D	1	23B0229-01	
30	01-MAR-2023	05:14	02282339ECD7.D	1	23B0229-02	
31	01-MAR-2023	05:35	02282340ECD7.D	1	23B0229-03	
32	01-MAR-2023	05:56	02282341ECD7.D	1	23B0229-04	
33	01-MAR-2023	06:17	02282342ECD7.D	4	23B0229-05RE1	
34	01-MAR-2023	06:39	02282343ECD7.D	3	23B0229-06RE1	
35	01-MAR-2023	07:00	02282344ECD7.D	1	23B0229-07	
36	01-MAR-2023	07:21	02282345ECD7.D	1	BLB0427-MS1	
37	01-MAR-2023	07:42	02282346ECD7.D	1	BLB0427-MSD1	
38	01-MAR-2023	08:03	02282347ECD7.D	3	23B0229-08RE1	
39	01-MAR-2023	08:24	02282348ECD7.D	1	AR1254CCV5	
40	01-MAR-2023	08:45	02282349ECD7.D	1	AR1660CCV6	

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230228.b

ARI Job No.: AR12 Method: PCB.m Instrument: ecd7.i Date: 28-FEB-2023

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
1904	02282310ECD7.D	AR1254ICV1		1	Aroclor-1254,
1925	02282311ECD7.D	AR1660ICV2		1	NO MANUAL INTEGRATION
1946	02282312ECD7.D	23A0556-01RE1		10	Aroclor-1262,
2007	02282313ECD7.D	AR1248CCV1		1	Aroclor-1248,
2028	02282314ECD7.D	AR1660CCV2		1	NO MANUAL INTEGRATION
2049	02282315ECD7.D	BLB0391-BLK1		1	NO MANUAL INTEGRATION
2110	02282316ECD7.D	BLB0391-BS1		1	NO MANUAL INTEGRATION
2131	02282317ECD7.D	BLB0391-BSD1		1	NO MANUAL INTEGRATION
2153	02282318ECD7.D	BLB0391-SRM1		1	NO MANUAL INTEGRATION
2214	02282319ECD7.D	23A0420-01		1	Aroclor-1254,
2235	02282320ECD7.D	23A0420-02RE1		3	NO MANUAL INTEGRATION
2256	02282321ECD7.D	23A0420-03RE1		3	NO MANUAL INTEGRATION
2317	02282322ECD7.D	23A0420-04		1	Aroclor-1254,
2338	02282323ECD7.D	BLB0391-MS1		1	NO MANUAL INTEGRATION
2359	02282324ECD7.D	BLB0391-MSD1		1	NO MANUAL INTEGRATION
0020	02282325ECD7.D	23A0420-05RE2		15	Aroclor-1254,
0041	02282326ECD7.D	23A0420-06RE2		25	NO MANUAL INTEGRATION

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230228.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
0102	02282327ECD7.D	23A0420-07RE1		3	Aroclor-1254,
0123	02282328ECD7.D	23A0420-08		1	NO MANUAL INTEGRATION
0144	02282329ECD7.D	23A0420-09RE1		5	Aroclor-1254,
0205	02282330ECD7.D	23A0276-01RE1		3	Aroclor-1254,
0226	02282331ECD7.D	AR1242CCV3		1	NO MANUAL INTEGRATION
0247	02282332ECD7.D	AR1660CCV4		1	NO MANUAL INTEGRATION
0308	02282333ECD7.D	BLB0427-BLK1		1	NO MANUAL INTEGRATION
0329	02282334ECD7.D	BLB0427-BS1		1	NO MANUAL INTEGRATION
0350	02282335ECD7.D	BLB0427-BSD1		1	NO MANUAL INTEGRATION
0411	02282336ECD7.D	BLB0427-SRML		1	NO MANUAL INTEGRATION
0432	02282337ECD7.D	23B0228-01RE1		3	NO MANUAL INTEGRATION
0453	02282338ECD7.D	23B0229-01		1	Aroclor-1254,
0514	02282339ECD7.D	23B0229-02		1	NO MANUAL INTEGRATION
0535	02282340ECD7.D	23B0229-03		1	Aroclor-1254,
0556	02282341ECD7.D	23B0229-04		1	Aroclor-1254,
0617	02282342ECD7.D	23B0229-05RE1		4	Aroclor-1254,
0639	02282343ECD7.D	23B0229-06RE1		3	Aroclor-1254,
0700	02282344ECD7.D	23B0229-07		1	Aroclor-1254,

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230228.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
0721	02282345ECD7.D	BLB0427-MS1		1	NO MANUAL INTEGRATION
0742	02282346ECD7.D	BLB0427-MSD1		1	NO MANUAL INTEGRATION
0803	02282347ECD7.D	23B0229-08RE1		3	Aroclor-1254,
0824	02282348ECD7.D	AR1254CCV5		1	NO MANUAL INTEGRATION
0845	02282349ECD7.D	AR1660CCV6		1	NO MANUAL INTEGRATION
1216	02282302ECD7.D	AR2162SCVICAL		1	NO MANUAL INTEGRATION
1319	02282303ECD7.D	AR1660		1	NO MANUAL INTEGRATION
1412	02282304ECD7.D	AR1660		1	NO MANUAL INTEGRATION
1719	02282305ECD7.D	0.25PPMAR1660		1	NO MANUAL INTEGRATION
1740	02282306ECD7.D	0.25PPMAR1242		1	NO MANUAL INTEGRATION
1801	02282307ECD7.D	0.25PPMAR1248		1	NO MANUAL INTEGRATION
1822	02282308ECD7.D	0.25PPMAR1254		1	NO MANUAL INTEGRATION
1843	02282309ECD7.D	DDTS		1	NO MANUAL INTEGRATION
1904	02282310ECD7.D	AR1254ICV1		1	NO MANUAL INTEGRATION
1925	02282311ECD7.D	AR1660ICV2		1	Aroclor-1260 [2C],
1946	02282312ECD7.D	23A0556-01RE1		10	NO MANUAL INTEGRATION
2007	02282313ECD7.D	AR1248CCV1		1	NO MANUAL INTEGRATION
2028	02282314ECD7.D	AR1660CCV2		1	Aroclor-1260 [2C],

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230228.b\230228.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
2049	02282315ECD7.D	BLB0391-BLK1		1	NO MANUAL INTEGRATION
2110	02282316ECD7.D	BLB0391-BS1		1	NO MANUAL INTEGRATION
2131	02282317ECD7.D	BLB0391-BSD1		1	NO MANUAL INTEGRATION
2153	02282318ECD7.D	BLB0391-SRM1		1	NO MANUAL INTEGRATION
2214	02282319ECD7.D	23A0420-01		1	Aroclor-1248 [2C],
2235	02282320ECD7.D	23A0420-02RE1		3	Aroclor-1248 [2C],
2256	02282321ECD7.D	23A0420-03RE1		3	Aroclor-1248 [2C],
2317	02282322ECD7.D	23A0420-04		1	Aroclor-1248 [2C],
2338	02282323ECD7.D	BLB0391-MS1		1	NO MANUAL INTEGRATION
2359	02282324ECD7.D	BLB0391-MSD1		1	NO MANUAL INTEGRATION
0020	02282325ECD7.D	23A0420-05RE2		15	Aroclor-1248 [2C],
0041	02282326ECD7.D	23A0420-06RE2		25	Aroclor-1248 [2C],
0102	02282327ECD7.D	23A0420-07RE1		3	Aroclor-1248 [2C],
0123	02282328ECD7.D	23A0420-08		1	Aroclor-1248 [2C],
0144	02282329ECD7.D	23A0420-09RE1		5	Aroclor-1248 [2C],
0205	02282330ECD7.D	23A0276-01RE1		3	Aroclor-1248 [2C],
0226	02282331ECD7.D	AR1242CCV3		1	NO MANUAL INTEGRATION
0247	02282332ECD7.D	AR1660CCV4		1	Aroclor-1260 [2C],

MANUAL INTEGRATION SUMMARY FOR DATABATCH - \\target\share\chem4\ecd7.i\230228.b\230228.b

Time	Filename	LabID	ClientId	DF	Manually Integrated Compounds
0308	02282333ECD7.D	BLB0427-BLK1		1	NO MANUAL INTEGRATION
0329	02282334ECD7.D	BLB0427-BS1		1	NO MANUAL INTEGRATION
0350	02282335ECD7.D	BLB0427-BSD1		1	NO MANUAL INTEGRATION
0411	02282336ECD7.D	BLB0427-SRM1		1	NO MANUAL INTEGRATION
0432	02282337ECD7.D	23B0228-01RE1		3	Aroclor-1248 [2C],
0453	02282338ECD7.D	23B0229-01		1	Aroclor-1248 [2C],
0514	02282339ECD7.D	23B0229-02		1	Aroclor-1248 [2C],
0535	02282340ECD7.D	23B0229-03		1	Aroclor-1248 [2C],
0556	02282341ECD7.D	23B0229-04		1	Aroclor-1248 [2C],
0617	02282342ECD7.D	23B0229-05RE1		4	Aroclor-1248 [2C],
0639	02282343ECD7.D	23B0229-06RE1		3	Aroclor-1248 [2C],
0700	02282344ECD7.D	23B0229-07		1	Aroclor-1248 [2C],
0721	02282345ECD7.D	BLB0427-MS1		1	NO MANUAL INTEGRATION
0742	02282346ECD7.D	BLB0427-MSD1		1	NO MANUAL INTEGRATION
0803	02282347ECD7.D	23B0229-08RE1		3	Aroclor-1248 [2C],
0824	02282348ECD7.D	AR1254CCV5		1	NO MANUAL INTEGRATION
0845	02282349ECD7.D	AR1660CCV6		1	Aroclor-1260 [2C],

Security Status Report

Date: 01-Mar-2023 12:52

02282302ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282303ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282304ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282305ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282306ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282307ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282308ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282309ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282310ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282311ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282312ECD7.D	Data Locked	richardl,	01-Mar-2023	12:52
02282313ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282314ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282315ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282316ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282317ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282318ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282319ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282320ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282321ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282322ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282323ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282324ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282325ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282326ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282327ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282328ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282329ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282330ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282331ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282332ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282333ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282334ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282335ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282336ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282337ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282338ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282339ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282340ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282341ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282342ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282343ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282344ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23
02282345ECD7.D	Data Locked	richardl,	01-Mar-2023	12:23



02282346ECD7.D  
02282347ECD7.D  
02282348ECD7.D  
02282349ECD7.D

Data Locked  
Data Locked  
Data Locked  
Data Locked

richardl, 01-Mar-2023 12:23  
richardl, 01-Mar-2023 12:23  
richardl, 01-Mar-2023 12:23  
richardl, 01-Mar-2023 12:23



**SURROGATE RECOVERY AND RT SUMMARY**  
**EPA 8082A**

Laboratory: Analytical Resources, LLC  
Client: Anchor OEA, LLC  
Sequence: SLB0342  
Calibration: GB00069

SDG/WO: 23B0228  
Project: AOC5 MR Phase 1  
Instrument: ECD7  
Calibration Date: 02/24/2023

Surrogate Compound	Spike Level ug/L	% Recovery	Recovery Limits	RT	Calibration Mean RT	RT Diff	RT Diff Limit	Q
<b>SLB0342-SCV1 (Water)</b>			Lab File ID: 02242313ECD7.D			Analyzed: 02/24/23 15:03		
Decachlorobiphenyl	40.000	85.8	80 - 120	13.894	13.89483	-0.0008	N/A	
Tetrachlorometaxylene	40.000	87.4	80 - 120	5.807	5.8095	-0.0025	N/A	
Decachlorobiphenyl [2C]	40.000	93.4	80 - 120	14.119	14.11917	-0.0002	N/A	
Tetrachlorometaxylene [2C]	40.000	89.4	80 - 120	5.685	5.687167	-0.0022	N/A	
<b>SLB0342-SCV2 (Water)</b>			Lab File ID: 02242314ECD7.D			Analyzed: 02/24/23 15:24		
Decachlorobiphenyl	40.000	92.5	80 - 120	13.895	13.89483	0.0002	N/A	
Tetrachlorometaxylene	40.000	84.1	80 - 120	5.808	5.8095	-0.0015	N/A	
Decachlorobiphenyl [2C]	40.000	101	80 - 120	14.12	14.11917	0.0008	N/A	
Tetrachlorometaxylene [2C]	40.000	86.3	80 - 120	5.686	5.687167	-0.0012	N/A	
<b>SLB0342-SCV3 (Water)</b>			Lab File ID: 02242315ECD7.D			Analyzed: 02/24/23 15:45		
Decachlorobiphenyl	40.000	82.8	80 - 120	13.893	13.89483	-0.0018	N/A	
Tetrachlorometaxylene	40.000	87.2	80 - 120	5.808	5.8095	-0.0015	N/A	
Decachlorobiphenyl [2C]	40.000	90.8	80 - 120	14.118	14.11917	-0.0012	N/A	
Tetrachlorometaxylene [2C]	40.000	91.0	80 - 120	5.687	5.687167	-0.0002	N/A	
<b>SLB0342-SCV4 (Water)</b>			Lab File ID: 02242316ECD7.D			Analyzed: 02/24/23 16:06		
Decachlorobiphenyl	40.000	86.6	80 - 120	13.894	13.89483	-0.0008	N/A	
Tetrachlorometaxylene	40.000	90.3	80 - 120	5.806	5.8095	-0.0035	N/A	
Decachlorobiphenyl [2C]	40.000	94.8	80 - 120	14.119	14.11917	-0.0002	N/A	
Tetrachlorometaxylene [2C]	40.000	92.7	80 - 120	5.685	5.687167	-0.0022	N/A	
<b>SLB0342-SCV5 (Water)</b>			Lab File ID: 02242317ECD7.D			Analyzed: 02/24/23 16:27		
Decachlorobiphenyl	40.000	86.1	80 - 120	13.894	13.89483	-0.0008	N/A	
Tetrachlorometaxylene	40.000	90.0	80 - 120	5.806	5.8095	-0.0035	N/A	
Decachlorobiphenyl [2C]	40.000	94.6	80 - 120	14.119	14.11917	-0.0002	N/A	
Tetrachlorometaxylene [2C]	40.000	91.6	80 - 120	5.685	5.687167	-0.0022	N/A	
<b>SLB0342-SCV6 (Water)</b>			Lab File ID: 02242318ECD7.D			Analyzed: 02/24/23 16:48		
Decachlorobiphenyl	40.000	128	80 - 120	13.893	13.89483	-0.0018	N/A	
Tetrachlorometaxylene	40.000	92.7	80 - 120	5.806	5.8095	-0.0035	N/A	
Decachlorobiphenyl [2C]	40.000	141	80 - 120	14.118	14.11917	-0.0012	N/A	
Tetrachlorometaxylene [2C]	40.000	95.4	80 - 120	5.685	5.687167	-0.0022	N/A	



**SURROGATE RECOVERY AND RT SUMMARY**  
**EPA 8082A**

Laboratory: Analytical Resources, LLC SDG/WO: 23B0228  
 Client: Anchor QEA, LLC Project: AOC5 MR Phase 1  
 Sequence: SLC0014 Instrument: ECD7  
 Calibration: GB00069 Calibration Date: 02/24/2023

Surrogate Compound	Spike Level ug/L	% Recovery	Recovery Limits	RT	Calibration Mean RT	RT Diff	RT Diff Limit	Q
<b>SLC0014-ICV1 (Solid)</b>			Lab File ID: 02282310ECD7.D			Analyzed: 02/28/23 19:04		
Decachlorobiphenyl	40.000	93.5	80 - 120	13.894	13.89483	-0.0008	N/A	
Tetrachlorometaxylene	40.000	100	80 - 120	5.809	5.8095	-0.0005	N/A	
Decachlorobiphenyl [2C]	40.000	104	80 - 120	14.119	14.11917	-0.0002	N/A	
Tetrachlorometaxylene [2C]	40.000	101	80 - 120	5.688	5.687167	0.0008	N/A	
<b>SLC0014-ICV2 (Solid)</b>			Lab File ID: 02282311ECD7.D			Analyzed: 02/28/23 19:25		
Decachlorobiphenyl	40.000	101	80 - 120	13.892	13.89483	-0.0028	N/A	
Tetrachlorometaxylene	40.000	105	80 - 120	5.807	5.8095	-0.0025	N/A	
Decachlorobiphenyl [2C]	40.000	101	80 - 120	14.118	14.11917	-0.0012	N/A	
Tetrachlorometaxylene [2C]	40.000	106	80 - 120	5.686	5.687167	-0.0012	N/A	
<b>SLC0014-CCV1 (Solid)</b>			Lab File ID: 02282313ECD7.D			Analyzed: 02/28/23 20:07		
Decachlorobiphenyl	40.000	91.5	80 - 120	13.894	13.89483	-0.0008	N/A	
Tetrachlorometaxylene	40.000	94.5	80 - 120	5.808	5.8095	-0.0015	N/A	
Decachlorobiphenyl [2C]	40.000	100	80 - 120	14.118	14.11917	-0.0012	N/A	
Tetrachlorometaxylene [2C]	40.000	96.3	80 - 120	5.686	5.687167	-0.0012	N/A	
<b>SLC0014-CCV2 (Solid)</b>			Lab File ID: 02282314ECD7.D			Analyzed: 02/28/23 20:28		
Decachlorobiphenyl	40.000	100	80 - 120	13.893	13.89483	-0.0018	N/A	
Tetrachlorometaxylene	40.000	102	80 - 120	5.808	5.8095	-0.0015	N/A	
Decachlorobiphenyl [2C]	40.000	102	80 - 120	14.118	14.11917	-0.0012	N/A	
Tetrachlorometaxylene [2C]	40.000	103	80 - 120	5.687	5.687167	-0.0002	N/A	
<b>SLC0014-CCV3 (Solid)</b>			Lab File ID: 02282331ECD7.D			Analyzed: 03/01/23 02:26		
Decachlorobiphenyl	40.000	99.5	80 - 120	13.892	13.89483	-0.0028	N/A	
Tetrachlorometaxylene	40.000	112	80 - 120	5.806	5.8095	-0.0035	N/A	
Decachlorobiphenyl [2C]	40.000	105	80 - 120	14.118	14.11917	-0.0012	N/A	
Tetrachlorometaxylene [2C]	40.000	117	80 - 120	5.686	5.687167	-0.0012	N/A	
<b>SLC0014-CCV4 (Solid)</b>			Lab File ID: 02282332ECD7.D			Analyzed: 03/01/23 02:47		
Decachlorobiphenyl	40.000	113	80 - 120	13.892	13.89483	-0.0028	N/A	
Tetrachlorometaxylene	40.000	99.3	80 - 120	5.806	5.8095	-0.0035	N/A	
Decachlorobiphenyl [2C]	40.000	110	80 - 120	14.118	14.11917	-0.0012	N/A	
Tetrachlorometaxylene [2C]	40.000	105	80 - 120	5.686	5.687167	-0.0012	N/A	



**SURROGATE RECOVERY AND RT SUMMARY**  
**EPA 8082A**

Laboratory: Analytical Resources, LLC  
Client: Anchor QEA, LLC  
Sequence: SLC0014  
Calibration: GB00069

SDG/WO: 23B0228  
Project: AOC5 MR Phase 1  
Instrument: ECD7  
Calibration Date: 02/24/2023

Surrogate Compound	Spike Level ug/kg wet	% Recovery	Recovery Limits	RT	Calibration Mean RT	RT Diff	RT Diff Limit	Q
<b>BLB0427-BLK1 (Solid)</b> Lab File ID: 02282333ECD7.D Analyzed: 03/01/23 03:08								
Decachlorobiphenyl	8.0000	93.4	40 - 126	13.891	13.89483	-0.0038	N/A	
Tetrachlorometaxylene	8.0000	74.2	44 - 120	5.806	5.8095	-0.0035	N/A	
Decachlorobiphenyl [2C]	8.0000	87.5	40 - 126	14.117	14.11917	-0.0022	N/A	
Tetrachlorometaxylene [2C]	8.0000	80.3	44 - 120	5.686	5.687167	-0.0012	N/A	
<b>BLB0427-BS1 (Solid)</b> Lab File ID: 02282334ECD7.D Analyzed: 03/01/23 03:29								
Decachlorobiphenyl	8.0000	97.2	40 - 126	13.892	13.89483	-0.0028	N/A	
Tetrachlorometaxylene	8.0000	80.7	44 - 120	5.807	5.8095	-0.0025	N/A	
Decachlorobiphenyl [2C]	8.0000	95.0	40 - 126	14.118	14.11917	-0.0012	N/A	
Tetrachlorometaxylene [2C]	8.0000	79.9	44 - 120	5.687	5.687167	-0.0002	N/A	
<b>BLB0427-BSD1 (Solid)</b> Lab File ID: 02282335ECD7.D Analyzed: 03/01/23 03:50								
Decachlorobiphenyl	8.0000	92.0	40 - 126	13.892	13.89483	-0.0028	N/A	
Tetrachlorometaxylene	8.0000	80.4	44 - 120	5.807	5.8095	-0.0025	N/A	
Decachlorobiphenyl [2C]	8.0000	90.5	40 - 126	14.117	14.11917	-0.0022	N/A	
Tetrachlorometaxylene [2C]	8.0000	78.0	44 - 120	5.686	5.687167	-0.0012	N/A	
<b>BLB0427-SRM1 (Solid)</b> Lab File ID: 02282336ECD7.D Analyzed: 03/01/23 04:11								
Decachlorobiphenyl	40.000	83.7	40 - 126	13.887	13.89483	-0.0078	N/A	
Tetrachlorometaxylene	40.000	70.1	44 - 120	5.805	5.8095	-0.0045	N/A	
Decachlorobiphenyl [2C]	40.000	82.1	40 - 126	14.114	14.11917	-0.0052	N/A	
Tetrachlorometaxylene [2C]	40.000	75.9	44 - 120	5.684	5.687167	-0.0032	N/A	
<b>23B0228-01 (Solid)</b> Lab File ID: 02282337ECD7.D Analyzed: 03/01/23 04:32								
Decachlorobiphenyl	7.9860	88.8	40 - 126	13.884	13.89483	-0.0108	N/A	
Tetrachlorometaxylene	7.9860	68.5	44 - 120	5.805	5.8095	-0.0045	N/A	
Decachlorobiphenyl [2C]	7.9860	83.9	40 - 126	14.112	14.11917	-0.0072	N/A	
Tetrachlorometaxylene [2C]	7.9860	76.3	44 - 120	5.683	5.687167	-0.0042	N/A	
<b>SLC0014-CCV5 (Solid)</b> Lab File ID: 02282348ECD7.D Analyzed: 03/01/23 08:24								
Decachlorobiphenyl	40.000	96.0	80 - 120	13.892	13.89483	-0.0028	N/A	
Tetrachlorometaxylene	40.000	93.8	80 - 120	5.807	5.8095	-0.0025	N/A	
Decachlorobiphenyl [2C]	40.000	107	80 - 120	14.118	14.11917	-0.0012	N/A	
Tetrachlorometaxylene [2C]	40.000	98.8	80 - 120	5.686	5.687167	-0.0012	N/A	





**INTERNAL STANDARD AREA AND RT SUMMARY**  
**EPA 8082A**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Sequence: SLB0342

Instrument: ECD7

Calibration: GB00069

Internal Standard	Response	RT	Reference Response	Reference RT	Area %	Area % Limits	RT Diff	RT Diff Limit	Q
<b>Secondary Cal Check (SLB0342-SCV1)</b>		(Water)	Lab File ID: 02242313ECD7.D			Analyzed: 02/24/23 15:03			
1-Bromo-2-Nitrobenzene	645975	3.489	673778	3.493	96	50 - 200	-0.004	+/-0.50	
Hexabromobiphenyl	1524245	14.268	1429847	14.268	107	50 - 200	0.000	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	316115	3.927	315256	3.928	100	50 - 200	-0.001	+/-0.50	
Hexabromobiphenyl [2C]	556950	15.007	513946	15.008	108	50 - 200	-0.001	+/-0.50	
<b>Secondary Cal Check (SLB0342-SCV2)</b>		(Water)	Lab File ID: 02242314ECD7.D			Analyzed: 02/24/23 15:24			
1-Bromo-2-Nitrobenzene	705650	3.493	673778	3.493	105	50 - 200	0.000	+/-0.50	
Hexabromobiphenyl	1555683	14.267	1429847	14.268	109	50 - 200	-0.001	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	340433	3.929	315256	3.928	108	50 - 200	0.001	+/-0.50	
Hexabromobiphenyl [2C]	565609	15.008	513946	15.008	110	50 - 200	0.000	+/-0.50	
<b>Secondary Cal Check (SLB0342-SCV3)</b>		(Water)	Lab File ID: 02242315ECD7.D			Analyzed: 02/24/23 15:45			
1-Bromo-2-Nitrobenzene	646554	3.49	673778	3.493	96	50 - 200	-0.003	+/-0.50	
Hexabromobiphenyl	1529451	14.268	1429847	14.268	107	50 - 200	0.000	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	316066	3.928	315256	3.928	100	50 - 200	0.000	+/-0.50	
Hexabromobiphenyl [2C]	557213	15.008	513946	15.008	108	50 - 200	0.000	+/-0.50	
<b>Secondary Cal Check (SLB0342-SCV4)</b>		(Water)	Lab File ID: 02242316ECD7.D			Analyzed: 02/24/23 16:06			
1-Bromo-2-Nitrobenzene	656887	3.488	673778	3.493	97	50 - 200	-0.005	+/-0.50	
Hexabromobiphenyl	1585505	14.267	1429847	14.268	111	50 - 200	-0.001	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	320936	3.925	315256	3.928	102	50 - 200	-0.003	+/-0.50	
Hexabromobiphenyl [2C]	570006	15.007	513946	15.008	111	50 - 200	-0.001	+/-0.50	
<b>Secondary Cal Check (SLB0342-SCV5)</b>		(Water)	Lab File ID: 02242317ECD7.D			Analyzed: 02/24/23 16:27			
1-Bromo-2-Nitrobenzene	661953	3.489	673778	3.493	98	50 - 200	-0.004	+/-0.50	
Hexabromobiphenyl	1574993	14.268	1429847	14.268	110	50 - 200	0.000	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	317807	3.926	315256	3.928	101	50 - 200	-0.002	+/-0.50	
Hexabromobiphenyl [2C]	565951	15.007	513946	15.008	110	50 - 200	-0.001	+/-0.50	
<b>Secondary Cal Check (SLB0342-SCV6)</b>		(Water)	Lab File ID: 02242318ECD7.D			Analyzed: 02/24/23 16:48			
1-Bromo-2-Nitrobenzene	656592	3.489	673778	3.493	97	50 - 200	-0.004	+/-0.50	
Hexabromobiphenyl	1584453	14.268	1429847	14.268	111	50 - 200	0.000	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	314741	3.926	315256	3.928	100	50 - 200	-0.002	+/-0.50	
Hexabromobiphenyl [2C]	568346	15.007	513946	15.008	111	50 - 200	-0.001	+/-0.50	



**INTERNAL STANDARD AREA AND RT SUMMARY**  
**EPA 8082A**

Laboratory: Analytical Resources, LLC  
Client: Anchor QEA, LLC  
Sequence: SLC0014

SDG: 23B0228  
Project: AOC5 MR Phase 1  
Instrument: ECD7  
Calibration: GB00069

Internal Standard	Response	RT	Reference Response	Reference RT	Area %	Area % Limits	RT Diff	RT Diff Limit	Q
<b>Initial Cal Check (SLC0014-ICV1)</b>		(Solid)	Lab File ID: 02282310ECD7.D			Analyzed: 02/28/23 19:04			
1-Bromo-2-Nitrobenzene	747385	3.492	747385	3.492	100	50 - 200	0.000	+/-0.50	
Hexabromobiphenyl	1654369	14.268	1654369	14.268	100	50 - 200	0.000	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	301252	3.928	301252	3.928	100	50 - 200	0.000	+/-0.50	
Hexabromobiphenyl [2C]	466597	15.008	466597	15.008	100	50 - 200	0.000	+/-0.50	
<b>Initial Cal Check (SLC0014-ICV2)</b>		(Solid)	Lab File ID: 02282311ECD7.D			Analyzed: 02/28/23 19:25			
1-Bromo-2-Nitrobenzene	751182	3.491	751182	3.491	100	50 - 200	0.000	+/-0.50	
Hexabromobiphenyl	1576447	14.266	1576447	14.266	100	50 - 200	0.000	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	306016	3.927	306016	3.927	100	50 - 200	0.000	+/-0.50	
Hexabromobiphenyl [2C]	465422	15.006	465422	15.006	100	50 - 200	0.000	+/-0.50	
<b>Blank (BLB0427-BLK1)</b>		(Solid)	Lab File ID: 02282333ECD7.D			Analyzed: 03/01/23 03:08			
1-Bromo-2-Nitrobenzene	787062	3.491	751182	3.491	105	50 - 200	0.000	+/-0.50	
Hexabromobiphenyl	1186214	14.265	1576447	14.266	75	50 - 200	-0.001	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	310177	3.927	306016	3.927	101	50 - 200	0.000	+/-0.50	
Hexabromobiphenyl [2C]	397133	15.005	465422	15.006	85	50 - 200	-0.001	+/-0.50	
<b>LCS (BLB0427-BS1)</b>		(Solid)	Lab File ID: 02282334ECD7.D			Analyzed: 03/01/23 03:29			
1-Bromo-2-Nitrobenzene	839905	3.491	751182	3.491	112	50 - 200	0.000	+/-0.50	
Hexabromobiphenyl	1248668	14.265	1576447	14.266	79	50 - 200	-0.001	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	342893	3.929	306016	3.927	112	50 - 200	0.002	+/-0.50	
Hexabromobiphenyl [2C]	424424	15.007	465422	15.006	91	50 - 200	0.001	+/-0.50	
<b>LCS Dup (BLB0427-BSD1)</b>		(Solid)	Lab File ID: 02282335ECD7.D			Analyzed: 03/01/23 03:50			
1-Bromo-2-Nitrobenzene	910442	3.49	751182	3.491	121	50 - 200	-0.001	+/-0.50	
Hexabromobiphenyl	1416430	14.264	1576447	14.266	90	50 - 200	-0.002	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	371249	3.927	306016	3.927	121	50 - 200	0.000	+/-0.50	
Hexabromobiphenyl [2C]	479766	15.005	465422	15.006	103	50 - 200	-0.001	+/-0.50	
<b>Reference (BLB0427-SRM1)</b>		(Solid)	Lab File ID: 02282336ECD7.D			Analyzed: 03/01/23 04:11			
1-Bromo-2-Nitrobenzene	845866	3.49	751182	3.491	113	50 - 200	-0.001	+/-0.50	
Hexabromobiphenyl	1138921	14.256	1576447	14.266	72	50 - 200	-0.010	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	341218	3.927	306016	3.927	112	50 - 200	0.000	+/-0.50	
Hexabromobiphenyl [2C]	424905	15	465422	15.006	91	50 - 200	-0.006	+/-0.50	



**INTERNAL STANDARD AREA AND RT SUMMARY**  
**EPA 8082A**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor OEA, LLC

Project: AOC5 MR Phase 1

Sequence: SLC0014

Instrument: ECD7

Calibration: GB00069

Internal Standard	Response	RT	Reference Response	Reference RT	Area %	Area % Limits	RT Diff	RT Diff Limit	Q
<b>LDW23-SC1009 (23B0228-01 )</b>		(Solid)	Lab File ID: 02282337ECD7.D			Analyzed: 03/01/23 04:32			
1-Bromo-2-Nitrobenzene	840131	3.49	751182	3.491	112	50 - 200	-0.001	+/-0.50	
Hexabromobiphenyl	857320	14.252	1576447	14.266	54	50 - 200	-0.014	+/-0.50	
1-Bromo-2-Nitrobenzene [2C]	337914	3.927	306016	3.927	110	50 - 200	0.000	+/-0.50	
Hexabromobiphenyl [2C]	373697	14.996	465422	15.006	80	50 - 200	-0.010	+/-0.50	







## HOLDING TIME SUMMARY

**Analysis: EPA 8082A**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Sample Name	Date Collected	Date Received	Date Prepared	Days to Prep	Max Days to Prep	Date Analyzed	Days to Analysis	Max Days to Analysis	Q
LDW23-SC1009 23B0228-01	02/08/23 16:30	02/09/23 10:55	02/17/23 13:53	8	14	03/01/23 04:32	12	40	

\* Indicates hold time exceedance.



## METHOD DETECTION AND REPORTING LIMITS

### EPA 8082A

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Matrix: Solid

Instrument: ECD7

Analyte	MDL	RL	Units
Aroclor 1016	1.6	4.0	ug/kg
Aroclor 1016 [2C]	1.6	4.0	ug/kg
Aroclor 1221	1.6	4.0	ug/kg
Aroclor 1221 [2C]	1.6	4.0	ug/kg
Aroclor 1232	1.6	4.0	ug/kg
Aroclor 1232 [2C]	1.6	4.0	ug/kg
Aroclor 1242	1.6	4.0	ug/kg
Aroclor 1242 [2C]	1.6	4.0	ug/kg
Aroclor 1248	1.6	4.0	ug/kg
Aroclor 1248 [2C]	1.6	4.0	ug/kg
Aroclor 1254	1.6	4.0	ug/kg
Aroclor 1254 [2C]	1.6	4.0	ug/kg
Aroclor 1260	0.6	4.0	ug/kg
Aroclor 1260 [2C]	0.6	4.0	ug/kg

# CERTIFICATE OF ANALYSIS

**Catalog No:** S-279N  
**Description:** Tetrachloro-m-xylene  
**Lot:** 0052481B-1  
**Solvent:** N/A  
**Hazards:** Refer to SDS for complete safety information

**Date Certified:** Jul 28, 2005  
**Expiration:** Jul 28, 2015  
**Sample Size:** 100 mg  
**Components:** 1  
**Storage Condition:** Ambient (>5 °C)



Signal Word: Warning


Certified Reference Material



Component	CAS #	Purity % (GC/FID)	Prepared Concentration	Certified Analyte Concentration <sup>1</sup>
Tetrachloro-meta-xylene	877-09-8	96.0	N/A	N/A

**Identification:**

Molecular formula: C<sub>8</sub>H<sub>6</sub>Cl<sub>4</sub>  
Molecular weight: 243.94



**C000147**

tetrachlorometaxylene  
Expires 1/15/2020

Prepared By Joshua Rains 1/15/2014

This Certified Reference Material was verified in accordance with ISO/IEC 17025

A product with a suffix (-1A, -2B, etc. or -01, -02, etc.) on its lot number has had its expiration date extended and is identical to the same lot number without the suffix.

<sup>1</sup> The Uncertainty calculated for this product is ±2.4%. These values are the expanded uncertainty and represent an estimated standard deviation equal to the positive square root of the total variation of the uncertainty of components. A normal distribution is assumed and a coverage factor of K=2 is chosen using approximately a 95% confidence level.

Labels and certificates follow U.S. Conventions in reporting numerical values: A comma (,) is used to separate units of one-thousand or greater. A period (.) is used as a decimal place marker.

Metrological traceability is established through in-house validated methods.

Purity, if stated, is equal to 100% minus found impurity components. Impurity components have not been identified.

The information on this certificate may not be reproduced without the express permission of the manufacturer. See reverse side for additional information

Hazard Information: Please refer to the SDS for information regarding the hazards associated with using this material.

This product was prepared according to in-house procedures and is guaranteed to be homogeneous.

Certified By:   
Larry Decker, Organic QC Manager



# AccuStandard

125 Market Street  
New Haven, CT 06513  
(203) 786-5290

## CERTIFICATE OF PRODUCT DATA

PRODUCT: C-209N

EXPIRATION: Jul 28, 2015

DESCRIPTION: 2,2',3,3',4,4',5,5',6,6'-Decachlorobiphenyl

LOT #: 990521LB-AC

SOLVENT: N/A

This product is guaranteed accurate to  $\pm 0.5\%$  of the Certified Analyte concentration through the Expiration Date on the Label.

Component	CAS #	Purity % (GC/MS)	Prepared Concentration <sup>1</sup>	Certified Analyte Concentration <sup>2</sup>
2,2',3,3',4,4',5,5',6,6'-Decachlorobiphenyl	2051-24-3	100	N/A	N/A

2;

**C000148**

decachlorobiphenyl  
Expires 1/15/2020

Prepared By Joshua Rains 1/15/2014

*\* I 1768 A*

Certified by: *R. Cooper*

Please note: AccuStandard follows the U.S. conventions in reporting numerical values, on both certificates and labels.

A comma (,) is used to separate units of one-thousand or greater.  
A period (.) is used as a decimal place marker.

1. All weights are traceable through National Institute of Standards & Technology, Test No. 822/254480  
 2. Certified Analyte Concentration = Purity x Prepared Concentration. The Uncertainty calculated for this product is  $\pm 0.5\%$  which is the Combined Uncertainty  $U_c(y)$ . It represents an estimated standard deviation equal to the positive square root of the total variance of the uncertainty of components. The Expanded Uncertainty is  $U$  which is  $U_c(y) * K$  where  $K$  is the coverage factor at the 95% confidence level ( $K=2$ ).  
 3. A product with a suffix (-1A, -2B, etc.) on its lot# has had its expiration date extended and is identical to the same lot# without the suffix.

This product was manufactured in accordance to quality system requirements of ISO 9001:2000 and ISO 17025

*\* Recertified ~ 4-6-09 (S)*



**Analytical Standard Record**  
**Standard ID: C000148**

Printed: 4/23/2015 11:54:44AM

Description:	decachlorobiphenyl	Expires:	15-Jan-2020
Standard Type:	Other	Prepared:	15-Jan-2014
Solvent:	na/a	Prepared By:	Joshua Rains
Final Volume (mls):	1	Department:	Organics
Vials:	1	Last Edit:	27-Feb-2015 13:03 by JGR
Vendor:	Accustandard	Lot #:	9905211b-ac
Vendor Catalog #:			

**Comments**

see i1768a  
SOM calibrations added 06/12/14 sdrd

Analyte	CAS Number	Concentration	Units
Decachlorobiphenyl [2C]	2051-24-3	1000000	ug/mL
Decachlorobiphenyl	2051-24-3	1000000	ug/mL
DCB 1660 [2C]	2051-24-3	1000000	ug/mL
DCB 1660	2051-24-3	1000000	ug/mL
DCB [2C]	2051-24-3	1000000	ug/mL
DCB (A) [2C]	2051-24-3	1000000	ug/mL
DCB (A)	2051-24-3	1000000	ug/mL
DCB	2051-24-3	1000000	ug/mL

Reviewed By

Date

# Certificate of Analysis



Phenova Certified Reference Materials are sold by Phenomenex.

411 Madrid Ave., Torrance, CA 90501 USA ■ Tel: 310-212-0555 ■ Fax: 310-328-7768 ■ info@phenomenex.com

Access your MSDS and digital C of A at [www.phenomenex.com/mysupport](http://www.phenomenex.com/mysupport). Re-order at [www.phenomenex.com/standards](http://www.phenomenex.com/standards)

## Certified Reference Material

This product is included in Phenova's ISO/IEC 17025 and ISO Guide 34 Scopes of Accreditation

**Catalog No.:** AL0-101461

**Lot Number:** CL13053

**Description:** Aroclor 1254

**Certification Date:** November 29, 2018

**Storage:** 4 °C

**Expiration Date:** November 30, 2026

**Provided As:** 1 mL in 2 mL Ampoule in Hexane

Andrea Gill, Certified Reference Materials Manager

Component	CAS #	Certified Value µg/mL	Expanded Uncertainty
Aroclor 1254	11097-69-1	1000	± 0.246%

I 09808  
Recd.   
02/24/20



Reference Material Producer  
Certificate No. 2427.02



Manufactured by Phenova, Inc.

Phenova's testing and calibration results are internationally recognized through the ILAC-MRA. Phenova is an accredited ISO Guide 34 Reference Material Provider and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

IL11110613\_US

# Certificate of Analysis

**Produced by Phenova**

6390 Joyce Drive STE 100, Golden, CO 80403 USA ■ Tel: 303-940-0033 ■ Fax: 303-940-0043 ■ info@phenova.com  
Access your Safety Data Sheets and digital Certificates at [www.phenova.com/documents](http://www.phenova.com/documents).

## Certified Reference Material

This product is certified in accordance with Phenova's ISO 17034 accreditation and supported by Phenova's ISO/IEC 17025 chemical testing accreditation

**Catalog No.:** AL0-101468

**Lot Number:** CL14017

**Description:** Aroclor 1221

**Certification Date:** August 20, 2019

**Storage:** 4 °C

**Expiration Date:** August 31, 2027

**Provided As:** 1 mL in 2 mL Ampoule in Isooctane

*Andrea Gill*

Andrea Gill, Certified Reference Materials Manager

Component	CAS #	Certified Value µg/mL	Expanded Uncertainty
Aroclor 1221	11104-28-2	1000	± 0.553%

J006466  
Recd of  
06/18/21



Reference Material Producer  
Certificate No. 2427.02



Phenova is an accredited ISO/IEC 17034 Reference Material  
Producer and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03



# Certificate of Analysis



Page 2 of 2

Produced by Phenova

6390 Joyce Drive STE 100, Golden, CO 80403 USA \* Tel: 303-940-0033 \* Fax: 303-940-0043 \* info@phenova.com

Access your Safety Data Sheets and digital Certificates at [www.phenova.com/documents](http://www.phenova.com/documents).

1. **Quality Document:** This Certificate of Analysis has been created in accordance with ISO Guide 31<sup>1</sup> and ISO Guide 35.<sup>2</sup>
2. **Quality Standards:** Phenova is accredited by A2LA to ISO 17034<sup>3</sup> and ISO/IEC 17025<sup>4</sup> as a producer of Certified Reference Materials and Reference Materials. This ensures that our manufacturing processes have been accredited to and meet strict international standards.
3. **Intended Use:** The product is manufactured for use in calibration, calibration verification, quantification, identification and other appropriate analytical control applications. The product is intended for routine laboratory analysis and research purposes only. Only trained personnel should handle this product.
4. **Handling and Usage Notes:** Store according to recommended conditions listed and avoid prolonged exposure to light. Visually inspect the solution inside the ampoule for any un-dissolved material. If particulate is visible, sonicate the unopened ampoule until material is fully dissolved. Dilute as required, use only class A glassware and diluents compatible with all analytes in the mixture. Considerations should be made related to repeated use of the opened product. Once opened, exposure to light, air, heat, objects, and additional transfer vessels may cause evaporation, degradation or contamination resulting in changes in concentration, uncertainty and stability duration. Store opened standards in a clean, tightly capped vessel under the recommended temperature. Appropriate controls, such as the use of additional verification standards should be used to confirm the opened product is fit for purpose under repeated use conditions.
5. **Hazardous Situation:** The product is intended for use by experienced professional personnel. A Safety Data Sheet (SDS) is available at [www.phenova.com/documents](http://www.phenova.com/documents).
6. **Level of Homogeneity:** The product has been certified to guarantee the certified values and their uncertainties at a volume of 2 µL.
7. **Certified Value:** Certified Value is based upon gravimetric and volumetric preparation using calibrated balances and Class A glassware.
8. **Raw Materials and Purity:** Phenova reference standard products are prepared from the highest quality starting materials. The purity of this material was verified using an ISO/IEC 17025 methodology.
9. **Expanded Uncertainty:** The expanded uncertainty (uCRM) as stated is determined in accordance with ISO/IEC Guide 98<sup>5</sup> and ISO Guide 35 incorporating Type A standard uncertainty at a 95% confidence level. The uncertainty contains elements of manufacturing (uM), homogeneity analysis (uH) and long-term stability testing (uLTS). The uncertainty is calculated based on the root-sum-of-squares equation times a coverage factor (k=2).

$$u_{CRM} = k \sqrt{u_M^2 + u_H^2 + u_{LTS}^2}$$

Transport conditions (short-term stability) have been tested such that there is no contribution to the uncertainty reported. The expanded uncertainty applies to the product as received.

10. **Metrological Traceability:** The property value (certified value and its uncertainty) are traceable through an unbroken chain of calibration to the SI base unit kg through a NIST traceable weight in accordance with ISO 17034. This is achieved through calibration of balances, verification of weights, use of national methodology for glassware calibration and product homogeneity and stability testing utilizing an ISO/IEC 17025 methodology.
11. **Values Obtained During Product Testing:** This product is subjected to verification, homogeneity and stability testing using an ISO/IEC 17025 chromatographic methodology. All values obtained during testing meet criteria in accordance with ISO 17034.
12. **Period of Validity:** The Certified Values, Uncertainties and Expiration Date are based on the unopened product being stored according to the recommended storage condition listed and are guaranteed until the expiration date. This product will be monitored during the period of validity and customers notified of any significant changes in stability.

## References:

- <sup>1</sup> ISO Guide 31 – Reference Materials – Contents of Certificates and Labels.
- <sup>2</sup> ISO Guide 35 – Reference Material – General and Statistical Principles for Certification.
- <sup>3</sup> ISO 17034 – General Requirements for the Competence of Reference Material Producers.
- <sup>4</sup> ISO/IEC 17025 – General Requirements for the Competence of Testing and Calibration Laboratories.
- <sup>5</sup> ISO/IEC Guide 98-3:2008(E) – Uncertainty of Measurement – Part 3: Guide to Expression of Uncertainty in Measurement (GUM: 1995)



Reference Material Producer  
Certificate No. 2427.02



phenova<sup>®</sup>  
Certified Reference Materials

A Phenomenex  
Company

Phenova is an accredited ISO/IEC 17034 Reference Material  
Producer and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

# Certificate of Analysis

**Produced by Phenova**

6390 Joyce Drive STE 100, Golden, CO 80403 USA ■ Tel: 303-940-0033 ■ Fax: 303-940-0043 ■ info@phenova.com  
Access your Safety Data Sheets and digital Certificates at [www.phenova.com/documents](http://www.phenova.com/documents).

## Certified Reference Material

This product is certified in accordance with Phenova's ISO 17034 accreditation and supported by Phenova's ISO/IEC 17025 chemical testing accreditation

**Catalog No.:** AL0-101469

**Lot Number:** CL14914

**Description:** Aroclor 1232

**Certification Date:** January 31, 2020

**Storage:** 4 °C

**Expiration Date:** January 31, 2028

**Provided As:** 1 mL in 2 mL Ampoule in Isooctane



Andrea Gill, Certified Reference Materials Manager

Component	CAS #	Certified Value µg/mL	Expanded Uncertainty
Aroclor 1232	11141-16-5	1000	± 0.738%

J 006467  
reed  
06/18/21



Reference Material Producer  
Certificate No. 2427.02



Phenova is an accredited ISO/IEC 17034 Reference Material Producer and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

# Certificate of Analysis



Page 2 of 2

Produced by Phenova

6390 Joyce Drive STE 100, Golden, CO 80403 USA ■ Tel: 303-940-0033 ■ Fax: 303-940-0043 ■ info@phenova.com  
Access your Safety Data Sheets and digital Certificates at [www.phenova.com/documents](http://www.phenova.com/documents).

1. **Quality Document:** This Certificate of Analysis has been created in accordance with ISO Guide 31<sup>1</sup> and ISO Guide 35.<sup>2</sup>
2. **Quality Standards:** Phenova is accredited by A2LA to ISO 17034<sup>3</sup> and ISO/IEC 17025<sup>4</sup> as a producer of Certified Reference Materials and Reference Materials. This ensures that our manufacturing processes have been accredited to and meet strict international standards.
3. **Intended Use:** The product is manufactured for use in calibration, calibration verification, quantification, identification and other appropriate analytical control applications. The product is intended for routine laboratory analysis and research purposes only. Only trained personnel should handle this product.
4. **Handling and Usage Notes:** Store according to recommended conditions listed and avoid prolonged exposure to light. Visually inspect the solution inside the ampoule for any un-dissolved material. If particulate is visible, sonicate the unopened ampoule until material is fully dissolved. Dilute as required, use only class A glassware and diluents compatible with all analytes in the mixture. Considerations should be made related to repeated use of the opened product. Once opened, exposure to light, air, heat, objects, and additional transfer vessels may cause evaporation, degradation or contamination resulting in changes in concentration, uncertainty and stability duration. Store opened standards in a clean, tightly capped vessel under the recommended temperature. Appropriate controls, such as the use of additional verification standards should be used to confirm the opened product is fit for purpose under repeated use conditions.
5. **Hazardous Situation:** The product is intended for use by experienced professional personnel. A Safety Data Sheet (SDS) is available at [www.phenova.com/documents](http://www.phenova.com/documents).
6. **Level of Homogeneity:** The product has been certified to guarantee the certified values and their uncertainties at a volume of 2 µL.
7. **Certified Value:** Certified Value is based upon gravimetric and volumetric preparation using calibrated balances and Class A glassware.
8. **Raw Materials and Purity:** Phenova reference standard products are prepared from the highest quality starting materials. The purity of this material was verified using an ISO/IEC 17025 methodology.
9. **Expanded Uncertainty:** The expanded uncertainty (uCRM) as stated is determined in accordance with ISO/IEC Guide 98<sup>5</sup> and ISO Guide 35 incorporating Type A standard uncertainty at a 95% confidence level. The uncertainty contains elements of manufacturing (uM), homogeneity analysis (uH) and long-term stability testing (uLTS). The uncertainty is calculated based on the root-sum-of-squares equation times a coverage factor (k=2).

$$uCRM = k\sqrt{uM^2 + uH^2 + uLTS^2}$$

Transport conditions (short-term stability) have been tested such that there is no contribution to the uncertainty reported. The expanded uncertainty applies to the product as received.

10. **Metrological Traceability:** The property value (certified value and its uncertainty) are traceable through an unbroken chain of calibration to the SI base unit kg through a NIST traceable weight in accordance with ISO 17034. This is achieved through calibration of balances, verification of weights, use of national methodology for glassware calibration and product homogeneity and stability testing utilizing an ISO/IEC 17025 methodology.
11. **Values Obtained During Product Testing:** This product is subjected to verification, homogeneity and stability testing using an ISO/IEC 17025 chromatographic methodology. All values obtained during testing meet criteria in accordance with ISO 17034.
12. **Period of Validity:** The Certified Values, Uncertainties and Expiration Date are based on the unopened product being stored according to the recommended storage condition listed and are guaranteed until the expiration date. This product will be monitored during the period of validity and customers notified of any significant changes in stability.

## References:

- <sup>1</sup> ISO Guide 31 – Reference Materials – Contents of Certificates and Labels.
- <sup>2</sup> ISO Guide 35 – Reference Material – General and Statistical Principles for Certification.
- <sup>3</sup> ISO 17034 – General Requirements for the Competence of Reference Material Producers.
- <sup>4</sup> ISO/IEC 17025 – General Requirements for the Competence of Testing and Calibration Laboratories.
- <sup>5</sup> ISO/IEC Guide 98-3:2008(E) – Uncertainty of Measurement – Part 3: Guide to Expression of Uncertainty in Measurement (GUM: 1995)



Reference Material Producer  
Certificate No. 2427.02



A Phenomenix  
Company

Phenova is an accredited ISO/IEC 17034 Reference Material  
Producer and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

# Certificate of Analysis

**Produced by Phenova**

6390 Joyce Drive STE 100, Golden, CO 80403 USA ■ Tel: 303-940-0033 ■ Fax: 303-940-0043 ■ info@phenova.com  
Access your Safety Data Sheets and digital Certificates at [www.phenova.com/documents](http://www.phenova.com/documents).

## Certified Reference Material

This product is certified in accordance with Phenova's ISO 17034 accreditation and supported by Phenova's ISO/IEC 17025 chemical testing accreditation

**Catalog No.:** AL0-101470

**Lot Number:** CL14018

**Description:** Aroclor 1242

**Certification Date:** August 20, 2019

**Storage:** 4 °C

**Expiration Date:** August 31, 2027

**Provided As:** 1 mL in 2 mL Ampoule in Isooctane



Andrea Gill, Certified Reference Materials Manager

Component	CAS #	Certified Value µg/mL	Expanded Uncertainty
Aroclor 1242	53469-21-9	1000	± 0.553%

J006468  
feed JR  
06/18/21



Reference Material Producer  
Certificate No. 2427.02



Phenova is an accredited ISO/IEC 17034 Reference Material  
Producer and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

# Certificate of Analysis



Page 2 of 2

Produced by Phenova

6390 Joyce Drive STE 100, Golden, CO 80403 USA \* Tel: 303-940-0033 \* Fax: 303-940-0043 \* info@phenova.com

Access your Safety Data Sheets and digital Certificates at [www.phenova.com/documents](http://www.phenova.com/documents).

1. **Quality Document:** This Certificate of Analysis has been created in accordance with ISO Guide 31<sup>1</sup> and ISO Guide 35.<sup>2</sup>
2. **Quality Standards:** Phenova is accredited by A2LA to ISO 17034<sup>3</sup> and ISO/IEC 17025<sup>4</sup> as a producer of Certified Reference Materials and Reference Materials. This ensures that our manufacturing processes have been accredited to and meet strict international standards.
3. **Intended Use:** The product is manufactured for use in calibration, calibration verification, quantification, identification and other appropriate analytical control applications. The product is intended for routine laboratory analysis and research purposes only. Only trained personnel should handle this product.
4. **Handling and Usage Notes:** Store according to recommended conditions listed and avoid prolonged exposure to light. Visually inspect the solution inside the ampoule for any un-dissolved material. If particulate is visible, sonicate the unopened ampoule until material is fully dissolved. Dilute as required, use only class A glassware and diluents compatible with all analytes in the mixture. Considerations should be made related to repeated use of the opened product. Once opened, exposure to light, air, heat, objects, and additional transfer vessels may cause evaporation, degradation or contamination resulting in changes in concentration, uncertainty and stability duration. Store opened standards in a clean, tightly capped vessel under the recommended temperature. Appropriate controls, such as the use of additional verification standards should be used to confirm the opened product is fit for purpose under repeated use conditions.
5. **Hazardous Situation:** The product is intended for use by experienced professional personnel. A Safety Data Sheet (SDS) is available at [www.phenova.com/documents](http://www.phenova.com/documents).
6. **Level of Homogeneity:** The product has been certified to guarantee the certified values and their uncertainties at a volume of 2 µL.
7. **Certified Value:** Certified Value is based upon gravimetric and volumetric preparation using calibrated balances and Class A glassware.
8. **Raw Materials and Purity:** Phenova reference standard products are prepared from the highest quality starting materials. The purity of this material was verified using an ISO/IEC 17025 methodology.
9. **Expanded Uncertainty:** The expanded uncertainty (uCRM) as stated is determined in accordance with ISO/IEC Guide 98<sup>5</sup> and ISO Guide 35 incorporating Type A standard uncertainty at a 95% confidence level. The uncertainty contains elements of manufacturing (uM), homogeneity analysis (uH) and long-term stability testing (uLTS). The uncertainty is calculated based on the root-sum-of-squares equation times a coverage factor (k=2).

$$u_{CRM} = k\sqrt{u_M^2 + u_H^2 + u_{LTS}^2}$$

Transport conditions (short-term stability) have been tested such that there is no contribution to the uncertainty reported. The expanded uncertainty applies to the product as received.

10. **Metrological Traceability:** The property value (certified value and its uncertainty) are traceable through an unbroken chain of calibration to the SI base unit kg through a NIST traceable weight in accordance with ISO 17034. This is achieved through calibration of balances, verification of weights, use of national methodology for glassware calibration and product homogeneity and stability testing utilizing an ISO/IEC 17025 methodology.
11. **Values Obtained During Product Testing:** This product is subjected to verification, homogeneity and stability testing using an ISO/IEC 17025 chromatographic methodology. All values obtained during testing meet criteria in accordance with ISO 17034.
12. **Period of Validity:** The Certified Values, Uncertainties and Expiration Date are based on the unopened product being stored according to the recommended storage condition listed and are guaranteed until the expiration date. This product will be monitored during the period of validity and customers notified of any significant changes in stability.

## References:

<sup>1</sup> ISO Guide 31 – Reference Materials – Contents of Certificates and Labels.

<sup>2</sup> ISO Guide 35 – Reference Material – General and Statistical Principles for Certification.

<sup>3</sup> ISO 17034 – General Requirements for the Competence of Reference Material Producers.

<sup>4</sup> ISO/IEC 17025 – General Requirements for the Competence of Testing and Calibration Laboratories.

<sup>5</sup> ISO/IEC Guide 98-3:2008(E) – Uncertainty of Measurement – Part 3: Guide to Expression of Uncertainty in Measurement (GUM: 1995)



Reference Material Producer  
Certificate No. 2427.02



phenova<sup>®</sup>  
Certified Reference Materials

A Phenomenex  
Company

Phenova is an accredited ISO/IEC 17034 Reference Material  
Producer and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

# Certificate of Analysis

**Produced by Phenova**

6390 Joyce Drive STE 100, Golden, CO 80403 USA ■ Tel: 303-940-0033 ■ Fax: 303-940-0043 ■ info@phenova.com  
Access your Safety Data Sheets and digital Certificates at [www.phenova.com/documents](http://www.phenova.com/documents).

## Certified Reference Material

This product is certified in accordance with Phenova's ISO 17034 accreditation and supported by Phenova's ISO/IEC 17025 chemical testing accreditation

**Catalog No.:** AL0-101471

**Lot Number:** CL15384

**Description:** Aroclor 1248

**Certification Date:** June 19, 2020

**Storage:** 4 °C

**Expiration Date:** June 30, 2028

**Provided As:** 1 mL in 2 mL Ampoule in Isooctane

*Andrea L Gill*

Andrea Gill, Certified Reference Materials Manager

Component	CAS #	Certified Value µg/mL	Expanded Uncertainty
Aroclor 1248	12672-29-6	1000	± 0.520%

*# J006469  
Reed, JR  
06/18/21*



Reference Material Producer  
Certificate No. 2427.02



Phenova is an accredited ISO/IEC 17034 Reference Material  
Producer and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

# Certificate of Analysis



Page 2 of 2

Produced by Phenova

6390 Joyce Drive STE 100, Golden, CO 80403 USA ■ Tel: 303-940-0033 ■ Fax: 303-940-0043 ■ info@phenova.com  
Access your Safety Data Sheets and digital Certificates at [www.phenova.com/documents](http://www.phenova.com/documents).

- 1. Quality Document:** This Certificate of Analysis has been created in accordance with ISO Guide 31<sup>1</sup> and ISO Guide 35.<sup>2</sup>
- 2. Quality Standards:** Phenova is accredited by A2LA to ISO 17034<sup>3</sup> and ISO/IEC 17025<sup>4</sup> as a producer of Certified Reference Materials and Reference Materials. This ensures that our manufacturing processes have been accredited to and meet strict international standards.
- 3. Intended Use:** The product is manufactured for use in calibration, calibration verification, quantification, identification and other appropriate analytical control applications. The product is intended for routine laboratory analysis and research purposes only. Only trained personnel should handle this product.
- 4. Handling and Usage Notes:** Store according to recommended conditions listed and avoid prolonged exposure to light. Visually inspect the solution inside the ampoule for any un-dissolved material. If particulate is visible, sonicate the unopened ampoule until material is fully dissolved. Dilute as required, use only class A glassware and diluents compatible with all analytes in the mixture. Considerations should be made related to repeated use of the opened product. Once opened, exposure to light, air, heat, objects, and additional transfer vessels may cause evaporation, degradation or contamination resulting in changes in concentration, uncertainty and stability duration. Store opened standards in a clean, tightly capped vessel under the recommended temperature. Appropriate controls, such as the use of additional verification standards should be used to confirm the opened product is fit for purpose under repeated use conditions.
- 5. Hazardous Situation:** The product is intended for use by experienced professional personnel. A Safety Data Sheet (SDS) is available at [www.phenova.com/documents](http://www.phenova.com/documents).
- 6. Level of Homogeneity:** The product has been certified to guarantee the certified values and their uncertainties at a volume of 2 µL.
- 7. Certified Value:** Certified Value is based upon gravimetric and volumetric preparation using calibrated balances and Class A glassware.
- 8. Raw Materials and Purity:** Phenova reference standard products are prepared from the highest quality starting materials. The purity of this material was verified using an ISO/IEC 17025 methodology.
- 9. Expanded Uncertainty:** The expanded uncertainty (uCRM) as stated is determined in accordance with ISO/IEC Guide 98<sup>5</sup> and ISO Guide 35 incorporating Type A standard uncertainty at a 95% confidence level. The uncertainty contains elements of manufacturing (uM), homogeneity analysis (uH) and long-term stability testing (uLTS). The uncertainty is calculated based on the root-sum-of-squares equation times a coverage factor (k=2).  
$$uCRM = k\sqrt{uM^2 + uH^2 + uLTS^2}$$

Transport conditions (short-term stability) have been tested such that there is no contribution to the uncertainty reported. The expanded uncertainty applies to the product as received.
- 10. Metrological Traceability:** The property value (certified value and its uncertainty) are traceable through an unbroken chain of calibration to the SI base unit kg through a NIST traceable weight in accordance with ISO 17034. This is achieved through calibration of balances, verification of weights, use of national methodology for glassware calibration and product homogeneity and stability testing utilizing an ISO/IEC 17025 methodology.
- 11. Values Obtained During Product Testing:** This product is subjected to verification, homogeneity and stability testing using an ISO/IEC 17025 chromatographic methodology. All values obtained during testing meet criteria in accordance with ISO 17034.
- 12. Period of Validity:** The Certified Values, Uncertainties and Expiration Date are based on the unopened product being stored according to the recommended storage condition listed and are guaranteed until the expiration date. This product will be monitored during the period of validity and customers notified of any significant changes in stability.

## References:

- <sup>1</sup> ISO Guide 31 – Reference Materials – Contents of Certificates and Labels.
- <sup>2</sup> ISO Guide 35 – Reference Material – General and Statistical Principles for Certification.
- <sup>3</sup> ISO 17034 – General Requirements for the Competence of Reference Material Producers.
- <sup>4</sup> ISO/IEC 17025 – General Requirements for the Competence of Testing and Calibration Laboratories.
- <sup>5</sup> ISO/IEC Guide 98-3:2008(E) – Uncertainty of Measurement – Part 3: Guide to Expression of Uncertainty in Measurement (GUM: 1995)



Reference Material Producer  
Certificate No. 2427.02



phenova<sup>®</sup>  
Certified Reference Materials

A Phenomenex  
Company

Phenova is an accredited ISO/IEC 17034 Reference Material  
Producer and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

# Certificate of Analysis

Phenova Certified Reference Materials are sold by Phenomenex.

411 Madrid Ave., Torrance, CA 90501 USA ■ Tel: 310-212-0555 ■ Fax: 310-328-7768 ■ info@phenomenex.com

Access your MSDS and digital C of A at [www.phenomenex.com/mysupport](http://www.phenomenex.com/mysupport). Re-order at [www.phenomenex.com/standards](http://www.phenomenex.com/standards)

## Certified Reference Material

This product is included in Phenova's ISO/IEC 17025 and ISO Guide 34 Scopes of Accreditation

**Catalog No.:** AL0-101474

**Lot Number:** CL11330

**Description:** Aroclor 1262

**Certification Date:** May 15, 2015

**Storage:** 4 °C

**Expiration Date:** April 30, 2023

**Provided As:** 1 mL in 2 mL Ampoule in Isooctane

**Revision Date:** April 2, 2018



Andrea Gill, Certified Reference Materials Manager

Component	CAS #	Certified Value µg/mL	Expanded Uncertainty
Aroclor 1262	37324-23-5	1000	± 0.516%

J 00647H  
Reed JK  
06/18/21



Reference Material Producer  
Certificate No. 2427.02



Manufactured by Phenova, Inc.

Phenova's testing and calibration results are internationally recognized through the ILAC-MRA. Phenova is an accredited ISO Guide 34 Reference Material Provider and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03



# Certificate of Analysis



Phenova Certified Reference Materials are sold by Phenomenex.

411 Madrid Ave., Torrance, CA 90501 USA ■ Tel: 310-212-0555 ■ Fax: 310-328-7768 ■ info@phenomenex.com

Access your MSDS and digital C of A at [www.phenomenex.com/mysupport](http://www.phenomenex.com/mysupport). Re-order at [www.phenomenex.com/standards](http://www.phenomenex.com/standards)

1. **Quality Document:** This Certificate of Analysis has been created in accordance with ISO Guide 31<sup>1</sup> and ISO Guide 35.<sup>2</sup>
  2. **Quality Standards:** Phenova is accredited by A2LA to ISO Guide 34<sup>3</sup> and ISO/IEC 17025<sup>4</sup> as a producer of Certified Reference Materials and Reference Materials. This ensures that our manufacturing processes have been accredited to and meet strict international standards.
  3. **Intended Use:** The product is manufactured for use in the calibration and calibration verification of chromatographic instrumentation performed in routine laboratory analysis.
  4. **Instruction:** Visually inspect the solution inside the ampoule for any un-dissolved material. If particulate is visible, sonicate the unopened ampoule until material is fully dissolved. Dilute as required, use only class A glassware and diluents compatible with all certified analytes in the mixture.
  5. **Hazardous Situation:** The product is intended for use by experienced professional personnel. A Material Safety Data Sheet (MSDS) is available at [www.phenomenex.com/mysupport](http://www.phenomenex.com/mysupport).
  6. **Level of Homogeneity:** The product has been certified to guarantee the certified values and their uncertainties at a volume of 2 µL.
  7. **Certified Value:** Certified Value is based upon gravimetric and volumetric preparation using calibrated balances and Class A glassware.
  8. **Raw Materials and Purity:** Phenova reference standard products are prepared from the highest quality starting materials. The purity of this material was verified using an ISO/IEC 17025 methodology.
  9. **Expanded Uncertainty:** The expanded uncertainty (uCRM) as stated is determined in accordance with ISO/IEC Guide 98<sup>5</sup> and ISO Guide 35 incorporating Type A standard uncertainty at a 95% confidence level. The uncertainty contains elements of manufacturing (uM), homogeneity analysis (uH) and long-term stability testing (uLTS). The uncertainty is calculated based on the root-sum-of-squares equation times a coverage factor (k=2).  
$$uCRM = k\sqrt{uM^2 + uH^2 + uLTS^2}$$
- Transport conditions (short-term stability) have been tested such that there is no contribution to the uncertainty reported. The expanded uncertainty applies to the product as received.
10. **Metrological Traceability:** The property value (certified value and its uncertainty) are traceable through an unbroken chain of calibration to the SI base unit kg through a NIST traceable weight in accordance with ISO Guide 34. This is achieved through calibration of balances, verification of weights, use of national methodology for glassware calibration and product homogeneity and stability testing utilizing an ISO/IEC 17025 methodology.
  11. **Values Obtained During Product Testing:** This product is subjected to verification, homogeneity and stability testing using an ISO/IEC 17025 chromatographic methodology. All values obtained during testing meet criteria in accordance with ISO Guide 34.
  12. **Period of Validity:** The Certified Values and their uncertainties are guaranteed until the expiration date. This product will be monitored during the period of validity and customers notified of any significant changes in stability.

## References:

- <sup>1</sup> ISO Guide 31:2000(E) – Reference Materials – Contents of Certificates and Labels.
- <sup>2</sup> ISO Guide 35:2006(E) – Reference Material – General and Statistical Principles for Certification.
- <sup>3</sup> ISO Guide 34:2009(E) – General Requirements for the Competence of Reference Material Producers.
- <sup>4</sup> ISO/IEC 17025:2005(E) – General Requirements for the Competence of Testing and Calibration Laboratories.
- <sup>5</sup> ISO/IEC Guide 98-3:2008(E) – Uncertainty of Measurement – Part 3: Guide to Expression of Uncertainty in Measurement (GUM: 1995)



Reference Material Producer  
Certificate No. 2427.02



Manufactured by Phenova, Inc.

Phenova's testing and calibration results are internationally recognized through the ILAC-MRA. Phenova is an accredited ISO Guide 34 Reference Material Provider and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

# Certificate of Analysis



Phenova Certified Reference Materials are sold by Phenomenex.

411 Madrid Ave., Torrance, CA 90501 USA ■ Tel: 310-212-0555 ■ Fax: 310-328-7768 ■ info@phenomenex.com  
Access your MSDS and digital C of A at [www.phenomenex.com/mysupport](http://www.phenomenex.com/mysupport). Re-order at [www.phenomenex.com/standards](http://www.phenomenex.com/standards)

## Certified Reference Material

This product is included in Phenova's ISO/IEC 17025 and ISO Guide 34 Scopes of Accreditation

**Catalog No.:** AL0-101475

**Lot Number:** CL11331

**Description:** Aroclor 1268

**Certification Date:** May 15, 2015

**Storage:** 4 °C

**Expiration Date:** April 30, 2023

**Provided As:** 1 mL in 2 mL Ampoule in Isooctane

**Revision Date:** April 2, 2018

Andrea Gill, Certified Reference Materials Manager

Component	CAS #	Certified Value µg/mL	Expanded Uncertainty
Aroclor 1268	11100-14-4	1000	± 0.516%

J006472  
Rec'd. JK  
06/18/21



Reference Material Producer  
Certificate No. 2427.02



Manufactured by Phenova, Inc.

Phenova's testing and calibration results are internationally recognized through the ILAC MRA. Phenova is an accredited ISO Guide 34 Reference Material Provider and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

# Certificate of Analysis



Phenova Certified Reference Materials are sold by Phenomenex.

411 Madrid Ave., Torrance, CA 90501 USA ■ Tel: 310-212-0555 ■ Fax: 310-328-7768 ■ info@phenomenex.com

Access your MSDS and digital C of A at [www.phenomenex.com/mysupport](http://www.phenomenex.com/mysupport). Re-order at [www.phenomenex.com/standards](http://www.phenomenex.com/standards)

1. **Quality Document:** This Certificate of Analysis has been created in accordance with ISO Guide 31<sup>1</sup> and ISO Guide 35.<sup>2</sup>
2. **Quality Standards:** Phenova is accredited by A2LA to ISO Guide 34<sup>3</sup> and ISO/IEC 17025<sup>4</sup> as a producer of Certified Reference Materials and Reference Materials. This ensures that our manufacturing processes have been accredited to and meet strict international standards.
3. **Intended Use:** The product is manufactured for use in the calibration and calibration verification of chromatographic instrumentation performed in routine laboratory analysis.
4. **Instruction:** Visually inspect the solution inside the ampoule for any un-dissolved material. If particulate is visible, sonicate the unopened ampoule until material is fully dissolved. Dilute as required, use only class A glassware and diluents compatible with all certified analytes in the mixture.
5. **Hazardous Situation:** The product is intended for use by experienced professional personnel. A Material Safety Data Sheet (MSDS) is available at [www.phenomenex.com/mysupport](http://www.phenomenex.com/mysupport).
6. **Level of Homogeneity:** The product has been certified to guarantee the certified values and their uncertainties at a volume of 2 µL.
7. **Certified Value:** Certified Value is based upon gravimetric and volumetric preparation using calibrated balances and Class A glassware.
8. **Raw Materials and Purity:** Phenova reference standard products are prepared from the highest quality starting materials. The purity of this material was verified using an ISO/IEC 17025 methodology.
9. **Expanded Uncertainty:** The expanded uncertainty (uCRM) as stated is determined in accordance with ISO/IEC Guide 98<sup>5</sup> and ISO Guide 35 incorporating Type A standard uncertainty at a 95% confidence level. The uncertainty contains elements of manufacturing (uM), homogeneity analysis (uH) and long-term stability testing (uLTS). The uncertainty is calculated based on the root-sum-of-squares equation times a coverage factor (k=2).

$$uCRM = k \cdot \sqrt{uM^2 + uH^2 + uLTS^2}$$

Transport conditions (short-term stability) have been tested such that there is no contribution to the uncertainty reported. The expanded uncertainty applies to the product as received.

10. **Metrological Traceability:** The property value (certified value and its uncertainty) are traceable through an unbroken chain of calibration to the SI base unit kg through a NIST traceable weight in accordance with ISO Guide 34. This is achieved through calibration of balances, verification of weights, use of national methodology for glassware calibration and product homogeneity and stability testing utilizing an ISO/IEC 17025 methodology.
11. **Values Obtained During Product Testing:** This product is subjected to verification, homogeneity and stability testing using an ISO/IEC 17025 chromatographic methodology. All values obtained during testing meet criteria in accordance with ISO Guide 34.
12. **Period of Validity:** The Certified Values and their uncertainties are guaranteed until the expiration date. This product will be monitored during the period of validity and customers notified of any significant changes in stability.

## References:

<sup>1</sup> ISO Guide 31:2000(E) – Reference Materials – Contents of Certificates and Labels.

<sup>2</sup> ISO Guide 35:2006(E) – Reference Material – General and Statistical Principles for Certification.

<sup>3</sup> ISO Guide 34:2009(E) – General Requirements for the Competence of Reference Material Producers.

<sup>4</sup> ISO/IEC 17025:2005(E) – General Requirements for the Competence of Testing and Calibration Laboratories.

<sup>5</sup> ISO/IEC Guide 98-3:2008(E) – Uncertainty of Measurement – Part 3: Guide to Expression of Uncertainty in Measurement (GUM: 1995)



Reference Material Producer  
Certificate No. 2427.02



Manufactured by Phenova, Inc.

Phenova's testing and calibration results are internationally recognized through the ILAC MRA. Phenova is an accredited ISO Guide 34 Reference Material Provider and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03



# Certificate of Analysis

## Aroclor 1016 Solution

Product Number: PP-282

Page: 1 of 1

Lot Number: CR-0761

Lot Issue Date: 28-Feb-2017

Expiration Date: 31-Mar-2025

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
Aroclor 1016	012674-11-2	NT01016	100.2 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage: Store at Room Temperature (15° to 30°C).

*K1254  
Rec'd JP  
02/05/17*

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001  
Registered  
TUV USA, Inc.

John Russo  
President

Monica Bourgeois  
Director of QA/RA



# Certificate of Analysis

**Product Name:** Aroclor 1260 Standard

**Product Number:** PP-362-1

**Lot Issue Date:** 20-Jan-2021

**Lot Number:** 0006582048

**Expiration Date:** 28-Feb-2025

**Description:**

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1260	011096-82-5	NT01023	100.4 ± 0.5 µg/mL

**Matrix:** isooctane (2,2,4-trimethylpentane)

K 1255

**Storage Conditions:** Store at Room Temperature (15° to 30°C).

**Traceability:**

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

**Homogeneity:**

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

**Intended Use:**

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

**Instructions for Use:**

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

**Hazards:**

Refer to the Safety Data Sheet on [www.agilent.com](http://www.agilent.com) for information regarding this RM.

**Expiration of Certification:**

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

**Maintenance of Certification:**

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

**Sample lot approver:**

Monica Bourgeois  
QMS Representative



ISO 17034 Cert  
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 1 of 1

[www.agilent.com/quality/](http://www.agilent.com/quality/)  
CSD-QA-015.1



ISO 17025 Cert  
No. AT-1937



# Certificate of Analysis ISO Guide 34

## Aroclor 1242 Solution

**Product Number:** PP-312

**Page:** 1 of 1

**Lot Number:** CS-6293

**Lot Issue Date:** 04-Jan-2019

**Expiration Date:** 31-Jan-2023

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
Aroclor 1242	053469-21-9	NT01020	100.4 ± 0.5 µg/mL

**Matrix:** isooctane (2,2,4-trimethylpentane)

**Storage:** Store at Room Temperature (15° to 30°C).

K1256

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

  
Monica Bourgeois  
QMS Representative



ISO Guide 34 Cert No.  
AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026  
registered ISO 9001 Quality Management System



ISO17025 Cert No.  
AT-1937

ISO 17034



Agilent

Trusted Answers

Reference Material Certificate

**Product Name:** Aroclor 1248 Standard **Lot Number:** 0006626997  
**Product Number:** PP-342-1 **Lot Issue Date:** 17-Aug-2021  
**Storage Conditions:** Store at Room Temperature (15° to 30°C). **Expiration Date:** 30-Sep-2025

Component Name	CERTIFIED VALUES			CAS#	Analyte Lot
	Concentration	Expanded Uncertainty			
Aroclor 1248	100.3	± 0.5 µg/mL		012672-29-6	NT01582

**Matrix:** isooctane (2,2,4-trimethylpentane)

K1257

**Description:**

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

**Traceability:**

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

**Homogeneity:**

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

**Instructions for Use:**

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

**Safety:**

Refer to the Safety Data Sheet on [www.agilent.com](http://www.agilent.com) for information regarding this analytical reference material.

**Intended Use:**

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

**Expiration of Certification:**

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.



# Certificate of Analysis

## Aroclor 1254 Solution

Product Number: PP-352

Page: 1 of 1

Lot Number: CS-2321

Lot Issue Date: 04-May-2018

Expiration Date: 31-May-2026

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
Aroclor 1254	011097-69-1	RM00922	100.4 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage: Store at Room Temperature (15° to 30°C).


K-1250

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001  
Registered  
TUV USA, Inc.

  
John Russo  
President

  
Monica Bourgeois  
Director of QA/RA





# Certificate of Analysis

**Product Name:** Aroclor 1221 Standard

**Product Number:** PP-292-1

**Lot Issue Date:** 28-Apr-2020

**Lot Number:** 0006535333

**Expiration Date:** 31-May-2024

**Description:**

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system, and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1221	011104-28-2	RM04278	100.2 ± 0.5 µg/mL

**Matrix:** isooctane (2,2,4-trimethylpentane)

**Storage Conditions:** Store at Room Temperature (15° to 30°C).

K1259

**Traceability:**

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

**Homogeneity:**

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

**Intended Use:**

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

**Instructions for Use:**

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

**Hazards:**

Refer to the Safety Data Sheet on [www.agilent.com](http://www.agilent.com) for information regarding this RM.

**Expiration of Certification:**

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

**Maintenance of Certification:**

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

**Sample lot approver:**

Monica Bourgeois  
QMS Representative



ISO 17034 Cert No.  
AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 1 of 1

[www.agilent.com/quality/](http://www.agilent.com/quality/)  
CSD-QA-015.1



ISO 17025 Cert  
No. AT-1937



# Certificate of Analysis ISO 17034

## Aroclor 1262 Standard

Product Number: PP-372-1

Page: 1 of 1

Lot Number: 0006499800

Lot Issue Date: 04-Nov-2019

Expiration Date: 30-Nov-2023

This ISO 17034 Reference Material (RM) was manufactured and verified in accordance with Agilent Technologies ISO 9001 registered quality system. A review of the gravimetric preparation data by our ISO 17025 accredited laboratory serves to verify the concentration of each analyte. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
Aroclor 1262	037324-23-5	RM14263	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage: Store at Room Temperature (15° to 30°C).

K1260

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

Monica Bourgeois  
QMS Representative



ISO 17034 Cert No.  
AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026  
registered ISO 9001 Quality Management System



ISO 17025 Cert No.  
AT-1937



# Certificate of Analysis ISO 17034

## Aroclor 1232 Standard

**Product Number:** PP-302-1

**Page:** 1 of 1

**Lot Number:** CF-2197A

**Lot Issue Date:** 05-Jul-2016

**Expiration Date:** 31-Aug-2023

This ISO 17034 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
Aroclor 1232	011141-16-5	NT01717	100.4 ± 0.5 µg/mL


**Matrix:** isooctane (2,2,4-trimethylpentane)

**Storage:** Store at Room Temperature (15° to 30°C).

K1261

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

  
Monica Bourgeois  
QMS Representative



ISO 17034 Cert No.  
AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026  
registered ISO 9001 Quality Management System



ISO17025 Cert No.  
AT-1937



# Certificate of Analysis

**Product Name:** Aroclor 1268 Standard

**Product Number:** PP-382-1

**Lot Issue Date:** 09-Feb-2021

**Lot Number:** 0006587800

**Expiration Date:** 31-Mar-2029

**Description:**

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

**Matrix:** isooctane (2,2,4-trimethylpentane)

**Storage Conditions:** Store at Room Temperature (15° to 30°C).

K1262

**Traceability:**

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

**Homogeneity:**

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

**Intended Use:**

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

**Instructions for Use:**

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

**Hazards:**

Refer to the Safety Data Sheet on [www.agilent.com](http://www.agilent.com) for information regarding this RM.

**Expiration of Certification:**

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

**Maintenance of Certification:**

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

**Sample lot approver:**

  
Monica Bourgeois  
QMS Representative



ISO 17034 Cert  
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 1 of 1

[www.agilent.com/quality/](http://www.agilent.com/quality/)  
CSD-QA-015.1



ISO 17025 Cert  
No. AT-1937

# Certificate of Analysis



Phenova Certified Reference Materials are sold by Phenomenex.

411 Madrid Ave., Torrance, CA 90501 USA ■ Tel: 310-212-0555 ■ Fax: 310-328-7768 ■ info@phenomenex.com

Access your MSDS and digital C of A at [www.phenomenex.com/mysupport](http://www.phenomenex.com/mysupport). Re-order at [www.phenomenex.com/standards](http://www.phenomenex.com/standards)

## Certified Reference Material

This product is included in Phenova's ISO/IEC 17025 and ISO Guide 34 Scopes of Accreditation

**Catalog No.:** AL0-101467

**Lot Number:** CL12975

**Description:** Aroclor 1016

**Certification Date:** November 19, 2018

**Storage:** 4 °C

**Expiration Date:** October 31, 2026

**Provided As:** 1 mL in 2 mL Ampoule in Isooctane

Andrea Gill, Certified Reference Materials Manager

Component	CAS #	Certified Value µg/mL	Expanded Uncertainty
Aroclor 1016	12674-11-2	1000	± 0.553%

125829



Reference Material Producer  
Certificate No. 2427.02



Manufactured by Phenova, Inc.

Phenova's testing and calibration results are internationally recognized through the ILAC-MRA. Phenova is an accredited ISO Guide 34 Reference Material Provider and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

# Certificate of Analysis



Phenova Certified Reference Materials are sold by Phenomenex.

411 Madrid Ave., Torrance, CA 90501 USA ■ Tel: 310-212-0555 ■ Fax: 310-328-7768 ■ info@phenomenex.com

Access your MSDS and digital C of A at [www.phenomenex.com/mysupport](http://www.phenomenex.com/mysupport). Re-order at [www.phenomenex.com/standards](http://www.phenomenex.com/standards)

1. **Quality Document:** This Certificate of Analysis has been created in accordance with ISO Guide 31<sup>1</sup> and ISO Guide 35.<sup>2</sup>
2. **Quality Standards:** Phenova is accredited by A2LA to ISO Guide 34<sup>3</sup> and ISO/IEC 17025<sup>4</sup> as a producer of Certified Reference Materials and Reference Materials. This ensures that our manufacturing processes have been accredited to and meet strict international standards.
3. **Intended Use:** The product is manufactured for use in the calibration and calibration verification of chromatographic instrumentation performed in routine laboratory analysis.
4. **Instruction:** Visually inspect the solution inside the ampoule for any un-dissolved material. If particulate is visible, sonicate the unopened ampoule until material is fully dissolved. Dilute as required, use only class A glassware and diluents compatible with all certified analytes in the mixture.
5. **Hazardous Situation:** The product is intended for use by experienced professional personnel. A Material Safety Data Sheet (MSDS) is available at [www.phenomenex.com/mysupport](http://www.phenomenex.com/mysupport).
6. **Level of Homogeneity:** The product has been certified to guarantee the certified values and their uncertainties at a volume of 2 µL.
7. **Certified Value:** Certified Value is based upon gravimetric and volumetric preparation using calibrated balances and Class A glassware.
8. **Raw Materials and Purity:** Phenova reference standard products are prepared from the highest quality starting materials. The purity of this material was verified using an ISO/IEC 17025 methodology.
9. **Expanded Uncertainty:** The expanded uncertainty (uCRM) as stated is determined in accordance with ISO/IEC Guide 98<sup>5</sup> and ISO Guide 35 incorporating Type A standard uncertainty at a 95% confidence level. The uncertainty contains elements of manufacturing (uM), homogeneity analysis (uH) and long-term stability testing (uLTS). The uncertainty is calculated based on the root-sum-of-squares equation times a coverage factor (k=2).

$$u_{CRM} = k \sqrt{uM^2 + uH^2 + uLTS^2}$$

Transport conditions (short-term stability) have been tested such that there is no contribution to the uncertainty reported. The expanded uncertainty applies to the product as received.

10. **Metrological Traceability:** The property value (certified value and its uncertainty) are traceable through an unbroken chain of calibration to the SI base unit kg through a NIST traceable weight in accordance with ISO Guide 34. This is achieved through calibration of balances, verification of weights, use of national methodology for glassware calibration and product homogeneity and stability testing utilizing an ISO/IEC 17025 methodology.
11. **Values Obtained During Product Testing:** This product is subjected to verification, homogeneity and stability testing using an ISO/IEC 17025 chromatographic methodology. All values obtained during testing meet criteria in accordance with ISO Guide 34.
12. **Period of Validity:** The Certified Values and their uncertainties are guaranteed until the expiration date. This product will be monitored during the period of validity and customers notified of any significant changes in stability.

## References:

<sup>1</sup> ISO Guide 31:2000(E) – Reference Materials – Contents of Certificates and Labels.

<sup>2</sup> ISO Guide 35:2006(E) – Reference Material – General and Statistical Principles for Certification.

<sup>3</sup> ISO Guide 34:2009(E) – General Requirements for the Competence of Reference Material Producers.

<sup>4</sup> ISO/IEC 17025:2005(E) – General Requirements for the Competence of Testing and Calibration Laboratories.

<sup>5</sup> ISO/IEC Guide 98-3:2008(E) – Uncertainty of Measurement – Part 3: Guide to Expression of Uncertainty in Measurement (GUM: 1995)



Reference Material Producer  
Certificate No. 2427.02



Manufactured by Phenova, Inc.

Phenova's testing and calibration results are internationally recognized through the ILAC MRA. Phenova is an accredited ISO Guide 34 Reference Material Provider and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

IL111063\_US

# Certificate of Analysis

**Produced by Phenova**

6390 Joyce Drive STE 100, Golden, CO 80403 USA ■ Tel: 303-940-0033 ■ Fax: 303-940-0043 ■ info@phenova.com  
Access your Safety Data Sheets and digital Certificates at [www.phenova.com/documents](http://www.phenova.com/documents).

## Certified Reference Material

This product is certified in accordance with Phenova's ISO 17034 accreditation and supported by Phenova's ISO/IEC 17025 chemical testing accreditation

**Catalog No.:** AL0-101462

**Lot Number:** CL18021

**Description:** Aroclor 1260

**Certification Date:** February 14, 2022

**Storage:** 4 °C

**Expiration Date:** February 28, 2030

**Provided As:** 1 mL in 2 mL Ampoule in Hexane

*Andrea L Gill*

Andrea Gill, Certified Reference Materials Manager

<b>Component</b>	<b>CAS #</b>	<b>Certified Value µg/mL</b>	<b>Expanded Uncertainty</b>
Aroclor 1260	11096-82-5	1000	± 0.553%

K005830



Reference Material Producer  
Certificate No. 2427.02



Phenova is an accredited ISO/IEC 17034 Reference Material Producer and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03

# Certificate of Analysis



Page 2 of 2

## Produced by Phenova

6390 Joyce Drive STE 100, Golden, CO 80403 USA ■ Tel: 303-940-0033 ■ Fax: 303-940-0043 ■ info@phenova.com  
Access your Safety Data Sheets and digital Certificates at [www.phenova.com/documents](http://www.phenova.com/documents).

1. **Quality Document:** This Certificate of Analysis has been created in accordance with ISO Guide 31<sup>1</sup> and ISO Guide 35.<sup>2</sup>
2. **Quality Standards:** Phenova is accredited by A2LA to ISO 17034<sup>3</sup> and ISO/IEC 17025<sup>4</sup> as a producer of Certified Reference Materials and Reference Materials. This ensures that our manufacturing processes have been accredited to and meet strict international standards.
3. **Intended Use:** The product is manufactured for use in calibration, calibration verification, quantification, identification and other appropriate analytical control applications. The product is intended for routine laboratory analysis and research purposes only. Only trained personnel should handle this product.
4. **Handling and Usage Notes:** Store according to recommended conditions listed and avoid prolonged exposure to light. Visually inspect the solution inside the ampoule for any un-dissolved material. If particulate is visible, sonicate the unopened ampoule until material is fully dissolved. Dilute as required, use only class A glassware and diluents compatible with all analytes in the mixture. Considerations should be made related to repeated use of the opened product. Once opened, exposure to light, air, heat, objects, and additional transfer vessels may cause evaporation, degradation or contamination resulting in changes in concentration, uncertainty and stability duration. Store opened standards in a clean, tightly capped vessel under the recommended temperature. Appropriate controls, such as the use of additional verification standards should be used to confirm the opened product is fit for purpose under repeated use conditions.
5. **Hazardous Situation:** The product is intended for use by experienced professional personnel. A Safety Data Sheet (SDS) is available at [www.phenova.com/documents](http://www.phenova.com/documents).
6. **Level of Homogeneity:** The product has been certified to guarantee the certified values and their uncertainties at a volume of 2 µL.
7. **Certified Value:** Certified Value is based upon gravimetric and volumetric preparation using calibrated balances and Class A glassware.
8. **Raw Materials and Purity:** Phenova reference standard products are prepared from the highest quality starting materials. The purity of this material was verified using an ISO/IEC 17025 methodology.
9. **Expanded Uncertainty:** The expanded uncertainty (uCRM) as stated is determined in accordance with ISO/IEC Guide 98<sup>5</sup> and ISO Guide 35 incorporating Type A standard uncertainty at a 95% confidence level. The uncertainty contains elements of manufacturing (uM), homogeneity analysis (uH) and long-term stability testing (uLTS). The uncertainty is calculated based on the root-sum-of-squares equation times a coverage factor (k=2).

$$u_{CRM} = k\sqrt{u_M^2 + u_H^2 + u_{LTS}^2}$$

Transport conditions (short-term stability) have been tested such that there is no contribution to the uncertainty reported. The expanded uncertainty applies to the product as received.

10. **Metrological Traceability:** The property value (certified value and its uncertainty) are traceable through an unbroken chain of calibration to the SI base unit kg through a NIST traceable weight in accordance with ISO 17034. This is achieved through calibration of balances, verification of weights, use of national methodology for glassware calibration and product homogeneity and stability testing utilizing an ISO/IEC 17025 methodology.
11. **Values Obtained During Product Testing:** This product is subjected to verification, homogeneity and stability testing using an ISO/IEC 17025 chromatographic methodology. All values obtained during testing meet criteria in accordance with ISO 17034.
12. **Period of Validity:** The Certified Values, Uncertainties and Expiration Date are based on the unopened product being stored according to the recommended storage condition listed and are guaranteed until the expiration date. This product will be monitored during the period of validity and customers notified of any significant changes in stability.

## References:

- <sup>1</sup> ISO Guide 31 – Reference Materials – Contents of Certificates and Labels.
- <sup>2</sup> ISO Guide 35 – Reference Material – General and Statistical Principles for Certification.
- <sup>3</sup> ISO 17034 – General Requirements for the Competence of Reference Material Producers.
- <sup>4</sup> ISO/IEC 17025 – General Requirements for the Competence of Testing and Calibration Laboratories.
- <sup>5</sup> ISO/IEC Guide 98-3:2008(E) – Uncertainty of Measurement – Part 3: Guide to Expression of Uncertainty in Measurement (GUM: 1995)



Reference Material Producer  
Certificate No. 2427.02



Phenova is an accredited ISO/IEC 17034 Reference Material  
Producer and ISO/IEC 17025 accredited Chemical Testing Laboratory.



Chemical Testing Laboratory  
Certificate No. 2427.03



# Recipient Copy

## CHAIN-OF-CUSTODY RECORD

COC No. 15570

Order Number: CB014985

Date Shipped: 12/12/2022


AirBill No(s):

From: QATS LABORATORY  
2700 CHANDLER AVENUE, BLDG. B  
LAS VEGAS, NV 89120  
PHONE: 1-702-895-8712



To: SUE DUNNIHOO  
ANALYTICAL RESOURCES INC.  
4611 S. 134TH PLACE SUITE 100  
TUKWILA WA 98168  
250-695-6207

519204142631

K011177  
K011178  
K011179

Sample ID	Sigma ID	Qty	Description/Remarks	→ Catalogue Number
PSRM0168	SR0431	1	PUGET SOUND SEDIMENT RM	PS-SRM
PSRM0169	SR0431	1	PUGET SOUND SEDIMENT RM	PS-SRM
PSRM0171	SR0431	1	PUGET SOUND SEDIMENT RM	PS-SRM
				
<p>12/12/2022</p>				
<p>PUGET SOUND SRM FOR DUWAMISH AOC4 PROJECT.</p>				

Please use the enclosed Sample Preparation Instructions. If catalogue number(s) are listed at the top of the Sample Preparation Instructions use the Sample Preparation Instructions with catalogue number(s) matching the catalogue number(s) of each of the samples listed above.

Relinquished by: (Signature) 	Date/Time (1400) 12/12/2022	Received by: (Signature) 	Date/Time 12/12/22 11:15
Custody Seal(s): <u>Present</u> /Absent	Remarks:		
Relinquished by: (Signature)	Date/Time	Received by: (Signature)	Date/Time



**Form I**  
**INORGANIC ANALYSIS DATA SHEET**  
**EPA 9060A m**

<b>LDW23-SC1009</b>
---------------------

Laboratory: Analytical Resources, LLC  
 Client: Anchor QEA, LLC  
 Project: AOC5 MR Phase 1  
 Matrix: Sediment      Laboratory ID: 23B0228-01 C      SDG: 23B0228  
 Sampled: 02/08/23 16:30      Prepared: 02/14/23 12:18      File ID: CubeData\_02162023@1136-019  
 % Solids: 67.04      Preparation: PSEP 1986 (modified)      Analyzed: 02/15/23 04:42  
 Batch: BLB0342      Sequence: SLB0179      Initial/Final: 0.1387 g Wet / 0.1387 mL  
 Instrument: TOC Cube      Calibration: FD00070

CAS NO.	Analyte	Concentration (% dry)	Dilution Factor	MDL	MRL	Q
	Total Organic Carbon	0.91	1	0.02	0.02	



## PREPARATION BATCH SUMMARY

### EPA 9060A m

Laboratory: Analytical Resources, LLC SDG: 23B0228  
Client: Anchor QEA, LLC Project: AOC5 MR Phase 1  
Batch: BLB0342 Batch Matrix: Solid Preparation: PSEP 1986 (modified)

SAMPLE NAME	LAB SAMPLE ID	LAB FILE ID	DATE PREPARED	OBSERVATIONS
LDW23-SC1009	23B0228-01	eData_02162023@1136-	02/14/23 12:18	
Blank	BLB0342-BLK1	eData_02162023@1136-	02/14/23 12:18	
LCS	BLB0342-BS1	eData_02162023@1136-	02/14/23 12:18	
MRL Check	BLB0342-MRL1	eData_02162023@1136-	02/14/23 12:18	
Reference	BLB0342-SRM1	eData_02162023@1136-	02/14/23 12:18	



**Form I**  
**METHOD BLANK DATA SHEET**  
**EPA 9060A m**  
TotalAnalytes

<b>Blank</b>
--------------

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Batch: BLB0342

Laboratory ID: BLB0342-BLK1

Prepared: 02/14/23 12:18

Matrix: Solid

Preparation: PSEP 1986 (modified)

Analyzed: 02/14/23 22:09

Sequence: SLB0179

Calibration: FD00070

Instrument: TOC Cube

CAS NO.	Analyte	Concentration (% wet)	Dilution Factor	MDL	MRL	Q
	Total Organic Carbon	ND	1	0.02	0.02	U



**LCS / LCS DUPLICATE RECOVERY**  
**EPA 9060A m**

Laboratory:	<u>Analytical Resources, LLC</u>	SDG:	<u>23B0228</u>
Client:	<u>Anchor QEA, LLC</u>	Project:	<u>AOC5 MR Phase 1</u>
Matrix:	<u>Solid</u>	Analyzed:	<u>02/14/23 22:39</u>
Batch:	<u>BLB0342</u>	Laboratory ID:	<u>BLB0342-BS1</u>
Preparation:	<u>PSEP 1986 (modified)</u>	Sequence Name:	<u>LCS</u>
Initial/Final:	<u>0.021 g / 0.021 mL</u>		

COMPOUND	SPIKE ADDED (% wet)	LCS CONCENTRATION (% wet)	Q	LCS % REC. #	QC LIMITS REC.
Total Organic Carbon	44.4	44.0		99.0	80 - 120

\* Indicates values outside of QC limits



## ANALYSIS BATCH (SEQUENCE) SUMMARY

### EPA 9060A m

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Sequence: SKD0371

Instrument: TOC Cube

Calibration: FD00070

Sample Name	Lab Sample ID	Lab File ID	Matrix	Analysis Date/Time
Cal Standard	SKD0371-CAL1	CubeData_04272022@1136-001	NA	04/26/22 12:30
Cal Standard	SKD0371-CAL2	CubeData_04272022@1136-002	NA	04/26/22 13:00
Cal Standard	SKD0371-CAL3	CubeData_04272022@1136-003	NA	04/26/22 13:30
Cal Standard	SKD0371-CAL4	CubeData_04272022@1136-004	NA	04/26/22 14:00
Cal Standard	SKD0371-CAL5	CubeData_04272022@1136-005	NA	04/26/22 14:30
Cal Standard	SKD0371-CAL6	CubeData_04272022@1136-006	NA	04/26/22 15:00
Cal Standard	SKD0371-CAL7	CubeData_04272022@1136-007	NA	04/26/22 15:30
Cal Standard	SKD0371-CAL8	CubeData_04272022@1136-008	NA	04/26/22 16:00
Cal Standard	SKD0371-CAL9	CubeData_04272022@1136-009	NA	04/26/22 16:30
Cal Standard	SKD0371-CALA	CubeData_04272022@1136-010	NA	04/26/22 17:00
Cal Standard	SKD0371-CALB	CubeData_04272022@1136-011	NA	04/26/22 17:30
Cal Standard	SKD0371-CALC	CubeData_04272022@1136-012	NA	04/26/22 18:00
Cal Standard	SKD0371-CALD	CubeData_04272022@1136-013	NA	04/26/22 18:30
Cal Standard	SKD0371-CALE	CubeData_04272022@1136-014	NA	04/26/22 19:00
Cal Standard	SKD0371-CALF	CubeData_04272022@1136-015	NA	04/26/22 19:31
Cal Standard	SKD0371-CALG	CubeData_04272022@1136-016	NA	04/26/22 20:01
Cal Standard	SKD0371-CALH	CubeData_04272022@1136-017	NA	04/26/22 20:31
Cal Standard	SKD0371-CALI	CubeData_04272022@1136-018	NA	04/26/22 21:01
Cal Standard	SKD0371-CALJ	CubeData_04272022@1136-019	NA	04/26/22 21:31
Cal Standard	SKD0371-CALK	CubeData_04272022@1136-020	NA	04/26/22 22:01
Initial Cal Check	SKD0371-ICV1	CubeData_04272022@1136-027	NA	04/27/22 02:03
Initial Cal Blank	SKD0371-ICB1	CubeData_04272022@1136-028	NA	04/27/22 02:33
Cal Standard	SKD0371-CALL	CubeData_04272022@1136-021	NA	04/27/22 11:08
Cal Standard	SKD0371-CALM	CubeData_04272022@1136-022	NA	04/27/22 11:08
Cal Standard	SKD0371-CALN	CubeData_04272022@1136-023	NA	04/27/22 11:09
Cal Standard	SKD0371-CALO	CubeData_04272022@1136-024	NA	04/27/22 11:09



## ANALYSIS BATCH (SEQUENCE) SUMMARY

### EPA 9060A m

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Sequence: SLB0179


Instrument: TOC Cube

Calibration: FD00070

Sample Name	Lab Sample ID	Lab File ID	Matrix	Analysis Date/Time
Initial Cal Blank	SLB0179-ICB1	CubeData_02162023@1136-003	NA	02/14/23 20:39
Initial Cal Check	SLB0179-ICV1	CubeData_02162023@1136-004	NA	02/14/23 21:09
MRL Check	BLB0342-MRL1	CubeData_02162023@1136-005	Solid	02/14/23 21:39
Blank	BLB0342-BLK1	CubeData_02162023@1136-006	Solid	02/14/23 22:09
LCS	BLB0342-BS1	CubeData_02162023@1136-007	Solid	02/14/23 22:39
Reference	BLB0342-SRM1	CubeData_02162023@1136-008	Solid	02/14/23 23:09
Calibration Check	SLB0179-CCV1	CubeData_02162023@1136-015	NA	02/15/23 02:41
Calibration Blank	SLB0179-CCB1	CubeData_02162023@1136-016	NA	02/15/23 03:11
LDW23-SC1009	23B0228-01	CubeData_02162023@1136-019	Solid	02/15/23 04:42
Calibration Check	SLB0179-CCV2	CubeData_02162023@1136-027	NA	02/15/23 08:44
Calibration Blank	SLB0179-CCB2	CubeData_02162023@1136-028	NA	02/15/23 09:15
Calibration Check	SLB0179-CCV3	CubeData_02162023@1136-039	NA	02/15/23 14:49
Calibration Blank	SLB0179-CCB3	CubeData_02162023@1136-040	NA	02/15/23 15:20
Calibration Check	SLB0179-CCV4	CubeData_02162023@1136-050	NA	02/15/23 20:54
Calibration Blank	SLB0179-CCB4	CubeData_02162023@1136-051	NA	02/15/23 21:25
Calibration Check	SLB0179-CCV5	CubeData_02162023@1136-061	NA	02/16/23 02:27
Calibration Blank	SLB0179-CCB5	CubeData_02162023@1136-062	NA	02/16/23 02:57



**Soli TOC Cube, Carbon**  
**Balance: BAL3**  
**Analyst: CDE**



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



**Soli TOC Cube, Carbon**  
**Balance: BAL3**  
**Analyst: CDE**



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



**Soli TOC Cube, Carbon**  
**Balance: BAL3**  
**Analyst: CDE**



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: soliTOC superuser


Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



**Soli TOC Cube, Carbon**  
**Balance: BAL3**  
**Analyst: CDE**



Name:

Access: soliTOC superuser


Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



**Soli TOC Cube, Carbon**  
**Balance: BAL3**  
**Analyst: CDE**



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: soliTOC superuser


Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



**Soli TOC Cube, Carbon**  
**Balance: BAL3**  
**Analyst: CDE**



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: soliTOC superuser


Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



**Soli TOC Cube, Carbon**  
**Balance: BAL3**  
**Analyst: CDE**



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



**Soli TOC Cube, Carbon**  
**Balance: BAL3**  
**Analyst: CDE**



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



**Soli TOC Cube, Carbon**  
**Balance: BAL3**  
**Analyst: CDE**



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023




soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC





**Soli TOC Cube, Carbon**  
**Balance: BAL3**  
**Analyst: CDE**



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: soliTOC superuser


Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



**Soli TOC Cube, Carbon**  
**Balance: BAL3**  
**Analyst: CDE**



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: soliTOC superuser


Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: solITOC superuser


Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC



**Soli TOC Cube, Carbon**  
**Balance: BAL3**  
**Analyst: CDE**



Name:


Access: soliTOC superuser

Date: Thu Feb 16 09:54:06 2023



soliTOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:


Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC

Soli TOC Cube, Carbon  
 Balance: BAL3  
 Analyst: CDE



Name:

Access: solITOC superuser

Date: Thu Feb 16 09:54:06 2023



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC







### INITIAL CALIBRATION DATA

#### EPA 9060A m

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Calibration: FD00070

Instrument: TOC Cube

Calibration Date: 04/26/2022 11:29

Compound	Level 07		Level 08		Level 09		Level 10		Level 11		Level 12	
	Conc	RF	Conc	RF	Conc	RF	Conc	RF	Conc	RF	Conc	RF
Total Organic Carbon	0.074075	1370638	0.08937	1351930	0.12056	2158544	0.14995	1559046	0.24	1346463	0.288	1430135
Total Carbon	0.074075	1370638	0.08937	1351930	0.12056	2158544	0.14995	1559046	0.24	1346463	0.288	1430135
Total Inorganic Carbon	0.074075	1370638	0.08937	1351930	0.12056	2158544	0.14995	1559046	0.24	1346463	0.288	1430135
% Soot	0.074075	1370638	0.08937	1351930	0.12056	2158544	0.14995	1559046	0.24	1346463	0.288	1430135



## INITIAL CALIBRATION DATA

### EPA 9060A m

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Calibration: FD00070

Instrument: TOC Cube

Calibration Date: 04/26/2022 11:29

Compound	Level 13		Level 14		Level 15		Level 16		Level 17		Level 18	
	Conc	RF	Conc	RF	Conc	RF	Conc	RF	Conc	RF	Conc	RF
Total Organic Carbon	0.414	1337053	0.606	1385937	0.894	1382774	1.188	1379790	1.5	1375927	1.818	1372882
Total Carbon	0.414	1337053	0.606	1385937	0.894	1382774	1.188	1379790	1.5	1375927	1.818	1372882
Total Inorganic Carbon	0.414	1337053	0.606	1385937	0.894	1382774	1.188	1379790	1.5	1375927	1.818	1372882
% Soot	0.414	1337053	0.606	1385937	0.894	1382774	1.188	1379790	1.5	1375927	1.818	1372882



**INITIAL CALIBRATION DATA**

**EPA 9060A m**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Calibration: FD00070

Instrument: TOC Cube


Calibration Date: 04/26/2022 11:29

Compound	Level 19		Level 20		Level 21		Level 22		Level 23		Level 24	
	Conc	RF	Conc	RF	Conc	RF	Conc	RF	Conc	RF	Conc	RF
Total Organic Carbon	2.49	1398606	2.982	1376871	4.188	1256057	4.818	1279542	5.406	1283358	7.2	1301408
Total Carbon	2.49	1398606	2.982	1376871	4.188	1256057	4.818	1279542	5.406	1283358	7.2	1301408
Total Inorganic Carbon	2.49	1398606	2.982	1376871	4.188	1256057	4.818	1279542	5.406	1283358	7.2	1301408
% Soot	2.49	1398606	2.982	1376871	4.188	1256057	4.818	1279542	5.406	1283358	7.2	1301408





Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser

Date: Wed Apr 27 11:07:12 2022




solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC





Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser

Date: Wed Apr 27 11:07:12 2022




solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC





Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser

Date: Wed Apr 27 11:07:12 2022




solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC





Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser

Date: Wed Apr 27 11:07:12 2022




solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC





Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:07:12 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:10:16 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:10:16 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser


Date: Wed Apr 27 11:10:16 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC



Soli TOC Cube, Carbon  
Balance: BAL3  
Analyst: DOE



Name:

Access: solITOC superuser

Date: Wed Apr 27 11:10:16 2022




solITOC V2.0.2 (31015f9) 2018-11-19  
Serial No: 0300.181017  
Mode CCC

### Calibration parameters TC, Whole range

a	-4.107546e-02
b	+1.548032e-06
c	+0.000000e+00
d	+0.000000e+00
e	+0.000000e+00
r	0.998372
r_old	0.998372
Proc.-SD	166.070255 µg

Calibration graph TC, Whole range



Name:

Access: solITOC superuser

Date: Wed Apr 27 11:19:56 2022



solITOC V2.0.2 (31015f9) 2018-11-19  
 Serial No: 0300.181017  
 Mode CCC





**INSTRUMENT BLANKS**  
**EPA 9060A m**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Instrument ID: TOC Cube

Calibration: FD00070

Sequence: SKD0371

Date Analyzed: 04/27/22 02:33

Lab Sample ID	Analyte	Found	MDL	MRL	Units	C
SKD0371-ICB1	Total Organic Carbon	0.00	0.02	0.02	%	



**INSTRUMENT BLANKS**  
**EPA 9060A m**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Instrument ID: TOC Cube

Calibration: FD00070

Sequence: SLB0179

Date Analyzed: 02/14/23 20:39

Lab Sample ID	Analyte	Found	MDL	MRL	Units	C
SLB0179-ICB1	Total Organic Carbon	0.00	0.02	0.02	%	
SLB0179-CCB1	Total Organic Carbon	0.00	0.02	0.02	%	
SLB0179-CCB2	Total Organic Carbon	0.00	0.02	0.02	%	
SLB0179-CCB3	Total Organic Carbon	0.00	0.02	0.02	%	
SLB0179-CCB4	Total Organic Carbon	0.00	0.02	0.02	%	
SLB0179-CCB5	Total Organic Carbon	0.004	0.02	0.02	%	



**INITIAL AND CONTINUING  
CALIBRATION CHECK  
EPA 9060A m**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Instrument ID: TOC Cube

Calibration: FD00070

Control Limit: +/- 10.00%

Sequence: SKD0371

Lab Sample ID	Analyte	True	Found	%R	Units	Method
SKD0371-ICV1	Total Organic Carbon	44.446	43.7	98.3	%	EPA 9060A m
	Total Carbon	44.446	44.1	99.2	%	EPA 9060A m
	Total Inorganic Carbon	0.0000	0.40		%	EPA 9060A m
	% Soot	0.0000	0.004		%	EPA 9060A m

\* Values outside of QC limits



**INITIAL AND CONTINUING  
CALIBRATION CHECK**  
**EPA 9060A m**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Instrument ID: TOC Cube

Calibration: FD00070

Control Limit: +/- 10.00%

Sequence: SLB0179

Lab Sample ID	Analyte	True	Found	%R	Units	Method
SLB0179-ICV1	Total Organic Carbon	44.446	44.5	100	%	EPA 9060A m
SLB0179-CCV1	Total Organic Carbon	44.446	44.1	99.2	%	EPA 9060A m
SLB0179-CCV2	Total Organic Carbon	44.446	43.9	98.7	%	EPA 9060A m
SLB0179-CCV3	Total Organic Carbon	44.446	43.9	98.9	%	EPA 9060A m
SLB0179-CCV4	Total Organic Carbon	44.446	44.3	99.6	%	EPA 9060A m
SLB0179-CCV5	Total Organic Carbon	44.446	44.5	100	%	EPA 9060A m

\* Values outside of QC limits



## STANDARD REFERENCE MATERIAL RECOVERY

EPA 9060A m

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Matrix: Solid

Laboratory ID: BLB0342-SRM1

Batch: BLB0342

Initial/Final: 0.251 g / 0.251 mL

Preparation: PSEP 1986 (modified)

Analyzed: 02/14/2023 23:09

Standard ID: L000790

Expires: 05/19/2024

Standard Lot#: NA

Description: 1941B - Organics in Marine Sediment (Conv

ANALYTE	TRUE (% wet)	FOUND (% wet)	MDL	MRL	Q	SRM % REC.	QC LIMITS REC.
Total Organic Carbon	2.9900	2.93	0.02	0.02		97.9	80 - 120

\* Values outside of QC limits



## HOLDING TIME SUMMARY

**Analysis: EPA 9060A m**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Sample Name	Date Collected	Date Received	Date Prepared	Days to Prep	Max Days to Prep	Date Analyzed	Days to Analysis	Max Days to Analysis	Q
LDW23-SC1009 23B0228-01	02/08/23 16:30	02/09/23 10:55	02/14/23 12:18	5	14	02/15/23 04:42			

\* Indicates hold time exceedance.



**METHOD DETECTION  
AND REPORTING LIMITS**

**EPA 9060A m**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Matrix: Solid

Instrument: TOC Cube

<b>Analyte</b>	<b>MDL</b>	<b>RL</b>	<b>Units</b>
Total Organic Carbon	0.02	0.02	%



# National Institute of Standards & Technology

## Certificate of Analysis

### Standard Reference Material® 1941b

#### Organics in Marine Sediment

This Standard Reference Material (SRM) is marine sediment collected at the mouth of the Baltimore (MD) Harbor. SRM 1941b is intended for use in evaluating analytical methods for the determination of selected polycyclic aromatic hydrocarbons (PAHs), polychlorinated biphenyl (PCB) congeners, and chlorinated pesticides in marine sediment and similar matrices. Information values are also provided for total organic carbon (TOC), total carbon, hydrogen, and nitrogen. All of the constituents for which certified, reference, and information values are provided in SRM 1941b were naturally present in the sediment before processing. A unit of SRM 1941b consists of a bottle containing 50 g of radiation-sterilized, freeze-dried sediment.

**Certified Mass Fraction Values:** Certified mass fraction values for PAHs, PCB congeners, and chlorinated pesticides are provided in Table 1 through Table 3. The certified values for the PAHs, PCB congeners, and chlorinated pesticides are based on the agreement of results obtained at NIST from two or more chemically independent analytical techniques along with results from an interlaboratory comparison study [1]. A NIST certified value is a value for which NIST has the highest confidence in its accuracy in that all known or suspected sources of bias have been investigated or taken into account [1].

**Reference Mass Fraction Values:** Reference mass fraction values for additional PAHs (some in combination), additional PCB congeners, and additional chlorinated pesticides are provided in Table 4 through Table 7. Reference values for alkylated PAH groups are provided in Table 8 and for selected hopanes and steranes in Table 9. A reference value for total organic carbon is provided in Table 10. Reference values are noncertified values that are the best estimate of the true value; however, the values do not meet the NIST criteria for certification and are provided with associated uncertainties that may reflect only measurement precision, may not include all sources of uncertainty, or may reflect a lack of sufficient statistical agreement among multiple analytical methods [1].

**Information Mass Fraction Values:** Information mass fraction values are provided in Table 11 for carbon, hydrogen, and nitrogen. An information value is considered to be a value that will be of use to the SRM user, but insufficient information is available to assess the uncertainty associated with the value [1]. Information values cannot be used to establish metrological traceability.

**Expiration of Certification:** The certification of SRM 1941b is valid, within the measurement uncertainty specified, until **01 October 2020**, provided the SRM is handled and stored in accordance with the instructions given in this certificate (see "Instructions for Handling, Storage, and Use"). This certification is nullified if the SRM is damaged, contaminated, or otherwise modified.

**Maintenance of SRM Certification:** NIST will monitor this SRM over the period of its certification. If substantive technical changes occur that affect the certification before the expiration of this certificate, NIST will notify the purchaser. Registration (see attached sheet or register online) will facilitate notification.

Coordination of the technical measurements leading to the certification of this material was under the leadership of M.M. Schantz and S.A. Wise of the NIST Chemical Sciences Division.

Analytical measurements for the certification of SRM 1941b were performed at NIST by J.R. Kucklick, B.J. Porter, D.L. Poster, M.M. Schantz, P. Schubert, S. Tutschku, and L.L. Yu of the NIST Chemical Sciences Division.

Carlos A. Gonzalez, Chief  
Chemical Sciences Division



Measurements for TOC were provided by a commercial laboratory and T.L. Wade of the Geochemical and Environmental Research Group, Texas A&M University (College Station, TX). The carbon, hydrogen, and nitrogen data were provided by a commercial laboratory. Results for the PAHs, PCBs, and chlorinated pesticides from 38 laboratories (see Appendix A) that participated in an interlaboratory comparison exercise coordinated by NIST were used. Results for the alkylated PAH groups, hopanes, and steranes from 33 laboratories (see Appendix B) that participated in another interlaboratory comparison exercise coordinated by NIST were also used.

Collection and preparation of SRM 1941b were performed by M.P. Cronise and C.N. Fales of the NIST Office of Reference Materials and B.J. Porter and M.M. Schantz of the NIST Chemical Sciences Division. The sediment material was collected with the assistance of G.G. Lauenstein, J. Collier, and J. Lewis (National Oceanic and Atmospheric Administration, Silver Spring, MD).

Consultation on the statistical design of the experimental work and evaluation of the data were provided by S.D. Leigh and J.H. Yen of the NIST Statistical Engineering Division.

Support aspects involved in the issuance of this SRM were coordinated through the NIST Office of Reference Materials.

## INSTRUCTIONS FOR HANDLING, STORAGE, AND USE

**Handling:** This material is naturally occurring marine sediment from an urban area and may contain constituents of unknown toxicities; therefore, caution and care should be exercised during its handling and use.

**Storage:** SRM 1941b must be stored in its original bottle at temperatures less than 30 °C and away from direct sunlight.

**Use:** Prior to removal of subsamples for analysis, the contents of the bottle should be mixed. The mass fractions of constituents in SRM 1941b are reported on a dry-mass basis. The SRM, as received, contains a mass fraction of approximately 2.4 % moisture (see "Conversion to Dry-Mass Basis"). The sediment sample should be dried to a constant mass before weighing for analysis; or a separate subsample of the sediment should be removed from the bottle at the time of analysis and dried to determine the mass fraction on a dry-mass basis. If the constituents of interest are volatile, then the moisture must be determined with a separate subsample.

## PREPARATION AND ANALYSIS<sup>(1)</sup>

**Sample Collection and Preparation:** The sediment used to prepare this SRM was collected from the Chesapeake Bay at the mouth of the Baltimore (MD) Harbor near the Francis Scott Key Bridge (39°12.3'N and 76°31.4'W). This location is very near the site where SRM 1941 and SRM 1941a were collected. The sediment was collected using a Kynar-coated modified Van Veen-type grab sampler. A total of approximately 3300 kg of wet sediment was collected from the site. The sediment was freeze-dried, sieved at 150 μm (100 % passing), homogenized in a cone blender, radiation sterilized (<sup>60</sup>Co), and then packaged in screw-capped amber glass bottles each containing approximately 50 g.

**Conversion to Dry-Mass Basis:** The results for the constituents in SRM 1941b are reported on a dry-mass basis; however, the material "as received" contains residual moisture. The amount of moisture in SRM 1941b was determined by measuring the mass loss after freeze-drying subsamples of 1.1 g to 1.3 g for four days at 1 Pa with a -10 °C shelf temperature and a -50 °C condenser temperature. The moisture content in SRM 1941b at the time of the certification analyses was 2.39 % ± 0.08 % (95 % confidence level). Analytical results for the organic constituents were determined on an as-received basis and then converted to a dry-mass basis by dividing by the conversion factor of 0.9761 (gram dry mass per gram as-received mass).

**Polycyclic Aromatic Hydrocarbons:** The general approach used for the value assignment of the PAHs in SRM 1941b was similar to that reported in detail elsewhere [2]. The approach consisted of combining results from analyses using various combinations of different extraction techniques and solvents, clean-up/isolation procedures, and chromatographic separation and detection techniques: Soxhlet extraction and pressurized-fluid extraction (PFE) using dichloromethane (DCM) or a hexane/acetone mixture, cleanup of the extracts using solid-phase extraction (SPE) or normal-phase liquid chromatography (LC), followed by analysis using the following techniques: (1) reversed-phase liquid chromatography with fluorescence detection (LC-FL) analysis of the total PAH fraction, (2) reversed-phase

<sup>(1)</sup> Certain commercial equipment, instruments or materials are identified in this certificate to adequately specify the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology.

LC-FL analysis of isomeric PAH fractions isolated by normal-phase LC (i.e., multidimensional LC), (3) gas chromatography/mass spectrometry (GC/MS) analysis of the PAH fraction on three stationary phases of different selectivity, i.e., a 5 % (all column compositions are given as mole fractions in %) phenyl-substituted methylpolysiloxane phase, a 50 % phenyl-substituted methylpolysiloxane phase, and a relatively non-polar proprietary phase.

Three sets of GC/MS results, designated as GC/MS (I), GC/MS (II), and GC/MS (III), were obtained using three columns with different selectivities for the separation of PAHs. For GC/MS (I) analyses, duplicate subsamples of approximately 1 g from ten bottles of SRM 1941b were extracted using PFE with DCM. Copper powder was added to the extract to remove elemental sulfur. The concentrated extract was passed through an aminopropyl SPE cartridge and eluted with 2 % DCM in hexane (all solvent concentrations are given as volume fractions in %). The processed extract was then analyzed by GC/MS using a 0.25 mm i.d.  $\times$  60 m fused silica capillary column with a 5 % phenyl-substituted methylpolysiloxane phase (0.25  $\mu$ m film thickness; DB-5 MS, J&W Scientific, Folsom, CA). The GC/MS (II) analyses were performed using 5 g subsamples from six bottles of SRM 1941b. These samples were extracted using PFE with DCM. The high molecular mass compounds were removed from the extracts using size exclusion chromatography (SEC) with a preparative-scale divinylbenzene-polystyrene column (10  $\mu$ m particle size with 10 nm diameter pores), and the sulfur was removed from the extracts by adding copper powder. The concentrated extract was passed through an aminopropyl SPE cartridge and eluted with 10 % DCM in hexane. The analysis was by GC/MS using a 0.25 mm i.d.  $\times$  60 m fused silica capillary column with a 50 % phenyl-substituted methylpolysiloxane phase (0.25  $\mu$ m film thickness; DB-17 MS, J&W Scientific). For the GC/MS (III), 9 g subsamples from six bottles of SRM 1941b were Soxhlet-extracted for 18 h with 250 mL of a mixture of 50 % hexane/50 % acetone. Copper powder was added to the extract to remove elemental sulfur, and the concentrated extract was passed through a silica SPE cartridge and eluted with 10 % DCM in hexane. The processed extract was then analyzed by GC/MS using a 0.25 mm i.d.  $\times$  60 m fused silica capillary column with a relatively non-polar proprietary phase (0.25  $\mu$ m film thickness; DB-XLB, J&W Scientific).

Two sets of LC-FL results, designated as LC-FL (total) and LC-FL (isomer), were used in the certification process. For the LC-FL (total), subsamples of approximately 1 g from six bottles of SRM 1941b were extracted using PFE with a mixture of 50 % hexane/50 % acetone. The extracts were concentrated and then processed through an aminopropylsilane SPE cartridge using 2 % DCM in hexane to obtain the total PAH fraction. For the LC-FL (isomer), a 5 g subsample from the six bottles was extracted using PFE with DCM and processed through an aminopropylsilane SPE cartridge using 10 % DCM in hexane; the PAH fraction was then fractionated further on a semi-preparative aminopropylsilane column ( $\mu$ Bondapak NH<sub>2</sub>, 9 mm i.d.  $\times$  30 cm, Waters Associates, Milford, MA) to isolate isomeric PAH fractions as described previously [3–6]. The total PAH fraction and the isomeric PAH fractions were analyzed using a 5  $\mu$ m particle-size polymeric octadecylsilane (C<sub>18</sub>) column (4.6 mm i.d.  $\times$  25 cm, Hypersil-PAH, Keystone Scientific, Inc., Bellefonte, PA) with wavelength-programmed fluorescence detection [4,5].

For the GC/MS and LC-FL measurements described above, selected perdeuterated PAHs were added to the sediment prior to solvent extraction for use as internal standards for quantification purposes.

In addition to the analyses performed at NIST, SRM 1941b was used in an interlaboratory comparison exercise in 1999 as part of the NIST Intercomparison Exercise Program for Organic Contaminants in the Marine Environment [7]. Results from 38 laboratories that participated in this exercise were used as the sixth data set in the determination of the certified values for PAHs in SRM 1941b. The laboratories participating in this exercise used the analytical procedures routinely used in their laboratories to measure the analytes of interest.

**Homogeneity Assessment for PAHs:** The homogeneity of SRM 1941b was assessed by analyzing duplicate samples of approximately 1 g from ten bottles selected by stratified random sampling. Samples were extracted, processed, and analyzed as described above for GC/MS (I). No statistically significant differences among bottles were observed for the PAHs at this sample size.

**PAH Isomers of Molecular Mass 300 and 302:** For the determination of the molecular mass 300 and 302 isomers, three subsamples of approximately 5 g each were extracted using PFE with DCM. The extracts were then concentrated with a solvent change to hexane and passed through an aminopropyl SPE cartridge and eluted with 10 % DCM in hexane. The processed extract was then analyzed by GC/MS using a 0.25 mm i.d.  $\times$  60 m fused silica capillary column with a 50 % phenyl-substituted methylpolysiloxane phase (0.25  $\mu$ m film thickness; DB-17MS, J&W Scientific). Perdeuterated dibenzo[*a,i*]pyrene was added to the sediment prior to extraction for use as an internal standard [8].

**PCBs and Chlorinated Pesticides:** The general approach used for the determination of PCBs and chlorinated pesticides in SRM 1941b consisted of combining results from analyses using various combinations of different extraction techniques and solvents, cleanup/isolation procedures, and chromatographic separation and detection techniques. Techniques and solvents included Soxhlet extraction and PFE using DCM or a hexane/acetone mixture.

clean-up/isolation using SPE or LC, followed by analysis using GC/MS and gas chromatography with electron capture detection (GC-ECD) on two columns with different selectivity for the separation of PCBs and chlorinated pesticides. The analytical methods are described in detail elsewhere [2].

Six sets of results were obtained and designated as GC-ECD (I) A and B, GC/MS (I) A and B, GC/MS (II), and Interlaboratory Comparison Exercise. For the GC-ECD (I) analyses, approximately 10 g subsamples from six bottles of SRM 1941b were extracted using PFE with DCM. Copper powder was added to the extract to remove elemental sulfur, and SEC, as described above, was used to remove the high molecular mass compounds. The concentrated extract was then fractionated on a semi-preparative aminopropylsilane column to isolate two fractions containing: (1) the PCBs and lower-polarity pesticides and (2) the more polar pesticides. GC-ECD analyses of the two fractions were performed on two columns of different selectivities for PCB separations: 0.25 mm × 60 m fused silica capillary column with a 5 % phenyl-substituted methylpolysiloxane phase (0.25 μm film thickness; DB-5, J&W Scientific), and a 0.25 mm × 60 m fused silica capillary column with a non-polar proprietary phase (0.25 μm film thickness; DB-XLB, J&W Scientific). The results from the 5 % phenyl phase are designated as GC-ECD (IA) and the results from the proprietary phase are designated as GC-ECD (IB). For the GC-ECD analyses, two PCB congeners that are not significantly present in the sediment extract (PCB 103 and PCB 198 [9,10]) and endosulfan I-*d*<sub>4</sub>, 4,4'-DDE-*d*<sub>8</sub>, 4,4'-DD-*d*<sub>8</sub>, and 4,4'-DDT-*d*<sub>8</sub> were added to the sediment prior to extraction for use as internal standards for quantification purposes.

Two sets of results were obtained by GC/MS. For GC/MS (I), approximately 9 g subsamples from six bottles were Soxhlet- extracted with a mixture of 50 % hexane/50 % acetone for approximately 18 h. Copper powder was added to the extract to remove elemental sulfur, and the concentrated extract was passed through a silica SPE cartridge and eluted with 10 % DCM in hexane. The processed extract was then analyzed by GC/MS with two ionization modes, electron impact (EI) and negative ion chemical ionization (NICI). The GC/MS EI method, GC/MS (IA), used a 0.25 mm i.d. × 60 m fused silica capillary column with a relatively non-polar proprietary phase (0.25 μm film thickness; DB-XLB, J&W Scientific). The GC/MS NICI method, GC/MS (IB), used a 0.25 mm i.d. × 60 m fused silica capillary column with a 5 % phenyl-substituted methylpolysiloxane phase (0.25 μm film thickness; DB-5MS, J&W Scientific). The GC/MS (II) results were obtained in the same manner as the GC/MS (IA) analyses except that three subsamples were Soxhlet-extracted with DCM for approximately 18 h. For the GC/MS analyses, selected carbon-13 labeled PCB congeners and chlorinated pesticides were added to the sediment prior to extraction for use as internal standards for quantification purposes.

In addition to the analyses performed at NIST, SRM 1941b was used in an interlaboratory comparison exercise in 1999 as part of the NIST Intercomparison Exercise Program for Organic Contaminants in the Marine Environment [7]. Results from 38 laboratories that participated in this exercise were used as the sixth data set in the determination of the certified values for PCB congeners and chlorinated pesticides in SRM 1941b. The laboratories participating in this exercise used the analytical procedures routinely used in their laboratories to measure the analytes of interest.

The reference value for PCB 77 was determined from a separate fraction. The samples were extracted and processed as for GC-ECD (I) above. The first (PCB and lower-polarity pesticide) fraction from the semi-preparative aminopropylsilane column was further fractionated using a Cosmosil PYE (pyrenylethyl group bonded) column (5 μm particle size, 4.6 mm i.d. × 25 cm; Phenomenex, Torrance, CA) [11]. Three fractions were collected: the first fraction contained the pesticides and multi-*ortho* PCBs, the second fraction contained the polychlorinated naphthalenes, non-*ortho* PCB congeners, and some mono-*ortho* PCB congeners, and the third fraction removed the residual planar compounds from the column. The second fraction was analyzed by GC/MS NICI using the same column as GC/MS (IB) above. Carbon-13 labeled PCB 77 was used as an internal standard for quantification purposes.

**Alkylated PAH Groups, Hopanes, and Steranes:** SRM 1941b was used in an interlaboratory comparison exercise in 2011 [12]. Results from 33 laboratories that participated in this exercise were used in the determination of the reference values for alkylated PAH groups, hopanes, and steranes in SRM 1941b. Note that not all laboratories returned data for each analyte. The laboratories participating in this exercise used the analytical procedures routinely used in their laboratories to measure the analytes of interest. For the alkylated PAHs, the majority of the laboratories (>90 %) used the parent PAH for determination of the response factor for the corresponding alkylated group.

**Total Organic Carbon (TOC):** Two laboratories provided results for TOC using similar procedures. Briefly, subsamples of approximately 200 mg were reacted with 6 mol/L hydrochloric acid and rinsed with deionized water prior to combustion in a gas fusion furnace. The carbon monoxide and carbon dioxide produced were measured and compared to a blank for calculation of the percent TOC. Each laboratory analyzed subsamples from three bottles of SRM 1941b. One of the laboratories also analyzed three subsamples from three bottles of SRM 1941b for carbon, hydrogen, and nitrogen.

Table 1. Certified Mass Fraction Values for PAHs in SRM 1941b

PAHs	Mass Fractions <sup>(a)</sup> ( $\mu\text{g}/\text{kg}$ )	
Naphthalene <sup>(b,c,d,e,f,g)</sup>	848	$\pm 95^{(h)}$
Fluorene <sup>(b,c,d,e,f,g)</sup>	85	$\pm 15^{(h)}$
Phenanthrene <sup>(b,c,d,e,f,g)</sup>	406	$\pm 44^{(h)}$
Anthracene <sup>(b,c,d,e,f,g)</sup>	184	$\pm 18^{(h)}$
3-Methylphenanthrene <sup>(b,c,d)</sup>	105	$\pm 13^{(h)}$
2-Methylphenanthrene <sup>(b,c,d)</sup>	128	$\pm 14^{(h)}$
1-Methylphenanthrene <sup>(b,c,d,g)</sup>	73.2	$\pm 5.9^{(h)}$
Fluoranthene <sup>(b,c,d,e,f,g)</sup>	651	$\pm 50^{(h)}$
Pyrene <sup>(b,c,d,e,f,g)</sup>	581	$\pm 39^{(h)}$
Benz[ <i>a</i> ]anthracene <sup>(b,c,d,e,f,g)</sup>	335	$\pm 25^{(h)}$
Chrysene <sup>(d,f)</sup>	291	$\pm 31^{(h)}$
Triphenylene <sup>(d,f)</sup>	108	$\pm 5^{(i)}$
Benzo[ <i>b</i> ]fluoranthene <sup>(c,e)</sup>	453	$\pm 21^{(h)}$
Benzo[ <i>k</i> ]fluoranthene <sup>(b,c,d,e)</sup>	225	$\pm 18^{(h)}$
Benzo[ <i>e</i> ]pyrene <sup>(b,c,d,g)</sup>	325	$\pm 25^{(h)}$
Benzo[ <i>a</i> ]pyrene <sup>(b,c,d,f,g)</sup>	358	$\pm 17^{(h)}$
Perylene <sup>(b,c,d,f,g)</sup>	397	$\pm 45^{(h)}$
Benzo[ <i>ghi</i> ]perylene <sup>(b,c,d,f,g)</sup>	307	$\pm 45^{(h)}$
Indeno[1,2,3- <i>cd</i> ]pyrene <sup>(b,c,d,f,g)</sup>	341	$\pm 57^{(h)}$
Dibenz[ <i>a,j</i> ]anthracene <sup>(b,c,d,f)</sup>	48.9	$\pm 4.6^{(h)}$
Dibenz[ <i>a,c</i> ]anthracene <sup>(c,f)</sup>	36.7	$\pm 5.2^{(h)}$
Dibenz[ <i>a,h</i> ]anthracene <sup>(c,f)</sup>	53	$\pm 10^{(h)}$
Benzo[ <i>b</i> ]chrysene <sup>(b,c,d,f)</sup>	53	$\pm 12^{(h)}$
Picene <sup>(b,c,d)</sup>	46.6	$\pm 4.7^{(h)}$

<sup>(a)</sup> Mass fractions reported on dry-mass basis; material as received contains approximately 2.4 % moisture.

<sup>(b)</sup> GC/MS (I) on 5 % phenyl-substituted methylpolysiloxane phase after PFE with DCM.

<sup>(c)</sup> GC/MS (II) on 50 % phenyl-substituted methylpolysiloxane phase after PFE with DCM.

<sup>(d)</sup> GC/MS (III) on a relatively non-polar proprietary phase after Soxhlet extraction with 50 % hexane/50 % acetone mixture.

<sup>(e)</sup> LC-FL (total) of total PAH fraction after PFE with DCM.

<sup>(f)</sup> LC-FL (isomer) of isomeric PAH fractions after PFE with DCM.

<sup>(g)</sup> 1999 Interlaboratory Comparison Study [7] with 21 to 29 laboratories submitting data for each PAH.

<sup>(h)</sup> Certified values are weighted means of the results from two to six analytical methods [13]. The uncertainty listed with each value is an expanded uncertainty about the mean, with coverage factor 2 (approximately 95 % confidence), calculated by combining a between-method variance incorporating inter-method bias with a pooled within-method variance following the ISO/JCGM Guide [14,15]. The measurand is the total mass fraction of the constituent listed and the values are metrologically traceable to the SI unit of mass, expressed as micrograms per kilogram on a dry-mass basis.

<sup>(i)</sup> The certified value is an unweighted mean of the results from two analytical methods. The uncertainty listed with the value is an expanded uncertainty about the mean, with coverage factor 2, calculated by combining a between-method variance [16] with a pooled, within-method variance following the ISO/JCGM Guide [14,15]. The measurand is the total mass fraction of the constituent listed and the value is metrologically traceable to the SI unit of mass, expressed as micrograms per kilogram on a dry-mass basis.

Table 2. Certified Mass Fraction Values for PCB Congeners<sup>(a)</sup> in SRM 1941b

PCB Congeners		Mass Fractions <sup>(b)</sup> ( $\mu\text{g}/\text{kg}$ )
PCB	8 (2,4'-Dichlorobiphenyl) <sup>(c,d,e,f,g)</sup>	1.65 $\pm$ 0.19 <sup>(h)</sup>
PCB	18 (2,2',5-Trichlorobiphenyl) <sup>(c,d,e,f,g)</sup>	2.39 $\pm$ 0.29 <sup>(h)</sup>
PCB	28 (2,4,4'-Trichlorobiphenyl) <sup>(c,d,e,f,g)</sup>	4.52 $\pm$ 0.57 <sup>(h)</sup>
PCB	31 (2,4',5-Trichlorobiphenyl) <sup>(c,e,f)</sup>	3.18 $\pm$ 0.41 <sup>(h)</sup>
PCB	44 (2,2',3,5'-Tetrachlorobiphenyl) <sup>(c,d,e,f,g)</sup>	3.85 $\pm$ 0.20 <sup>(i)</sup>
PCB	49 (2,2',4,5'-Tetrachlorobiphenyl) <sup>(c,d,e,f)</sup>	4.34 $\pm$ 0.28 <sup>(i)</sup>
PCB	52 (2,2',5,5'-Tetrachlorobiphenyl) <sup>(c,d,e,f,g)</sup>	5.24 $\pm$ 0.28 <sup>(i)</sup>
PCB	66 (2,3',4,4'-Tetrachlorobiphenyl) <sup>(c,e,f,g,j)</sup>	4.96 $\pm$ 0.53 <sup>(i)</sup>
PCB	87 (2,2',3,4,5'-Pentachlorobiphenyl) <sup>(c,d,f,j)</sup>	1.14 $\pm$ 0.16 <sup>(h)</sup>
PCB	95 (2,2',3,5',6-Pentachlorobiphenyl) <sup>(c,e,f,g)</sup>	3.93 $\pm$ 0.62 <sup>(i)</sup>
PCB	99 (2,2',4,4',5-Pentachlorobiphenyl) <sup>(c,d,e,f,g)</sup>	2.90 $\pm$ 0.36 <sup>(i)</sup>
PCB	101 (2,2',4,5,5'-Pentachlorobiphenyl) <sup>(c,e,f,g,j)</sup>	5.11 $\pm$ 0.34 <sup>(i)</sup>
PCB	105 (2,3,3',4,4'-Pentachlorobiphenyl) <sup>(c,d,e,f,g,j)</sup>	1.43 $\pm$ 0.10 <sup>(i)</sup>
PCB	110 (2,3,3',4',6-Pentachlorobiphenyl) <sup>(c,e,f,j)</sup>	4.62 $\pm$ 0.36 <sup>(i)</sup>
PCB	118 (2,3',4,4',5-Pentachlorobiphenyl) <sup>(c,d,e,f,g,j)</sup>	4.23 $\pm$ 0.19 <sup>(i)</sup>
PCB	128 (2,2',3,3',4,4'-Hexachlorobiphenyl) <sup>(c,d,e,f,g,j)</sup>	0.696 $\pm$ 0.044 <sup>(i)</sup>
PCB	138 (2,2',3,4,4',5'-Hexachlorobiphenyl) <sup>(c,e,f,j)</sup>	3.60 $\pm$ 0.28 <sup>(i)</sup>
PCB	149 (2,2',3,4',5,6-Hexachlorobiphenyl) <sup>(c,d,e,j)</sup>	4.35 $\pm$ 0.26 <sup>(h)</sup>
PCB	153 (2,2',4,4',5,5'-Hexachlorobiphenyl) <sup>(c,d,e,f,g,j)</sup>	5.47 $\pm$ 0.32 <sup>(i)</sup>
PCB	156 (2,3,3',4,4',5-Hexachlorobiphenyl) <sup>(c,d,e,f,j)</sup>	0.507 $\pm$ 0.090 <sup>(h)</sup>
PCB	170 (2,2',3,3',4,4',5-Heptachlorobiphenyl) <sup>(c,d,e,f,g,j)</sup>	1.35 $\pm$ 0.09 <sup>(i)</sup>
PCB	180 (2,2',3,4,4',5,5'-Heptachlorobiphenyl) <sup>(c,d,e,f,g,j)</sup>	3.24 $\pm$ 0.51 <sup>(i)</sup>
PCB	183 (2,2',3,4,4',5,6-Heptachlorobiphenyl) <sup>(c,d,e,j)</sup>	0.979 $\pm$ 0.087 <sup>(h)</sup>
PCB	187 (2,2',3,4',5,5',6-Heptachlorobiphenyl) <sup>(c,d,e,f,g,j)</sup>	2.17 $\pm$ 0.22 <sup>(i)</sup>
PCB	194 (2,2',3,3',4,4',5,5'-Octachlorobiphenyl) <sup>(c,d,e,j)</sup>	1.04 $\pm$ 0.06 <sup>(h)</sup>
PCB	195 (2,2',3,3',4,4',5,6-Octachlorobiphenyl) <sup>(c,e,g,j)</sup>	0.645 $\pm$ 0.060 <sup>(i)</sup>
PCB	201 (2,2',3,3',4,5',6,6'-Octachlorobiphenyl) <sup>(c,e,j)</sup>	0.777 $\pm$ 0.034 <sup>(h)</sup>
PCB	206 (2,2',3,3',4,4',5,5',6-Nonachlorobiphenyl) <sup>(c,e,f,g,j)</sup>	2.42 $\pm$ 0.19 <sup>(i)</sup>
PCB	209 Decachlorobiphenyl <sup>(c,d,e,f,g,j)</sup>	4.86 $\pm$ 0.45 <sup>(i)</sup>

<sup>(a)</sup> PCB congeners are numbered according to the scheme proposed by Ballschmiter and Zell [9] and later revised by Schulte and Malisch [10] to conform to IUPAC rules, except PCB 201. Under the Ballschmiter and Zell numbering system, the IUPAC PCB 201 is listed as PCB 200.

<sup>(b)</sup> Mass fractions reported on dry-mass basis; material as received contains approximately 2.4 % moisture.

<sup>(c)</sup> GC/MS (IA) on a relatively non-polar proprietary phase after Soxhlet extraction with 50 % hexane/50 % acetone mixture.

<sup>(d)</sup> GC-ECD (IA) on 5 % phenyl-substituted methylpolysiloxane phase after PFE extraction with DCM.

<sup>(e)</sup> GC-ECD (IB) on a relatively non-polar proprietary phase; same extracts analyzed as in GC-ECD (IA).

<sup>(f)</sup> GC/MS (II) on a relatively non-polar proprietary phase after Soxhlet extraction with DCM.

<sup>(g)</sup> 1999 Interlaboratory Comparison Study [7] with 13 to 31 laboratories submitting data for each PCB congener.

<sup>(h)</sup> Certified values are unweighted means of the results from three to five analytical methods. The uncertainty listed with each value is an expanded uncertainty about the mean, with coverage factor 2, calculated by combining a between-method variance [16] with a pooled, within method variance following the ISO/JCGM Guide [14,15]. The measurand is the total mass fraction of the constituent listed and the values are metrologically traceable to the SI unit of mass, expressed as micrograms per kilogram on a dry-mass basis.

<sup>(i)</sup> Certified values are weighted means of the results from three to six analytical methods [13]. The uncertainty listed with each value is an expanded uncertainty about the mean, with coverage factor 2 (approximately 95 % confidence), calculated by combining a between-method variance incorporating inter-method bias with a pooled within-method variance following the ISO/JCGM Guide [14,15]. The measurand is the total mass fraction of the constituent listed and the values are metrologically traceable to the SI unit of mass, expressed as micrograms per kilogram on a dry-mass basis.

<sup>(j)</sup> GC/MS (IB) on 5 % phenyl-substituted methylpolysiloxane phase; same extracts analyzed as in GC/MS (IA).

Table 3. Certified Mass Fraction Values for Chlorinated Pesticides in SRM 1941b

Chlorinated Pesticides	Mass Fractions <sup>(a)</sup> ( $\mu\text{g}/\text{kg}$ )
Hexachlorobenzene <sup>(b,c,d,e)</sup>	5.83 $\pm$ 0.38 <sup>(f)</sup>
<i>cis</i> -Chlordane <sup>(b,c,d,e,g)</sup>	0.85 $\pm$ 0.11 <sup>(h)</sup>
<i>trans</i> -Chlordane <sup>(b,c,e)</sup>	0.566 $\pm$ 0.093 <sup>(f)</sup>
<i>cis</i> -Nonachlor <sup>(b,e,g)</sup>	0.378 $\pm$ 0.053 <sup>(h)</sup>
<i>trans</i> -Nonachlor <sup>(b,c,d,e,g)</sup>	0.438 $\pm$ 0.073 <sup>(f)</sup>
4,4'-DDE <sup>(b,d,e,g)</sup>	3.22 $\pm$ 0.28 <sup>(h)</sup>
4,4'-DDD <sup>(b,d,e,g)</sup>	4.66 $\pm$ 0.46 <sup>(h)</sup>

<sup>(a)</sup> Mass fractions reported on dry-mass basis; material as received contains approximately 2.4 % moisture.

<sup>(b)</sup> GC/MS (IA) on a relatively non-polar proprietary phase after Soxhlet extraction with 50 % hexane/50 % acetone mixture.

<sup>(c)</sup> GC/MS (IB) on 5 % phenyl-substituted methylpolysiloxane phase; same extracts analyzed as in GC/MS (IA).

<sup>(d)</sup> GC/MS (II) on a relatively non-polar proprietary phase after Soxhlet extraction with DCM.

<sup>(e)</sup> 1999 Interlaboratory Comparison Study [7] with 13 to 31 laboratories submitting data for each pesticide.

<sup>(f)</sup> Certified values are unweighted means of the results from three to five analytical methods. The uncertainty listed with each value is an expanded uncertainty about the mean, with coverage factor 2, calculated by combining a between-method variance [16] with a pooled, within-method variance following the ISO/JCGM Guide [14,15]. The measurand is the total mass fraction of the constituent listed and the values are metrologically traceable to the SI unit of mass, expressed as micrograms per kilogram on a dry-mass basis.

<sup>(g)</sup> GC-ECD (IA) on 5 % phenyl-substituted methylpolysiloxane phase after PFE extraction with DCM.

<sup>(h)</sup> Certified values are weighted means of the results from three to five analytical methods [13]. The uncertainty listed with each value is an expanded uncertainty about the mean, with coverage factor 2 (approximately 95 % confidence), calculated by combining a between-method variance incorporating inter-method bias with a pooled within-method variance following the ISO/JCGM Guide [14,15]. The measurand is the total mass fraction of the chlorinated pesticides listed and the values listed are metrologically traceable to the SI unit of mass, expressed as micrograms per kilogram on a dry-mass basis.

Table 4. Reference Mass Fraction Values for PAHs in SRM 1941b

PAHs	Mass Fractions <sup>(a)</sup>		
	(µg/kg)		
1-Methylnaphthalene <sup>(b,c,d,e)</sup>	127	±	14 <sup>(f)</sup>
2-Methylnaphthalene <sup>(b,c,d,e)</sup>	276	±	53 <sup>(f)</sup>
2,6-Dimethylnaphthalene <sup>(b,c,d,e)</sup>	75.9	±	4.5 <sup>(f)</sup>
2,3,5-Trimethylnaphthalene <sup>(b,c,d,e)</sup>	25.5	±	5.1 <sup>(f)</sup>
Biphenyl <sup>(b,c,d,e)</sup>	74.0	±	8.0 <sup>(f)</sup>
Acenaphthylene <sup>(b,c,d,e)</sup>	53.3	±	6.4 <sup>(f)</sup>
Acenaphthene <sup>(b,c,d,e)</sup>	38.4	±	5.2 <sup>(f)</sup>
9-Methylphenanthrene <sup>(c)</sup>	63.5	±	2.5 <sup>(g)</sup>
4-Methylphenanthrene and 9-Methylphenanthrene <sup>(b,d)</sup>	80.1	±	4.8 <sup>(f)</sup>
2-Methylanthracene <sup>(c,d)</sup>	36	±	15 <sup>(f)</sup>
8-Methylfluoranthene <sup>(b)</sup>	49.5	±	2.7 <sup>(g)</sup>
7-Methylfluoranthene <sup>(b)</sup>	45.4	±	1.5 <sup>(g)</sup>
1-Methylfluoranthene <sup>(b)</sup>	42.4	±	2.1 <sup>(g)</sup>
3-Methylfluoranthene <sup>(b)</sup>	28.8	±	1.3 <sup>(g)</sup>
2-Methylpyrene <sup>(b)</sup>	78.7	±	4.0 <sup>(g)</sup>
4-Methylpyrene <sup>(b)</sup>	66.4	±	2.6 <sup>(g)</sup>
1-Methylpyrene <sup>(b)</sup>	52.5	±	2.3 <sup>(g)</sup>
Acephenanthrene <sup>(d)</sup>	30.5	±	1.9 <sup>(g)</sup>
Benzo[ <i>c</i> ]phenanthrene <sup>(b,c,d)</sup>	58	±	15 <sup>(f)</sup>
Benzo[ <i>a</i> ]fluoranthene <sup>(b,c,d)</sup>	73	±	18 <sup>(f)</sup>
Benzo[ <i>j</i> ]fluoranthene <sup>(c)</sup>	217	±	5 <sup>(g)</sup>
Indeno[1,2,3- <i>cd</i> ]fluoranthene <sup>(d)</sup>	9.63	±	0.34 <sup>(g)</sup>
Pentaphene <sup>(d)</sup>	25.3	±	1.0 <sup>(g)</sup>

<sup>(a)</sup> Mass fractions reported on dry-mass basis; material as received contains approximately 2.4 % moisture.

<sup>(b)</sup> GC/MS (I) on 5 % phenyl-substituted methylpolysiloxane phase after PFE with DCM.

<sup>(c)</sup> GC/MS (II) on 50 % phenyl-substituted methylpolysiloxane phase after PFE with DCM.

<sup>(d)</sup> GC/MS (III) on a relatively non-polar proprietary phase after Soxhlet extraction with 50 % hexane/50 % acetone mixture.

<sup>(e)</sup> 1999 Interlaboratory Comparison Study [7] with 14 to 26 laboratories submitting data for each PAH.

<sup>(f)</sup> Reference values are weighted means of the results from two to four analytical methods [13]. The uncertainty listed with each value is an expanded uncertainty about the mean, with coverage factor 2 (approximately 95 % confidence), calculated by combining a between-method variance incorporating inter-method bias with a pooled within-method variance following the ISO/JCGM Guide [14,15]. The measurand is the total mass fraction of PAHs listed as determined by the methods indicated. The values are metrologically traceable to the SI unit of mass, expressed as micrograms per kilogram on a dry-mass basis.

<sup>(g)</sup> Reference values are the means of results obtained by NIST using one analytical technique. The expanded uncertainty,  $U$ , is calculated as  $U = k u_c$ , where  $u_c$  is one standard deviation of the analyte mean, and the coverage factor,  $k$ , is determined from the Student's  $t$ -distribution for the associated degrees of freedom (19 for footnote b and 5 for footnotes c and d) and 95 % confidence level for each analyte. The measurand is the total mass fraction of the PAHs listed as determined by the method indicated. The values listed are metrologically traceable to the SI unit of mass, expressed as micrograms per kilogram on a dry-mass basis.

Table 5. Reference Mass Fraction Values for PAHs of Molecular Mass 300 and 302 in SRM 1941b

PAHs of Molecular Mass 300 and 302	Mass Fractions <sup>(a,b,c)</sup> ( $\mu\text{g}/\text{kg}$ )
Coronene	72.6 $\pm$ 4.7
Dibenzo[ <i>b,e</i> ]fluoranthene	10.3 $\pm$ 0.3
Naphtho[1,2- <i>b</i> ]fluoranthene	91.0 $\pm$ 3.1
Naphtho[1,2- <i>k</i> ]fluoranthene and Naphtho[2,3- <i>j</i> ]fluoranthene	79.8 $\pm$ 2.5
Naphtho[2,3- <i>b</i> ]fluoranthene	23.5 $\pm$ 0.3
Dibenzo[ <i>b,k</i> ]fluoranthene	95.6 $\pm$ 3.1
Dibenzo[ <i>a,k</i> ]fluoranthene	26.6 $\pm$ 0.4
Dibenzo[ <i>j,l</i> ]fluoranthene	63.8 $\pm$ 1.8
Dibenzo[ <i>a,l</i> ]pyrene	11.1 $\pm$ 1.0
Naphtho[2,3- <i>k</i> ]fluoranthene	10.7 $\pm$ 0.6
Naphtho[1,2- <i>a</i> ]pyrene	16.7 $\pm$ 1.4
Naphtho[2,3- <i>e</i> ]pyrene	33.2 $\pm$ 2.3
Dibenzo[ <i>a,e</i> ]pyrene	76.1 $\pm$ 3.6
Naphtho[2,1- <i>a</i> ]pyrene	59.2 $\pm$ 1.8
Dibenzo[ <i>e,i</i> ]pyrene	35.0 $\pm$ 2.4
Naphtho[2,3- <i>a</i> ]pyrene	16.5 $\pm$ 0.6
Benzo[ <i>b</i> ]perylene	38.2 $\pm$ 1.2
Dibenzo[ <i>a,i</i> ]pyrene	25.5 $\pm$ 1.0
Dibenzo[ <i>a,h</i> ]pyrene	6.94 $\pm$ 0.29

<sup>(a)</sup> Mass fractions reported on dry-mass basis; material as received contains approximately 2.4 % moisture.

<sup>(b)</sup> Reference values are the means of results obtained by NIST using one analytical technique. The expanded uncertainty,  $U$ , is calculated as  $U = k u_c$ , where  $u_c$  is one standard deviation of the analyte mean, and the coverage factor,  $k$ , is determined from the Student's  $t$ -distribution for two degrees of freedom and 95 % confidence level for each analyte. The measurand is the total mass fraction of the constituent listed as determined by the method indicated. The values are metrologically traceable to the SI unit of mass, expressed as micrograms per kilogram on a dry-mass basis.

<sup>(c)</sup> GC/MS on 50 % phenyl-substituted methylpolysiloxane phase after PFE with DCM [8].



Table 6. Reference Mass Fraction Values for PCB Congeners<sup>(a)</sup> in SRM 1941b

PCB Congeners			Mass Fractions <sup>(b,c)</sup> ( $\mu\text{g}/\text{kg}$ )		
PCB	45	(2,2',3,6-Tetrachlorobiphenyl) <sup>(d,e)</sup>	0.73	±	0.12
PCB	56	(2,3,3',4'-Tetrachlorobiphenyl) <sup>(d,f,g)</sup>	1.21	±	0.11
PCB	63	(2,3,4',5-Tetrachlorobiphenyl) <sup>(e,f,g)</sup>	0.213	±	0.040
PCB	70	(2,3',4',5-Tetrachlorobiphenyl) <sup>(e,f,g)</sup>	4.99	±	0.29
PCB	74	(2,4,4',5-Tetrachlorobiphenyl) <sup>(e,f,g)</sup>	2.04	±	0.15
PCB	77	(3,3',4,4'-Tetrachlorobiphenyl) <sup>(h)</sup>	0.31	±	0.03
PCB	107	(2,3,3',4',5-Pentachlorobiphenyl) <sup>(d,e,f,g)</sup>	0.628	±	0.028
PCB	132	(2,2',3,3',4,6'-Hexachlorobiphenyl) <sup>(d,f,g)</sup>	1.28	±	0.27
PCB	146	(2,2',3,4',5,5'-Hexachlorobiphenyl) <sup>(e,f,g)</sup>	1.22	±	0.12
PCB	158	(2,3,3',4,4',6-Hexachlorobiphenyl) <sup>(d,e,f,g)</sup>	0.65	±	0.15
PCB	163	(2,3,3',4',5,6-Hexachlorobiphenyl) <sup>(e,f,g)</sup>	1.28	±	0.06
PCB	174	(2,2',3,3',4,5,6'-Heptachlorobiphenyl) <sup>(d,e,f,g)</sup>	1.51	±	0.39
PCB	193	(2,3,3',4',5,5',6-Heptachlorobiphenyl) <sup>(d,e,f,g)</sup>	0.292	±	0.075

<sup>(a)</sup> PCB congeners are numbered according to the scheme proposed by Ballschmiter and Zell [9] and later revised by Schulte and Malisch [10] to conform with IUPAC rules, except PCB 107. Under the Ballschmiter and Zell numbering system, the IUPAC PCB 107 is listed as PCB 108.

<sup>(b)</sup> Mass fractions reported on dry-mass basis; material as received contains approximately 2.4 % moisture.

<sup>(c)</sup> For these PCB congeners except PCB 77, the reference values are unweighted means of the results from two to four analytical methods. The uncertainty listed with each value is an expanded uncertainty about the mean, with coverage factor 2, calculated by combining a between-method variance [16] with a pooled within-method variance following the ISO/JCGM Guide [14,15]. For PCB 77, the reference value is the mean of results obtained by NIST using one analytical technique. The expanded uncertainty,  $U$ , is calculated as  $U = k u_c$ , where  $u_c$  is one standard deviation of the analyte mean, and the coverage factor,  $k$ , is determined from the Student's  $t$ -distribution corresponding to two degrees of freedom and 95 % confidence level for PCB 77. The measurand is the total mass fraction of the PCB Congeners listed as determined by the method or methods indicated. The values listed are metrologically traceable to the SI unit of mass, expressed as microgram per kilogram on a dry-mass basis.

<sup>(d)</sup> GC-ECD (IA) on 5 % phenyl-substituted methylpolysiloxane phase after PFE extraction with DCM.

<sup>(e)</sup> GC-ECD (IB) on a relatively non-polar proprietary phase; same extracts analyzed as in GC-ECD (IA).

<sup>(f)</sup> GC/MS (IA) on a relatively non-polar proprietary phase after Soxhlet extraction with 50 % hexane/50 % acetone mixture.

<sup>(g)</sup> GC/MS (IB) on 5 % phenyl-substituted methylpolysiloxane phase; same extracts analyzed as in GC/MS (IA).

<sup>(h)</sup> GC/MS NICI on a 5 % phenyl-substituted methylpolysiloxane phase; same extracts analyzed as in GC-ECD (I) fractionated using a PYE column.

Table 7. Reference Mass Fraction Values for Selected Chlorinated Pesticides in SRM 1941b

Chlorinated Pesticides	Mass Fractions <sup>(a,b)</sup> ( $\mu\text{g}/\text{kg}$ )
2,4'-DDE <sup>(c,d)</sup>	0.38 $\pm$ 0.12
4,4'-DDT <sup>(e,f)</sup>	1.12 $\pm$ 0.42

<sup>(a)</sup> Mass Fractions reported on dry-mass basis; material as received contains approximately 2.4 % moisture.

<sup>(b)</sup> The reference values are unweighted means of the results from two analytical methods. The uncertainty listed with each value is an expanded uncertainty about the mean, with coverage factor 2, calculated by combining a between-method variance [16] with a pooled, within-method variance following the ISO/JCGM Guide [14,15]. The measurand is the total mass fraction of the chlorinated pesticides listed as determined by the methods indicated. The values listed are metrologically traceable to the SI unit of mass, expressed as micrograms per kilogram on a dry-mass basis.

<sup>(c)</sup> GC/MS (IB) on 5 % phenyl-substituted methylpolysiloxane phase; same extracts analyzed as in GC/MS (IA).

<sup>(d)</sup> GC-ECD (IB) on a relatively non-polar proprietary phase; same extracts analyzed as in GC-ECD (IA).

<sup>(e)</sup> GC/MS (II) on a relatively non-polar proprietary phase after Soxhlet extraction with DCM.

<sup>(f)</sup> 1999 Interlaboratory Comparison Study [7] with 10 laboratories submitting data for 4,4'-DDT.

Table 8. Reference Mass Fraction Values for Alkylated PAH Groups in SRM 1941b

Alkylated PAH Group	Mass Fraction <sup>(a,b)</sup> ( $\mu\text{g}/\text{kg}$ )
C2-decalins	18 $\pm$ 5
C4-decalins	41 $\pm$ 4
C2-naphthalenes	187 $\pm$ 53
C3-naphthalenes	158 $\pm$ 42
C1-benzothiophenes	25 $\pm$ 14
C2-benzothiophenes	20 $\pm$ 11
C3-benzothiophenes	22 $\pm$ 13
C4-benzothiophenes	18 $\pm$ 5
C1-fluorenes	57 $\pm$ 18
C2-fluorenes	122 $\pm$ 43
C3-fluorenes	128 $\pm$ 31
C1-phenanthrenes/anthracenes	313 $\pm$ 99
C2-phenanthrenes/anthracenes	247 $\pm$ 62
C3-phenanthrenes/anthracenes	165 $\pm$ 46
C4-phenanthrenes/anthracenes	87 $\pm$ 36
C1-dibenzothiophenes	54 $\pm$ 13
C2-dibenzothiophenes	91 $\pm$ 18
C3-dibenzothiophenes	84 $\pm$ 15
C4-dibenzothiophenes	57 $\pm$ 13
C1-fluoranthenes/pyrenes	252 $\pm$ 48
C2-fluoranthenes/pyrenes	205 $\pm$ 38
C3-fluoranthenes/pyrenes	102 $\pm$ 22
C4-fluoranthenes/pyrenes	121 $\pm$ 59
C1-benzanthracenes/chrysenes/triphenylenes	208 $\pm$ 43
C2-benzanthracenes/chrysenes/triphenylenes	120 $\pm$ 24
C3-benzanthracenes/chrysenes/triphenylenes	73 $\pm$ 31
C4-benzanthracenes/chrysenes/triphenylenes	41 $\pm$ 11

<sup>(a)</sup> The reference mass fraction value reported on a dry-mass basis is the median of results using one analytical technique. The expanded uncertainty,  $U$ , is calculated as  $U = ku_c$ , where  $u_c$  is one standard deviation of the median, and the coverage factor,  $k = 2$ . The measurand is the total mass fraction of the alkylated PAH groups listed as determined by the interlaboratory study methods. The values listed are metrologically traceable to the SI unit of mass fraction, expressed as micrograms per kilogram on a dry-mass basis.

<sup>(b)</sup> Data from the interlaboratory study [12].

Table 9. Reference Mass Fraction Values for Hopanes and Steranes in SRM 1941b

Hopane or Sterane	Mass Fraction <sup>(a,b)</sup> (µg/kg)
17α(H)-22,29,30-Trisnorhopane	54 ± 18
17α(H)-21β(H)-30-Norhopane	137 ± 21
17α(H)-21β(H)-30-Hopane	215 ± 44
17α(H)-21β(H)-22R-Homohopane	44 ± 10
17α(H)-21β(H)-22S-Homohopane	48 ± 13
5α(H)-14α(H),17α(H)-Cholestane 20R	41 ± 11
5α(H)-14β(H),17β(H)-Cholestane 20R	27 ± 6
5α(H)-14β(H),17β(H)-24-Methylcholestane 20R	21 ± 8
5α(H)-14α(H),17α(H)-24-Ethylcholestane 20R	19 ± 5
5α(H)-14β(H),17β(H)-24-Ethylcholestane 20R	41 ± 9

- <sup>(a)</sup> The reference mass fraction value reported on a dry-mass basis is the median of results using one analytical technique. The expanded uncertainty,  $U$ , is calculated as  $U = k u_c$ , where  $u_c$  is one standard deviation of the median, and the coverage factor,  $k = 2$ . The measurand is the total mass fraction of the constituent listed as determined by the methods used during the interlaboratory study. The values are metrologically traceable to the SI unit of mass, expressed as micrograms per kilogram on a dry-mass basis.
- <sup>(b)</sup> Data from the interlaboratory study [12].

Table 10. Reference Mass Fraction Value for Total Organic Carbon in SRM 1941b

Total Organic Carbon (TOC)	2.99 % ± 0.24 % <sup>(a,b)</sup>
----------------------------	----------------------------------

- <sup>(a)</sup> Mass fraction is reported on a dry-mass basis; material as received contains approximately 2.4 % moisture.
- <sup>(b)</sup> The reference value for total organic carbon is a weighted mean value from routine measurements made by two laboratories [21]. The uncertainty listed is an expanded uncertainty about the mean, with coverage factor 2 (approximately 95 % confidence), calculated by combining a between-method variance incorporating inter-method bias with a pooled within-method variance. The reporting follows the ISO/JCGM Guides [2]. The measurand is the total mass fraction of TOC listed as determined by the methods indicated. The values listed are metrologically traceable to the SI unit of mass, expressed as a percent on a dry-mass basis.

Table 11. Information Mass Fraction Values for Carbon, Hydrogen, and Nitrogen in SRM 1941b

Elements	Mass Fractions <sup>(a)</sup> (%)
Carbon	3.3
Hydrogen	1.2
Nitrogen	<0.5

- <sup>(a)</sup> Mass fraction is reported on a dry-mass basis; material as received contains approximately 2.4 % moisture.

## REFERENCES

- [1] May, W.E.; Parris, R.M.; Beck II, C.M.; Fassett, J.D.; Greenberg, R.R.; Guenther, F.R.; Kramer, G.W.; Wise, S.A.; Gills, T.E.; Colbert, J.C.; Gettings, R.J.; MacDonald, B.R.; *Definition of Terms and Modes Used at NIST for Value-Assignment of Reference Materials for Chemical Measurements*; NIST Special Publication 260-136 (2000); available at <http://www.nist.gov/srm/publications.cfm> (accessed Jan 2015).
- [2] Wise, S.A.; Poster, D.L.; Schantz, M.M.; Kucklick, J.R.; Sander, L.C.; Lopez de Alda, M.; Schubert, P.; Parris, R.M.; Porter, B.J.; *Two New Marine Sediment Standard Reference Materials (SRMs) for the Determination of Organic Contaminants*; Anal. Bioanal. Chem., Vol. 378, pp. 1251–1264 (2004).
- [3] Wise, S.A.; Chesler, S.N.; Hertz, H.S.; Hilpert, L.R.; May, W.E.; *Chemically-Bonded Aminosilane Stationary Phase for the High Performance Liquid Chromatographic Separation of Polynuclear Aromatic Hydrocarbons*; Anal. Chem., Vol. 49, pp. 2306–2310 (1977).
- [4] May, W.E.; Wise, S.A.; *Liquid Chromatographic Determination of Polycyclic Aromatic Hydrocarbons in Air Particulate Extracts*; Anal. Chem., Vol. 56, pp. 225–232 (1984).
- [5] Wise, S.A.; Benner, B.A.; Byrd, G.D.; Chesler, S.N.; Rebbert, R.E.; Schantz, M.M.; *Determination of Polycyclic Aromatic Hydrocarbons in a Coal Tar Standard Reference Material*; Anal. Chem., Vol. 60, pp. 887–894 (1988).
- [6] Wise, S.A.; Deissler, A.; Sander, L.C.; *Liquid Chromatographic Determination of Polycyclic Aromatic Hydrocarbon Isomers of Molecular Weight 278 and 302 in Environmental Standard Reference Materials*; Polycyclic Aromat. Compd., Vol. 3, pp. 169–184 (1993).
- [7] Schantz, M.M.; Parris, R.M.; Wise, S.A.; *NIST/NOAA Intercomparison Exercise Program for Organic Contaminants in the Marine Environment: Description and Results of 1999 Organic Intercomparison Exercises*; NOAA Technical Memorandum NOS NCCOS CCMA 146, Silver Spring, MD (2000).
- [8] Schubert, P.; Schantz, M.M.; Sander, L.C.; Wise, S.A.; *Determination of Polycyclic Aromatic Hydrocarbons with Molecular Mass 300 and 302 in Environmental-Matrix Standard Reference Materials by Gas Chromatography-Mass Spectrometry*; Anal. Chem., Vol. 75, pp. 234–246 (2003).
- [9] Ballschmiter, K.; Zell, M.; *Analysis of Polychlorinated Biphenyls (PCB) by Glass Capillary Gas Chromatography - Composition of Technical Aroclor- and Clophen-PCB Mixtures*; Fresenius' Z. Anal. Chem., Vol. 302, pp. 20–31 (1980).
- [10] Schulte, E.; Malisch, R.; *Calculation of the Real PCB Content in Environmental Samples. I. Investigation of the Composition of Two Technical PCB Mixtures*; Fresenius' Z. Anal. Chem., Vol. 314, pp. 545–551 (1983).
- [11] Brubaker, W.W., Jr.; Schantz, M.M.; Wise, S.A.; *Determination of Non-ortho Polychlorinated Biphenyls in Environmental Standard Reference Materials*; Fresenius' J. Anal. Chem., Vol. 367, pp. 401–406 (2000).
- [12] Schantz, M.M.; Kucklick, J.R.; *NIST Interlaboratory Analytical Comparison Study to Support Deepwater Horizon Natural Resource Damage Assessment: Description and Results for Crude Oil QA10OIL01*; NISTIR 7793 (2011).
- [13] Ruhkin, A.L.; Vangel, M.G.; *Estimation of a Common Mean and Weighted Means Statistics*; J. Am. Statist. Assoc., Vol. 93, pp. 303–308 (1998).
- [14] JCGM 100:2008; *Evaluation of Measurement Data — Guide to the Expression of Uncertainty in Measurement (GUM 1995 with Minor Corrections)*; Joint Committee for Guides in Metrology (2008); available at [http://www.bipm.org/utis/common/documents/jcgm/JCGM\\_100\\_2008\\_E.pdf](http://www.bipm.org/utis/common/documents/jcgm/JCGM_100_2008_E.pdf) (accessed Jan 2015); see also Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297; U.S. Government Printing Office: Washington, DC (1994); available at <http://www.nist.gov/pml/pubs/tn1297/index.cfm> (accessed Jan 2015).
- [15] JCGM 101:2008; *Evaluation of measurement data – Supplement 1 to the “Guide to the expression of uncertainty in measurement” - Propagation of distributions using a Monte Carlo method*; JCGM (2008); available at [http://www.bipm.org/utis/common/documents/jcgm/JCGM\\_101\\_2008\\_E.pdf](http://www.bipm.org/utis/common/documents/jcgm/JCGM_101_2008_E.pdf) (accessed Jan 2015).
- [16] Levenson, M.S.; Banks, D.L.; Eberhardt, K.R.; Gill, L.M.; Guthrie, W.F.; Liu, H.-K.; Vangel, M.G.; Yen, J.H.; Zhang, N.F.; *An Approach to Combining Results from Multiple Methods Motivated by the ISO GUM*; J. Res. Natl. Inst. Stand. Technol., Vol. 105, pp. 571–579 (2000).

**Certificate Revision History:** 16 January 2015 (Corrected IUPAC name for PCB-56 and PCB-107; editorial changes); 10 June 2014 (Units corrected from mg/kg to µg/kg in Tables 8 and 9; editorial changes); 10 April 2012 (Reference value added for alkylated PAH groups, hopanes, and steranes; extension of certification period; editorial changes); 16 August 2004 (Reference values for the butyl tins removed; editorial changes); 15 July 2002 (Original certificate date).

Users of this SRM should ensure that the Certificate of Analysis in their possession is current. This can be accomplished by contacting the SRM Program: telephone (301) 975-2200; fax (301) 948-3730; e-mail [srminfo@nist.gov](mailto:srminfo@nist.gov); or via the Internet at <http://www.nist.gov/srm>.

## APPENDIX A

The laboratories listed below performed measurements that contributed to the certification of PAHs, PCBs, and chlorinated pesticides in SRM 1941b Organics in Marine Sediment.

Arthur D. Little, Inc; Cambridge, MA  
Axys Analytical Services; Sidney, BC, Canada  
B & B Laboratories; College Station, TX  
Battelle Ocean Sciences; Duxbury, MA  
Bedford Institute of Oceanography; Dartmouth, NS, Canada  
California Department of Fish and Game; Rancho Cordova, CA  
Central Contra Costa Sanitary District; Martinez, CA  
Chesapeake Biological Laboratory; Solomons, MD  
Centro de Investigaciones Energeticas Medioambientales y Tecnologicas; Madrid, Spain  
City of Los Angeles Environmental Monitoring Division; Playa del Rey, CA  
City of San Jose Environmental Services Department; San Jose, CA  
Columbia Analytical Services; Kelso, WA  
East Bay Municipal Utility District; Oakland, CA  
Florida Department of Environmental Protection; Tallahassee, FL  
Manchester Environmental Laboratory; Port Orchard, WA  
Murray State University; Murray, KY  
Massachusetts Water Resources Authority Central Lab; Winthrop, MA  
National Research Council of Canada; Ottawa, Ontario, Canada  
National Oceanic and Atmospheric Association (NOAA), National Marine Fisheries Service (NMFS), Auke Bay Laboratory; Juneau, AK  
NOAA, National Ocean Service/Center for Coastal Environmental Health and Biomolecular Research; Charleston, SC  
NOAA, NMFS, Sandy Hook Marine Laboratory; Highlands, NJ  
NOAA, NMFS, Northwest Fisheries Science Center; Seattle, WA  
Orange County Sanitation District; Fountain Valley, CA  
Philip Analytical Services; Burlington, Ontario, Canada  
Serv de Hidrografia Naval; Buenos Aires, Argentina  
Skidaway Institute of Technology; Savannah, GA  
Southwest Laboratory of Oklahoma; Broken Arrow, OK  
Severn Trent Knoxville Laboratory; Knoxville, TN  
Texas A&M University, Geochemical and Environmental Research Group; College Station, TX  
Texas Parks and Wildlife Department; San Marcos, TX  
University of California at Los Angeles, Institute of Geophysics and Planetary Physics; Los Angeles, CA  
University of Connecticut, Environmental Research Institute; Storrs, CT  
University of Rhode Island, Graduate School of Oceanography; Narragansett, RI  
US Department of Agriculture, Environmental Chemistry Laboratory; Beltsville, MD  
US Environmental Protection Agency, Atlantic Ecology Division; Narragansett, RI  
US Geological Survey, National Water Quality Laboratory; Denver, CO  
Woods Hole Group Environmental Lab; Raynham, MA  
Wright State University; Dayton, OH

## APPENDIX B

The laboratories listed below performed measurements that contributed to the certification of alkylated PAH groups, hopanes, and steranes in SRM 1941b Organics in Marine Sediment.

Alpha Analytical, Inc.; Mansfield, MA  
Analytical Resources, Inc.; Tukwila, WA  
Axy's Analytical Services; Sydney, BC, Canada  
Battelle Analytical & Environmental Chemistry Laboratory; Duxbury, MA  
Center for Laboratory Sciences; Pasco, WA  
Columbia Analytical Services; Jacksonville, FL  
Columbia Analytical Services; Rochester, NY  
Columbia Analytical Services, Kelso, WA  
Florida Department of Environmental Protection; Tallahassee, FL  
Florida International University; North Miami, FL  
Michigan Department of Natural Resources and Environment; Lansing, MI  
Mississippi State Chemical Laboratory; Mississippi State, MS  
NIST; Charleston, SC  
NIST; Gaithersburg, MD  
NOAA/NCCOS/NOS; Charleston, SC  
NOAA/NMFS/Alaska Fisheries Science Center; Juneau, AK  
NY State Department of Health; Albany, NY  
Pace Analytical Services, Inc. Minneapolis; Minneapolis, MN  
RJ Lee Group, Inc; Monroeville, PA  
TDI/B&B Laboratories, Inc.; College Station, TX  
TestAmerica Laboratories; Mobile, AL  
TestAmerica Laboratories; West Sacramento, CA  
TestAmerica Laboratories; University Park, IL  
TestAmerica Laboratories; Schriever, LA  
TestAmerica Laboratories; Edison, NJ  
TestAmerica Laboratories; Knoxville, TN  
TestAmerica Laboratories; Pittsburgh, PA  
TestAmerica Laboratories; South Burlington, VT  
TestAmerica Laboratories; Tacoma, WA  
US Army Engineer Research and Development Center; Vicksburg, MS  
USGS Columbia Environmental Research Center; Columbia, MO  
University of Iowa, State Hygienic Laboratory; Iowa City, IO  
Washington State Public Health Laboratories; Shoreline, WA



Date of Issue:  
31 March 2014

## SAFETY DATA SHEET

### 1. SUBSTANCE AND SOURCE IDENTIFICATION

#### Product Identifier

**SRM Number:** 1941b  
**SRM Name:** Organics in Marine Sediment  
**Other Means of Identification:** Not applicable.

#### Recommended Use of This Material and Restrictions of Use

This Standard Reference Material (SRM) is marine sediment collected at the mouth of the Baltimore (MD) Harbor. SRM 1941b is intended for use in evaluating analytical methods for the determination of selected polycyclic aromatic hydrocarbons (PAHs), polychlorinated biphenyl (PCB) congeners, and chlorinated pesticides in marine sediment and similar matrices. All of the constituents for which certified, reference, and information values are provided in SRM 1941b were naturally present in the sediment before processing. A unit of SRM 1941b consists of a bottle containing 50 g of radiation-sterilized, freeze-dried sediment.

#### Company Information

National Institute of Standards and Technology  
Standard Reference Materials Program  
100 Bureau Drive, Stop 2300  
Gaithersburg, Maryland 20899-2300

Telephone: 301-975-2200  
FAX: 301-948-3730  
E-mail: SRMMSDS@nist.gov  
Website: <http://www.nist.gov/srm>

Emergency Telephone ChemTrec:  
1-800-424-9300 (North America)  
+1-703-527-3887 (International)

### 2. HAZARDS IDENTIFICATION

#### Classification

**Physical Hazard:** Not classified.  
**Health Hazard:** Not classified.

#### Label Elements

**Symbol**  
No Symbol/Pictogram

**Signal Word**  
Not applicable.

**Hazard Statement(s):** Not applicable.

**Precautionary Statement(s):** Not applicable.

**Hazards Not Otherwise Classified:** Not applicable.

**Ingredients(s) with Unknown Acute Toxicity:** Not applicable.

### 3. COMPOSITION AND INFORMATION ON HAZARDOUS INGREDIENTS

**Substance:** Marine sediment

**Other Designations:** Sediment.

This material is naturally occurring marine sediment from an urban area. The material contains trace amounts of polycyclic aromatic hydrocarbons (PAHs), polychlorinated biphenyl (PCB) congeners, and should be handled with care. Components are listed in compliance with OSHA's 29 CFR 1910.1200; for the actual values see the Certificate of Analysis.

Hazardous Component(s)	CAS Number	EC Number (EINECS)	Nominal Mass Concentration (%)
Marine Sediment	Not available	Not available	23B0228 CLPLIKE (Rev1) - Page 479 of 498 100

---

#### 4. FIRST AID MEASURES

---

##### Description of First Aid Measures:

**Inhalation:** If adverse effects occur, remove to uncontaminated area. If not breathing, give artificial respiration or oxygen by qualified personnel. Seek immediate medical attention.

**Skin Contact:** Wash skin with soap and water.

**Eye Contact:** Flush eyes with water for at least 15 minutes. If necessary, seek medical attention.

**Ingestion:** If adverse effects occur after ingestion, seek medical treatment.

**Most Important Symptoms/Effects, Acute and Delayed:** May cause irritation.

**Indication of any immediate medical attention and special treatment needed, if necessary:** If any of the above symptoms are present, seek medical attention if needed.

---

#### 5. FIRE FIGHTING MEASURES

---

**Fire and Explosion Hazards:** Negligible fire hazard. Avoid generating dust. See Section 9, "Physical and Chemical Properties" for flammability properties.

##### Extinguishing Media:

Suitable: Use extinguishing media appropriate for surrounding fire.

Unsuitable: None listed.

**Specific Hazards Arising from the Chemical:** None listed.

**Special Protective Equipment and Precautions for Fire-Fighters:** Avoid inhalation of material or combustion byproducts. Wear full protective clothing and NIOSH approved self-contained breathing apparatus (SCBA).

**NFPA Ratings** (0 = Minimal; 1 = Slight; 2 = Moderate; 3 = Serious; 4 = Severe)

Health = 1

Fire = 0

Reactivity = 0

---

#### 6. ACCIDENTAL RELEASE MEASURES

---

**Personal Precautions, Protective Equipment and Emergency Procedures:** Any accumulated material on surfaces should be removed and properly disposed of. Use suitable protective equipment; see Section 8, "Exposure Controls and Personal Protection".

**Methods and Materials for Containment and Clean up:** Collect spilled material in appropriate container for disposal. Keep out of water supplies and sewers. Keep unnecessary people away, isolate hazard area and deny entry.

---

#### 7. HANDLING AND STORAGE

---

**Safe Handling Precautions:** Minimize dust generation and accumulation on surfaces. Routine housekeeping should be instituted to ensure that dusts do not accumulate on surfaces. See Section 8, "Exposure Controls and Personal Protection".

**Storage:** Store and handling in accordance with all current regulations and standards.

---

#### 8. EXPOSURE CONTROLS AND PERSONAL PROTECTION

---

**Exposure Limits:** No occupational exposure limits have been established for marine sediment. This material is a particulate matter and adequate inhalation/respiratory protection should be used to minimize exposure. The exposure limits for Particulates Not Otherwise Regulated (PNOR) are applicable.

OSHA (PEL): 15 mg/m<sup>3</sup> (TWA, total particulates not otherwise regulated)

OSHA (PEL) 5 mg/m<sup>3</sup> (TWA, respirable particulates not otherwise regulated)

NIOSH (REL): 10 mg/m<sup>3</sup> (TWA, total particulates not otherwise regulated, 8 h)

NIOSH (REL): 5 mg/m<sup>3</sup> (TWA, respirable particulates not otherwise regulated)

**Engineering Controls:** Provide local exhaust or process enclosure ventilation system. Ensure compliance with applicable exposure limits.



**Personal Protection:** In accordance with OSHA 29 CFR 1910.132, subpart I, wear appropriate Personal Protective Equipment (PPE) to minimize exposure to this material.

**Respiratory Protection:** If workplace conditions warrant a respirator, a respiratory protection program that meets OSHA 29CFR 1910.134 must be followed. Refer to NIOSH 42 CFR 84 for applicable certified respirators.

**Eye/Face Protection:** Wear splash resistant safety goggles with a face shield. An eye wash station should be readily available near areas of use.

**Skin and Body Protection:** Personal protective equipment for the body should be selected based on the task being performed and the risks involved and should be approved by a specialist before handling this product. Chemical-resistant gloves should be worn at all times when handling chemicals.

---

## 9. PHYSICAL AND CHEMICAL PROPERTIES

---

### Descriptive Properties:

<b>Appearance</b> (physical state, color, etc.):	amorphous powder
<b>Molecular Formula:</b>	not applicable
<b>Molar Mass (g/mol):</b>	not applicable
<b>Odor:</b>	not available
<b>Odor threshold:</b>	not available
<b>pH:</b>	not available
<b>Evaporation rate:</b>	not applicable
<b>Melting point/freezing point (°C):</b>	not available
<b>Specific Gravity (water=1)</b>	not available
<b>Vapor Pressure (mmHg):</b>	not applicable
<b>Vapor Density (air = 1):</b>	not applicable
<b>Viscosity (cP):</b>	not applicable
<b>Solubility(ies):</b>	not available
<b>Partition coefficient (n-octanol/water):</b>	not available
<b>Particle Size:</b>	<150 µm

### Thermal Stability Properties:

<b>Autoignition Temperature (°C):</b>	not available
<b>Thermal Decomposition (°C):</b>	not available
<b>Initial boiling point and boiling range (°C):</b>	not available
<b>Explosive Limits, LEL (Volume %):</b>	not available
<b>Explosive Limits, UEL (Volume %):</b>	not available
<b>Flash Point (°C):</b>	not available
<b>Flammability (solid, gas):</b>	not available

---

## 10. STABILITY AND REACTIVITY

---

**Reactivity:** Stable at normal temperatures and pressure.

**Stability:**   X   Stable        Unstable

**Possible Hazardous Reactions:** None listed.

**Conditions to Avoid:** Avoid generating dust.

**Incompatible Materials:** None listed.

**Fire/Explosion Information:** See Section 5, "Fire Fighting Measures".

**Hazardous Decomposition:** Thermal decomposition will produce oxides of carbon.

**Hazardous Polymerization:**        Will Occur   X   Will Not Occur

---

## 11. TOXICOLOGICAL INFORMATION

---

**Route of Exposure:**  Inhalation  Skin  Ingestion

**Symptoms Related to the Physical, Chemical and Toxicological Characteristics:** Generated dust may cause irritation if inhaled.

**Potential Health Effects (Acute, Chronic and Delayed):**

**Inhalation:** Generated dust may cause irritation.

**Skin Contact:** May cause mechanical irritation.

**Eye Contact:** May cause mechanical irritation.

**Ingestion:** No data available.

**Numerical Measures of Toxicity:**

**Acute Toxicity:** Not classified; no data available.

**Skin Corrosion/Irritation:** Not classified; no data available.

**Serious Eye damage/ Eye irritation:** Not classified; no data available.

**Respiratory Sensitization:** Not classified; no data available.

**Skin Sensitization:** Not classified; no data available.

**Germ Cell Mutagenicity:** Not classified; no data available.

**Carcinogenicity:** Not classified.

**Listed as a Carcinogen/Potential Carcinogen**  Yes  No  
Marine sediment is not listed by NTP, IARC or OSHA as a carcinogen.

**Reproductive Toxicity:** Not classified; no data available.

**Specific Target Organ Toxicity, Single Exposure:** Not classified; no data available.

**Specific Target Organ Toxicity, Repeated Exposure:** Not classified; no data available.

**Aspiration Hazard:** Not classified; no data available.

---

## 12. ECOLOGICAL INFORMATION

---

**Ecotoxicity Data:** No data available.

**Persistence and Degradability:** No data available.

**Bioaccumulative Potential:** No data available.

**Mobility in Soil:** No data available.

**Other Adverse effects:** No data available.

---

## 13. DISPOSAL CONSIDERATIONS

---

**Waste Disposal:** Dispose of waste in accordance with all applicable federal, state, and local regulations.

---

## 14. TRANSPORTATION INFORMATION

---

**U.S. DOT and IATA:** Not regulated by DOT or IATA.

---

## 15. REGULATORY INFORMATION

---

**U.S. Regulations:**

CERCLA Sections 102a/103 (40 CFR 302.4): Not regulated.

SARA Title III Section 302 (40 CFR 355.30): Not regulated.

SARA Title III Section 304 (40 CFR 355.40): Not regulated.

SARA Title III Sections 311/312 Hazardous Categories (40 CFR 370.21):

ACUTE HEALTH: No.  
CHRONIC HEALTH: No.  
FIRE: No.  
REACTIVE: No.  
PRESSURE: No.

**State Regulations:**

California Proposition 65: Not listed.

**U.S. TSCA Inventory:** Not listed.

**TSCA 12(b), Export Notification:** Not listed.

**Canadian Regulations:**

WHMIS Information: Not provided for this material.

---

**16. OTHER INFORMATION**

---

**Issue Date:** 31 March 2014

**Sources:** 29 CFR Occupational Health and Safety Office (OSHA) 1910.1000, *Limits for Air Contaminants*, Table Z-1; available at [http://www.osha.gov/pls/oshaweb/owadisp.show\\_document?p\\_table=STANDARDS&p\\_id=9992](http://www.osha.gov/pls/oshaweb/owadisp.show_document?p_table=STANDARDS&p_id=9992) (accessed Mar 2014).

Center for Disease Control (CDC) NIOSH Pocket Guide to Chemical Hazards, *Particulates not otherwise regulated*; available at <http://www.cdc.gov/niosh/npg/npgd0480.html> (accessed Mar 2014).

**Key of Acronyms:**

ACGIH	American Conference of Governmental Industrial Hygienists	NRC	Nuclear Regulatory Commission
ALI	Annual Limit on Intake	NTP	National Toxicology Program
CAS	Chemical Abstracts Service	OSHA	Occupational Safety and Health Administration
CERCLA	Comprehensive Environmental Response, Compensation, and Liability Act	PEL	Permissible Exposure Limit
CFR	Code of Federal Regulations	RCRA	Resource Conservation and Recovery Act
DOT	Department of Transportation	REL	Recommended Exposure Limit
EC50	Effective Concentration, 50 %	RM	Reference Material
EINECS	European Inventory of Existing Commercial Chemical Substances	RQ	Reportable Quantity
EPCRA	Emergency Planning and Community Right-to-Know Act	RTECS	Registry of Toxic Effects of Chemical Substances
IARC	International Agency for Research on Cancer	SARA	Superfund Amendments and Reauthorization Act
IATA	International Air Transportation Agency	SCBA	Self-Contained Breathing Apparatus
IDLH	Immediately Dangerous to Life and Health	SRM	Standard Reference Material
LC50	Lethal Concentration, 50 %	STEL	Short Term Exposure Limit
LD50	Lethal Dose, 50 %	TLV	Threshold Limit Value
LEL	Lower Explosive Limit	TPQ	Threshold Planning Quantity
MSDS	Material Safety Data Sheet	TSCA	Toxic Substances Control Act
NFPA	National Fire Protection Association	TWA	Time Weighted Average
NIOSH	National Institute for Occupational Safety and Health	UEL	Upper Explosive Limit
NIST	National Institute of Standards and Technology	WHMIS	Workplace Hazardous Materials Information System

**Disclaimer:** Physical and chemical data contained in this SDS are provided only for use in assessing the hazardous nature of the material. The SDS was prepared carefully, using current references; however, NIST does not certify the data in the SDS. The certified values for this material are given in the NIST Certificate of Analysis.

Users of this SRM should ensure that the SDS in their possession is current. This can be accomplished by contacting the SRM Program: telephone (301) 975-2200; fax (301) 948-3730; [23802280@nist.gov](mailto:23802280@nist.gov) or <http://www.nist.gov/srm>

Bill to:

68455

23B0228 CLPLIKE (Rev1) - Page 484 of 498

Ship to:

68456

DAVE MITCHELL  
ANALYTICAL RESOURCES INC  
4611 S 134TH PLACE  
SUITE 100  
TUKWILA, WA 98168-3240

1 (206) 695-6205

DAVE MITCHELL  
ANALYTICAL RESOURCES INC  
4611 S 134TH PLACE  
SUITE 100  
TUKWILA, WA 98168-3240

1 (206) 695-6205


Ship via	UPS Ground	Description	
Salesmen	MCMIDM2	Instructions	
Contact	DAVE MITCHELL	Prof	

Order discrepancies (other than back ordered items) must be reported to our Customer Relations Department at 301-975-6776 within 5 days of receipt of shipment or this order will be considered complete. NIST SRMs/RMs are generally not returnable - with the exception of defective goods or shipments made in error by NIST. To return a SRM/RM, please call for instructions and a Return SRM/RM Authorization Number before shipment. Returns WILL NOT BE ACCEPTED without prior authorization.

Order	UOM	Ship	UOM	B/O	UOM	Item	Description
1	EACH	1	EACH	0	EACH	1941B	Organics in Marine Sediment
							Total qty:
							1 / EACH
<b>NOT FOR HUMAN CONSUMPTION, LABORATORY USE ONLY.</b>							

Picked by	
Packed by	
# of pieces	
Weight	

9/21/16 04:04 PM



1200101



**MP Biomedicals, LLC**

29525 Fountain Parkway  
Solon, Ohio 44139

Telephone: 440/337-1200  
Toll Free: 800/854-0530

Fax: 440/337-1180  
web: www.mpbio.com

## Certificate of Analysis

**Product Description:** Microcrystalline Cellulose Powder\_  
**Catalog Number:** 191499\_  
**Lot:** Q9483\_


<b>Formula:</b> (C <sub>6</sub> H <sub>10</sub> O <sub>5</sub> ) <sub>n</sub> <b>CAS #:</b> 9004-34-6 <b>Physical Description:</b> White Powder	<b>Formula Weight:</b> N/A <b>Storage:</b> 15 - 30°C
---	---

Test	Specification	Result
Identity Test	Passes	Passes
Purity	97.0 - 102.0%	97.0 - 102.0%
Moisture	<5.0%	3.4%
Particle Size/Mesh	Wt %	
+60 mesh	<8%	<1%
+200 mesh	>45%	55%
pH	5 - 7	6.73
Residue on Ignition	<0.05%	<0.05%
Water Soluble Substances	<12.0 mg/5 g	4.5 mg/5 g
Heavy Metals	<10 ppm	<10 ppm

**H001822**  
Microcrystalline Cellulose Powder (TOC)  
Expires 11/30/2022  
*Prepared By Casey English 2/22/2019*

Identification A & B: Passes  
Bulk Density: 0.29 g/ml  
Bulk Density (graduated cylinder): 0.31 g/ml  
Conductivity: 18 µS/cm  
Starch: Negative  
Ether Soluble Substances: 0.01%  
Total Aerobic microbial Count: 100 cfu/g  
Total Mold and Yeast Count: 20 cfu/g  
Staphylococcus aureus: Absent/1 g  
Pseudomonas aeruginosa: Absent/1 g  
E. coli: Absent/1 g  
Salmonella: Absent/10 g  
Particle size:

- 450 mesh: 77%  
- d10: 37 um  
- d50: 139 um  
- d90: 271 um  
TUP: <9/600 cm<sup>2</sup>  
Degree of brightness: >88%  
Powder flow-angle of repose: <42°  
Recommended Retest Date: 11/30/2022



07/26/2018 - John Huang, PhD  
MP Biomedicals, LLC.  
Quality Control Manager

This is an electronically generated document  
<mailto:biotech@mpbio.com>  
<http://www.mpbio.com>

Online Ordering, MSDSs, certificates of analysis and data sheets now available on our web site  
Technical Service: 1-800-279-5490 (440-337-1200) Customer Service: 1-800-854-0530 (440-337-1200)



**MP Biomedicals, LLC**

29525 Fountain Parkway  
Solon, Ohio 44139

Telephone: 440/337-1200  
Toll Free: 800/854-0530

Fax: 440/337-1180  
web: www.mpbio.com

## Certificate of Analysis

**Product Description:** Microcrystalline Cellulose Powder\_  
**Catalog Number:** 191499\_  
**Lot:** Q9483\_

**Formula:** (C<sub>6</sub>H<sub>10</sub>O<sub>5</sub>)<sub>n</sub>

**CAS #:** 9004-34-6

**Physical Description:** White Powder

**Formula Weight:** N/A

**Storage:** 15 - 30°C

Test	Specification	Result
Identity Test	Passes	Passes
Purity	97.0 - 102.0%	97.0 - 102.0%
Moisture	<5.0%	3.4%
Particle Size/Mesh	Wt %	
+60 mesh	<8%	<1%
+200 mesh	>45%	55%
pH	5 - 7	6.73
Residue on Ignition	<0.05%	<0.05%
Water Soluble Substances	<12.0 mg/5 g	4.5 mg/5 g
Heavy Metals	<10 ppm	<10 ppm

**H001822**

Microcrystalline Cellulose Powder (TOC)

Expires 11/30/2022

Prepared By Casey English 2/22/2019

Identification A & B: Passes

Bulk Density: 0.29 g/ml

Bulk Density (graduated cylinder): 0.31 g/ml

Conductivity: 18 µS/cm

Starch: Negative

Ether Soluble Substances: 0.01%

Total Aerobic microbial Count: 100 cfu/g

Total Mold and Yeast Count: 20 cfu/g

Staphylococcus aureus: Absent/1 g


Pseudomonas aeruginosa: Absent/1 g

E. coli: Absent/1 g

Salmonella: Absent/10 g

Particle size:

- 450 mesh: 77%  
- d10: 37 um  
- d50: 139 um  
- d90: 271 um  
TUP: <9/600 cm<sup>2</sup>  
Degree of brightness: >88%  
Powder flow-angle of repose: <42°  
Recommended Retest Date: 11/30/2022



07/26/2018 - John Huang, PhD  
MP Biomedicals, LLC.  
Quality Control Manager

This is an electronically generated document  
<mailto:biotech@mpbio.com>  
<http://www.mpbio.com>

Online Ordering, MSDSs, certificates of analysis and data sheets now available on our web site  
Technical Service: 1-800-279-5490 (440-337-1200) Customer Service: 1-800-854-0530 (440-337-1200)





**MP Biomedicals, LLC**

29525 Fountain Parkway  
Solon, Ohio 44139

Telephone: 440/337-1200  
Toll Free: 800/854-0530

Fax: 440/337-1180  
web: www.mpbio.com

## Certificate of Analysis

**Product Description:** Microcrystalline Cellulose Powder\_  
**Catalog Number:** 191499\_  
**Lot:** Q9483\_

**Formula:** (C<sub>6</sub>H<sub>10</sub>O<sub>5</sub>)<sub>n</sub>

**CAS #:** 9004-34-6

**Physical Description:** White Powder

**Formula Weight:** N/A

**Storage:** 15 - 30°C

Test	Specification	Result
Identity Test	Passes	Passes
Purity	97.0 - 102.0%	97.0 - 102.0%
Moisture	<5.0%	3.4%
Particle Size/Mesh	Wt %	
+60 mesh	<8%	<1%
+200 mesh	>45%	55%
pH	5 - 7	6.73
Residue on Ignition	<0.05%	<0.05%
Water Soluble Substances	<12.0 mg/5 g	4.5 mg/5 g
Heavy Metals	<10 ppm	<10 ppm

**H001822**

Microcrystalline Cellulose Powder (TOC)

Expires 11/30/2022

*Prepared By Casey English 2/22/2019*

Identification A & B: Passes

Bulk Density: 0.29 g/ml

Bulk Density (graduated cylinder): 0.31 g/ml

Conductivity: 18 µS/cm

Starch: Negative

Ether Soluble Substances: 0.01%

Total Aerobic microbial Count: 100 cfu/g

Total Mold and Yeast Count: 20 cfu/g

Staphylococcus aureus: Absent/1 g


Pseudomonas aeruginosa: Absent/1 g

E. coli: Absent/1 g

Salmonella: Absent/10 g

Particle size:

- 450 mesh: 77%  
- d10: 37 um  
- d50: 139 um  
- d90: 271 um  
TUP: <9/600 cm<sup>2</sup>  
Degree of brightness: >88%  
Powder flow-angle of repose: <42°  
Recommended Retest Date: 11/30/2022



07/26/2018 - John Huang, PhD  
MP Biomedicals, LLC.  
Quality Control Manager

This is an electronically generated document  
<mailto:biotech@mpbio.com>  
<http://www.mpbio.com>

Online Ordering, MSDSs, certificates of analysis and data sheets now available on our web site  
Technical Service: 1-800-279-5490 (440-337-1200) Customer Service: 1-800-854-0530 (440-337-1200)



**Form I**  
**INORGANIC ANALYSIS DATA SHEET**  
**SM 2540 G-97**

<b>LDW23-SC1009</b>
---------------------

Laboratory: Analytical Resources, LLC  
 Client: Anchor QEA, LLC  
 Project: AOC5 MR Phase 1  
 Matrix: Sediment      Laboratory ID: 23B0228-01 C      SDG: 23B0228  
 Sampled: 02/08/23 16:30      Prepared: 02/14/23 15:24      File ID:  
 % Solids: 67.04      Preparation: No Prep Wet Chem      Analyzed: 02/14/23 15:26  
 Batch: BLB0363      Sequence:  
 Instrument: BAL2      Calibration:

CAS NO.	Analyte	Concentration (%)	Dilution Factor	MDL	MRL	Q
	Total Solids	67.04	1	0.04	0.04	



**TOTAL SOLIDS/VOLATILE SOLIDS (TS / TVS) BENCHSHEET for Solid samples**  
**Method: PSEP 1986, SM2540, EPA 160.1**  
 (dry at 104 (12-24 hr) then combust at 550 (30 min))

Batch: BLB0363  
 Date: 2/14/2023 15:26  
 Analyst: UW

Instrumentation: Drying Ovens: 12; Muffle Furnace: 2; Analytical Balance: BAL2

**Batch drying time**  
 record times as mm/dd/yy hh:mm  
 date/time in oven: 2/14/2023 18:02  
 date/time out: 2/15/2023 9:10  
 elapsed hrs = 15.1 OK

TS (%) calculated as:  
 Final dry wt (g) = (Dry Wt - Tare Wt)  
 TS = (Final Dry Wt)/ (grams Sample-Tare)

Oven Temps, °C  
 Start Temp 111  
 Dry Cycle 1 101  
 Dry Cycle 2  
 Dry Cycle 3

TVS (mg/kg dry wt) calculated as:  
 Final ash wt (g) = (min ash wt - tare wt)  
 TVS (mg/kg) = [(Dry wt-Ash wt)/ (dry weight)] \*1,000,000  
 if ash wt > dry wt, "Chk for Err"  
 if dry wt-ash wt < 0.001 g, "< (1/dry wt) \*1,000,000

**Balance Calibration Check**  
 Record weights to 4 places

Cal Weight ID:	CV-02	CV-02	CV-02	CV-02	CV-02
Date & Time:	2/14/23 16:57	2/14/23 17:14	2/15/23 10:15		
Cal Wt (g):	10.0000	10.0000	10.0000		
	Cal OK!	Cal OK!	Cal OK!	Cal OK!	

Sample ID	Dish #	Tare Wt. (g)	Dish & Sample (g)	Dry Wt 104C (grams)			dry Wt (g)	TS (%)	Notes	ASH WT 550C (grams)			Ash Wt (g)	TVS		Notes	Fixed (%)
				1	2	3				1	2	3		(mg/kg)	(%)		
BLB0363-BLK1	1	0.8165	0.0000	0.8164			-0.0001	0.01%		0.8163	0.8164	STOP	-0.0002	(1,000,000)	-100.00%		200.00
23B0132-02	2	0.8080	7.7567	6.2917			5.4837	78.92%									
23B0186-02	3	0.7976	7.6119	2.1886			1.3910	20.41%									
23B0201-04	4	0.8178	8.7379	0.8578			0.0400	0.51%									
23B0217-02	5	0.7970	9.9528	7.8559			7.0589	77.10%		7.7296	7.7242	STOP	6.9272	18.657	1.87%		98.13
BLB0363-DUP1	6	0.8191	8.8818	6.9830			6.1639	76.45%	RPD=0.8	6.8768	6.8740	STOP	6.0549	17.684	1.77%	RPD=5.4	98.23
BLB0363-DUP2	7	0.8274	9.7658	7.6926			6.8652	76.81%	RSD=0.4	7.5774	7.5745	STOP	6.7471	17.203	1.72%	RSD=4.2	98.28
23B0217-03	8	0.7801	8.5405	6.4880			5.7079	73.55%									
23B0217-04	9	0.7882	8.2467	6.0610			5.2728	70.70%									
23B0217-05	10	0.7791	9.9661	7.8837			7.1046	77.33%									
23B0217-06	11	0.8327	8.8414	6.9735			6.1408	76.68%									
23B0228-01	12	0.8155	9.5531	6.6730			5.8575	67.04%									
23B0229-01	13	0.8069	6.7036	3.7901			2.9832	50.59%									
23B0229-02	14	0.7832	6.4483	3.8997			3.1165	55.01%									
23B0229-03	15	0.8288	7.6607	4.4594			3.6306	53.14%									
23B0229-04	16	0.8370	7.3723	4.2251			3.3881	51.84%									
23B0229-05	17	0.8489	6.9526	3.7079			2.8590	46.84%									
23B0229-06	18	0.8065	6.6629	3.6930			2.8865	49.29%									
23B0229-07	19	0.8210	9.6116	6.9955			6.1745	70.24%									
23B0229-08	20	0.7915	6.8197	3.8895			3.0980	51.39%									
23B0276-01	21	0.8121	9.6147	6.3198			5.5077	62.57%									
23B0278-01	22	0.8155	6.9050	3.4122			2.5967	42.64%		3.1587	3.1583	STOP	2.3428	97.778	9.78%		90.22
23B0278-02	23	0.7992	7.0762	3.4903			2.6911	42.87%		3.2454	3.2440	STOP	2.4448	91.524	9.15%		90.85



Form I  
METHOD BLANK DATA SHEET  
SM 2540 G-97  
TotalAnalytes

Blank

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Batch: BLB0363

Laboratory ID: BLB0363-BLK1

Prepared: 02/14/23 15:24

Matrix: Solid

Preparation: No Prep Wet Chem

Analyzed: 02/14/23 15:26

Sequence:

Calibration:

Instrument: BAL2

CAS NO.	Analyte	Concentration (%)	Dilution Factor	MDL	MRL	Q
	Total Solids	ND	1	0.04	0.04	U



## HOLDING TIME SUMMARY

Analysis: SM 2540 G-97

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Sample Name	Date Collected	Date Received	Date Prepared	Days to Prep	Max Days to Prep	Date Analyzed	Days to Analysis	Max Days to Analysis	Q
LDW23-SC1009 23B0228-01	02/08/23 16:30	02/09/23 10:55	02/14/23 15:24	5	28	02/14/23 15:26	6	28	

\* Indicates hold time exceedance.



**Analytical Resources, LLC**  
Analytical Chemists and Consultants

## METHOD DETECTION AND REPORTING LIMITS

**SM 2540 G-97**

Laboratory: Analytical Resources, LLC

SDG: 23B0228

Client: Anchor QEA, LLC

Project: AOC5 MR Phase 1

Matrix: Solid

Instrument:

<b>Analyte</b>	<b>MDL</b>	<b>RL</b>	<b>Units</b>
Total Solids	0.04	0.04	%



TOTAL SOLIDS BENCHSHEET					Batch:	BLB0340
Method: PSEP 1986					Date:	2/14/2023 11:36
(dry at 103-105 C)					Analyst:	CR
Instrumentation					Drying Oven:	15
					Analytical Balance:	B139298002
Batch drying time			Oven Temp, C		TS (%) calculated as:	
Record times as mm/dd/yy hh:mm			109		Final dry wt (g) = (Dry Wt - Tare Wt)	
Date/time in oven:	2/15/2023 12:23		99		TS = (Final Dry Wt X 100)/ (sample & dish -dish tare)	
Date/time out:	2/16/2023 9:31					
Elapsed hrs:	21.1					
					Oven Temps, °C	
					Start Temp:	109
					End Temp:	99
SAMPLE ID	Dish Tare Wt (g)	Dish with Sample (g)	Dry Wt (g)	Solids Wt (g)	TS (%)	Sample Decanted
23B0077-05	0.7900	12.2000	9.8200	9.03	79.14%	No
23B0217-02	0.8000	12.3400	9.5800	8.78	76.08%	Yes
23B0217-03	0.8100	12.8800	9.8600	9.05	74.98%	Yes
23B0217-04	0.7900	12.1200	9.2500	8.46	74.67%	Yes
23B0217-05	0.8000	12.2800	9.8300	9.03	78.66%	Yes
23B0217-06	0.7900	11.9000	9.4500	8.66	77.95%	Yes
23B0228-01	0.7800	12.4700	8.0300	7.25	62.02%	No
23B0229-01	0.8200	11.7100	6.4700	5.65	51.88%	Yes
23B0229-02	0.8000	11.2300	6.6400	5.84	55.99%	Yes
23B0229-03	0.7900	12.9400	7.4500	6.66	54.81%	Yes
23B0229-04	0.7800	12.3400	6.8100	6.03	52.16%	Yes
23B0229-05	0.7700	11.3200	5.5100	4.74	44.93%	Yes
23B0229-06	0.8100	12.4300	6.4600	5.65	48.62%	Yes
23B0229-07	0.8000	11.6500	8.2500	7.45	68.66%	Yes
23B0229-08	0.7600	11.6300	6.1600	5.40	49.68%	Yes
23B0263-01	0.8000	12.8500	11.6300	10.83	89.88%	No
23B0278-01	0.8100	12.4300	5.8600	5.05	43.46%	Yes
23B0278-02	0.7700	12.1400	5.9100	5.14	45.21%	Yes

TOTAL SOLIDS BENCHSHEET

Method: PSEP 1986

(dry at 103-105 C)

Instrumentation

Batch: BLB0340

Date: 2/14/2023 11:36

Analyst: OR

Drying Oven: 015

Analytical Balance: B139298002

Batch drying time

Record times as mm/dd/yy hh:mm

Date/time in oven: 2/15/23 12:23  
 Date/time out: 2/16/23 9:31  
 Elapsed hrs: 0.0

Oven Temp, C

109  
99

TS (%) calculated as:

Final dry wt (g) = (Dry Wt - Tare Wt)

TS = (Final Dry Wt X 100)/(sample & dish -dish tare)

Oven Temps, °C

Start Temp: 109  
 End Temp: 99

SAMPLE ID	Dish Tare Wt (g)	Dish with Sample (g)	Dry Wt (g)	Solids Wt (g)	TS (%)	Sample Decanted
23B0077-05	0.79	12.20	9.82			No
23B0217-02	0.80	12.34	9.58			No YES
23B0217-03	0.81	12.88	9.86			No YES
23B0217-04	0.79	12.12	9.25			No YES
23B0217-05	0.80	12.28	9.83			No YES
23B0217-06	0.79	11.90	9.45			No YES
23B0228-01	0.78	12.47	8.03			No
23B0229-01	0.82	11.71	6.47			No YES
23B0229-02	0.80	11.23	6.64			No YES
23B0229-03	0.79	12.94	7.45			No YES
23B0229-04	0.78	12.34	6.81			No YES
23B0229-05	0.77	11.32	5.51			No YES
23B0229-06	0.81	12.43	6.46			No YES
23B0229-07	0.80	11.65	8.25			No YES
23B0229-08	0.76	11.63	6.16			No YES
23B0263-01	0.80	12.85	11.63			No
23B0278-01	0.81	12.43	8.86			No YES
23B0278-02	0.77	12.14	5.91			No YES

T/S + Screens  
 1 copy - 23B0217  
 23B0263  
 2 copies - 23B0229  
 23B0278  
 3 copies - 23B0077  
 No copies - 23B0228

Done!!